

Date <b>01/24/2017</b>	PSC Device Code <b>3V4</b>	No. of Sections <b>4</b>	Complete each item with Y (Yes), N (No), or NA (Not Applicable)
Name of Business <b>FML Terminal Logistics, LLC</b>			<input checked="" type="checkbox"/> New Installation (w/ RFI Check)
Mailing Address <b>3 Sugar Creek Center Blvd Ste 550 Sugar Land TX 77478</b>			<input checked="" type="checkbox"/> Performed Calibration
Device Location <b>FML Terminal Logistics - Manitou</b>			<input checked="" type="checkbox"/> Use as a Reference Scale
City <b>Manitou</b>	State <b>ND</b>	Zip Code <b>58776</b>	<input checked="" type="checkbox"/> Replaced Existing Equipment
County			<input checked="" type="checkbox"/> Non-Commercial
Device Contact/Manager <b>null</b>			<input checked="" type="checkbox"/> Scale Indoors
Email Address <b>null</b>			<input checked="" type="checkbox"/> Variance Permit Posted; Expiration Date: <b>06/01/2017</b>
Telephone Number <b>null</b>			<input checked="" type="checkbox"/> Stored/Recalled Weights meet NDAC 69-10-02-23 & 69-10-02-24
Cell Number			<input checked="" type="checkbox"/> Software is NTEP Approved
Cell Number			<input checked="" type="checkbox"/> Built-In Standards; Date Certified:
Cell Number			<input checked="" type="checkbox"/> Multiple Decks/Single Indicator
Cell Number			<input checked="" type="checkbox"/> Customer Has Clear View
Cell Number			<input checked="" type="checkbox"/> Clearance Below Scale Clear
Cell Number			<input checked="" type="checkbox"/> Video Camera Working
Cell Number			<input checked="" type="checkbox"/> Approach Requirements Met
Cell Number			<input checked="" type="checkbox"/> View Distance > 200'
Cell Number			<input checked="" type="checkbox"/> Pit Coping and Crush Strip Good
Cell Number			<input checked="" type="checkbox"/> Two-Way Audio Working
Cell Number			<input checked="" type="checkbox"/> In response to PSC Quality Assurance Inspection

**Scale Information - RLWS ATV**

Scale Manufacturer <b>RLWS ATV</b>	Indicator Manufacturer/Model <b>RLWS 720i-2A</b>	Indicator Serial No. <b>1768700072</b>	Weighing Elements <b>8 Cell</b>
Capacity/Divisions/Units <b>160,000 x 20 lbs</b>	Legible Label <b>X Y _ N</b>	Class <b>III</b>	Printer Manufacturer/Model <b>Epson TM-U590</b>
Deck Size <b>11' x 70'</b>	Clearance (inches) <b>17"</b>	Approach - 12' Concrete Level <b>Gravel</b>	Printer Serial No. <b>J9SF012199</b>
SR or Discrimination Test Zero Load = NA Loaded = NA	Motion Detection Range = 3d	Hard Surface Approach (Length/Slope/Condition) <b>Gravel</b>	
AZSM (Auto Zero)		Range = 3d	

**Test Data**

LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-	
			As Found	As Left				As Found	As Left
Description: <b>Corner Test</b>					Description: <b>Section Test</b>				
1	10,000 lbs		+20	0	1	20,000 lbs			
2	"		-200	0	2	"			
3	"		+80	0	3	"			
4	"		+40	0	4	20,000 lbs			
5	"		+60	0					
6	"		+100	0					
7	"		-900	0					
8	10,000 lbs		+40	0					
	BC=0								

**Strain Load Test**

Section (Increasing or Decreasing)	<b>Increasing 2</b>	Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary.  <b>Set up RLWS ATV EZ7011-ST-100-ATV S/N- 6VRK. Scale has a variance to be used legal for trade. Set up scale as needed. While testing scale found corner seven to be bad. Replaced load cell as needed. Tested scale and adjusted into acceptance tolerance. Scale has been granted until 06/01/2017 or until concrete is pour which ever comes first. Also a copy of the scale ticket must be filled with the commission. Once every three months scale must be tested and must maintain acceptance tolerance. Permit number 515. Adjusted print formats as per Bryan.</b>
Full Truck Weight	<b>60,020</b>	
Test Weight	<b>20,000</b>	
Empty Truck Weight	<b>40,020</b>	
Errors	<b>0</b>	
Physical Seal <input checked="" type="checkbox"/> Y <input checked="" type="checkbox"/> N	Electronic Audit Trail: <input type="checkbox"/> Y <input type="checkbox"/> N	
Seal Date: 01/24/2017	Audit Trail Information:	
Seal Type: Lead		
Meets tolerances in PSC adopted NIST Handbook 44? <input checked="" type="checkbox"/> Maintenance <input checked="" type="checkbox"/> Acceptance <input checked="" type="checkbox"/>		
<input checked="" type="checkbox"/> Approved <input type="checkbox"/> Rejected <input type="checkbox"/> Taken out of Service <input checked="" type="checkbox"/> Sticker Applied		
By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.		
Cody Bannick	1718	
Permit Holder Signature	Permit Number	
	01/24/2017	
Operator Signature	Date	