



Date 08/02/2017	PSC Device Code 3V4	No. of Sections 4	Complete each item with Y (Yes), N (No), or NA (Not Applicable)	
Name of Business FML Terminal Logistics, LLC			<input type="checkbox"/> New Installation (w/ RFI Check)	<input type="checkbox"/> Performed Calibration
Mailing Address 3 Sugar Creek Center Blvd Ste 550 Sugar Land TX 77478			<input checked="" type="checkbox"/> Modified Equipment	<input type="checkbox"/> Use as a Reference Scale
Device Location FML Terminal Logistics - Manitou 6353 93rd Ave NW			<input type="checkbox"/> Replaced Existing Equipment	<input type="checkbox"/> Non-Commercial
City Manitou	State ND	Zip Code 58776	<input type="checkbox"/> Scale Indoors	
County			NA Variance Permit Posted; Expiration Date:	
Device Contact/Manager Bryan Jolley			NA Stored/Recalled Weights meet NDAC 69-10-02-23 & 69-10-02-24	
Cell Number			NA Software is NTEP Approved	
Telephone Number			NA Built-In Standards; Date Certified:	
Email Address			NA Multiple Decks/Single Indicator	
			NA Customer Has Clear View	
			NA Clearance Below Scale Clear	
			NA Video Camera Working	
			NA Approach Requirements Met	
			NA View Distance > 200'	
			NA Pit Coping and Crush Strip Good	
			NA Two-Way Audio Working	
			NA In response to PSC Quality Assurance Inspection	

Scale Information - RLWS ATV

Scale Manufacturer RLWS ATV	Indicator Manufacturer/Model RLWS 720i-2A	Indicator Serial No. 1768700072	Weighing Elements 8 Cell
Capacity/Divisions/Units 160,000 x 20 lbs	Legible Label X Y _ N	Class III	Printer Manufacturer/Model Epson TM-U590
Deck Size 11' x 70'	Clearance (inches) 12"	Approach - 12' Concrete Level 12" concrete	Printer Serial No. J9SF012199
SR or Discrimination Test Zero Load = NA Loaded = NA	Motion Detection Range = 3d	Hard Surface Approach (Length/Slope/Condition) Gravel	
		AZSM (Auto Zero) Range = 3d	

Test Data

LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-	
			As Found	As Left				As Found	As Left
Description:					Description:				
1	10,000 lbs			0	1	20,000 lbs			0
2	"			0	2	"			0
3	"		+20/40	0	3	"		+40/60	0
4	"			0	4	20,000 lbs			0
5	"		0	0					
6	"		+20/40	0					
7	"			0					
8	10,000 lbs			0					
	BC=0								

Strain Load Test

Section (Increasing or Decreasing)	Sec 3 Increasing	Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary. Arrived on site to move Rice Lake ATV to new concrete foundation. We prepped scale by putting it on the shipping brackets and un wiring the expansion cable as well as the home run cable. We moved the scale and got it set back on load cells and re-wired and anchored to the foundation. We then proceeded to test the scale to find Section 3 to be 40/60lbs positive we then adjusted as needed. Found scale to be accurate no further adjustments were needed at this time.
Full Truck Weight	59,980	
Test Weight	20,000	
Empty Truck Weight	39,980	
Errors	0	
Physical Seal <input checked="" type="checkbox"/> Y <input type="checkbox"/> N	Electronic Audit Trail: <input type="checkbox"/> Y <input type="checkbox"/> N	
Seal Date: 08/02/2017	Audit Trail Information:	
Seal Type: Lead		
Meets tolerances in PSC adopted NIST Handbook 44? <input type="checkbox"/> Maintenance <input checked="" type="checkbox"/> Acceptance <input checked="" type="checkbox"/> Approved <input type="checkbox"/> Rejected <input type="checkbox"/> Taken out of Service <input checked="" type="checkbox"/> Sticker Applied		
By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.		
<i>[Signature]</i>	1802	
Permit Holder Signature	Permit Number	
<i>[Signature]</i>	08/02/2017	
Operator Signature	Date	