

Date 08/22/2017	PSC Device Code 3V6	No. of Sections 6	Complete each item with Y (Yes), N (No), or NA (Not Applicable)	
Name of Business C & C Transload, LLC			<input type="checkbox"/> New Installation (w/ RFI Check)	<input checked="" type="checkbox"/> Performed Calibration
Mailing Address 404 N Street Suite 306 Rock Springs WY 82901			<input type="checkbox"/> Modified Equipment	<input checked="" type="checkbox"/> Use as a Reference Scale
Device Location Blaisdell-Fairmount Minerals 6350 67th Ave NW			<input type="checkbox"/> Replaced Existing Equipment	<input type="checkbox"/> Non-Commercial
City Blaisdell	State ND	Zip Code 58718	<input checked="" type="checkbox"/> Variance Permit Posted; Expiration Date: 10/31/2017	
County Mountrail			<input checked="" type="checkbox"/> Software is NTEP Approved	
Telephone Number 701-313-0218			<input checked="" type="checkbox"/> Built-In Standards; Date Certified:	
Cell Number			<input type="checkbox"/> Multiple Decks/Single Indicator	<input checked="" type="checkbox"/> Customer Has Clear View
Device Contact/Manager Bryan			<input checked="" type="checkbox"/> Clearance Below Scale Clear	<input checked="" type="checkbox"/> Video Camera Working
Email Address bryan@cctransload.com			<input type="checkbox"/> Approach Requirements Met	<input checked="" type="checkbox"/> View Distance > 200'
			<input checked="" type="checkbox"/> Pit Coping and Crush Strip Good	<input checked="" type="checkbox"/> Two-Way Audio Working
<input checked="" type="checkbox"/> In response to PSC Quality Assurance Inspection				

Scale Information - South Cardinal - 007396

Scale Manufacturer Cardinal EPR	Indicator Manufacturer/Model Cardinal 225	Indicator Serial No. E10515-0015	Weighing Elements 12 Cell 75K- DB-75000S
Capacity/Divisions/Units 160,000 x 20 lbs	Legible Label X Y _ N	Class IIIL	Printer Manufacturer/Model Computer System
Deck Size 11' x 120'	Clearance (inches) 5"	Approach - 12' Concrete Level 10'	Printer Serial No. Page Printer
SR or Discrimination Test Zero Load = NA Loaded = NA		Motion Detection Range = 60 lbs.	Hard Surface Approach (Length/Slope/Condition) 10', Now Paved
			AZSM (Auto Zero) Range = 60 lbs.

Test Data

LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-	
			As Found	As Left				As Found	As Left
Description: Corner Test					Description: Section Test				
1	10,000 lbs		-	0	1	20,000 lbs		+40	0
2	10,000 lbs		-	0	2	20,000 lbs		-20	0
3	10,000 lbs		-	0	3	20,000 lbs		-20	0
4	10,000 lbs		-	0	4	20,000 lbs		-40	0
5	10,000 lbs		-	0	5	20,000 lbs		-20	0
6	10,000 lbs		-	0	6	20,000 lbs		-20/40	0
7	10,000 lbs		-	0					
8	10,000 lbs		-	0		BC=0			
9	10,000 lbs		-	0					
10	10,000 lbs		-	0/-					
11	10,000 lbs		-	0					
12	10,000 lbs		-	0					
	BC=0								


Strain Load Test


Section (Increasing or Decreasing)	1 Increasing	4 Increasing	Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary.
Full Truck Weight	60,460 lbs	60,460 lbs	
Test Weight	20,000 lbs	20,000 lbs	
Empty Truck Weight	40,460 lbs	40,460 lbs	
Errors	0	0	

Physical Seal Y _ N Electronic Audit Trail: Y _ N
 Seal Date: 08/22/2017 Audit Trail Information:
 Seal Type: Lead/Software

Meets tolerances in PSC adopted NIST Handbook 44? _ Maintenance Acceptance
 Approved _ Rejected _ Taken out of Service Sticker Applied

By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.


 Permit Holder Signature _____ 1768
 Permit Number


 Operator Signature _____ 08/22/2017
 Date

Performed service and test of scale per Variance. found scale running minus. calibrated and adjusted corners and sections as needed. scale approved.

Date 08/22/2017	PSC Device Code 3V7	No. of Sections 7	Complete each item with Y (Yes), N (No), or NA (Not Applicable)	
Name of Business C & C Transload, LLC			<input type="checkbox"/> New Installation (w/ RFI Check)	<input checked="" type="checkbox"/> Performed Calibration
Mailing Address 404 N Street Suite 306 Rock Springs WY 82901			<input type="checkbox"/> Modified Equipment	<input checked="" type="checkbox"/> Use as a Reference Scale
Device Location Blaisdell-Fairmount Minerals 6350 67th Ave NW			<input type="checkbox"/> Replaced Existing Equipment	<input type="checkbox"/> Non-Commercial
City Blaisdell	State ND	Zip Code 58718	<input checked="" type="checkbox"/> Variance Permit Posted; Expiration Date: 10/31/2017	
County Mountrail	Telephone Number 701-313-0218		<input checked="" type="checkbox"/> Software is NTEP Approved	
Device Contact/Manager Bryan	Cell Number		<input checked="" type="checkbox"/> Built-In Standards; Date Certified:	
Email Address bryan@cctransload.com			<input type="checkbox"/> Multiple Decks/Single Indicator	<input checked="" type="checkbox"/> Customer Has Clear View
			<input checked="" type="checkbox"/> Clearance Below Scale Clear	<input checked="" type="checkbox"/> Video Camera Working
			<input type="checkbox"/> Approach Requirements Met	<input checked="" type="checkbox"/> View Distance > 200'
			<input checked="" type="checkbox"/> Pit Coping and Crush Strip Good	<input checked="" type="checkbox"/> Two-Way Audio Working
<input type="checkbox"/> In response to PSC Quality Assurance Inspection				

Scale Information - North Cardinal - 007397

Scale Manufacturer Cardinal EPR	Indicator Manufacturer/Model Cardinal 225	Indicator Serial No. E19714-0098	Weighing Elements 14 Cell 75K DB-7500S
Capacity/Divisions/Units 160,000 x 20 lbs	Legible Label X Y _ N	Class IIIL	Printer Manufacturer/Model Computer System
Deck Size 11' x 130'	Clearance (inches) 5"	Approach - 12' Concrete Level 10'	Printer Serial No. Page Printer
SR or Discrimination Test Zero Load = NA Loaded = NA	Motion Detection Range = 60 lbs.		Hard Surface Approach (Length/Slope/Condition) 10' Now Paved
		AZSM (Auto Zero) Range = 60 lbs.	

Test Data

LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-	
			As Found	As Left				As Found	As Left
Description: Corner Test					Description:				
1	10,000 lbs		-	0	14	10,000 lbs		-	0
2	10,000 lbs		-	0		BC=0			
3	10,000 lbs		-	0	Section				
4	10,000 lbs		-	0	1	20,000 lbs		+20	0
5	10,000 lbs		-	0	2	20,000 lbs		-20	0
6	10,000 lbs		-	0	3	20,000 lbs		-20	0
7	10,000 lbs		-	0	4	20,000 lbs		-20	0
8	10,000 lbs		-	0	5	20,000 lbs		-20	0
9	10,000 lbs		-	0	6	20,000 lbs		-20/40	0
10	10,000 lbs		-	0	7	20,000 lbs		-40	0
11	10,000 lbs		-	0					
12	10,000 lbs		-	0		BC=0			
13	10,000 lbs		-	0					

Strain Load Test

Section (Increasing or Decreasing)	4 Increasing	7 Increasing	Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary.
Full Truck Weight	60,480 lbs	60,460 lbs	
Test Weight	20,000 lbs	20,000 lbs	
Empty Truck Weight	40,480/60 lbs	40,460 lbs	
Errors	0/-	0	
Physical Seal <input checked="" type="checkbox"/> Y <input type="checkbox"/> N	Electronic Audit Trail: <input type="checkbox"/> Y <input type="checkbox"/> N		Performed service and test of scale per posted variance permit. Found scale running slightly minus. calibrated and adjusted sections and corners as needed. Wind had picked up by the time we strained scale. about a 15 mph blowing in the west door. east door was calm.
Seal Date: 08/22/2017	Audit Trail Information:		
Seal Type: Lead			
Meets tolerances in PSC adopted NIST Handbook 44? <input type="checkbox"/> Maintenance <input checked="" type="checkbox"/> Acceptance			
<input checked="" type="checkbox"/> Approved <input type="checkbox"/> Rejected <input type="checkbox"/> Taken out of Service <input checked="" type="checkbox"/> Sticker Applied			
By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.			
Permit Holder Signature		1768	Permit Number
Operator Signature		08/22/2017	Date