



**ROCKY MOUNTAIN
SCALE WORKS™**

800-700-5048

Date 01-19-2015	PSC Device Code 3V4	No. of Sections 4	Complete each item with Y (Yes), N (No), or NA (Not Applicable)	
Name of Business Environmental Materials			<input type="checkbox"/> New Installation (w/ RFI Check)	<input type="checkbox"/> Performed Calibration
Mailing Address 357 102 x Ave			<input type="checkbox"/> Modified Equipment	<input type="checkbox"/> Use as a Reference Scale
City Killdeer	State ND	Zip Code 58640	<input checked="" type="checkbox"/> Replaced Existing Equipment	<input type="checkbox"/> Non-Commercial
County Dunn	Telephone Number 701-572-5017		<input checked="" type="checkbox"/> Variance Permit Posted; Expiration Date: 10-31-2016	
Device Contact/Manager Karey Reynolds	Cell Number 406-697-3585		<input checked="" type="checkbox"/> Stored/Recalled Weights Meet NDAC §§ 69-10-02-23 & 69-10-02-24	
Email Address Kreynolds	Device Location 102 x Ave Killdeer, ND 58640 Main Truck Scale		<input type="checkbox"/> Software is NTEP Approved	
			<input type="checkbox"/> Built-In Standards; Date Certified: _____	
			<input checked="" type="checkbox"/> Multiple Decks/Single Indicator	<input checked="" type="checkbox"/> Customer Has Clear View
			<input checked="" type="checkbox"/> Clearance Below Scale Clear	<input type="checkbox"/> Video Camera Working
			<input type="checkbox"/> Approach Requirements Met	<input checked="" type="checkbox"/> View Distance > 200'
			<input checked="" type="checkbox"/> Pit Coping and Crush Strip Good	<input type="checkbox"/> Two-Way Audio Working
			<input type="checkbox"/> In Response to PSC Quality Assurance Inspection	

Scale Information

Scale Manufacturer Avery Weigh Tronix	Indicator Manufacturer/Model Avery Weigh Tronix /	Indicator Serial No. 141215661	Weighing Elements Electronic 120,000 lb
Capacity/Divisions/Units 120,000 x 20 lb	Legible Label Y N III L	Printer Manufacturer/Model EPSON TM 295	Printer Serial No. J9KF091964
Deck Size 70'	Clearance (inches) Variance Posted 7"	Approach - 12' Concrete Level Variance Posted 10' Level	Hard Surface Approach (Length/Slope/Condition) 10' Flat / Condition Good
SR or Discrimination Test Zero Load = _____ lb. Loaded = _____ lb.	Motion Detection Range = 60 lb.	AZSM (Auto Zero) Range = 60 lb.	

Test Data

LBP / Section / Product Wt.	Value of Test Weights Used	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Errors +/-	
		As Found	As Left			As Found	As Left
2	0	0	0	2	0	0	0
C-1	10,500 lb	10,500 lb	10,500 lb	S-1	20,500	20500	20,500
2	"	"	"	2	"	"	"
3	"	"	"	3	"	20480	20480
4	"	"	"	4	"	20,500	20,500
5	"	-20	-20	BC	0	0	0
6	"	-20	-20				
7	"	10,500	10,500				
8	"	"	"				
BC	0	0	0				

Strain Load Test

Section (Increasing or Decreasing)	1/O	4/I	Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary. • Variance for clearance + approach / Exit ramps posted on 8-16-13. Expired 10-31-16 • Electronically Added Scale • All scale equipment is in good working order • Indicator Replaced • Slight Winds • Most stable weight recorded
Full Truck Weight	49280	49280	
Test Weight	20500	20500	
Empty Truck Weight	28780	28780	
Errors			
Physical Seal <input checked="" type="checkbox"/> Y <input type="checkbox"/> N Seal Date: 01-19-15 Seal Type: Seals Sticker	Electronic Audit Trail: <input type="checkbox"/> Y <input type="checkbox"/> N Audit Trail Information:		
Meets tolerances in PSC adopted NIST Handbook 44? <input checked="" type="checkbox"/> Maintenance <input type="checkbox"/> Acceptance			
<input checked="" type="checkbox"/> Approved <input type="checkbox"/> Rejected <input type="checkbox"/> Taken out of Service <input checked="" type="checkbox"/> Sticker Applied			
By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.			
T. Dylan Moses Permit Holder Signature	1724 Permit No.		
Karey Reynolds Operator Signature	1/19/15 Date		