



# ROCKY MOUNTAIN SCALE WORKS™

## 800-700-5048

Date: <u>4-27-15</u>	PSC Device Code: <u>3V4</u>	No. of Sections: <u>4</u>	Complete each item with Y (Yes), N (No), or NA (Not Applicable)	
Name of Business: <u>Environmental Materials</u>			<input checked="" type="checkbox"/> New Installation (w/ RFI Check)	<input checked="" type="checkbox"/> Performed Calibration
Mailing Address: <u>357 102X Ave</u>			<input checked="" type="checkbox"/> Modified Equipment	<input checked="" type="checkbox"/> Use as a Reference Scale
City: <u>Killdeer</u>	State: <u>ND</u>	Zip Code: <u>58640</u>	<input checked="" type="checkbox"/> Replaced Existing Equipment	<input checked="" type="checkbox"/> Non-Commercial
County: <u>Dunn</u>	Telephone Number: <u>701-572</u>		<input checked="" type="checkbox"/> Variance Permit Posted; Expiration Date: <u>10-31-16</u>	
Device Contact/Manager: <u>Karey Reynolds</u>	Cell Number: <u>406-697-3585</u>		<input checked="" type="checkbox"/> Stored/Recalled Weights Meet NDAC §§ 69-10-02-23 & 69-10-02-24	
Email Address: <u>Kreynolds@EMI-Ash.com</u>	Device Location: <u>102A Ave Killdeer, ND 58640 Main Truck Scale</u>		<input checked="" type="checkbox"/> Software is NTEP Approved	
			<input checked="" type="checkbox"/> Built-In Standards; Date Certified: _____	
			<input checked="" type="checkbox"/> Multiple Decks/Single Indicator	<input checked="" type="checkbox"/> Customer Has Clear View
			<input checked="" type="checkbox"/> Clearance Below Scale Clear	<input checked="" type="checkbox"/> Video Camera Working
			<input checked="" type="checkbox"/> Approach Requirements Met	<input checked="" type="checkbox"/> View Distance > 200'
			<input checked="" type="checkbox"/> Pit Coping and Crush Strip Good	<input checked="" type="checkbox"/> Two-Way Audio Working
			<input checked="" type="checkbox"/> In Response to PSC Quality Assurance Inspection	

Scale Manufacturer: <u>Andry Weigh-Tronix</u>	Indicator Manufacturer/Model: <u>AWIX/E1310</u>	Indicator Serial No.: <u>14125661</u>	Weighing Elements: <u>Electronic 120,000lb</u>
Capacity/Divisions/Units: <u>120,000lb</u>	Legible Label: <u>X Y N</u>	Class: <u>III L</u>	Printer Manufacturer/Model: <u>EPSON/TM245</u>
Deck Size: <u>70'</u>	Clearance (inches): <u>Variance Posted 7"</u>	Printer Serial No.: <u>J9KFB091964</u>	Hard Surface Approach (Length/Slope/Condition): <u>10' Flat / Condition Good</u>
SR or Discrimination Test	Motion Detection	AZSM (Auto Zero)	
Zero Load = _____ lb. Loaded = _____ lb.	Range = <u>60</u> lb.	Range = <u>60</u> lb.	

LBP / Section / Product Wt.	Value of Test Weights Used	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Errors +/-	
		As Found	As Left			As Found	As Left
<u>2</u>	<u>0</u>	<u>0</u>	<u>0</u>	<u>2</u>	<u>0</u>	<u>0</u>	<u>0</u>
<u>C-1</u>	<u>10,500lb</u>	<u>0</u>	<u>0</u>	<u>5-1</u>	<u>20500 lb</u>	<u>0</u>	<u>0</u>
<u>2</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>2</u>	<u>"</u>	<u>"</u>	<u>"</u>
<u>3</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>3</u>	<u>"</u>	<u>+20</u>	<u>+20</u>
<u>4</u>	<u>"</u>	<u>"</u>	<u>"</u>	<u>4</u>	<u>"</u>	<u>0</u>	<u>0</u>
<u>5</u>	<u>"</u>	<u>+20</u>	<u>+20</u>	<u>BC</u>	<u>0</u>	<u>0</u>	<u>0</u>
<u>6</u>	<u>"</u>	<u>0</u>	<u>0</u>				
<u>7</u>	<u>"</u>	<u>"</u>	<u>"</u>				
<u>8</u>	<u>"</u>	<u>"</u>	<u>"</u>				
<u>BC</u>	<u>0</u>	<u>0</u>	<u>0</u>				

Section (Increasing or Decreasing)		<u>110</u>	<u>4/F</u>	Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary. <u>Variance For Clearance &amp; Approach / Exit Ramps Posted.</u> <u>Electronically Added Scale</u> <u>All Scale Elements in good working order</u> <u>Other than Variance elements, Certification meets all ND W&amp;M Requirements</u>
Full Truck Weight		<u>49120</u>	<u>49120</u>	
Test Weight		<u>20500</u>	<u>20500</u>	
Empty Truck Weight		<u>28620</u>	<u>28620</u>	
Errors		<u>0</u>	<u>0</u>	
<input checked="" type="checkbox"/> Physical Seal <u>X</u> Y <u>N</u> Seal Date: <u>4-27-15</u> Seal Type: <u>Seal + Sticker</u>	Electronic Audit Trail: <u>Y</u> <input checked="" type="checkbox"/> N Audit Trail Information: _____			
Meets tolerances in PSC adopted NIST Handbook 44? <input checked="" type="checkbox"/> Maintenance <u>X</u> Acceptance _____				
<input checked="" type="checkbox"/> Approved <u>X</u> Rejected _____ Taken out of Service _____ Sticker Applied _____				
By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.				
Permit Holder Signature: <u>[Signature]</u>		Permit No.: <u>1724</u>		
Operator Sr				