



**ROCKY MOUNTAIN
SCALE WORKS™**

800-700-5048

Date: <u>7-30-15</u>	PSC Device Code	No. of Sections	Complete each item with Y (Yes), N (No), or NA (Not Applicable)	
Name of Business: <u>Environmental Materials</u>			<input checked="" type="checkbox"/> New Installation (w/ RFI Check)	<input checked="" type="checkbox"/> Performed Calibration
Mailing Address: <u>357 102X Ave.</u>			<input checked="" type="checkbox"/> Modified Equipment	<input checked="" type="checkbox"/> Use as a Reference Scale
City: <u>Kinderhook</u> State: <u>ND</u> Zip Code: <u>58440</u>			<input checked="" type="checkbox"/> Replaced Existing Equipment	<input checked="" type="checkbox"/> Non-Commercial
County: <u>Dunn</u> Telephone Number: <u>701-572-5017</u>			<input checked="" type="checkbox"/> Variance Permit Posted; Expiration Date: <u>10-31-15</u>	
Device Contact/Manager: <u>Korey Reynolds</u> Cell Number: <u>406-697-3585</u>			<input checked="" type="checkbox"/> Stored/Recalled Weights Meet NDAC §§ 69-10-02-23 & 69-10-02-24	
Email Address: <u>KReynolds@EMI-Ash.com</u>			<input checked="" type="checkbox"/> Software is NTEP Approved	
Device Location: <u>102X Ave Kinderhook ND 58440 Main Truck Scale</u>			<input checked="" type="checkbox"/> Built-In Standards; Date Certified: _____	
			<input checked="" type="checkbox"/> Multiple Decks/Single Indicator	<input checked="" type="checkbox"/> Customer Has Clear View
			<input checked="" type="checkbox"/> Clearance Below Scale Clear	<input checked="" type="checkbox"/> Video Camera Working
			<input checked="" type="checkbox"/> Approach Requirements Met	<input checked="" type="checkbox"/> View Distance > 200'
			<input checked="" type="checkbox"/> Pit Coping and Crush Strip Good	<input checked="" type="checkbox"/> Two-Way Audio Working
			<input checked="" type="checkbox"/> In Response to PSC Quality Assurance Inspection	

Scale Information		Indicator Manufacturer/Model	Indicator Serial No.	Weighing Elements
Capacity/Divisions/Units		Legible Label	Printer Manufacturer/Model	Printer Serial No.
Deck Size		Clearance (inches)	Approach - 12' Concrete Level	Hard Surface Approach (Length/Slope/Condition)
SR or Discrimination Test		Motion Detection	AZSM (Auto Zero)	Range = _____ lb.

Capacity/Divisions/Units: 130,000lb x 20lb
 Deck Size: 70'
 SR or Discrimination Test: Zero Load = _____ lb. Loaded = _____ lb.
 Indicator Manufacturer/Model: Avery Weigh-Tronix-WI-125 GTN
 Indicator Serial No.: 029616
 Weighing Elements: Electronic 120,000lb
 Legible Label: X Y N III L
 Printer Manufacturer/Model: EPSON/TM295
 Printer Serial No.: 59KFA91964
 Clearance (inches): Variance Permitted 7"
 Approach - 12' Concrete Level: Variance Permitted 6' level
 Hard Surface Approach (Length/Slope/Condition): Correct 10' Flat condition good
 Motion Detection: 60 lb.
 AZSM (Auto Zero): 60 lb.

LBP / Section / Product Wt.	Value of Test Weights Used	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Errors +/-	
		As Found	As Left			As Found	As Left
<u>Z</u>	<u>0</u>	<u>0</u>	<u>0</u>	<u>Z</u>	<u>0</u>	<u>0</u>	<u>0</u>
<u>C1</u>	<u>10,500</u>	<u>0</u>	<u>0</u>	<u>S1</u>	<u>20,500</u>	<u>0</u>	<u>0</u>
<u>2</u>	<u>11</u>	<u>0</u>	<u>0</u>	<u>2</u>	<u>11</u>	<u>0</u>	<u>0</u>
<u>3</u>	<u>11</u>	<u>0</u>	<u>0</u>	<u>3</u>	<u>11</u>	<u>0</u>	<u>0</u>
<u>4</u>	<u>11</u>	<u>0</u>	<u>0</u>	<u>4</u>	<u>11</u>	<u>+20</u>	<u>+20</u>
<u>5</u>	<u>11</u>	<u>0</u>	<u>0</u>	<u>BC</u>	<u>0</u>	<u>0</u>	<u>0</u>
<u>6</u>	<u>11</u>	<u>0</u>	<u>0</u>				
<u>7</u>	<u>11</u>	<u>+20</u>	<u>+20</u>				
<u>8</u>	<u>11</u>	<u>0</u>	<u>0</u>				
<u>BC</u>	<u>0</u>	<u>0</u>	<u>0</u>				

Strain Load Test		Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary.
Section (Increasing or Decreasing)	<u>1 / D</u> <u>4 / I</u>	<u>Variance for Clearance + approach / exit ramps posted on 10-31-15. Expires 10-31-15</u>
Full Truck Weight	<u>49,980</u> <u>49,980</u>	
Test Weight	<u>20,500</u> <u>20,500</u>	
Empty Truck Weight	<u>29,480</u> <u>29,480</u>	
Errors	<u>0</u> <u>0</u>	<u>Electrically Added Scale.</u>
<input checked="" type="checkbox"/> Physical Seal <input checked="" type="checkbox"/> Y <input type="checkbox"/> N	Electronic Audit Trail: <input type="checkbox"/> Y <input type="checkbox"/> N	<u>All scale elements are in good working condition.</u>
Seal Date: <u>7-30-15</u>	Audit Trail Information:	
Seal Type: <u>lead + sticker</u>		<u>Other Than Variance elements, certification meets all ND Weights + Measures Requirements</u>
Meets tolerances in PSC adopted NIST Handbook 44? <input type="checkbox"/> Maintenance <input type="checkbox"/> Acceptance		
<input checked="" type="checkbox"/> Approved <input type="checkbox"/> Rejected <input type="checkbox"/> Taken out of Service <input checked="" type="checkbox"/> Sticker Applied		
By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.		
Permit Holder Signature: <u>T. Dylan Moses</u>		
Operator Signature: <u>[Signature]</u>		
Permit No.: <u>1724</u>		
Date: <u>7-30-15</u>		