

ROCKY MOUNTAIN SCALE WORKS™

800-700-5048

Date 10-7-15		PSC Device Code 3V4		No. of Sections 4		Complete each item with Y (Yes), N (No), or NA (Not Applicable)					
Name of Business Environmental Materials						<input checked="" type="checkbox"/> New Installation (w/ RFI Check)		<input checked="" type="checkbox"/> Performed Calibration			
Mailing Address 357 102x Ave						<input checked="" type="checkbox"/> Modified Equipment		<input checked="" type="checkbox"/> Use as a Reference Scale			
City Killdeer State ND Zip Code 58640						<input checked="" type="checkbox"/> Replaced Existing Equipment		<input checked="" type="checkbox"/> Non-Commercial			
County Dunn Telephone Number 701-572-5017						<input checked="" type="checkbox"/> Variance Permit Posted; Expiration Date: 10-31-2016					
Device Contact/Manager Karey Reynolds Cell Number 406-697-3585						<input checked="" type="checkbox"/> Stored/Recalled Weights Meet NDAC §§ 69-10-02-23 & 69-10-02-24					
Email Address Kreynolds@emi-ash.com						<input checked="" type="checkbox"/> Software is NTEP Approved					
Device Location 102x Ave Killdeer, ND 58640 Moria Truck Scale						<input checked="" type="checkbox"/> Built-In Standards; Date Certified: _____					
						<input checked="" type="checkbox"/> Multiple Decks/Single Indicator		<input checked="" type="checkbox"/> Customer Has Clear View			
						<input checked="" type="checkbox"/> Clearance Below Scale Clear		<input checked="" type="checkbox"/> Video Camera Working			
						<input checked="" type="checkbox"/> Approach Requirements Met		<input checked="" type="checkbox"/> View Distance > 200'			
						<input checked="" type="checkbox"/> Pit Coping and Crush Strip Good		<input checked="" type="checkbox"/> Two-Way Audio Working			
						<input checked="" type="checkbox"/> In Response to PSC Quality Assurance Inspection					

Scale Manufacturer Avery Weigh-Tronix		Indicator Manufacturer/Model Avery Weigh-Tronix		Indicator Serial No. 141215661		Weighing Elements Electronic 120,000lb	
Capacity/Divisions/Units 120,000 lb x 20 lb		Legible Label Class X Y N III C		Printer Manufacturer/Model Epson TM-U285		Printer Serial No. J9KE091964	
Deck Size 70'		Clearance (inches) Variance Posted 7"		Approach - 12' Concrete Level Variance Posted 10' (20=1)		Hard Surface Approach (Length/Slope/Condition) 10' Flat / Condition Good	
SR or Discrimination Test Zero Load = _____ lb. Loaded = _____ lb.		Motion Detection Range = 60 lb.		AZSM (Auto Zero) Range = 60 lb.			

LBP / Section / Product Wt.	Value of Test Weights Used	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Errors +/-	
		As Found	As Left			As Found	As Left
2	0	0	0	2	0	0	0
C-1	10,500	10,500	10,500	S-1	20,500	20,500	20,500
2	"	"	"	2	"	"	"
3	"	"	"	3	"	"	"
4	"	"	"	4	"	"	"
5	"	"	"	BC	0	0	0
6	"	"	"				
7	"	-20	-20				
8	"	10,500	10,500				
BC	0	0	0				

Section (Increasing or Decreasing)				Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary.			
Full Truck Weight		49800		<ul style="list-style-type: none"> • Variance for Clearance & Approach / Exit Ramps Posted on 8-16-13. Expires 10-31-15 • Electronically Added Scale • All Scale equipment in good order • Slight Winds • Most Stable Weight Was recorded during testing 			
Test Weight		20500					
Empty Truck Weight		29300					
Errors		0					
Physical Seal <input checked="" type="checkbox"/> Y <input type="checkbox"/> N		Seal Date: 10-7-15		Electronic Audit Trail: <input type="checkbox"/> Y <input type="checkbox"/> N			
Seal Type: Seal + Sticker		Audit Trail Information:					
Meets tolerances in PSC adopted NIST Handbook 44? <input checked="" type="checkbox"/> Maintenance <input type="checkbox"/> Acceptance							
<input checked="" type="checkbox"/> Approved <input type="checkbox"/> Rejected <input type="checkbox"/> Taken out of Service <input checked="" type="checkbox"/> Sticker Applied							
By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.							
T. Dylan Moses				1724			
Permit Holder Signature				Permit No.			
				10-7-15			
Operator Signature				Date			