



**ROCKY MOUNTAIN
SCALE WORKS™**
800-700-5048

Date	1-14-16	PSC Device Code	3V4	No. of Sections	4	Complete each item with Y (Yes), N (No), or NA (Not Applicable)	
Name of Business	Environmental Materials			<input checked="" type="checkbox"/> New Installation (w/ RFI Check)	<input checked="" type="checkbox"/> Performed Calibration	<input checked="" type="checkbox"/> Modified Equipment	<input checked="" type="checkbox"/> Use as a Reference Scale
Mailing Address	357 102x Ave			<input checked="" type="checkbox"/> Replaced Existing Equipment	<input checked="" type="checkbox"/> Non-Commercial	Y Variance Permit Posted; Expiration Date: 10-31-2016	
City	Killdeer	State	ND	Zip Code	58640	Y Stored/Recalled Weights Meet NDAC §§ 69-10-02-23 & 69-10-02-24	
County	Dunn	Telephone Number	701-572-5017			N/A Software is NTEP Approved	
Device Contact/Manager	Karey Reynolds	Cell Number	406-697-3585			N/A Built-In Standards; Date Certified: _____	
Email Address	KReynolds@EMI-Ash.Com			<input checked="" type="checkbox"/> Multiple Decks/Single Indicator	<input checked="" type="checkbox"/> Customer Has Clear View	<input checked="" type="checkbox"/> Clearance Below Scale Clear	<input checked="" type="checkbox"/> Video Camera Working
Device Location	102x Ave Killdeer, ND 58640 Main Truck Scale			<input checked="" type="checkbox"/> Approach Requirements Met	<input checked="" type="checkbox"/> View Distance > 200'	<input checked="" type="checkbox"/> Pit Coping and Crush Strip Good	<input checked="" type="checkbox"/> Two-Way Audio Working
				N/A In Response to PSC Quality Assurance Inspection			

Scale Manufacturer	Avery Weigh-Tronix	Indicator Manufacturer/Model	Avery Weigh-Tronix	Indicator Serial No.	141215661	Weighing Elements	Electronic 120,000 lb
Capacity/Divisions/Units	120,000 lb x 20lb	Legible Label	Y	Printer Manufacturer/Model	EPSON TM-V295	Printer Serial No.	J9KF891964
Deck Size	70'	Clearance (inches)	Variance Posted 7"	Approach - 12' Concrete Level	Variance Posted 10' level	Hard Surface Approach (Length/Slope/Condition)	10' Flat / Condition Good
SR or Discrimination Test		Motion Detection		AZSM (Auto Zero)		Range =	60 lb.
Zero Load =		lb. Loaded =		Range =	60	lb.	

LBP / Section / Product Wt.	Value of Test Weights Used	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Errors +/-	
		As Found	As Left			As Found	As Left
2	0	0	0	2	0	0	0
C-1	10,500	10,500	10,500	5-1	20,500	20500	0
2	"	"	"	2	"	20500	0
3	"	"	"	3	"	20480	-20
4	"	"	"	4	"	20500	0
5	"	"	"	BC	0	0	0
6	"	"	"				
7	"	10,500	-				
8	"	-	-				
BC	0	0	0				

Section (Increasing or Decreasing)	I/D	4/I	Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary. • Variance For Clearance & approach/Exit Ramps Posted on 8-16-13. Expires 10-31-16 • Electronically Added Scale • All Scale equipment in good order • Most stable weight recorded • Weather is windy with snow flurries
Full Truck Weight	49640	49640	
Test Weight	20500	20500	
Empty Truck Weight	29160	29160	
Errors	-20	-20	
<input checked="" type="checkbox"/> Physical Seal <input checked="" type="checkbox"/> Y <input type="checkbox"/> N Seal Date: 1-14-16 Seal Type: Seal/Sticker	Electronic Audit Trail: <input type="checkbox"/> Y <input type="checkbox"/> N Audit Trail Information:		
Meets tolerances in PSC adopted NIST Handbook 44? <input checked="" type="checkbox"/> Maintenance <input type="checkbox"/> Acceptance			
<input checked="" type="checkbox"/> Approved <input type="checkbox"/> Rejected <input type="checkbox"/> Taken out of Service <input checked="" type="checkbox"/> Sticker Applied			
By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.			
J. Dylek Moses Permit Holder Signature	1724 Permit No.		
<i>Karey Reynolds</i> Operator Signature	1-14-16 Date		