



# CAPITAL SCALE

PO BOX 2021 · Bismarck, ND 58502-2021  
Phone: 701-255-1556

Date <b>7-23-15</b>	PSC Device Code <b>3VS</b>	No. of Sections <b>5</b>	Complete each item with Y (Yes), N (No), or NA (Not Applicable)	
Name of Business <b>Brenntag</b>			<input type="checkbox"/> New Installation (w/ RFI Check)	<input type="checkbox"/> Performed Calibration
Mailing Address			<input type="checkbox"/> Modified Equipment	<input type="checkbox"/> Use as a Reference Scale
City <b>Dickinson</b> State <b>ND</b> Zip Code			<input type="checkbox"/> Replaced Existing Equipment	<input type="checkbox"/> Non-Commercial
County			<input checked="" type="checkbox"/> Variance Permit Posted; Expiration Date: _____	
Device Contact/Manager			<input checked="" type="checkbox"/> Stored/Recalled Weights Meet NDAC §§ 69-10-02-23 & 69-10-02-24	
Email Address			<input checked="" type="checkbox"/> Software is NTEP Approved	
Device Location			<input type="checkbox"/> Built-In Standards; Date Certified: _____	
			<input checked="" type="checkbox"/> Multiple Decks/Single Indicator	<input checked="" type="checkbox"/> Customer Has Clear View
			<input type="checkbox"/> Clearance Below Scale Clear	<input checked="" type="checkbox"/> Video Camera Working
			<input checked="" type="checkbox"/> Approach Requirements Met	<input checked="" type="checkbox"/> View Distance > 200'
			<input checked="" type="checkbox"/> Pit Coping and Crush Strip Good	<input checked="" type="checkbox"/> Two-Way Audio Working
			<input checked="" type="checkbox"/> In Response to PSC Quality Assurance Inspection	

### Scale Information

Scale Manufacturer <b>Toledo</b>	Indicator Manufacturer/Model <b>TNO 510</b>	Indicator Serial No. <b>1523818658</b>	Weighing Elements <b>FLL (10)</b>
Capacity/Divisions/Units <b>200,000 x 200</b>	Legible Label <b>Y N</b>	Class <b>III</b>	Printer Manufacturer/Model <b>J91CE092604</b>
Deck Size <b>30'</b>	Clearance (inches) <b>48</b>	Approach - 12' Concrete Level <b>OK</b>	Hard Surface Approach (Length/Slope/Condition) <b>Requires Variance</b>
SR or Discrimination Test	Motion Detection	AZSM (Auto Zero)	
Zero Load = _____ lb. Loaded = _____ lb.	Range = <b>60</b> lb.	Range = <b>60</b> lb.	

### Test Data

LBP / Section / Product Wt.	Value of Test Weights Used	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Errors +/-	
		As Found	As Left			As Found	As Left
LBP 1	10000#		0	Sect 1	20000#		0
2			0	2			0
3			0	3			0
4			0	4			-20
5			0	5	6		0
6			0	BU	0		0
7			0	TRUCK 1	48200#		20#
8			0				
9			0				
10	0		0				
BU	0		0		5		0
				BU	0		0

### Strain Load Test

Section (Increasing or Decreasing)	<b>Y</b>			Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary.  <b>Variance is missing</b>
Full Truck Weight	<b>48200</b>			
Test Weight	<b>20000#</b>			
Empty Truck Weight	<b>28240</b>			
Errors	<b>-20</b>			
<input type="checkbox"/> Physical Seal <b>Y</b> <input type="checkbox"/> N Seal Date: Seal Type:	<input type="checkbox"/> Electronic Audit Trail: <b>Y</b> <input type="checkbox"/> N Audit Trail Information:			
Meets tolerances in PSC adopted NIST Handbook 44? <input checked="" type="checkbox"/> Maintenance <input type="checkbox"/> Acceptance				
<input checked="" type="checkbox"/> Approved <input type="checkbox"/> Rejected <input type="checkbox"/> Taken out of Service <input checked="" type="checkbox"/> Sticker Applied				

By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and including NIST Handbook 44, for use of the device in commerce.

Permit Holder Signature 	Permit No. <b>1670</b>
Operator Signature 	Date <b>7-23-15</b>