

Date 12/21/2015	PSC Device Code 3V6	Number of Sections 6	Complete each item with Y (Yes), N (No), or NA (Not Applicable)	
Name of Business CHS, Inc. dbs CHS - Milton Group			<input type="checkbox"/> N New Installation (w/ RFI Check)	<input type="checkbox"/> Y Performed Calibration
Mailing Address 515 W Montrose Ave			<input type="checkbox"/> N Modified Equipment	<input type="checkbox"/> N Use as a Reference Scale
City Milton			<input type="checkbox"/> Y Replaced Existing Equipment	<input type="checkbox"/> N Non-Commercial
State ND			<input type="checkbox"/> Y Variance Permit Posted; Exp. Date: 8/31/2016	
Zip Code 58260			<input type="checkbox"/> NA Stored/Recalled Weights Meet NDAC §§ 69-10-02-23 & 69-10-02-24	
County Cavalier			<input type="checkbox"/> NA Software is NTEP Approved	
Telephone Number 701-496-3141			<input type="checkbox"/> NA Built-In Standards; Date Certified: NA	
Device Contact Person/Manager Josh Roppel			<input type="checkbox"/> N Multiple Decks/Single Indicator	<input type="checkbox"/> Y Customer Has Clear View
Cell Phone Number 701-256-2462			<input type="checkbox"/> Y Clearance Below Scale Clear	<input type="checkbox"/> Y Video Camera Working
E-mail Address josh.roppe@chsinc.com			<input type="checkbox"/> Y Approach Requirements Met	<input type="checkbox"/> Y View Distance > 200'
Device Location Langdon IN Scale - 10617 92nd Street NE; Langdon, ND 58249			<input type="checkbox"/> Y Pit Coping & Crush Strip Good	<input type="checkbox"/> Y Two-Way Audio Working
			<input type="checkbox"/> N In Response to PSC Quality Assurance Inspection	

SCALE INFORMATION

Scale Manufacturer Rice Lake Weighing Systems	Indicator Manufacturer/Model RLWS / 920i	Indicator Serial Number Scale:56E2 / Ind:1739600032	Weighing Elements 12 - 75K Load Cells
Capacity/Divisions/Units 160,000 x 20 lbs	Legible Label <input checked="" type="checkbox"/> Y <input type="checkbox"/> N	Printer Manufacturer/Model Computer System	Printer Serial Number NA
Deck Size 117' x 12'	Class IIIL	Approach (12' Concrete Level) 12' Concrete, Good	Hard Surface Approach (Length/Slope/Condition) 27' Concrete, Good
SR (Sensitivity Response) OR Discrimination Test ZERO LOAD = 1.4d = 2d lb.	Clearance (Inches) 12" (Pitless)	Motion Detection Range = 60 lb.	AZSM (Auto Zero) Range = 60 lb.

SCALE TEST DATA

LBP / Section / Product Weight	Value of Test Weights Used	Errors + / -		LBP / Section / Product Weight	Value of Test Weights Used	Errors + / -	
		As Found	As Left			As Found	As Left
LBP #1	10,000 lbs	0	0	Section #1	20,000 lbs	0	0
LBP #2	10,000 lbs	-0/-20	0	Section #2	20,000 lbs	0	0
LBP #3	10,000 lbs	0	0	Section #3	20,000 lbs	0	0
LBP #4	10,000 lbs	0	0	Section #4	20,000 lbs	0	0
LBP #5	10,000 lbs	0	0	Section #5	20,000 lbs	0	0
LBP #6	10,000 lbs	0	0	Section #6	20,000 lbs	0	0
LBP #7	10,000 lbs	0	0				
LBP #8	10,000 lbs	0	0	BC	00	0	0
LBP #9	10,000 lbs	0	0				
LBP #10	10,000 lbs	0	0				
LBP #11	10,000 lbs	0	0				
LBP #12	10,000 lbs	0	0				
BC	00	0	0				

STRAIN LOAD TEST


Section (Increasing or Decreasing)	3/1	6/1	NA	Remarks (Include environmental conditions, if applicable.) Attach additional sheets as necessary.
Full Truck Weight	56,600	56,600	NA	
Test Weight	20,000	20,000	NA	
Empty Truck Weight	36,620	36,600	NA	
Errors	-20	0	NA	

Physical Seal: Y N
Date: 2015 Type: Lead

Electronic Audit Trail: Y N
Audit Trail Information: NA

Meets tolerance in PSC adopted NIST Handbook 44? Maintenance Acceptance
 Approved Rejected Taken Out of Service Sticker Applied

By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks", the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.

Permit Holder Signature:  1398
Permit No.

Operator Signature: _____ Date: _____

Replaced digital indicator after power was shorted out to remote display. Also replaced main cpu on remote display. Remved 920i serial # 1670700081