



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY
 Public Service Commission
 SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company Kennedy Scales, Inc	Email Address RBaker@KennedyScales.com	Application Date 12/29/2014	
Mailing Address 11485 Xeon Street NW	City Coon Rapids	State MN	Zip Code 55448
Telephone Number 763-784-1022	Cell Phone Number 612-965-3914	Fax Number 763-784-1206	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input checked="" type="checkbox"/> 1. Rail <input checked="" type="checkbox"/> 2. Truck <input checked="" type="checkbox"/> 3. Livestock <input checked="" type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):


Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
<i>e.g. 1001</i>	<i>e.g. John Doe</i>	<i>e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6</i>
1295	Tim Seifert	Scales 1, 2, 4, 6, 7

Continued on Page 2



List below all field standards (attach current calibration reports):

Class F 25 LB Cast Weight	25121-25160
30 LB Kit Decimal/Fractional	A6
Metric Kit	M6
1000 LB Cast Cube	KSHD 1 - KSHD 18
Tiffin Weight Cart - 3000 LB	W60885



Coon Rapids, MN 800-735-3038
TESTED & APPROVED

Additional Application Items (initial where appropriate):

M O N T H	1	2	3	4	5	6	Y E A R	15
	7	8	9	10	11	12		16

Standardized Test Report	<input checked="" type="checkbox"/> Copy enclosed
	<input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input checked="" type="checkbox"/> Copy enclosed
	<input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input checked="" type="checkbox"/> Copy enclosed
	<input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am Rob Baker, and have authority to represent this company.
 By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.

[Signature]
Signature

Send Completed Application and Related Documents To:

Public Service Commission
 600 E Boulevard Ave Dept 408
 Bismarck ND 58505-0480
 Telephone: (701) 328-2400
 Fax: (701) 328-2410



Receipt Date: November 5, 2014
Test Date: November 5, 2014
Report Date: November 5, 2014

State Test No.: 333366
Set Serial No.: W60885
Bar Code: 202397

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448

Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 07

Item(s) Submitted: 3000 lb Weight Cart
Manufacturer: Tiffin
ASTM E617 Type: II
Equipment ID#: None
Condition: Good
Temperature: 20.7°C
Pressure: 734.9 mmHg
Relative Humidity: 46. %

Nominal Value	Serial No.	Correction (g)		NIST HB105-8 Tol		Unc. (g) (k=2)
		As Found	As Left	As Found	As Left	
3000 lb		840	85	*	Meets	60

* Weight Cart as found exceeds NIST HB 105-8 tolerance.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Mark Nicollet

Quality Manager

Reviewed by:

Kari Anderson

Metrologist



Receipt Date: November 3, 2014
Test Date: November 4, 2014
Report Date: November 5, 2014

State Test No.: 333348
Set Serial No.: KSHD 1 to 18
Bar Code: 200647

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448

Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 07

Item(s) Submitted: Cast Cube Weights
Manufacturer: Toledo
ASTM E617 Type: II
Equipment ID#: None
Condition: Good
Temperature: 20.8°C
Pressure: 736. mmHg
Relative Humidity: 45. %

Nominal Value	lb	Serial No.	Correction (g)		NIST HB105-1 Class		Unc. (g) (k=2)
			As Found	As Left	As Found	As Left	
1000	lb	KSDH - 1	-1	-1	F	F	5
1000	lb	KSDH - 2	-7	-7	F	F	5
1000	lb	KSDH - 3	-13	-13	F	F	5
1000	lb	KSDH - 4	-12	-12	F	F	5
1000	lb	KSDH - 5	-15	-15	F	F	5
1000	lb	KSDH - 6	-6	-6	F	F	5
1000	lb	KSDH - 7	4	4	F	F	5
1000	lb	KSDH - 8	-4	-4	F	F	5
1000	lb	KSDH - 9	-24	-24	F	F	5
1000	lb	KSDH - 10	-5	-5	F	F	5
1000	lb	KSDH - 11	-30	-30	F	F	5
1000	lb	KSDH - 12	-11	-11	F	F	5
1000	lb	KSDH - 13	-15	-15	F	F	5
1000	lb	KSDH - 14	-4	-4	F	F	5
1000	lb	KSDH - 15	6	6	F	F	5
1000	lb	KSDH - 16	-6	-6	F	F	5
1000	lb	KSDH - 17	-12	-12	F	F	5
1000	lb	KSDH - 18	-55	6	*	F	5

* Weight(s) as found exceed NIST HB 105-1 Class F tolerance.

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Mark Nicollet

Mark Nicollet
Quality Manager

Reviewed by:

Heidi Jones

Heidi Jones
Laboratory Administrator



Receipt Date: January 20, 2015
Test Date: January 26, 2015
Report Date: January 26, 2015

State Test No.: 333673
Set Serial No.: L1-L20
Barcode: 200648

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448

Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 09

Item(s) Submitted: Cast handle weights
Manufacturer: Assorted
ASTM E617 Type: II
Equipment ID#: None
Condition: Excellent
Temperature: 18.7°C
Pressure: 730.1 mmHg
Relative Humidity: 41. %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Unc. (mg) (k=2)
		As Found	As Left	As Found	As Left	
50 lb	L1	8290	-40	*	F	110.
50 lb	L2	6130	20	*	F	110.
50 lb	L3	7300	70	*	F	110.
50 lb	L4	-1220	-1220	F	F	110.
50 lb	L5	70	70	F	F	110.
50 lb	L6	290	290	F	F	110.
50 lb	L7	4580	-10	*	F	110.
50 lb	L8	4650	-210	*	F	110.
50 lb	L9	3360	-140	*	F	110.
50 lb	L10	2030	-490	F	F	110.
50 lb	L11	-1120	-1120	F	F	110.
50 lb	L12	630	630	F	F	110.
50 lb	L13	130	130	F	F	110.
50 lb	L14	-150	-150	F	F	110.
50 lb	L15	1800	1800	F	F	110.
50 lb	L16	2680	250	*	F	110.
50 lb	L17	9710	20	*	F	110.
50 lb	L18	330	330	F	F	110.
50 lb	L19	-20	-20	F	F	110.
50 lb	L20	-660	-660	F	F	110.

* Weight(s) as found exceed NIST HB 105-1 Class F tolerance.

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Heidi Jones
Heidi Jones
Laboratory Administrator

Reviewed by:
Mark Nicollet
Mark Nicollet
Quality Manager



Receipt Date: January 28, 2015
Test Date: February 9, 2015
Report Date: February 11, 2015

State Test No.: 333719
Set Serial No.: 25121 - 25160
Barcode: 200651

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448
Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 11

Item(s) Submitted: Cast Hand Weights
Manufacturer: Rice Lake
ASTM E617 Type: I & II
Equipment ID#: None
Condition: Good
Temperature: 18.6°C
Pressure: 743.9 mmHg
Relative Humidity: 34. %

Nominal Value	lb	Serial No.	Correction (mg)		NIST HB105-1 Class		Unc. (mg) (k=2)
			As Found	As Left	As Found	As Left	
25	lb	25121	-70.	-70.	F	F	75.
25	lb	25122	520.	520.	F	F	75.
25	lb	25123	410.	410.	F	F	75.
25	lb	25124	380.	380.	F	F	75.
25	lb	25125	410.	410.	F	F	75.
25	lb	25126	370.	370.	F	F	75.
25	lb	25127	110.	110.	F	F	75.
25	lb	25128	400.	400.	F	F	75.
25	lb	25129	1000.	-20.	F	F	75.
25	lb	25130	1010.	260.	F	F	75.
25	lb	25131	570.	570.	F	F	75.
25	lb	25132	710.	710.	F	F	75.
25	lb	25133	500.	500.	F	F	75.
25	lb	25134	800.	800.	F	F	75.
25	lb	25135	510.	510.	F	F	75.
25	lb	25136	880.	880.	F	F	75.
25	lb	25137	760.	760.	F	F	75.
25	lb	25138	1090.	-190.	*	F	75.
25	lb	25139	810.	810.	F	F	75.
25	lb	25140	1170.	-40.	*	F	75.
25	lb	25141	-70.	-70.	F	F	75.
25	lb	25142	100.	100.	F	F	75.
25	lb	25143	590.	590.	F	F	75.
25	lb	25144	300.	300.	F	F	75.
25	lb	25145	400.	400.	F	F	75.
25	lb	25146	880.	880.	F	F	75.
25	lb	25147	160.	160.	F	F	75.
25	lb	25148	240.	240.	F	F	75.
25	lb	25149	-50.	-50.	F	F	75.
25	lb	25150	930.	-30.	F	F	75.



Continued,

Receipt Date: January 28, 2015
Test Date: February 9, 2015
Report Date: February 11, 2015

State Test No.: 333719
Set Serial No.: 25121 - 25160
Barcode: 200651

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448
Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 11

Item(s) Submitted: Cast Hand Weights
Manufacturer: Rice Lake
ASTM E617 Type: I & II
Equipment ID#: None
Condition: Good
Temperature: 18.6°C
Pressure: 743.9 mmHg
Relative Humidity: 34. %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Unc. (mg) (k=2)
		As Found	As Left	As Found	As Left	
25 lb	25151	-180.	-180.	F	F	75.
25 lb	25152	750.	750.	F	F	75.
25 lb	25153	150.	150.	F	F	75.
25 lb	25154	700.	700.	F	F	75.
25 lb	25155	370.	370.	F	F	75.
25 lb	25156	1060.	240.	*	F	75.
25 lb	25157	1110.	-250.	*	F	75.
25 lb	25158	1130.	90.	*	F	75.
25 lb	25159	720.	720.	F	F	75.
25 lb	25160	1090.	-130.	*	F	75.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Pete Whebbe

Metrologist

Reviewed by:

Mark Nicollet

Quality Manager



Receipt Date: January 20, 2015
 Test Date: January 27, 2015
 Report Date: January 27, 2015

State Test No.: 333662
 Set Serial No.: A6
 Barcode: 200640

Calibration Report

KENNEDY SCALE, INC.
 11485 XEON STREET NW
 COON RAPIDS, MN 55448
 Contact: ROB BAKER
 Phone: 763-784-1022
 PO Number: NONE
 SOP: 12
 Technician ID: 09


Item(s) Submitted: 30 lb kit w/decimals & fractions
 Manufacturer: Rice Lake
 ASTM E617 Type: I & II
 Equipment ID#: None
 Condition: Good
 Temperature: 19.6°C
 Pressure: 740. mmHg
 Relative Humidity: 46. %


Kit contents:

Nominal Value	Quantity		
10 lb	2		
5 lb	1		
1 lb	5		
Decimals	7	From 0.01 lb	To 0.2 lb
Fractions	10	From 1/4 oz	To 4 oz

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These Standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculation methods conform to NIST Technical Note 1297. Actual uncertainties will be provided upon request. Results apply to items identified in this report only.

Heidi Jones

 Laboratory Administrator

Reviewed by:
 Mark Nicollet

 Quality Manager



Receipt Date: January 20, 2015
Test Date: January 24, 2015
Report Date: January 24, 2015

State Test No.: 333672
Set Serial No.: M6
Barcode: 200650

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448

Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 09

Item(s) Submitted: Metric weight set
Manufacturer: Rice Lake
ASTM E617 Type: I & II
Equipment ID#: None
Condition: Good
Temperature: 20.°C
Pressure: 726.4 mmHg
Relative Humidity: 47. %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Unc. (mg) (k=2)
		As Found	As Left	As Found	As Left	
5000	09	161.	161.	F	F	25.
2000	09	68.	68.	F	F	10.
2000	09	81.	81.	F	F	10.
1000	09	38.	38.	F	F	6.
500	09	25.2	25.2	F	F	4.6
200	09	7.	7.	F	F	1.
200	09	6.	6.	F	F	1.
100	09	3.18	3.18	F	F	0.45
50	09	4.14	4.14	F	F	0.25
20	09	1.67	1.67	F	F	0.25
20	09	0.65	0.65	F	F	0.25
10	09	0.67	0.67	F	F	0.12
5	09	0.88	0.88	F	F	0.1
2	09	0.48	0.48	F	F	0.07
2	09	0.50	0.50	F	F	0.07
1	09	0.50	0.50	F	F	0.07

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Heidi Jones
Heidi Jones
Laboratory Administrator

Reviewed by:
Mark Nicollet
Mark Nicollet
Quality Manager



Receipt Date: November 5, 2014
 Test Date: November 6, 2014
 Report Date: November 7, 2014

State Test No.: 333369
 Set Serial No.: NONE
 Bar Code: 013286

Calibration Report

KENNEDY SCALE, INC.
 11485 XEON STREET NW
 COON RAPIDS, MN 55448
 Contact: ROB BAKER
 Phone: 763-784-1022
 PO Number: NONE
 SOP: 12
 Technician ID: 09


Item(s) Submitted: Fuel substitution weights
 Manufacturer: Rice Lake
 ASTM E617 Type: II
 Equipment ID#: None
 Condition: Good
 Temperature: 20.8°C
 Pressure: 743.6 mmHg
 Relative Humidity: 45. %


Kit contents:

Nominal Value	Quantity
8 oz	14

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These Standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculation methods conform to NIST Technical Note 1297. Actual uncertainties will be provided upon request. Results apply to items identified in this report only.

Heidi Jones

 Laboratory Administrator

Reviewed by:
 Mark Nicollet

 Quality Manager

United States Department of Commerce National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Minnesota

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.



2015

Scope

Mass Echelon II

50 kg to 1 mg
1000 lb to 0.001 lb
4 oz to 0.03125 oz

Mass Echelon III

50 kg to 1 mg
5000 lb to 0.001 lb
4 oz to 0.03125 oz

Weight Carts

10 000 lb to 2000 lb

Wheel Load Weighers

20 000 lb to 2000 lb

Railroad Test Cars

110 000 lb to 80 000 lb

Volume Gravimetric, I

20 L to 1 mL
100 gal to 0.25 qt

Volume Transfer, II

1500 gal to 5 gal
100 gal to 25 gal LPG

A handwritten signature in blue ink that reads "Carol T. Hockert".

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2015-01-01 to 2015-12-31



SECRETARY OF STATE NORTH DAKOTA



[Home](#) | [Business Records Search](#)

KENNEDY SCALES, INC.

Corporation Details

System ID: 15125100 **Phone:** (763) 784-1022
Type: FOREIGN BUSINESS CORPORATION
Status: Active & Good Standing
Original File Date: 11/01/1999 **Effective Date:** 11/01/1999
State of Origin: Minnesota

Nature of Business

SERVICE & SALES OF WEIGHING EQUIPMENT

Principal Office

11485 XEON ST NW COON RAPIDS, MN 55448-3150

Registered Agent

SEARCH COMPANY OF NORTH DAKOTA LLC
1501 N 12TH ST STE 1
BISMARCK, ND 58501-2713
Established Date: Jan 24, 2014

Generate an Annual Report To File


To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2014](#) [2015](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)

[Disclaimer](#) | [Privacy Policy](#)

 We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

 Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.