



**APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY**  
 Public Service Commission  
 SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

<b>Name of Company</b> Rocky Mtn Scale Works	<b>Email Address</b> Dylan@scals911.com	<b>Application Date</b> 5-27-14	
<b>Mailing Address</b> 6360 US Hwy 12 W PO Box 1056	<b>City</b> Lolo	<b>State</b> MT	<b>Zip Code</b> 59847-1056
<b>Telephone Number</b> 406-543-5181	<b>Cell Phone Number</b> 406-239-0228	<b>Fax Number</b> 406-543-5182	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input checked="" type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
e.g. 1001	e.g. John Doe	e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6
1724	Dylan Moses	Scales - 2

Continued on Page 2



List below all field standards (attach current calibration reports):

See attached	


Additional Application Items (initial where appropriate):

Standardized Test Report	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.  
 Yes     No

I am Dylan Moses, and have authority to represent this company.  
By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.

  
Signature

Send Completed Application and Related Documents To:

Public Service Commission  
600 E Boulevard Ave Dept 408  
Bismarck ND 58505-0480  
Telephone: (701) 328-2400  
Fax: (701) 328-2410



# SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

## ROCKY MOUNTAIN SCALE WORKS

### Trade Name

**System ID:** 33561000      **Phone:** (406) 543-5181  
**Status:** Active  
**Original File Date:** 02/05/2013      **Last Renewal Date:**

### Nature of Business

SCALE SALES AND SERVICE

### Owners

**RATHER ENTERPRISES, INC.**  
6360 HWY 12 W  
PO BOX 1056  
LOLO, MT 59847-1056

[Return to Search Results](#)[Contact Us](#)[Disclaimer](#) | [Privacy Policy](#)

 We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

 Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.

North Dakota

nd.gov Official Portal for  
North Dakota State GovernmentNorth Dakota  
LEGENDARY

# SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

## RATHER ENTERPRISES, INC.

### Corporation Details

**System ID:** 33560900**Phone:** (800) 700-5048**Type:** FOREIGN BUSINESS CORPORATION**Status:** Active & Good Standing**Original File Date:** 02/05/2013**Effective Date:** 02/05/2013**State of Origin:** Montana

### Nature of Business

SCALE SALES AND SERVICE

### Principal Office

6360 HWY 12 W PO BOX 1056 LOLO, MT 59847-1056

### Registered Agent

**ENVIRONMENTAL MATERIALS INC.**

4506 18TH AVE W

WILLISTON, ND 58801-

Established Date: Feb 05, 2013

### Generate an Annual Report To File


To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2014](#) (generates a forms-fillable pdf in a new pop-up window)[Return to Search Results](#)[Contact Us](#)[Disclaimer](#) | [Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



**SOUTH DAKOTA DEPARTMENT OF PUBLIC SAFETY**

**Office of Weights and Measures  
Metrology Lab**

Lab: 1500 N Garfield – E. Truck Bypass Phone: 605-773-3170  
Office: 118 West Capitol Avenue Phone: 605-773-3697

**REPORT OF CALIBRATION**

**LAB TEST NUMBER: MP3324**  
**DATE OF REPORT: 05/06/2014**  
**DATE RECEIVED: 05/05/2014**  
**DATE OF TEST: 05/06/2014**



**Submitted By:** ROCKY MOUNTAIN SCALE WORKS  
**Contact:** Todd Rather  
**Mailing Address:** 5535 Interstate Ave  
**City, State, Zip:** Billings MT, 59101  
**Phone:** 800-700-5048  
**S/A Number:**

**Standards Submitted:**

- 1 -WEIGHTS CARTS
- 16 -1000 LB TEST WEIGHTS
- 500 LB TEST WEIGHTS
- 50 LB TEST WEIGHTS
- 25 LB TEST WEIGHTS
- AVOIRDUPOIS WEIGHT KITS
- METRIC WEIGHT KITS
- 5 GALLON TEST MEASURES

**Uncertainty Statement:** The combined standard uncertainty includes the standard uncertainty reported for the standard, the standard uncertainty for the measurement process, and a component of uncertainty to account for any observed deviations from values that are less than surveillance limits. The combined standard uncertainty is multiplied by a coverage factor of  $k = 2$  to provide an expanded uncertainty, which defines an interval having a level of confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the ISO/IEC Guide to the Expression of Uncertainty in Measurement. The expanded uncertainty is not to be confused with a tolerance limit for the user during application. All established Uncertainties are less than 1/3 applicable Class "F" tolerances.

**Traceability statement:**


The Standards of the SD Metrology Laboratory are traceable to the International System of Units (SI) through the National Institute of Standards and Technology recognized/traceable lab in the State of Minnesota, (Test Reports: 327678, 327681, 327682, 327683, 327684, 327685, 327686, 327687, 327813) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

The artifacts submitted for calibration have been examined by the State of South Dakota and found to be appropriate for the intended use and to be accurate within Class "F" Tolerances as established by the National Institute of Standards and Technology-Weights and Measures Division. Test methods are in accordance with NIST Handbook 145 and NIST IR 6969.

This document does not represent or imply endorsement by NIST Office of Weights and Measures, NMI, or any agency of the State and/or national governments. The reported test values relate only to the observations made at the time and conditions of the test. This report may not be reproduced, except in full without the written approval of this laboratory. The client must not use this document to claim product endorsement by this laboratory.

  
**Ron Peterson, Metrologist**



  
**Date**

**SOUTH DAKOTA WEIGHTS AND MEASURES / METROLOGY LAB**

Lab: 1500 N. Garfield-E. Truck Bypass Phone: 605-773-3170  
 Office: 118 West Capitol Avenue Phone: 605-773-3697  
 Pierre, SD 57501




<b>Submitted by:</b>	ROCKY MOUNTAIN SCALE WORKS	<b>Report Number:</b>	MP3324
<b>Mailing Address:</b>	5535 Interstate Ave	<b>Date Received:</b>	05/05/14
<b>City, State, Zip:</b>	Billings MT, 59101	<b>Date tested:</b>	05/06/14
<b>Manufacturer:</b>	B-TEK Scales	<b>Condition of Cart:</b>	GOOD
<b>Serial Number:</b>	16278B	<b>Temperature (c):</b>	21.4
<b>Test Method Used</b>	SOP 33/ Double Sub.	<b>Humidity:</b>	43.1%
<b>Nominal (lb):</b>	4500	<b>Pressure (mm/Hg):</b>	706.7
<b>Tolerance (lb):</b>	1.25		

*The values reported below relate only to those observations made at the time and conditions of the test. This test report, so numbered, may not be reproduced, except in full, without approval of the laboratory.*

As Found (lb)	As Left (lb)	Uncertainty-lb. (K=2)
0.27	0.27	0.15

The weight cart was cleaned and painted (if needed) and allowed to come to environmental equilibrium in the laboratory prior to calibration. The weight cart was adjusted, as needed and noted above, as close as possible to zero error. All fluid levels were adjusted as close as possible to the full/reference marks. Liquid levels must be maintained as close to reference levels as possible during use. Any maintenance, repairs or damage to weight cart or its components will likely result in an out-of-tolerance condition; therefore, maintenance or replacement of components such as batteries, tires, filters, etc. will require calibration of the weight cart prior to subsequent use.

*The above weight cart was compared with standards of the State of South Dakota, which are traceable the National Institute of Standards and Technology(NIST) Weights and Measures Division and have known values. The assigned test number provides documented evidence for measurement traceability*

  
 Ron Peterson, Metrologist

05/06/2014  
 Date of Report

Office of Weights and Measures  
 118 W. Capitol Ave.  
 Pierre, SD 57501

Phone: 605-773-3697  
 Fax: 605-773-6631  
 www.dps.sd.gov

**SOUTH DAKOTA WEIGHTS AND MEASURES / METROLOGY LAB**

Lab: 1500 N. Garfield-E. Truck Bypass Phone: 605-773-3170

Office: 118 West Capitol Avenue Phone: 605-773-3697

Pierre, SD 57501



<b>Submitted by:</b>	ROCKY MOUNTAIN SCALE WORKS	<b>Report Number:</b>	MP3324
<b>Mailing Address:</b>	5535 Interstate Ave	<b>Date Received:</b>	05/05/14
<b>City, State, Zip:</b>	Billings MT, 59101	<b>Date tested:</b>	05/06/14
<b>Artifacts Submitted</b>	1000 lb Bulk weights	<b>Condition of Weights:</b>	GOOD
<b>Manufacturer:</b>	NA	<b>Temperature (c):</b>	21.9
<b>Test Method Used:</b>	SOP 8/ MODIFIED SUB	<b>Humidity:</b>	42.8%
<b>Equipment Used:</b>	Russell Balance/ Vaisala PTU301	<b>Pressure (mm/Hg):</b>	706.8

Treatment of artifacts prior to testing: Thermal equilibrium time/conditions were obtained by placing the artifacts in the lab overnight.

Compliance Statement: These weights and associated uncertainties were evaluated against NIST Handbook 105-1 NIST Class F tolerances and the weights were within tolerance at the time of calibration.

Standards Used: SD Lab 1000 Lb Working Standard.

The values reported below relate only to those observations made at the time and conditions of the test. This test report, so numbered, may not be reproduced, except in full, without approval of the laboratory

Nominal Value	Serial Number	Tolerance=0.10 lb	Uncertainty (lb)= 0.025 (K=2.09)
		As Received lb	As Left lb
1000 lb	07	0.010	0.010
1000 lb	08	0.019	0.019
1000 lb	4N87	0.042	0.042
1000 lb	4N8W	0.048	0.048
1000 lb	4N8X	0.022	0.022
1000 lb	4N8Y	0.094	-0.008
1000 lb	4N90	0.084	-0.002
1000 lb	4N91	0.053	0.053
1000 lb	54EK	0.055	0.055
1000 lb	54EL	0.057	0.057
1000 lb	54EM	0.048	0.048
1000 lb	54EN	0.025	0.025
1000 lb	54EP	0.030	0.030
1000 lb	54EQ	0.040	0.040
1000 lb	54ER	0.027	0.027
1000 lb	54ES	0.042	0.042

END OF REPORT

  
 Ron Peterson, Metrologist

05/06/2014  
 Date of Report

# United States Department of Commerce

## National Institute of Standards and Technology

Certificate of Metrological Traceability For:

# South Dakota

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

### *Scope*

#### **Mass Echelon III**

30 kg to 1 mg

1000 lb to 0.001 lb

8 oz to 0.03125 oz

Weight Carts

5000 lb to 2000 lb

#### **Volume Transfer, II**

5 gal



2014

A handwritten signature in blue ink that reads "Carol T. Hockert".

Carol T. Hockert, Chief  
NIST Office of Weights and Measures

Effective Dates: 2014-01-01 to 2014-12-31



ALDINGER CO  
1440 PRUDENTIAL  
DALLAS, TX 75235

ROCKY MOUNTAIN SCALE WORKS  
5535 INTERSTATE AVE, UNIT 1  
BILLINGS, MT 59101

I.D.: 11141301  
Description: WEIGHT SET  
Manufacturer: UNKNOWN  
Gage Type: WEIGHT SET  
Temp./RH: 20.6 C / 22 %  
Cal Date: 11/14/2013

Serial Number: 11141301  
Model Number: 50LB TO 50LB CLASS F  
Performed By: 671  
As Found Condition: In Tolerance  
As Left Condition: In Tolerance  
Cal. Due Date: 11/14/2014

Weights painted prior to As Left data.

50lb	F	208	7.2	49.9949000	50.0051000	50.0005740	50.0005740	0.0005100	lb
50lb	F	39	7.2	49.9949000	50.0051000	49.9990310	49.9990310	0.0005100	lb
50lb	F	2010	7.2	49.9949000	50.0051000	50.0010150	50.0010150	0.0005100	lb
50lb	F	38	7.2	49.9949000	50.0051000	50.0005740	50.0005740	0.0005100	lb
50lb	F	1021	7.2	49.9949000	50.0051000	50.0005740	50.0005740	0.0005100	lb
50lb	F	1022	7.2	49.9949000	50.0051000	50.0001330	50.0001330	0.0005100	lb
50lb	F	36	7.2	49.9949000	50.0051000	50.0010150	50.0010150	0.0005100	lb
50lb	F	43	7.2	49.9949000	50.0051000	50.0005740	50.0005740	0.0005100	lb
50lb	F	40	7.2	49.9949000	50.0051000	50.0007940	50.0007940	0.0005100	lb
50lb	F	1005	7.2	49.9949000	50.0051000	50.0032190	50.0032190	0.0005100	lb
50lb	F	1062	7.2	49.9949000	50.0051000	50.0018970	50.0018970	0.0005100	lb
50lb	F	31	7.2	49.9949000	50.0051000	49.9996920	49.9996920	0.0005100	lb
50lb	F	205	7.2	49.9949000	50.0051000	50.0014560	50.0014560	0.0005100	lb
50lb	F	2055	7.2	49.9949000	50.0051000	50.0010150	50.0010150	0.0005100	lb
50lb	F	270	7.2	49.9949000	50.0051000	50.0005740	50.0005740	0.0005100	lb
50lb	F	20	7.2	49.9949000	50.0051000	50.0010150	50.0010150	0.0005100	lb
50lb	F	1088	7.2	49.9949000	50.0051000	50.0021170	50.0021170	0.0005100	lb
50lb	F	1020	7.2	49.9949000	50.0051000	49.9992510	49.9992510	0.0005100	lb
50lb	F	1008	7.2	49.9949000	50.0051000	50.0005740	50.0005740	0.0005100	lb
50lb	F	218	7.2	49.9949000	50.0051000	50.0007940	50.0007940	0.0005100	lb

\*In "As Left" column "A" denotes an adjusted test point and "F" denotes a failed test point. Blank denotes an acceptable value.

TROEMNER	52001	WEIGHT - 50 LB	2/15/2013	2/28/2014	241738
VWR	671T	THERMOHYGROMETER	11/6/2012	11/30/2014	236991
ALDINGER CO	PPCA501-A	MASS	6	8/22/2013	

Technician Signature Eric Rogers

Calibration values are based on conventional mass defined by OIML IR 33 as the mass that the item would have weighed in air under conventional conditions, (Temp 20C; Air 1.2 mg/cm<sup>3</sup>) vs a reference density of 8.0 g/cm<sup>3</sup>. Magnetism met the requirements of ASTM E 617-97 except where otherwise noted. The calibration interval has been specified by the customer. Aldinger Co. certifies that the instrument listed above has been tested, calibrated (if necessary), and meets the criteria established in the associated test procedure unless otherwise noted. The standards used are traceable to the National Institute of Standards and Technology (NIST). Aldinger Co. calibration and control system meets the general requirements for the competence of calibration and testing laboratories (ISO/IEC 17025-2005 & ANSI/NCSL Z540-1-1994). The uncertainty calculation includes the UUT. In tolerance conditions are based on test results falling with specified limits with no reduction by the uncertainty of the measurement. This uncertainty represents an expanded uncertainty expressed at approximately 95% confidence level using a coverage factor of k=2. This report shall not be reproduced, except in full, without the written approval of Aldinger Co. \*The uncertainty value is listed in the same unit as the test point.

User Login

Contact Us

Call Us. (888)822-1299 toll free

Calibration

Repair

Sales

Force & Torque

Resources



View, Email & Print  
**Online Instrument Tracking**


**Shop Online**

**Need a Quote? Click Here**



**ISO 17025** accredited and backed by over **230 years** of combined experience in the weighing and measuring industry, our factory trained and highly **knowledgeable technicians** can do more than just calibrate nearly all of your equipment - they can help you **avoid downtime** by leaning on experience to identify and correct potential problems.

**Learn More**



- Calibration
- Repair
- Products
- Systems
- Rentals

**Join Our Newsletter**

Sign up now to receive important industry updates, special offers, and discounts.

Email Address:

[Join Now](#)

Check it out!

**Green Initiative**



**What type of company are you looking for?**

**Aldinger is an accredited** [calibration](#), certification and [repair](#) center for test and measurement instrumentation. Imagine a business partner that you can rely on to not only keep your equipment working and accurate, but to go above and beyond to help your company make consistently better products. Aldinger even offers GPS dispatched 24/7 emergency repair service to our business partners - because you never plan a breakdown.

**In addition** to being the leader in these services Aldinger is also accredited by A2LA to ISO/IEC 17025 standards. Couple a huge scope of calibration capabilities with 40 years of business and technical experience you will understand why we are a primary top tier supplier of calibration and repair services.

**As well as** calibration and repair services we also [sell](#) scales, balances and custom weighing systems. We provide counting scale [rental equipment](#) for inventory, saving you time and money. In addition we can test overhead lifting equipment with proof load testing that will keep you OSHA compliant.

**How can we help** your company? [Contact us](#) today and let us know.





SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005  
& ANSI/NCSL Z540-1-1994

ALDINGER COMPANY  
1440 Prudential Drive  
Dallas, TX 75235  
Timothy R. Detten Phone: 214 638 1808

CALIBRATION

Valid To: June 30, 2014

Certificate Number: 1509.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2,4</sup> ( $\pm$ )	Comments
Calipers <sup>3,5</sup>	Up to 6 in Up to 12 in Up to 18 in Up to 24 in	0.6R 0.6R 0.6R 0.6R	Gage blocks
Dial and Test Indicators <sup>3,5</sup>	Up to 3 in	0.6R	Gage blocks
Gage Blocks	Up to 1 in (2 to 4) in	(2 + 1.5L) $\mu$ in (3 + 1.5L) $\mu$ in	Master gage block set
Height Gages <sup>3,5</sup>	Up to 48 in	0.6R	Gage blocks
Length Standards	Up to 18 in Up to 48 in	(37 + 1.5L) $\mu$ in (65 + 1.5L) $\mu$ in	ULM, gage blocks SuperMic, gage blocks
Micrometers <sup>3,5</sup>	Up to 36 in	0.6R	Gage blocks grade 2

Parameter/Equipment	Range	CMC <sup>2,4</sup> ( $\pm$ )	Comments
Plain Ring Gages	(0.125 to 4) in	(22 + 1.5L) $\mu$ in	ULM, gage blocks
Threads – Pitch Diameter Major Diameter	(0.040 to 4) in (0.040 to 4) in	(69 + 1.5L) $\mu$ in (18 + 1.5L) $\mu$ in	Three-wire method Direct measure
Wire Gages – TPI 60 Degree	(4 to 80) TPI	(11 + 1.5L) $\mu$ in	VK master thread wire set

## II. Electrical – DC/Low Frequency

Range/Equipment	Frequency	CMC <sup>2,6,7,9</sup> ( $\pm$ )	Comments
AC Current – Measure 1 A 3 A	60 Hz to 1 kHz 60 Hz to 1 kHz	0.1 % + 0.4 mA 0.1 % + 1.8 mA	HP 34401
AC Current – Generate (32 to 320) $\mu$ A 320 $\mu$ A to 3.2 mA (3.2 to 32) mA (32 to 320) mA 320 mA to 3.2 A (3.2 to 10) A  (3.2 to 32) A  (32 to 200) A  (16 to 160) A  (160 to 1000) A	60 Hz to 3 kHz 60 Hz to 3 kHz 60 Hz to 3 kHz 60 Hz to 3 kHz 60 Hz to 3 kHz 60 Hz to 3 kHz  (60 to 100) Hz (100 to 440) Hz  (60 to 100) Hz (100 to 440) Hz  (60 to 100) Hz  (60 to 100) Hz	0.072 % + 0.9 $\mu$ A 0.07 % + 0.6 $\mu$ A 0.07 % + 3.2 $\mu$ A 0.083 % + 32 $\mu$ A 0.1 % + 480 $\mu$ A 0.21 % + 3 mA  0.2 % + 5.5 mA 0.78 % + 27 mA  0.21 % + 90 mA 0.67 % + 0.25 A  0.21 % + 28 mA  0.21 % + 0.45 A	Wavetek 9100       using 10 turn coil       using 50 turn coil

*Peter Noyce*

Range/Equipment	Frequency	CMC <sup>2,6,7,9</sup> (±)	Comments
AC Voltage – Generate			
(32 to 320) mV	60 Hz to 3 kHz	0.041 % + 20 μV	Wavetek 9100
320 mV to 3.2 V	60 Hz to 3 kHz	0.041 % + 200 μV	
(3.2 to 32) V	60 Hz to 3 kHz	0.041 % + 2.0 mV	
(32 to 105) V	60 Hz to 3 kHz	0.043 % + 6.3 mV	
(105 to 320) V	60 Hz to 1 kHz (1 to 3) kHz	0.066 % + 20 mV 0.09 % + 20 mV	
(320 to 800) V	(60 to 100) Hz (1 to 3) kHz	0.079 % + 63 mV 0.1 % + 63 mV	
AC Voltage – Measure			
100 mV	60 Hz to 3 kHz	0.065 % + 40 μV	HP 34401
1 V	60 Hz to 3 kHz	0.064 % + 300 μV	
10 V	60 Hz to 3 kHz	0.064 % + 3 mV	
100 V	60 Hz to 3 kHz	0.064 % + 30 mV	
750 V	60 Hz to 3 kHz	0.064 % + 230 mV	

Parameter/Equipment	Range	CMC <sup>2,6,9</sup> (±)	Comments
Capacitance – Generate			
Low	(0.5 to 4) nF (4 to 40) nF (40 to 400) nF 400 nF to 4 μF	0.61 % + 15 pF 0.6 % + 60 pF 0.61 % + 320 pF 0.8 % + 3.2 nF	Wavetek 9100
High	(4 to 40) μF (40 to 400) μF 400 μF to 4 mF (4 to 40) mF	1 % + 32 nF 1 % + 320 nF 1 % + 3.2 μF 2 % + 120 μF	

*Peter Nhyu*

Parameter/Equipment	Range	CMC <sup>2,6,7,9</sup> (±)	Comments
DC Current – Generate	(0 to 320) $\mu$ A 320 $\mu$ A to 3.2 mA (3.2 to 32) mA (32 to 320) mA 320 mA to 3.2 A (3.2 to 11) A  (3.2 to 32) A (32 to 100) A (100 to 160) A  (160 to 520) A (520 to 1000) A	0.014 % + 11 nA 0.014 % + 83 nA 0.014 % + 900 nA 0.016 % + 9.6 $\mu$ A 0.06 % + 120 $\mu$ A 0.057 % + 940 $\mu$ A  0.19 % + 1.2 mA 0.072 % + 9.4 mA 0.068 % + 45 mA  0.056 % + 47 mA 0.055 % + 230 mA	Wavetek 9100     using 10 turn coil   using 50 turn coil
DC Current – Measure, Fixed Points	10 mA 100 mA 1 A 3 A	0.076 % + 6 $\mu$ A 0.05 % + 5 $\mu$ A 0.1 % + 100 $\mu$ A 0.14 % + 600 $\mu$ A	HP 34401
DC Voltage – Generate	(0 to 320) mV 320 mV to 3.2 V (3.2 to 32) V (32 to 320) V (320 to 1050) V	0.0063 % + 4.2 $\mu$ V 0.0062 % + 42 $\mu$ V 0.0066 % + 420 $\mu$ V 0.0069 % + 4.5 mV 0.0084 % + 20 mV	Wavetek 9100
DC Voltage – Measure, Fixed Points	100 mV 1 V 10 V 100 V 1000 V	0.0055 % + 3.5 $\mu$ V 0.0042 % + 7 $\mu$ V 0.0036 % + 50 $\mu$ V 0.0047 % + 600 $\mu$ V 0.0047 % + 10 mV	HP 34401
Electrical Calibration of Temperature Controllers <sup>3,5</sup>	(-200 to 1371) $^{\circ}$ C	0.56 $^{\circ}$ C	Omega CL27
Resistance – Generate	(0 to 40) $\Omega$ (40 to 400) $\Omega$ (0.4 to 4) k $\Omega$ (4 to 40) k $\Omega$ (40 to 400) k $\Omega$ (0.4 to 4) M $\Omega$ (4 to 40) M $\Omega$ (40 to 400) M $\Omega$	0.11 % + 50 m $\Omega$ 0.051 % + 100 m $\Omega$ 0.037 % + 200 m $\Omega$ 0.052 % + 2 $\Omega$ 0.053 % + 20 $\Omega$ 0.049 % + 200 $\Omega$ 0.15 % + 2 k $\Omega$ 0.14 % + 40 k $\Omega$	Wavetek 9100

*Peter Noyes*

Parameter/Equipment	Range	CMC <sup>2,6,7,9</sup> (±)	Comments
Resistance – Measure, Fixed Points	100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ 10 MΩ 100 MΩ	0.01 % + 4 mΩ 0.01 % + 10 mΩ 0.01 % + 100 mΩ 0.01 % + 1 Ω 0.01 % + 10 Ω 0.04 % + 100 Ω 0.8 % + 10 kΩ	HP 34401
RTD Simulation –  PT385, 100 Ω	(-200 to -100) °C (-100 to 100) °C (100 to 630) °C (630 to 850) °C	0.26 °C 0.17 °C 0.35 °C 0.53 °C	Wavetek 9100
Thermocouple Simulation –  Type E  Type J  Type K  Type T	(-250 to -200) °C (-200 to -100) °C (-100 to 100) °C (100 to 1000) °C  (-210 to -100) °C (-100 to 800) °C (800 to 1000) °C (1000 to 1200) °C  (-250 to -200) °C (-200 to -100) °C (-100 to 100) °C (100 to 600) °C (600 to 1372) °C  (-250 to -200) °C (-200 to -100) °C (-100 to 0) °C (0 to 400) °C	0.45 °C 0.23 °C 0.18 °C 0.22 °C  0.26 °C 0.20 °C 0.22 °C 0.24 °C  0.57 °C 0.28 °C 0.20 °C 0.24 °C 0.28 °C  0.59 °C 0.28 °C 0.23 °C 0.18 °C	Wavetek 9100

*Peter Noyes*

III. Mechanical

Parameter/Equipment	Range <sup>8</sup>	CMC <sup>2</sup> (±)	Comments
Balances <sup>3,5</sup>	(0 to 20) g (0 to 200) g  (0 to 200) g (0 to 1000) g (0 to 5000) g  (0 to 20) g (0 to 200) g (0 to 1000) g (0 to 5000) g (0 to 10 000) g (0 to 20 000) g	0.11 mg 0.69 mg  1.3 mg 6.6 mg 34 mg  0.81 mg 4.8 mg 24 mg 130 mg 370 mg 460 mg	Handbook 44 with: class 1 weights  class 2 weights  class 4 weights
Force – Universal Testing Machines <sup>3,5</sup>  Tension/Compression  Compression      Tension	(0 to 100) lbf  (100 to 1000) lbf (1000 to 10 000) lbf (10 000 to 50 000) lbf (10 000 to 100 000) lbf (50 000 to 500 000) lbf  (100 to 1000) lbf (1000 to 10 000) lbf (10 000 to 50 000) lbf	0.0064 lb  0.26 % Indication 0.24 % Indication 0.33 % Indication 0.21 % Indication 0.21 % Indication  0.23 % Indication 0.23 % Indication 0.29 % Indication	Class F weights ASTM E4 w/ load cells
Mass	1 mg 2 mg 5 mg 10 mg 20 mg 30 mg 50 mg 100 mg 200 mg 300 mg 500 mg 1 g 2 g 5 g 10 g	1.1 µg 1.1 µg 1.1 µg 3.7 µg 1.3 µg 1.1 µg 1.8 µg 1.7 µg 1.5 µg 1.4 µg 1.8 µg 9.4 µg 6.4 µg 11 µg 16 µg	Modified double substitution

*Peter Nkyu*

Parameter/Equipment	Range <sup>8</sup>	CMC <sup>2</sup> (±)	Comments
Mass (cont)	20 g 50 g 100 g 200 g 500 g	16 µg 25 µg 44 µg 120 µg 460 µg	Modified double substitution
	1 kg 2 kg 5 kg 10 kg 20 kg  0.8859 g (1/32 oz) 1.772 g (1/16 oz) 3.544 g (1/8 oz) 7.087 g (1/4 oz) 14.17 g (1/2 oz) 28.35 g (1 oz) 56.7 g (2 oz) 113.4 g (4 oz) 226.8 g (8 oz)  0.4536 g (0.001 lb) 0.9072 g (0.002 lb) 2.27 g (0.005 lb) 4.54 g (0.01 lb) 9.07 g (0.02 lb) 22.68 g (0.05 lb) 45.36 g (0.1 lb) 90.72 g (0.2 lb)  453.6 g (1 lb) 907.2 g (2 lb) 2267.96 g (5 lb) 4535.9 g (10 lb) 9071.8 g (20 lb) 22679.62 g (50 lb) 45359.237 g (100 lb) 90718.474 g (200 lb) 226796.185g (500 lb) 453592.37g (1000 lb)	1.2 mg 1.4 mg 1.9 mg 94 mg 94 mg  12 µg 94 µg 240 µg 170 µg 170 µg 370 µg 150 µg 740 µg 5.1 mg  130 µg 56 µg 26 µg 16 µg 56 µg 79 µg 37 µg 87 µg  3.8 mg 7.9 mg 8.3 mg 7.4 mg 94 mg 95 mg 0.43 g 0.43 g 0.45 g 0.47 g	Direct comparison
Pipettes <sup>3,5</sup>	(1 to 50) µL (50 to 100) µL (100 to 500) µL (500 to 1000) µL (1000 to 5000) µL	0.31 µL 0.31 µL 0.33 µL 0.39 µL 0.38 µL	Gravimetric method

*Peter Nkyu*

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Optical Tachometer <sup>3,5</sup>	200 FPM 3000 FPM 29 999 FPM	1.3 FPM 1.7 FPM 3.9 FPM	Calibrated strobe
Pressure Gauges, Transducers and Transmitters <sup>3,5</sup>	(0 to 5000) psig (0 to 300) psig  (0 to 100) psig (0 to 500) psig	3.9 psig 0.18 psig  0.03 psig 0.19 psig	Druck transducer  Transmation 195
Scales <sup>3,5</sup>	(0 to 10) lb (0 to 20) lb  (0 to 50) lb (0 to 100) lb (0 to 200) lb (0 to 500) lb (0 to 1000) lb (0 to 5000) lb (0 to 20 000) lb	0.000 25 lb 0.000 25 lb  0.006 3 lb 0.006 3 lb 0.006 4 lb 0.006 4 lb 0.019 lb 0.019 lb 0.021 lb	Handbook 44 with: class 4 weights  class F weights
Torque Wrenches and Transducers <sup>3,5</sup>	(5 to 50) in·lb (30 to 400) in·lb (80 to 1000) in·lb (240 to 3000) in·lb	0.17 in·lb 1.8 in·lb 3.9 in·lb 21 in·lb	CDI torque calibrator
Vacuum Gauges <sup>3,5</sup>	(0 to 15) psig (0 to 15) psig	0.03 psig 0.03 psig	Transmation 195 Druck DPI 610

*Peter Noyes*

IV. Thermodynamic

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Temperature – Measuring Equipment <sup>3,5</sup>	(30 to 300) °C (86 to 572) °F	0.090 °C 0.16 °F	Hart 1502A with 5628 PRT in dry block
	(-20 to 200) °C (-4 to 392) °F	0.061 °C 0.11 °F	Hart 1502A/5628 PRT in wet bath
Temperature – Measure <sup>3,5</sup>	(20 to 260) °C (68 to 500) °F	0.41 °C 0.74 °F	Fluke 51 with K type thermocouple

V. Time & Frequency

Parameter/Equipment	Range	CMC <sup>2,7,9</sup> (±)	Comments
Frequency – Measure	40 Hz to 300 kHz	0.014 % Indication	HP 34401
Timers and Stopwatches <sup>3,5</sup>	≤ 24 hr	0.14 s / 24 hr	Reference stopwatch

<sup>1</sup> This laboratory offers commercial calibration service and field calibration service.

<sup>2</sup> Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.



- <sup>4</sup> In the statement of CMC,  $L$  is the numerical value of the nominal length of the device measured in inches and  $R$  is the numerical value of the resolution of the device in microinches unless otherwise noted.
- <sup>5</sup> The CMC stated for calibrations performed in the laboratory is applicable for calibrations performed in the field.
- <sup>6</sup> CMC for the Wavetek 9100A is based on 1-year specifications within a temperature range of  $23\text{ }^{\circ}\text{C} \pm 5\text{ }^{\circ}\text{C}$ . Field calibrations will be performed within  $23\text{ }^{\circ}\text{C} \pm 5\text{ }^{\circ}\text{C}$  either in the customer-supplied area or in Aldinger's Mobile Lab.
- <sup>7</sup> CMC for the HP 34401 is based on 1-year specifications within a temperature range of  $18\text{ }^{\circ}\text{C}$  to  $28\text{ }^{\circ}\text{C}$ . Field calibrations will be performed within  $18\text{ }^{\circ}\text{C}$  to  $28\text{ }^{\circ}\text{C}$ , 30 % to 55 % humidity either in the customer-supplied area or in Aldinger's Mobile Lab.
- <sup>8</sup> Where ranges are not specified, the CMC stated is for the cardinal points only.
- <sup>9</sup> In the statement of CMC, percentage (%) refers to percent of reading, unless otherwise noted.

*Peter Ahya*



The American Association for Laboratory Accreditation

World Class Accreditation

# Accredited Laboratory

A2LA has accredited

## ALDINGER COMPANY

*Dallas, TX*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *G the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/N additional program requirements in the field of calibration. This accreditation demonstrates technical competence operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated*

Presented this 16<sup>th</sup> day of July 2012.



  
\_\_\_\_\_  
President & CEO  
For the Accreditation Council  
Certificate Number 1509.01  
Valid to June 30, 2014

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Ac*