



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY

Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company Seneca Companies	Email Address khankins@senecaco.com	Application Date 9-10-14	
Mailing Address 4140 E. 14 th St.	City Des Moines	State IA	Zip Code 50313
Telephone Number 515-262-5000	Cell Phone Number	Fax Number 816-761-8351	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input checked="" type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input checked="" type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
e.g. 1001	e.g. John Doe	e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6
1747	William Rogers	Liquid - 1, 2

Continued on Page 2



List below all field standards (attach current calibration reports):

5 gallon Seraphin # 12-90054	
100 gallon prover # 11-539 49-01	
100 gallon prover # 6871998	

Additional Application Items (initial where appropriate):

Standardized Test Report	<input type="checkbox"/> Copy enclosed
	<input type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed
	<input type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed
	<input type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:


Include my company information on your registered service company list for public contact.
 Yes No

I am Keri Hankins, and have authority to represent this company.
 By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.

Keri Hankins
Signature

Send Completed Application and Related Documents To:

Public Service Commission
 600 E Boulevard Ave Dept 408
 Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410



SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

SENECA COMPANIES, INC.

Corporation Details

System ID: 26034400**Phone:** (515) 264-4330**Type:** FOREIGN BUSINESS CORPORATION**Status:** Active & Good Standing**Original File Date:** 01/17/2012**Effective Date:** 01/17/2012**State of Origin:** Iowa

Nature of Business

GENERAL CONTRACTING SERVICES, PETROLEUM EQUIPMENT SALES

Principal Office

4140 E 14TH ST DES MOINES, IA 50313-3804

Registered Agent

C T CORPORATION SYSTEM

314 E THAYER AVE


BISMARCK, ND 58501-4018

Established Date: Jan 17, 2012

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2014](#) (generates a forms-fillable pdf in a new pop-up window)[Return to Search Results](#)[Contact Us](#)[Disclaimer](#)[Privacy Policy](#)

 We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

 Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



**SOUTH DAKOTA
DEPARTMENT OF PUBLIC SAFETY**

**Office of Weights and Measures
Metrology Lab**

Lab: 1500 N Garfield – E. Truck Bypass Phone: 605-773-3170
Office: 118 West Capitol Avenue Phone: 605-773-3697



REPORT OF CALIBRATION

LAB TEST NUMBER: MP3334
DATE OF REPORT: 09/17/2014
DATE RECEIVED: 09/16/2014
DATE OF TEST: 09/17/2014

Submitted By: Seneca Companies
Contact: Accounts Billable
Mailing Address: 4140 E, 14th St.
City, State, Zip: Des Moines, IA 50313
Phone: 515-337-3402
S/A Number:

Standards Submitted:

2 -5 GALLON TEST MEASURES

Uncertainty Statement: The combined standard uncertainty includes the standard uncertainty reported for the standard, the standard uncertainty for the measurement process, and a component of uncertainty to account for any observed deviations from values that are less than surveillance limits. The combined standard uncertainty is multiplied by a coverage factor k to provide an expanded uncertainty, which defines an interval having a level of confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the ISO/IEC Guide to the Expression of Uncertainty in Measurement. The expanded uncertainty is not to be confused with a tolerance limit for the user during application. All established Uncertainties are less than 1/3 applicable Class "F" tolerances.

Traceability statement:


The Standards of the SD Metrology Laboratory are traceable to the International System of Units (SI) through the National Institute of Standards and Technology. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

The artifacts submitted for calibration have been examined by the State of South Dakota and found to be appropriate for the intended use and to be accurate within Class "F" Tolerances as established by the National Institute of Standards and Technology-Weights and Measures Division. Test methods are in accordance with NIST Handbook 145 and NIST IR 6969 and/or NIST IR 7383.

This document does not represent or imply endorsement by NIST Office of Weights and Measures, NMI, or any agency of the State and/or national governments. The reported test values relate only to the observations made at the time and conditions of the test. This report may not be reproduced, except in full without the written approval of this laboratory. The client must not use this document to claim product endorsement by this laboratory.


Ron Peterson, Metrologist




Date



SOUTH DAKOTA WEIGHTS AND MEASURES / METROLOGY LAB

Lab: 1500 N. Garfield-E. Truck Bypass Phone: 605-773-3170
Office: 118 West Capitol Avenue Phone: 605-773-3697
Pierre, SD 57501

CALIBRATION REPORT

Submitted by: Seneca Companies
Mailing Address: 4140 E, 14th St.
City, State, Zip: Des Moines, IA 50313

Report Number: MP3334

Date Received: 09/16/14

Date tested: 09/17/14

Temperature (°C): 21.7

Humidity: 47.4

Pressure (mm/Hg): 713.0%

Water Temperature(°C): 16.8 - 16.9

Equipment Used: 5 Gal Slicker SN 7178/ Fluke 1560 Probe Thermometer

Condition of artifacts: GOOD

5 Gallon Test Measure

		Nominal Volume (gal)	Volume as Found (in ³ at 60 °F)	Volume as Found (gal at 60 °F)	Volume as Left (in ³ at 60 °F)	Volume as Left (gal at 60 °F)	unc ±(in ³)	k Factor	Adjusted
Manufacturer	SN								
Seraphin	12-90054	5	1154.92	5.000	1154.92	5.000	0.30	2.18	No
Seraphin	07-03372	5	1155.02	5.000	1155.02	5.000	0.30	2.18	No

W.L.
ROSEN →
BAYLE →
LANDER

Treatment of artifacts prior to calibration: Artifacts were stored in the Metrology lab at the test area.

Compliance Statement: The test measures and associated uncertainties were evaluated against NIST Handbook 105-3, Specifications and Tolerances for Reference Standards and Field Standards Weights and Measures, February 2012. This evaluation indicates that the test measures were in compliance at the time of the test.

Procedure used: NISTIR 7383, SOP 19, Standard Operation Procedure for Calibration of Graduated Neck Type Metal Provers (Volume Transfer Method). Water was used as the test medium.

Supplemental information:

No Neck Scale verification test was completed.

Ron Peterson, Metrologist

End of Report

09/17/2014
Date of Report



Pat Quinn, Governor – Robert Flider, Director
METROLOGY STANDARDS LABORATORY
NIST TRACEABLE Calibration Certificate

• P.O. Box 19281 □ Springfield, IL 62794-9281 • 217/785-8480

Business Number: 900565
 W&M Invoice Number: 400806
 Purchase Order Number: N/A

Report of Test Number: 14-262
 Observer: 018, 364
 Date Issued: 12-12-2014



Customer: Seneca Companies
 Address: 4140 East 14th Street
 City & State: Des Moines IA 50313

Date Received: 12-9-2014
 Date of Test: 12-9-2014
 Temperature: 22.3 °C
 Relative Humidity: 59 %
 Barometric Pressure: 770.49 mmhg

Date of Primary Standard Calibration: 2011
 Date of Working Standard Calibration: N/A

Test Method: NISTIR 7383 - SOP 19
 Standards Used: 8806-A

Description of Test Items: Brownie 100 gallon volumetric prover - Model 100 USG, Serial number 6871998. Prover was sealed, decaled and impressed IL-14. Attention Ed Rogers. Trailer Unit 6520 NK.

Condition of Test Items: Acceptable

<u>Nominal Capacity</u>	<u>As Found Data</u>	<u>As Left Data</u>	<u>Uncertainty</u>
100 gallon	0.8 in ³	0.8 in ³	+/-5.0 in ³

Michael W. Rockford
 Metrologist

The Illinois Standards have demonstrated metrological traceability to the international system of units (SI), which may be to the SI through a National Metrology Institute such as NIST and the laboratory has demonstrated measurement proficiency. The laboratory calculates measurement uncertainties in accordance with NIST-OWM Standard Operating Procedure 29, which is in conformance with ISO "Guide to the Expression of Uncertainty in Measurement." The confidence level is 95 % with a coverage factor of two, (k = 2). This document shall not be reproduced, except in full, without the approval of the Illinois Laboratory.

United States Department of Commerce National Institute of Standards and Technology

Certificate of Metrological Traceability For:

South Dakota

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.



2014

Scope

Mass Echelon III

30 kg to 1 mg
1000 lb to 0.001 lb
8 oz to 0.03125 oz

Weight Carts
5000 lb to 2000 lb

Volume Transfer, II
5 gal

A handwritten signature in blue ink that reads "Carol T. Hockert".

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2014-01-01 to 2014-12-31

United States Department of Commerce National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Illinois

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.



2013-2014

Scope

Mass Echelon I 5 kg to 1 mg	Volume Transfer, II 1500 gal to 1 gal
Mass Echelon II 5 kg to 1 mg	100 gal to 20 gal LPG
Mass Echelon III 30 kg to 1 mg 2500 lb to 0.001 lb 8 oz to 0.03125 oz	
Weight Carts 5000 lb to 2500 lb	

A handwritten signature in blue ink that reads "Carol T. Hockert".

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2013-01-01 to 2014-12-31