



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY
 Public Service Commission
 SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company <i>Cloverdale Foods Company</i>		Email Address <i>Larryh@cloverdalefoods.com</i>		Application Date	
Mailing Address <i>3015 34th St. N.W.</i>		City <i>Mandan</i>	State <i>ND</i>	Zip Code <i>58554</i>	
Telephone Number <i>701-667-8512</i>		Cell Phone Number <i>701-400-7135</i>		Fax Number <i>701-667-8513</i>	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input checked="" type="checkbox"/> 5. Belt <input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: <i>200#</i> <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (Indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
<i>1488</i>	<i>Larry Huether</i>	<i>5, 6, 7</i>

Continued on Page 2



List below all field standards (attach current calibration reports):

500 #	2 ea.	
50 #	3 ea.	
20 #	7 ea.	
10 #	8 ea.	
5 #	7 ea.	
2 #	7 ea.	
1 #	6 ea.	
802.	1 ea.	
402.	1 ea.	

Additional Application Items (initial where appropriate):


Standardized Test Report	<input type="checkbox"/> Copy enclosed
	<input type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed
	<input type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed
	<input type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.

Yes No

I am Larry Huebner, and have authority to represent this company. By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.



 Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410



SECRETARY OF STATE NORTH DAKOTA



[Home](#) | [Business Records Search](#)

CLOVERDALE FOODS COMPANY

Corporation Details

System ID: 1632000 **Phone:** (701) 663-9511
Type: BUSINESS CORPORATION
Status: Active & Good Standing
Original File Date: 09/12/1925 **Effective Date:** 09/12/1925
State of Origin: North Dakota

Nature of Business

FOOD MANUFACTURING

Principal Office

3015 34TH ST NW PO BOX 667 MANDAN, ND 58554-0667

Registered Agent

T J RUSSELL
3015 34TH ST NW
PO BOX 667
MANDAN, ND 58554-0667
Established Date: Jul 07, 2000

Authorized Shares

Class	Number	Par Value
COMMON	100.000000	\$5000.000000

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2014](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)

[Disclaimer](#)

[Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



Receipt Date: February 24, 2015
Test Date: February 25, 2015
Report Date: February 26, 2015

State Test No.: 333878
Set Serial No.: None
Barcode: 201064

Calibration Report

CLOVERDALE FOODS COMPANY
3015 34TH ST NW
MANDAN, ND 58554
Contact: LARRY HUETHER
Phone: 701-667-8512
PO Number: NONE
SOP: 12
Technician ID: 07

Item(s) Submitted: Cast Cube Weights
Manufacturer: Rice Lake
ASTM E617 Type: II
Equipment ID#: None
Condition: Good
Temperature: 18.7°C
Pressure: 738. mmHg
Relative Humidity: 35. %

Nominal Value	Serial No.	Correction (g)		NIST HB105-1 Class		Unc. (g) (k=2)
		As Found	As Left	As Found	As Left	
500 lb		-1.3	-1.3	F	F	1.1
500 lb		20.9	20.9	F	F	1.1

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Mark Nicollet

Quality Manager

Reviewed by:
Pete Whebbe

Metrologist



Receipt Date: February 24, 2015
Test Date: February 25, 2015
Report Date: February 25, 2015

State Test No.: 333880
Set Serial No.: None
Barcode: 201065

Calibration Report

CLOVERDALE FOODS COMPANY
3015 34TH ST NW
MANDAN, ND 58554

Item(s) Submitted: Cast Handle Weights
Manufacturer: Rice Lake
ASTM E617 Type: II
Equipment ID#: None
Condition: Good
Temperature: 18.7°C
Pressure: 738. mmHg
Relative Humidity: 35. %

Contact: LARRY HUETHER
Phone: 701-667-8512
PO Number: NONE
SOP: 12
Technician ID: 07

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Unc. (mg) (k=2)
		As Found	As Left	As Found	As Left	
50 lb		1920	1920	F	F	110
50 lb		2450	1210	*	F	110
50 lb		1290	1290	F	F	110
20 lb		740	740	F	F	75
20 lb		320	320	F	F	75
20 lb		1290	80	*	F	75
20 lb		630	630	F	F	75
20 lb		1400	440	*	F	75
20 lb		1060	810	*	F	75
20 lb		720	720	F	F	75

* Weight(s) as found exceed NIST HB 105-1 Class F tolerance.

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Mark Nicollet

Mark Nicollet
Quality Manager

Page 1 of 1

Reviewed by:
Pete Whebbe

Pete Whebbe
Metrologist



Receipt Date: February 24, 2015
Test Date: March 3, 2015
Report Date: March 3, 2015

State Test No.: 333879
Set Serial No.: None
Barcode: 201752

Calibration Report

CLOVERDALE FOODS COMPANY
3015 34TH ST NW
MANDAN, ND 58544
Contact: LARRY HUETHER
Phone: 701-667-8512
PO Number: NONE
SOP: 12
Technician ID: 11

Item(s) Submitted: 50 lb set/loose weights
Manufacturer: Unknown
ASTM E617 Type: II
Equipment ID#: None
Condition: Fair
Temperature: 20.1°C
Pressure: 729.2 mmHg
Relative Humidity: 52. %

Nominal Value	Serial No.	Correction (mg)		ASTM E617 Class		Unc. (mg) (k=2)
		As Found	As Left	As Found	As Left	
10 lb		26.	26.	4	4	15.
10 . lb		-91.	-91.	5	5	15.
10 .. lb		-25.	-25.	4	4	15.
5 lb	1	75.	75.	5	5	10.
5 lb		43.	43.	5	5	10.
5 lb		27.	27.	4	4	10.
2 lb		9.2	9.2	5	5	6.
2 lb		36.6	36.6	6	6	6.
2 lb		54.4	22.8	7	5	6.
2 lb		17.5	17.5	5	5	6.
2 ... lb		16.7	16.7	5	5	6.
1 lb		-16.9	-16.9	6	6	6.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Pete Whebbe

Pete Whebbe
Metrologist

Reviewed by:

Mark Nicollet

Mark Nicollet
Quality Manager

Page 1 of 1



NYLAP LAB CODE 105003-0

Accredited by the National Voluntary Laboratory Accreditation Program for the specific scope of accreditation under lab code 105003. This report may not be used to claim product endorsement by NVLAP or any other government agency, and may not be reproduced, except in full, without written approval from the laboratory.



Receipt Date: March 17, 2015
Test Date: March 18, 2015
Report Date: March 19, 2015

State Test No.: 334019
Set Serial No.: None/Individual
Barcode: 201078

Calibration Report

CLOVERDALE FOODS COMPANY
3015 34TH ST NW
MANDAN, ND 58554

Item(s) Submitted: Satin stainless weights
Manufacturer: Rice Lake
ASTM E617 Type: II
Equipment ID#: None
Condition: Good
Temperature: 20.4°C
Pressure: 744.7 mmHg
Relative Humidity: 48. %

Contact: LARRY HUETHER
Phone: 701-667-8512
PO Number: NONE
SOP: 12
Technician ID: 09

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Unc. (mg) (k=2)
		As Found	As Left	As Found	As Left	
5 lb		12	12	F	F	10.
5 lb		31	31	F	F	10.
2 lb		-9	-9	F	F	6.
2 lb	0161	45	45	F	F	6.
2 lb	6355	40	40	F	F	6.
1 lb	5	22	22	F	F	6.
1 lb	0487	20	20	F	F	6.
8 oz		-12	-12	F	F	6.

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Heidi Jones
Heidi Jones
Laboratory Administrator

Reviewed by:
Mark Nicollet
Mark Nicollet
Quality Manager



Receipt Date: March 17, 2015
 Test Date: March 18, 2015
 Report Date: March 18, 2015

State Test No.: 334018
 Set Serial No.: Pellet
 Barcode: 201077

Calibration Report

CLOVERDALE FOODS COMPANY
 3015 34TH ST NW
 MANDAN, ND 58554
 Contact: LARRY HUETHER
 Phone: 701-667-8512
 PO Number: NONE
 SOP: 12
 Technician ID: 11

Item(s) Submitted: Cast Hand Weight
 Manufacturer: Rice Lake
 ASTM E617 Type: II
 Equipment ID#: None
 Condition: Good
 Temperature: 18.8°C
 Pressure: 744.8 mmHg
 Relative Humidity: 32. %

Kit contents:

Nominal Value	Quantity
20 lb	1

The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These Standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculation methods conform to NIST Technical Note 1297. Actual uncertainties will be provided upon request. Results apply to items identified in this report only.

Pete Whebbe

Pete Whebbe
 Metrologist

Reviewed by:

Mark Nicollet

Mark Nicollet
 Quality Manager



Receipt Date: February 24, 2015
Test Date: March 3, 2015
Report Date: March 3, 2015

State Test No.: 333877
Set Serial No.: None
Barcode: 202016

Calibration Report

CLOVERDALE FOODS COMPANY
3015 34TH ST NW
MANDAN, ND 58544
Contact: LARRY HUETHER
Phone: 701-667-8512
PO Number: NONE
SOP: 12
Technician ID: 11


Item(s) Submitted: 30 lb Kit
Manufacturer: Rice Lake
ASTM E617 Type: II
Equipment ID#: None
Condition: Good
Temperature: 19.9°C
Pressure: 728.9 mmHg
Relative Humidity: 50. %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Unc. (mg) (k=2)
		As Found	As Left	As Found	As Left	
10 . lb		92.	92.	F	F	15.
10 .. lb		83.	83.	F	F	15.
5 lb		54.	54.	F	F	10.
2 lb		35.9	35.9	F	F	6.
1 lb		17.5	17.5	F	F	6.

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.


Pete Whebbe



Metrologist

Reviewed by:

Mark Nicollet



Quality Manager

Page 1 of 1



NYLAP LAB CODE 105003-0

Accredited by the National Voluntary Laboratory Accreditation Program for the specific scope of accreditation under lab code 105003. This report may not be used to claim product endorsement by NVLAP or any other government agency, and may not be reproduced, except in full, without written approval from the laboratory.



Receipt Date: February 24, 2015
Test Date: March 3, 2015
Report Date: March 3, 2015

State Test No.: 333876
Set Serial No.: None
Barcode: 201059

Calibration Report

CLOVERDALE FOODS COMPANY
3015 34TH ST NW
MANDAN, ND 58544

Contact: LARRY HUETHER
Phone: 701-667-8512
PO Number: NONE
SOP: 12
Technician ID: 11

Item(s) Submitted: 50 lb kit
Manufacturer: Rice Lake, Troemner
ASTM E617 Type: II
Equipment ID#: None
Condition: Good
Temperature: 19.9°C
Pressure: 728.5 mmHg
Relative Humidity: 50. %

Nominal Value		Serial No.	Correction (mg)		NIST HB105-1 Class		Unc. (mg) (k=2)
			As Found	As Left	As Found	As Left	
10	lb	1	119.	119.	F	F	15.
10	lb	2	137.	137.	F	F	15.
10	lb	3	133.	133.	F	F	15.
5	lb	None	45.	45.	F	F	10.
5	lb	1	37.	37.	F	F	10.
5	lb	2	1.	1.	F	F	10.
2	lb	None	50.1	50.1	F	F	6.
1	lb		25.1	25.1	F	F	6.
1	lb		13.4	13.4	F	F	6.
1	lb	2	24.9	24.9	F	F	6.
1	lb	3	23.4	23.4	F	F	6.
0.5	lb	None	5.6	5.6	F	F	6.
4	oz	3	6.37	6.37	F	F	0.07

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Pete Whebbe

Metrologist

Reviewed by:

Mark Nicollet

Quality Manager

Page 1 of 1



NYLAP LAB CODE 105003-0

Accredited by the National Voluntary Laboratory Accreditation Program for the specific scope of accreditation under lab code 105003. This report may not be used to claim product endorsement by NVLAP or any other government agency, and may not be reproduced, except in full, without written approval from the laboratory.