



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY

Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company <i>FAIRBANKS SCALES</i>	Email Address <i>MCONWAY@FAIRBANKS</i>	Application Date <i>3-1-15</i>	
Mailing Address <i>4850 BROADWAY</i>	City <i>DENVER</i>	State <i>CO</i>	Zip Code <i>80216</i>
Telephone Number <i>720 839-1792</i>	Cell Phone Number <i>720 839-1792</i>	Fax Number <i>303 296-0269</i>	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input checked="" type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input checked="" type="checkbox"/> 3. Livestock <input checked="" type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: <i>UP TO TRUCK</i> <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
<i>e.g. 1001</i>	<i>e.g. John Doe</i>	<i>e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6</i>
	<i>MARK CONWAY</i>	<i>1, 2, 3, 4, 6, 7</i>
	<i>GLEN MEYER</i>	<i>1, 2, 3, 4, 6, 7</i>
	<i>JOHN ROBINSON</i>	<i>1, 2, 3, 4, 6, 7</i>

Continued on Page 2



List below all field standards (attach current calibration reports):

5- 2500 LB WEIGHTS	1 - 3K' CART
4- 1000 LB WEIGHTS	4 - 2500 LB WEIGHTS
1- 500 LB WEIGHTS	9 - 1000 LB WEIGHTS
1 - 3000 LB CART	10
500 LB - 50 LB TEST WEIGHTS	


Additional Application Items (initial where appropriate):

Standardized Test Report	<input checked="" type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input checked="" type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input checked="" type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am MARK CONWAY, and have authority to represent this company. By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.



 Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410

North Dakota

nd.gov Official Portal for
North Dakota State GovernmentNorth Dakota
LEGENDARY

SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

FAIRBANKS SCALES INC.

Corporation Details

System ID: 2987100 **Phone:** (816) 471-0231
Type: FOREIGN BUSINESS CORPORATION
Status: Active & Good Standing
Original File Date: 03/31/1988 **Effective Date:** 03/31/1988
State of Origin: Kansas

Nature of Business

SALES AND SERVICE OF SCALES AND RELATED EQUIPMENT

Principal Office

821 LOCUST ST KANSAS CITY, MO 64106-1908

Registered Agent

CORPORATION SERVICE COMPANY
1501 N 12TH ST STE 1
BISMARCK, ND 58501-2713
Established Date: Mar 01, 2010

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2014](#) [2015](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)[Disclaimer](#) | [Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



Wyoming Department of Agriculture
Weights and Measures Laboratory
6607 Campstool Rd
Cheyenne, WY 82002
(307)777-7556



Report of Calibration
For
3000 lb Motorized Test Cart

Test Number: 15004

Date of Calibration: January 28, 2015

Submitted by: Glen Meyer
Fairbanks Service
4850 Broadway
Denver, CO 80216
(800)435-4660

Manufacturer: Summit Truck Bodies

Date of Manufacture: June 2011

Model Number: 3000

Serial Number: 3K0006

Calibration Values:

As Found Conventional Mass* (lb)	As Left Conventional Mass (lb)	Uncertainty (lb)
2997.07	2999.98	0.14

* The Conventional Mass is the mass determined/calculated by weighing in air of reference density 1.2 mg/cm³ against reference weights of density 8.0 g/cm³. Buoyancy corrections are considered negligible and were not made unless otherwise stated.

Conditions of Test & Traceability:

The above weight cart was compared with standards of the State of Wyoming, which are traceable to the National Institute of Standards and Technology (NIST) and have known values. The assigned test number shown above provides documented evidence for measurement traceability.

Test Method: NIST IR6969 SOP 7
Balance: PR6246
Temperature: 20.6 °C
Pressure: 614.2 mm Hg
Relative Humidity: 49.4%

Inspection Checklist for Weight Carts



Test Number: 15004 Date of Inspection: 1/27/2015
Manufacturer: Summit Truck Bodies Date of Manufacture: 6/1/2011
Model Number: 3000 ID/SN Number: 3K0006

Check All That Apply (X-Passes Inspection, N-Fails Inspection, or N/A-Not Applicable)

X Nominal Mass of Weight Cart 3000 lb Suitably Marked (Y/N): Y
X Powered by: Electric/Generator Diesel Gasoline X
X Fluid Levels: Engine Oil: X Hydraulic Fluid: X Sealed (Y/N): N
Battery: X Sealed (Y/N): N
Liquid Fuel: X Reference line present (Y/N): Y

X Fluid drain tubes extend beyond the body of the cart (Y/N): Y

X Number of axles: 2

X Number/Size of Tires: 4-21"x9"x15"

X Sealed wheel bearings (Y/N): Y

X Drain holes present in locations where water may accumulate (Y/N): Y

X Weight restraint railing permanently fixed and solid (Y/N): Y

X Adjusting cavity accessible (Y/N): Y

X Adjusting cavity approximate capacity: 75 lb

X Adjusting cavity sealed (Y/N): Y

N/A Service brakes functioning properly (Y/N): Hydraulic

X Parking brake functioning properly (Y/N): Y

N/A Remote control functioning properly (Y/N):

X General condition at time of calibration (note any accumulated dirt/debris, loose parts of evidence of tampering or unauthorized entry of seals).

Fair condition. Hydraulic cap was not sealed upon arrival; sealed upon return to the customer. Battery is not securely mounted (bungee cord) and no security seal was affixed as the hold down is missing.

X List and report any repair and maintenance performed, parts replaced, leaks repaired, new battery, carburetor, exhaust system, wheels changed, welding performed, etc.. Include any comments or changes since the last calibration.

No known repairs.

Robert Weidler, WDA State Metrologist

1-29-15 Date



Wyoming Department of Agriculture
Weights and Measures Laboratory
6607 Campstool Rd
Cheyenne, WY 82002
(307)777-7556



Calibration Certificate For

Five – 2500 lb Class F Weights,
Four – 1000 lb Class F Weights,
One – 500 lb Class F Weight,
and
One – 30 lb Class F Weight Kit (21 Weights)

Manufacturer: Listed in Table A and B
Serial No.: Listed in Table A and B

Submitted by
Glen Meyer
Fairbanks Service
4850 Broadway
Denver, CO 80216
(800)435-4660

Serial Number	Nominal (lb)	Conventional Mass Correction (g)		Tolerance (g)	Expanded Uncertainty (g)
		As Found	As Left		
2	2500	5	5	110	25
FW1-2500	2500	8	8	110	25
FW2-2500	2500	-32	-32	110	25
FW4-2500	2500	-81	33	110	25
FW8-2500	2500	-69	13	110	25
4	1000	-31.4	20.8	45	7.4
5	1000	-3.5	-3.5	45	7.6
6	1000	-12.9	-12.9	45	7.6
FW2	1000	-32	3.7	45	7.6
FB11	500	3.6	3.6	23	6.4
7	50	-2.68	0.13	2.3	0.20
614-304	50	-0.18	-0.18	2.3	0.20
CSS-E-21	50	-0.82	-0.82	2.3	0.20
FWS-D-4	50	-0.08	-0.08	2.3	0.20
FWS-D-11	50	-0.73	-0.73	2.3	0.20
FWS-D-21	50	-0.72	-0.72	2.3	0.20
FWS-D-22	50	-0.25	-0.25	2.3	0.20
FWS-D-24	50	-1.61	0.37	2.3	0.20
FWS-D-31	50	-0.31	-0.31	2.3	0.20
FWS-D-76	50	-0.84	-0.84	2.3	0.20

The data in this table applies only to those items specifically listed on this report.

Table A – Fairbanks 2500 lb, 1000 lb, 500 lb and 50 lb Class F Weights. Condition: good; assumed density 7.20 g/cm³.



Uncertainty Statement:

The combined standard uncertainty includes the standard uncertainty reported for the standard and the standard uncertainty for the measurement process. The combined standard uncertainty is multiplied by a coverage factor of 2.07 to 2.02 to give an expanded uncertainty, which defines an interval having a level of confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the 1993 ISO Guide to the Expression of Uncertainty in Measurement. The expanded uncertainty is not to be confused with a tolerance limit for the user during application.

Traceability Statement:

Standards used for comparison are traceable to United States national standards at NIST and are part of a comprehensive measurement assurance program for ensuring continued accuracy and traceability reported by this laboratory. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only. Traceability is maintained to the SI using the following conversion: 1 lb = 0.45359237 kg.

Supplemental Information

Description of artifacts submitted for testing:

Thirty one Class F stainless steel, cast iron and aluminum test weights. Assumed densities are listed in the appropriate table.

Conditions of artifacts submitted for testing:

Condition is noted in the title of the appropriate table.

Treatment of artifacts prior to testing:

Artifacts were clean upon arrival; no additional treatment needed. Thermal equilibrium: 24 hours next to balance in the laboratory.

Equipment and Standards:

<u>Balance</u>	<u>Range</u>	<u>Standards Used</u>
Mettler AX26	0 g – 22 g	WY WS
Mettler XP505	0 g – 520 g	WY WS
Mettler XP5003	0 kg – 5.1 kg	WY WS
Mettler XP32003	0 kg – 32.1 kg	WY WS
Mettler XP604	0 kg – 600 kg	WY WS
Mettler XP2003	0 kg – 2000 kg	WY WS

Procedure used:

Single Substitution Method (NISTIR 6969, SOP 7)


Environmental conditions are maintained within the following parameters:

<u>Temperature</u>	<u>Relative Humidity</u>
18 °C to 27 °C	40.0% to 60%

Date Artifacts Received: January 26, 2015

Date of test: January 28, 2015

Test performed by:


Robert Weidler

Date of Report Preparation: January 29, 2015

WDA State Metrologist

This document does not represent or imply endorsement by the Wyoming Department of Agriculture, NIST or any agency of the State and/or national governments. This document may not be reproduced, except in full, without the written permission of the Wyoming Department of Agriculture Weights and Measures Laboratory.

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Wyoming



This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

2014 - 2015

Scope

Mass Echelon III

3000 lb to 0.001 lb

8 oz to 0.03125 oz

Weight Carts

4500 lb to 2000 lb

Volume Transfer, II

1000 gal to 5 gal

100 gal to 25 gal LPG

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2014-01-01 to 2015-12-31

Amended: 2014-03-01

Expanded Scope to increase lower Mass Echelon III Weight Cart limit to 2000 lb.



CALIBRATION CERTIFICATE # 15-067



Customer Name Fairbanks Scales
Customer Address 4850 Broadway, Denver CO 80216-6344
Submitted By Mark Conway
Certificate Date February 20, 2015
Calibration Date February 12, 2015
Calibration Due This certificate expires at 12:01 A.M. on February 12, 2016, per State of Colorado Revised Statute Title 35 Article 14 Section 123:2009.
Serial Number(s) See table
Manufacturer See table
Description Class F, 1000 lb, 500 lb and 50 lb cast iron weights
Condition Received Good
Procedure Used National Institute of Standards and Technology (NIST) NISTIR 6969:2012 SOP 8, Weighing by Modified Substitution

Traceability This certificate has been issued under the authority of the Commissioner of the Colorado Department of Agriculture, Inspection and Consumer Services Division, pursuant to the State of Colorado Revised Statute Title 35 Article 14:2009. Standards used for comparison are traceable to the International System of Units (SI) through standards at NIST, by way of the certificate number above.

Uncertainty The combined standard uncertainty of the measurements is multiplied by the coverage factor *k* listed in the table on page 2 (based on available degrees of freedom) to give an expanded uncertainty which defines an interval having a level of confidence of 95.45 %. The expanded uncertainty presented in this report was calculated according to the BIPM JCGM 100:2008 *Evaluation of measurement data – Guide to the expression of uncertainty in measurement* (GUM 1995 with 2010 minor corrections). Uncertainty components evaluated may include, but are not limited to, standard deviation of the process, mass standard uncertainties, the uncertainty for magnitude of the air buoyancy correction and/or for any uncorrected errors associated with air buoyancy corrections, uncertainties associated with densities of the standards and test items, and a component to account for any observed deviations from mass standard values that are less than surveillance limits.

Magnetism None of the weights used for this calibration have been tested for magnetic properties, and no magnetism components are included in the uncertainty budget.

Metrologist(s) performing calibration

Kate Smetana, Metrologist
 Authorized Signatory

Environmental Conditions at Time of Calibration

	minimum	units	maximum	units
Temperature:	18.1	°C	19.8	°C
Barometric Pressure:	628.0	mmHg	634.4	mmHg
Relative Humidity:	44.9	%	47.0	%

Conversion Factors from NIST Special Publication 811:2008, *Guide for the Use of the International System of Units (SI)*
 1 pound (avoirdupois) (lb) = 0.45359237 kilogram (kg)
 1 ounce (avoirdupois) (oz) = 0.02834952 kilogram (kg)
 MS Invoice # 1531

CO Dept. of Agriculture, Metrology Laboratory

Calibration Certificate # 15-067

Fairbanks Scales

Certificate Date: February 20, 2015



Conventional Mass Correction

Serial Number	Manufacturer	Nominal Value (lb)	Conventional Mass Correction		Tolerance ± g	Uncertainty ± g	k
			As Found g	As Left g			
614 80	Fairbanks Morse	1000		- 6.6	45	7.0	2.06
614 88	Fairbanks Morse	1000		4.3	45	7.0	2.06
614 89	Fairbanks Morse	1000		4.4	45	7.0	2.06
FM-B-039 14	Fairbanks Morse	1000		- 6.5	45	7.0	2.06
FM-B-064 5	Fairbanks Morse	1000		- 23.5	45	7.0	2.06
FW-B10	Fairbanks Morse	500	- 33.8	- 1.0	23	5.36	2.12
FW-B12	Fairbanks Morse	500	- 51.1	CONDEMNED			
FW-B17	Fairbanks Morse	500	- 30.2	- 2.1	23	5.36	2.12
FW-B18	Fairbanks Morse	500		- 2.2	23	5.36	2.12
FW-B19	Fairbanks Morse	500	- 37.8	0.9	23	5.36	2.12
FW-B3	Fairbanks Morse	500	- 24.7	CONDEMNED			
FW-B6	Fairbanks Morse	500	- 20.6	CONDEMNED			
3	Fairbanks Morse	50		- 1.17	2.3	0.30	2.06
35091	HRS CO	50	- 1.90	- 0.04	2.3	0.30	2.06
614 300	Fairbanks Morse	50	- 2.75	- 0.08	2.3	0.30	2.06
614 301	Fairbanks Morse	50	- 2.79	0.18	2.3	0.30	2.06
CSS-D-27	Rice Lake	50	- 2.11	- 0.30	2.3	0.30	2.06
F-108	Fairbanks Morse	50	- 1.96	CONDEMNED			
FB3	Fairbanks Morse	50		0.75	2.3	0.30	2.06
FM-D-192	Fairbanks Morse	50		- 1.26	2.3	0.30	2.06
FM-D-260	Fairbanks Morse	50	- 2.26	0.25	2.3	0.30	2.06
FWS-D-36	Fairbanks Morse	50		- 1.26	2.3	0.30	2.06
FWS-D-43	Rice Lake	50		- 0.83	2.3	0.30	2.06

Supplemental Information:

All items were left "as found", except as listed in the "As Found" column. The items except those marked "CONDEMNED" were found at time of test, or adjusted, to meet the specifications and tolerances stated in NIST Handbook 105-1:1990, *Specifications and Tolerances for Reference Standards and Field Standard Weights and Measures (Field Standard Weights (NIST Class F))*, and are approved for use in the State of Colorado. Those marked "CONDEMNED" were found to be out of tolerance and not adjustable. They are not approved for commercial use in the State of Colorado

END OF DOCUMENT

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Colorado

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon II

10 kg to 1 mg

Mass Echelon III

30 kg to 1 mg

2500 lb to 0.001 lb

8 oz to 0.03125 oz

Weight Carts

5000 lb to 2000 lb

Volume Transfer, II

100 gal to 5 gal

100 gal to 25 gal LPG

Length,

Steel Tapes, Bench, Tapes

200 ft to 1 ft

Rigid Rules

10 in to 1 in

Frequency

Tuning Forks ≤ 80 mph

Grain Moisture

≤ 20 %



2015 - 2016


A handwritten signature in blue ink, reading "Carol T. Hockert".

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2015-01-01 to 2016-12-31



Colorado Department of Agriculture
Inspection and Consumer Services Division
Metrology Laboratory
3125 Wyandot Street
Denver, CO 80211-3824
(303) 867-4270 FAX (303) 477-4248
www.colorado.gov/ag/ics



CALIBRATION CERTIFICATE # 14-147A


Customer Name Fairbanks Scales
Customer Address 4850 Broadway, Denver CO 80216-6344
Submitted By Arnie Otterstetter
Certificate Date April 18, 2014
Calibration Date April 11, 2014
Calibration Due This certificate expires at 12:01 A.M. on April 11, 2015, per State of Colorado Revised Statute Title 35 Article 14:2009.
Serial Number(s) See page 2
Manufacturer See page 2
Description Summit 3000 lb gasoline powered weight cart
Condition Received Class F, 2500 lb, 1000 lb, and 50 lb cast iron weights
Procedure Used Water found in cavity of one 1000 lb weight
National Institute of Standards and Technology (NIST) NISTIR 6969:2012 SOP 8,
Weighing by Modified Substitution

Traceability This certificate has been issued under the authority of the Commissioner of the Colorado Department of Agriculture, Inspection and Consumer Services Division, pursuant to the State of Colorado Revised Statute Title 35 Article 14. Standards used for comparison are traceable to the International System of Units (SI) through standards at NIST, by way of the certificate number above.

Uncertainty The combined standard uncertainty of the measurements is multiplied by the coverage factor k listed in the table on page 2 (based on available degrees of freedom) to give an expanded uncertainty which defines an interval having a level of confidence of 95.45 %. The expanded uncertainty presented in this report was calculated according to the BIPM JCGM 100:2008 *Evaluation of measurement data – Guide to the expression of uncertainty in measurement* (GUM 1995 with 2010 minor corrections). Uncertainty components evaluated may include, but are not limited to, standard deviation of the process, mass standard uncertainties, the uncertainty for magnitude of the air buoyancy correction and/or for any uncorrected errors associated with air buoyancy corrections, uncertainties associated with densities of the standards and test items, and a component to account for any observed deviations from mass standard values that are less than surveillance limits.

Magnetism None of the weights used for this calibration have been tested for magnetic properties, and no magnetism components are included in the uncertainty budget.

Metrologist(s) performing calibration


Diane Wise, Metrologist
Authorized Signatory


Kate Smetana, Metrologist

Environmental Conditions at Time of Calibration

	minimum	units	maximum	units
Temperature:	20.5	°C	20.7	°C
Barometric Pressure:	625.2	mm Hg	628.0	mm Hg
Relative Humidity:	50.0	%	51.6	%

Conversion Factors from NIST Special Publication 811:2005, *Guide for the Use of the International System of Units (SI)*
1 pound (avoirdupois) (lb) = 0.45359237 kilogram (kg)

MS Invoice # 1149

CO Dept. of Agriculture, Metrology Laboratory
 Calibration Certificate # 14-147A

Fairbanks Scales

Certificate Date: April 18, 2014



Serial Number	Item Mfr.	Nominal Value (lb)	Conventional Mass Correction		Tolerance (\pm lb)	Uncertainty (\pm lb)	k
			As Found (lb)	As Left (lb)			
3K0005	Summit	3000	- 3.232	0.118	0.300	0.059	
FBS-09	Rice Lake	2500		0.016	0.250	0.066	
FBS-10	Rice Lake	2500		0.041	0.250	0.066	
FW-04	Rice Lake	2500		0.060	0.250	0.066	
FW5B	Rice Lake	2500		0.012	0.250	0.066	
			(g)	(g)	(\pm g)	(\pm g)	
1	Fairbanks Morse	1000	377.83	- 3.37	45.4	5.16	2.13
3	Fairbanks Morse	1000		2.73	45.4	5.16	2.13
008	Fairbanks Morse	1000		12.23	45.4	5.16	2.13
10	Fairbanks Morse	1000		27.53	45.4	5.16	2.13
6	Rice Lake	50		1.00	2.3	0.31	2.03
16	Rice Lake	50		0.56	2.3	0.31	2.03
CSS-D-28	Rice Lake	50		0.89	2.3	0.31	2.03
CSS-D-3	Rice Lake	50		0.11	2.3	0.31	2.03
D-0135	Webb	50		1.68	2.3	0.31	2.03
FM-D-245	Rice Lake	50		0.01	2.3	0.31	2.03
FM-D-257	Fairbanks	50		1.42	2.3	0.31	2.03
FM-D-264	Rice Lake	50		1.15	2.3	0.31	2.03
FS28	Fairbanks	50		0.75	2.3	0.31	2.03
FS-D-1	Rice Lake	50		0.98	2.3	0.31	2.03
FS-D-14	Rice Lake	50		0.83	2.3	0.31	2.03
FS-D-16	Fairbanks	50		1.36	2.3	0.31	2.03
FW-044	Fairbanks Morse	50		1.02	2.3	0.31	2.03
FW-D-81	Rice Lake	50		- 0.37	2.3	0.31	2.03
FWS-D-17	Rice Lake	50		1.02	2.3	0.31	2.03
FWS-D-25	Rice Lake	50		1.32	2.3	0.31	2.03
FWS-D-34	Rice Lake	50		0.40	2.3	0.31	2.03
FWS-D-41	Fairbanks	50		1.20	2.3	0.31	2.03
MPC-D-010	Fairbanks Morse	50		0.24	2.3	0.31	2.03
MPC-D-33	Rice Lake	50		1.08	2.3	0.31	2.03

Supplemental Information:

All items were left "as found", no adjustments were made. The item(s) were found at time of test, or adjusted, to meet the tolerances stated in NIST Handbook 105-1:1990, *Specifications and Tolerances for Reference Standards and Field Standard Weights and Measures (Field Standard Weights (NIST Class F))*, and are approved for use in the State of Colorado.

The weight cart was adjusted to meet the tolerances stated in NIST Handbook 105-8:2003, *Specifications and Tolerances for Reference Standards and Field Standard Weights and Measures (Field Standard Weight Carts)*, and is approved for use in the State of Colorado.

Standards used for testing

- CO-V-1
- CO-C-1
- B-001
- CO-DSTD-1

END OF DOCUMENT



Fairbanks Scales

2500 Cleveland Ave N
St. Paul, MN 55113

651-631-9287

Upon Completion email this document to Shelly Bauske

sbauske@nd.gov

Date	PSC Device Code	Number of Sections	Complete each item with Y (Yes), N (No), or NA (Not Applicable)	
4/15/2015	3V7	7	<input type="checkbox"/> New Installation (w/RFI Check)	<input checked="" type="checkbox"/> Performed Calibration
Name of Business			<input type="checkbox"/> Modified Equipment	<input type="checkbox"/> Use as a Reference Scale
Cat Scale #1540			<input type="checkbox"/> Replaced Existing Equipment	<input type="checkbox"/> Non-Commercial
Mailing Address			Expiration Date: <input type="text"/>	
Loves Country Store I-29 32nd ave.			<input type="checkbox"/> Variance Permit Posted;	69-10-02-23 & 69-10-02-24
City	State	Zip Code	<input type="checkbox"/> Stored/Recalled Weights Meet NDAC	
Fargo	ND	58104	<input checked="" type="checkbox"/> Software is NTEP Approved	Date Certified: <input type="text"/>
County	Telephone Number		<input checked="" type="checkbox"/> Built-in Standards;	
Cass	800-336-9889		<input checked="" type="checkbox"/> Multiple Decks/Single Indicator	<input type="checkbox"/> Customer has Clear View
Device Contact Manager	Cell Number		<input checked="" type="checkbox"/> Clearance Below Scale Clear	<input checked="" type="checkbox"/> Video Camera Working
Linda	none		<input checked="" type="checkbox"/> Approach Requirements Met	<input type="checkbox"/> View Distance > 200'
Email Address			<input checked="" type="checkbox"/> Pit Coping and Crush Strip Good	<input checked="" type="checkbox"/> Two-Way Audio Working
Catscale.com			<input type="checkbox"/> In Response to PSC Quality Assurance Inspection	
Device Location				
Southeast of fuel pumps				

Scale Information	
Scale Manufacturer	Indicator Manufacturer/Model
Fairbanks	Fairbanks 2600
Capacity/Divisions/Units	Indicator Serial Numbers
120,000x20lb	102990050020
Deck Size	Legible Label
80x10	x Yes No
Clearance (inches)	Class
5/8"	III/IIIIL
SR or Discrimination Test	Printer Manufacturer/Model
Zero Load= NA lb.	Citizens/DP3550
Loaded= NA lb.	Approach - 12' Concrete Level
Motion Detection	Hard Surface Approach (Length/Slope/Cond)
Range= 3.....60 lb.	NA
AZSM (auto zero)	
Range= 3.....60 lb.	

LBP/Section/ Product Wt.	Value of Test Weights Used	Errors +/-		LBP/Section/ Product Wt.	Value of Test Weights Used	Errors +/-	
		As Found	As Left			As Found	As Left
1 LBP	21000	21000	21000	Zero Load=	0	0	0
2	21000	21000	21000	S-1	21000	21000	21000
3	21000	21000	21000	2	21000	21000	21000
4	21000	21000	21000	3	21000	21000	21000
5	21000	21000	21000	4	21000	21000	21000
6	21000	21000	21000	5	21000	21000	21000
7	21000	21000	21000	6	21000	20980	21000
8	21000	21000	21000	7	21000	21000	21000
9	21000	21000	21000				
10	21000	21000	21000				
11	21000	20980	21000				
12	21000	20980	21000				
13	21000	21000	21000				
14	21000	21000	21000				

Strain Load Test		Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary. replaced SSC board in section 6 of scale, test and calibrate as needed.
Section (Increasing or Decreasing)	D	
Full Truck Weight	55800	
Test Weight	21000	
Empty Truck Weight	34800	
Errors	0	
Physical Seal	<input checked="" type="checkbox"/> Y <input type="checkbox"/> N	
Seal Date:	4/15/2015	
Seal Type:	lead	
Electronic Audit Trail:	<input checked="" type="checkbox"/> Y <input type="checkbox"/> N	
Meets tolerances in PSC adopted NIST Handbook 44? <input type="checkbox"/> Maintenance <input checked="" type="checkbox"/> Acceptance <input checked="" type="checkbox"/>		
x Approved <input type="checkbox"/> Rejected <input type="checkbox"/> Taken out of Service <input checked="" type="checkbox"/> Sticker Applied		
By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State Laws and rules, including NIST Handbook 44, for use of the device in commerce.		
Permit Holder Signature	Permit #	
	984	
Operator Signature	Date	
	4/15/2015	
Blank Signature-Customer Signature		