



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY
 Public Service Commission
 SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company Montana Dakota Scale Service Inc. #111155, 1976@ycom.com	Email Address	Application Date 8-21-15	
Mailing Address 13217 Hwy 200	City Fairview	State MT	Zip Code 59221
Telephone Number 406 742 8944	Cell Phone Number 406 480 2264	Fax Number 406 742 8944	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input checked="" type="checkbox"/> 1. Rail <input checked="" type="checkbox"/> 2. Truck <input checked="" type="checkbox"/> 3. Livestock <input checked="" type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
e.g. 1001	e.g. John Doe	e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6
1454	Lance Walker	2, 3, 4, 6, 7

Application for Registration as a Registered Service Company
Page 2



List below all field standards (attach current calibration reports):

18 - 50lb weights	
4 - 25lb weights	
8 - 1000lb weights	
3 - 2500lb weights	
1 - 2000lb cart	
1 - 2500lb cart	
1 - 32lb kit	
1 - 2kg kit	


Additional Application Items (initial where appropriate):

Standardized Test Report	<input checked="" type="checkbox"/> Copy enclosed <input type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input checked="" type="checkbox"/> Copy enclosed <input type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input checked="" type="checkbox"/> Copy enclosed <input type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am president, and have authority to represent this company.
By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.



Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410



Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163

Company Name & Address:		Date of Test:	Test Number:
Lance Waller		6/29/2015	2015-080
Montana-Dakota Scale Service		Serial Number:	1000's
Route 1, Box 1640			
Fairview, MT 59221			

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 6/29/2015

Description and condition of artifacts received: Items were in good condition.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
23.54	25.65	670.56	42.65	45.71

Conventional Mass Value:		Assumed Density of Artifacts:		7.2 g/cm ³		
Nominal	Serial No.	As Found (g)	As Left (g)	Uncertainty ± (g)	NIST 105-1 Class F ± (g)	k factor
1000 lb	4	-8.9	-8.9	5.9	45	2.28
1000 lb	5	-1.9	-1.9	5.9	45	2.28
1000 lb	6	47.1	-10.9	5.9	45	2.28
2500 lb	03X6	30.4	30.4	7.7	91	2.28
2500 lb	1	-9.6	-9.6	7.7	91	2.28
2500 lb	2	-5.6	-5.6	7.7	91	2.28
2500 lb	Cart	-423.5	42.7	7.7	227	2.28

Standards and Procedures used for testing:

The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7

All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, Issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:

The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:

The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows NISTIR 6969, Issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: Dave Fraser

David Fraser

Email: dafraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163

Company Name & Address:	Date of Test:	Test Number:
Lance Waller	6/29/2015	2015-077
Montana-Dakota Scale Service		
Route 1, Box 1640	Serial Number:	1000's
Fairview, MT 59221		

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 6/29/2015

Description and condition of artifacts received: Items were in good condition.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
23.54	25.65	670.56	42.65	45.71

Conventional Mass Value: Assumed Density of Artifacts: 7.2 g/cm³

Nominal	Serial No.	As Found (g)	As Left (g)	Uncertainty ± (g)	NIST 105-1 Class F ± (g)	k factor
1000 lb	8	-37.0	-0.9	5.9	45	2.28
1000 lb	1	-4.9	-4.9	5.9	45	2.28
1000 lb	5JIL	5.1	5.1	5.9	45	2.28
1000 lb	5JIK	25.1	25.1	5.9	45	2.28
1000 lb	2	-57.9	-10.9	5.9	45	2.28
2000 lb	Cart	-124.9	-124.9	6.7	227	2.28

Standards and Procedures used for testing:

The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7

All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, Issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:

The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:

The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows NISTIR 6969, Issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: Dave Fraser

David Fraser

Email: dfraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163

Company Name & Address:	Date of Test:	Test Number:
Lance Waller	6/29/2015	2015-075
Montana-Dakota Scale Service	Kit Number	
Route 1, Box 1640	No Kit #	
Fairview, MT 59221		

All results contained within this report only relate to the Item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 6/29/2015
Description and condition of artifacts received: Items were in good condition.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
22.98	24.23	667.51	40.1	42.73

Conventional Mass Value:		Assumed Density of Artifacts:			7.2 g/cm ³	
Nominal	Serial No.	As Found ± (g)	As Left ± (g)	Uncertainty ± (g)	NIST 105-1 Class F ± (g)	k factor
50 lb	2	-5.79	0.23	0.30	2.3	2.28
50 lb	20	-6.86	0.37	0.30	2.3	2.28
50 lb	4	-4.82	0.26	0.30	2.3	2.28
50 lb	12	-7.30	0.24	0.30	2.3	2.28
50 lb	8	-5.82	0.27	0.30	2.3	2.28
50 lb	10	-5.40	0.17	0.30	2.3	2.28
50 lb	6	-4.95	0.18	0.30	2.3	2.28
50 lb	21	-4.37	0.24	0.30	2.3	2.28
50 lb	16	-5.77	0.21	0.30	2.3	2.28

Standards and Procedures used for testing:

The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7

All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, Issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:

The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:

The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows *NISTIR 6969*, issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: *Dave Fraser*

David Fraser

Email: dafrazer@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163

Company Name & Address:		Date of Test:	Test Number:
Lance Waller		6/29/2015	2015-075
Montana-Dakota Scale Service		Kit Number	
Route 1, Box 1640		No Kit #	
Fairview, MT 59221			

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 6/29/2015
Description and condition of artifacts received: Items were in good condition.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
22.98	24.23	667.51	40.1	42.73

Conventional Mass Value:		Assumed Density of Artifacts:			7.2 g/cm ³	
Nominal	Serial No.	As Found ± (g)	As Left ± (g)	Uncertainty ± (g)	NIST 105-1 Class F ± (g)	k factor
50 lb	18	-6.16	0.25	0.30	2.3	2.28
50 lb	13	-5.63	0.61	0.30	2.3	2.28
50 lb	19	-5.51	0.40	0.30	2.3	2.28
50 lb	17	-5.39	-0.35	0.30	2.3	2.28
50 lb	15	-5.89	0.23	0.30	2.3	2.28
50 lb	22	-4.77	0.27	0.30	2.3	2.28
50 lb	11	-5.68	0.52	0.30	2.3	2.28
50 lb	1	-5.77	0.28	0.30	2.3	2.28
50 lb	14	-4.75	0.26	0.30	2.3	2.28

Standards and Procedures used for testing:

The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7

Traceability Statement:

The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:

The expanded uncertainty presented in this report is consistent with the 1993 ISO Guide to Expression of Uncertainty in Measurement and follows NISTIR 6969, Issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: Dove Fraser

David Fraser

Email: dfraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163

Company Name & Address:	Date of Test:	Test Number:
Lance Waller	6/30/2015	2015-076
Montana-Dakota Scale Service	Kit Number	
Route 1, Box 1640	25 lb	
Fairview, MT 59221		

All results contained within this report only relate to the Item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 6/29/2015
Description and condition of artifacts received: Items were in good condition with no discernable defects.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
22.56	23.58	668.53	42.54	44.72

Conventional Mass Value:			Assumed Density of Artifacts:		7.2 g/cm ³	
Nominal	Serial No.	As Found ±(g)	As Left ±(g)	Uncertainty ± (g)	NIST 105-1 Class F ± (g)	k factor
25 lb	1	-4.81	0.07	0.30	1.1	2.28
25 lb	2	-5.01	0.05	0.30	1.1	2.28
25 lb	3	9.97	0.11	0.30	1.1	2.28
25 lb	4	-4.11	0.13	0.30	1.1	2.28

Standards and Procedures used for testing:
The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7
All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, Issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:
The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:
The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows NISTIR 6969, issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: Dave Fraser

David Fraser

Email: dafraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
 2801 N Cooke St. Helena, Montana 59601
 (406)449-2582 FAX (406)443-8163

Company Name & Address: Lance Waller Montana-Dakota Scale Service Route 1, Box 1640 Fairview, MT 59221	Date of Test: 7/1/2015	Test Number: 2015-078
	Kit Number 200	

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 6/29/2015
 Description and condition of artifacts received: Items were in good condition with no discernable defects.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
24.56	25.48	665.23	42.26	44.54

Conventional Mass Value:		Assumed Density of Artifacts:		7.84 g/cm ³		
Nominal	Serial No.	As Found ±(g)	As Left ±(g)	Uncertainty ±(g)	NIST 105-1 Class F ±(g)	k factor
5 lb	A	0.068	0.068	0.017	0.23	2.28
5 lb	B	0.036	0.036	0.017	0.23	2.28
5 lb	C	0.043	0.043	0.017	0.23	2.28
5 lb	D	0.045	0.045	0.017	0.23	2.28
5 lb	E	0.069	0.069	0.017	0.23	2.28
1 lb	A	0.023	0.023	0.002	0.07	2.28
1 lb	B	0.022	0.022	0.002	0.07	2.28
1 lb	C	0.023	0.023	0.002	0.07	2.28
1 lb	D	0.030	0.030	0.002	0.07	2.28

Standards and Procedures used for testing:

The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7

All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, Issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:

The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:

The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows *NISTIR 6969*, Issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: Dave Fraser

David Fraser

Email: dfraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163

Company Name & Address:

Lance Waller
Montana-Dakota Scale Service
Route 1, Box 1640
Fairview, MT 59221

Date of Test:

7/1/2015

Test Number:

2015-078

Kit Number

200

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 6/29/2015

Description and condition of artifacts received: Items were in good condition with no discernable defects.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
24.56	25.48	665.23	42.26	44.54

Conventional Mass Value:
Assumed Density of Artifacts:
7.84 g/cm³

Nominal	Serial No.	As Found ±(g)	As Left ±(g)	Uncertainty ± (g)	NIST 105-1 Class F ± (g)	k factor
1 lb	E	0.017	0.017	0.0022	0.07	2.28
8 oz	1	0.009	0.009	0.0022	0.045	2.28
4 oz	1	0.005	0.005	0.00074	0.023	2.28
2 oz	1	0.003	0.003	0.00074	0.011	2.28
1 oz	1 oz	0.002	0.002	0.00074	0.0054	2.28
1/2 oz	1/2 oz	0.0010	0.0010	0.0003	0.003	2.28
1/4 oz	1/4 oz	0.0007	0.0007	0.0003	0.002	2.28
1/8 oz	1/8 oz	0.0003	0.0003	0.0003	0.0013	2.28
1/16 oz	1	0.00031	0.00031	0.00027	0.0011	2.28

Standards and Procedures used for testing:

The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7

All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:

The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:

The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows *NISTIR 6969*, issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: *Dave Fraser*

David Fraser

Email: dafraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163

Company Name & Address:

Lance Waller
Montana-Dakota Scale Service
Route 1, Box 1640
Fairview, MT 59221

Date of Test:

7/1/2015

Test Number:

2015-078

Kit Number

200

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received:

6/29/2015

Description and condition of artifacts received:

Items were in good condition with no discernable defects.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
20.34	21.77	665.23	42.10	42.60

Conventional Mass Value:

Assumed Density of Artifacts:

7.84 g/cm³

Nominal	Serial No.	As Found ±(g)	As Left ±(g)	Uncertainty ±(g)	NIST 105-1 Class F ±(g)	k factor
1/32 oz	1	0.00045	0.00045	0.00004	0.00087	2.28
1/32 oz	2	-0.00001	-0.00001	0.00004	0.00087	2.28

Standards and Procedures used for testing:

The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7

All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:

The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:

The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows *NISTIR 6969*, issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: Dave Fraser

David Fraser

Email: dafraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163

Company Name & Address:	Date of Test:	Test Number:
Lance Waller	7/1/2015	2015-081
Montana-Dakota Scale Service	Kit Number	
Route 1, Box 1640	10	
Fairview, MT 59221		

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 6/29/2015
Description and condition of artifacts received: Items were in good condition.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
23.66	24.15	664.72	49.5	47.3

Conventional Mass Value:			Assumed Density of Artifacts:		7.84 g/cm ³	
Nominal	Serial No.	As Found ±(g)	As Left ±(g)	Uncertainty ± (g)	NIST 105-1 Class F ± (g)	k factor
1 kg	10	0.024	0.024	0.012	0.1	2.28
500 g	10	0.032	0.032	0.0083	0.07	2.28
200 g	10	0.0102	0.0102	0.0051	0.04	2.28
200 g *	10	0.0055	0.0055	0.0051	0.04	2.28
100 g	10	0.00451	0.00451	0.00074	0.02	2.28
50 g	10	0.00323	0.00323	0.00073	0.01	2.28
20 g	10	0.00061	0.00061	0.00014	0.004	2.28
20 g *	10	-0.00001	-0.00001	0.00014	0.004	2.28
10 g	10	0.00055	0.00055	0.00014	0.002	2.28

Standards and Procedures used for testing:

The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7

All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:

The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:

The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows *NISTIR 6969*, issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: *Dave Fraser*

David Fraser

Email: dfraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163

Company Name & Address:	Date of Test:	Test Number:
Lance Waller	7/1/2015	2015-081
Montana-Dakota Scale Service		Kit Number
Route 1, Box 1640		10
Fairview, MT 59221		

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 6/29/2015

Description and condition of artifacts received: Items were in good condition.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
23.66	24.15	664.72	49.5	47.3

Conventional Mass Value: Assumed Density of Artifacts: 7.84 g/cm³

Nominal	Serial No.	As Found ±(g)	As Left ±(g)	Uncertainty ±(g)	NIST 105-1 Class F ±(g)	k factor
5 g	10	-0.00019	-0.00019	0.00027	0.0015	2.28
2 g	10	0.00041	0.00041	0.00027	0.0011	2.28
2 g*	10	0.00011	0.00011	0.00027	0.0011	2.28
1 g	10	0.000409	0.000409	0.000044	0.0009	2.28

Standards and Procedures used for testing:

The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7

All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:

The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:

The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows NISTIR 6969, Issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: Dave Fraser

David Fraser

Email: dfraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Montana

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.



2015

Scope

Mass Echelon III	Volume Transfer, II
30 kg to 1 mg	1500 gal to 5 gal
3000 lb to 0.001 lb	100 gal LPG to 25 gal LPG
8 oz to 0.03125 oz	
Weight Carts	
5000 lb to 2000 lb	

A handwritten signature in blue ink, appearing to read "Carol T. Hockert".

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2015-01-01 to 2015-12-31