



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY
 Public Service Commission
 SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company Zebra Technologies	Email Address JCPX46@zebra.com	Application Date 12-24-14	
Mailing Address 2N470 Prairie Ave	City Glen Ellyn	State IL	Zip Code 60137
Telephone Number 630-795-9788	Cell Phone Number 630-795-9788	Fax Number	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
e.g. 1001	e.g. John Doe	e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6
1739	Adrian Caraballo	Scales - 7

Continued on Page 2





List below all field standards (attach current calibration reports):

Rice Lake 12582	Serial # TS-1933



FOR SERVICE CALL 1-800-653-5350

REG # _____

TESTED & APPROVED _____ TECHNICIAN _____

MONTH			DAY						YEAR
JAN	FEB	MAR	1	2	3	4	5	6	15
APR	MAY	JUN	7	8	9	10	11	12	16
JUL	AUG	SEP	13	14	15	16	17	18	17
OCT	NOV	DEC	19	20	21	22	23	24	18
			25	26	27	28	29	30	31

Additional Application Items (initial where appropriate):


Standardized Test Report	<u> X </u>	Copy enclosed
	<u> </u>	No change in report filed previously
Tested and Approved Sticker	<u> X </u>	Copy enclosed
	<u> </u>	No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<u> NA </u>	Copy enclosed
	<u> NA </u>	No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.

Yes No

I am Adrian Caraballo, and have authority to represent this company. By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.




 Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410

North Dakota

nd.gov Official Portal for
North Dakota State GovernmentNorth Dakota
LEGISLATURE

SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

ZEBRA TECHNOLOGIES INTERNATIONAL, LLC

Corporation Details

System ID: 17942500**Phone:** (847) 634-6700**Type:** FOREIGN LIMITED LIABILITY COMPANY**Status:** Active & Good Standing**Original File Date:** 04/10/2002**Effective Date:** 04/10/2002**State of Origin:** Illinois

Nature of Business

SALES OF BAR CODE PRINTERS AND PRINTING MATERIALS

Principal Office

333 CORPORATE WOOD PKWY VERNON HILLS, IL 60061-3109

Registered Agent

C T CORPORATION SYSTEM

314 E THAYER AVE

BISMARCK, ND 58501-4018

Established Date: Aug 27, 2008

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2014](#) (generates a forms-fillable pdf in a new pop-up window)[Return to Search Results](#)[Contact Us](#)[Disclaimer](#) | [Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Illinois



2013-2014

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon I	Volume Transfer, II
5 kg to 1 mg	1500 gal to 1 gal
Mass Echelon II	100 gal to 20 gal LPG
5 kg to 1 mg	
Mass Echelon III	
30 kg to 1 mg	
2500 lb to 0.001 lb	
8 oz to 0.03125 oz	
Weight Carts	
5000 lb to 2500 lb	

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2013-01-01 to 2014-12-31



Pat Quinn, Governor – Robert Flider, Director
METROLOGY STANDARDS LABORATORY
NIST TRACEABLE Calibration Certificate

P.O. Box 19281 Springfield, IL 627949281 • 217/785-8480

Business Number: 800010
 W&M Invoice Number: 400809
 Purchase Order Number: N/A

Report of Test Number: 14-274
 Observer: 018
 Date Issued: 12-23-2014



Customer: Zebra Technologies
 Address: 2N470 Praire Avenue
 City & State: Glen Ellyn IL 60137

Date Received: 11-25-2014
 Date of Test: 12-18-2014
 Temperature: 20.9 °C
 Relative Humidity: 64.0%
 Barometric Pressure: 753.4 mmhg

Date of Primary Standard Calibration: N/A
 Date of Working Standard Calibration: 2010, 2012

Test Method: NISTIR 6969 - SOP 8
 Standards Used: Working Standards
 Description of Test Items: Rice Lake customary test weight kit - Serial number TS-1933, containing twenty-two items ranging 5 pound to 0.001 pound. Kit was sealed and decaled IL-14. Attention Adrian Caraballo.
 Condition of Test Items: Acceptable


<u>Nominal Mass Value</u>	<u>As Found Conventional Mass (mg)</u>	<u>As Left Conventional Mass (mg)</u>	<u>Uncertainty (mg)</u>	<u>Tolerance Class F</u>
5 lb-1	55.7	55.7	+/-26.7	In
5 lb-2	61.7	61.7	+/-26.7	In
5 lb-3	99.7	99.7	+/-26.7	In
5 lb-4	68.9	68.9	+/-26.7	In
5 lb-5	73.7	73.7	+/-26.7	In
1 lb-1	23.7	23.7	+/-8.5	In
1 lb-2	17.5	17.5	+/-8.5	In
1 lb-3	20.5	20.5	+/-8.5	In
1 lb-4	12.3	12.3	+/-8.5	In
1 lb-5	26.9	26.9	+/-8.5	In
0.5 lb	17.16	17.16	+/-5.32	In
0.2 lb	4.12	4.12	+/-2.10	In
0.2 lb*	6.16	6.16	+/-2.10	In
0.1 lb	3.05	3.05	+/-1.06	In
0.05 lb	0.71	0.71	+/-0.53	In
0.02 lb	0.49	0.49	+/-0.21	In
0.02 lb*	0.68	0.68	+/-0.21	In
0.01 lb	0.60	0.60	+/-0.18	In
0.005 lb	0.15	0.15	+/-0.15	In
0.002 lb	0.46	0.46	+/-0.11	In
0.002 lb*	0.43	0.43	+/-0.11	In
0.001 lb	0.28	0.28	+/-0.10	In

Michael W. Rockford

Michael W. Rockford
 Metrologist

The Illinois Standards have demonstrated metrological traceability to the international system of units (SI), which may be to the SI through a National Metrology Institute such as NIST and the laboratory has demonstrated measurement proficiency. The laboratory calculates measurement uncertainties in accordance with NIST-OWM Standard Operating Procedure 29, which is in conformance with ISO "Guide to the Expression of Uncertainty in Measurement." The confidence level is 95 % with a coverage factor of two (k = 2). This document shall not be reproduced, except in full, without the approval of the Illinois Laboratory.

US Service Test Report

 <p>2N470 Prairie Ave Glen Ellyn, IL 60137 Services Division SCALE TEST REPORT Service Technician Contact Information:</p>	SCALE MANUFACTURER Motorola Solutions	MODEL OF SCALE	APPROVAL No. 12-122A2
	SCALE INTERVAL <input checked="" type="checkbox"/> SINGLE <input type="checkbox"/> DUAL	SCALE UNITS US lb	DISPLAY TYPE <input type="checkbox"/> POS <input checked="" type="checkbox"/> Motorola
	MAX CAPACITY 30 lb	MIN DIVISION SIZE 0.01 lb	DEVICE SERIAL NO.
	SCALE TYPE <input checked="" type="checkbox"/> Bench Counter Class III Scale		LANE NUMBER
Name: Adrian Caraballo	<p>Check Applicable Box:</p> <input type="checkbox"/> New Install <input type="checkbox"/> Repair <input type="checkbox"/> Rejected Scale		
W&M ID No:	<p>Test Weight Information</p> Service Weight Kit No. Weight Kit Certification No. TS-1933 14-274		
Phone Number: 630-795-9788			
E-Mail: JCPX46@zebra.com			

CUSTOMER / STORE NAME: _____

STORE (street, city, state, zip): _____



TEST RESULTS

SCALE CONDITION AS FOUND: _____

ZERO BALANCE AS FOUND: _____

Increasing Load, Over Capacity, Decreasing Load, and Shift Test

Applied Load Single Interval	Scale Indication	All Tolerances are +/-		Applied Load Dual Interval	Scale Indication	All Tolerances are +/-	
		ACCEPT SINGLE INT	MAINT SINGLE INT			ACCEPT DUAL INT	MAINT DUAL INT
0.00		0.00 lb	.01 lb	0.000		0.000	.005 lb
0.10		.005 lb	.01 lb	0.100		.0025 lb	.005 lb
5.00		.005 lb	.01 lb	5.000		.005 lb	.01 lb
10.00		.01 lb	.02 lb	10.000		.005 lb	.01 lb
20.00		.01 lb	.02 lb	20.00		.01 lb	.02 lb
30.00		.015 lb	.03 lb	30.00		.015 lb	.03 lb
Over Capacity 30.20		N/A	N/A	Over Capacity 30.20		N/A	N/A
30.00		.015 lb	.03 lb	30.00		.015 lb	.03 lb
20.00		.01 lb	.02 lb	20.00		.01 lb	.02 lb
10.00		.01 lb	.02 lb	10.000		.005 lb	.01 lb
5.00		.005 lb	.01 lb	5.000		.005 lb	.01 lb
0.10		.005 lb	.01 lb	0.100		.0025 lb	.005 lb
0.00		0.00 lb	.01 lb	0.000		0.000	.005 lb
10.00	Pos. 1	.01 lb	.02 lb	10.00	Pos. 1	.005 lb	.01 lb
10.00	Pos. 2	.01 lb	.02 lb	10.00	Pos. 2	.005 lb	.01 lb
10.00	Pos. 3	.01 lb	.02 lb	10.00	Pos. 3	.005 lb	.01 lb
10.00	Pos. 4	.01 lb	.02 lb	10.00	Pos. 4	.005 lb	.01 lb

CHECK APPLICABLE TOLERANCE Acceptance - within 30 days of Calibration Maintenance - after 30 days from Calibration <input type="checkbox"/> Acceptance <input type="checkbox"/> Maintenance	THE ERRORS IN THIS SCALE AS INDICATED ABOVE <input type="checkbox"/> ARE <input type="checkbox"/> ARE NOT WITHIN THE ACCURACY REQUIREMENTS PRESCRIBED BY THE TOLERANCES IN NIST HANDBOOK-44
--	---

REPAIRS, ADJUSTMENTS, MODIFICATIONS OR RECOMMENDATIONS MADE AT THIS TIME: _____

Copy of This form to be submitted to Device Owner, Department of W&M, and Service Technician to retain a copy

APPLICABLE SEALING METHOD <input type="checkbox"/> PHYSICAL SEAL <input checked="" type="checkbox"/> ELECTRONIC AUDIT TRAIL	RECORD AUDIT TRAIL INFORMATION: CAL # _____ PAR # _____	DATE: _____	REJECTION TAG NO. _____
---	--	-------------	-------------------------

Service Technician Signature: _____ Customer / Store Manager Signature: _____