



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY

Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company <i>Main Line Measurement</i>	Email Address <i>Audie.Pankhurst@</i>	Application Date <i>6-4-15</i>	
Mailing Address <i>P.O. Box 4344</i>	City <i>Minot</i>	State <i>ND</i>	Zip Code <i>58702</i>
Telephone Number <i>337 936 1034</i>	Cell Phone Number <i>337 936 1034</i>	Fax Number	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input checked="" type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input checked="" type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
<i>e.g. 1001</i>	<i>e.g. John Doe</i>	<i>e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6</i>
<i>1731</i>	<i>Audie Pankhurst</i>	<i>1000 and 200 Gal Proveh</i>
<i>1728</i>	<i>Justin Pankhurst</i>	<i>1000 and 200 Gal Proveh</i>

Continued on Page 2



List below all field standards (attach current calibration reports):

1000 Cotton Bulk Prover	
200 Cotton Bulk Prover	

Additional Application Items (initial where appropriate):

Standardized Test Report	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No


I am Audie Parkhurst, and have authority to represent this company.
By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.

Audie Parkhurst
Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410

North Dakota

nd.gov Official Portal for
North Dakota State GovernmentNorth Dakota
LEGENDARY

SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

MAIN LINE MEASUREMENT L.L.C.

Corporation Details

System ID: 31557400 **Phone:** (337) 936-1034
Type: LIMITED LIABILITY COMPANY
Status: Active & Good Standing
Original File Date: 05/15/2012 **Effective Date:** 05/15/2012
State of Origin: North Dakota

Nature of Business

METER CALIBRATION SERVICES

Principal Office

2717 5TH ST NW MW MINOT, ND 58703-0722

Registered Agent

JACOB C MAXSON
TOWN & COUNTRY CENTER
1015 S BROADWAY STE 15
MINOT, ND 58701-4667
Established Date: May 15, 2012

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2014](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)

[Disclaimer](#)

[Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



Test Number
 G-000005184

REPORT OF CALIBRATION

FOR
 1-1000 Gallon Bulk Prover
 1-200 Gallon Bulk Prover



SUBMITTED BY
 Main Line Measurement LLC
 P.O. Box 4344
 Minot, North Dakota 58702

The standards of Texas are traceable to the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The data below applies only to the artifacts identified in this report at the time of test.

Test Date: 07/22/2015
Calibration Due: 07/31/2016

Temperature Range: 15 °C to 30 °C
Relative Humidity Range: 40 % to 60 %

Procedure: NISTIR 7383, SOP 19, Volume Transfer Method
Standard(s): Giddings Metrology Laboratory Echelon II Volume Transfer Standards

The expanded standard uncertainty includes the standard uncertainty reported for the standard, the standard uncertainty for the measurement process, and a component of uncertainty to account for any observed deviations that have a significant effect on the calibration. The expanded uncertainty given here is in compliance with NIST Technical Note 1297 ("Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results") with a $k = 2$, representing a 95.45 % confidence level.

This report is not to be used to claim product endorsement by the Texas Department of Agriculture or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of the Texas Department of Agriculture Metrology Laboratory.

Note:

- To convert from gallon to cubic inch: Multiply gallon by 231
- To convert from cubic inch to gallon: Divide cubic inch by 231
- To convert from cubic inch to cubic meter: Multiply cubic inch by 0.000016387064

for Philip Wright
 Director For Consumer Product Protection

Daniel Gibbons
 Metrologist



TEXAS DEPARTMENT OF AGRICULTURE
COMMISSIONER SID MILLER

Metrology Laboratory - 1258 CR 226 / P.O. Box 1518 - Giddings, Texas 78942
Phone: (979) 542-3231 - Fax: (888) 205-7741

REPORT OF CALIBRATION


Test Completed:
07/22/2015

For
1-1000 Gallon Bulk Prover
1-200 Gallon Bulk Prover

Test Number
G-000005184

Calibration Due:
07/31/2016

Submitted by
Main Line Measurement LLC
P.O. Box 4344
Minot, North Dakota 58702



The volumetric provers described below have been compared to the standards of the State of Texas and were found to deliver as follows:

Temperature Range: 15 °C to 30 °C
Relative Humidity Range: 40 % to 60 %
Procedure: NISTIR 7383, SOP 19, Volume Transfer Method

Standard(s): Giddings Metrology Laboratory Echelon II
Volume Transfer Standards

Volume (Gallon)	Serial / ID #	As Found Volume Delivered @ 60 °F (in ³)	As Left Volume Delivered @ 60 °F (in ³)	Expanded Uncertainty ± (in ³)	Coefficient of Expansion (1 / °F)	Average Water Temp. at Time of Test (°F)	Manufacturer
1000	1459	231002.47	230999.07	34.37	0.0000265	81.75	GS&S
200	1460	46197.61	46199.61	7.93	0.0000265	72.73	GS&S

To convert from gallon to cubic inch: multiply by 231
To convert from cubic inch to gallon: divide by 231

Above values apply when prover is level and the 30-second drainage method is used after cessation of the main flow. The above values do not apply if the calibration seals are tampered with or broken.

The expanded uncertainty given here is in compliance with NIST Technical Note 1297 ("Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results") with a coverage factor of two, representing a 95.45 % confidence level.

This report is not to be used to claim product endorsement by the Texas Department of Agriculture or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of the Texas Department of Agriculture Metrology Laboratory.

for Philip Wright
Director For Consumer Product Protection

Daniel Gibbons
Metrologist

United States Department of Commerce National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Texas



This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

2015-2016

Mass Echelon III
1000 kg to 1 mg
5000 lb to 0.001 lb
12 oz to 0.03125 oz

Weight Carts
6000 lb to 2500 lb

Volume Transfer, II
1000 gal to 5 gal
300 gal to 25 gal LPG

A handwritten signature in blue ink, appearing to read "Carol T. Hockerl", is written over a horizontal line.

Carol T. Hockerl, Chief
NIST Office of Weights and Measures

Effective Dates: 2015-01-01 to 2016-12-31