

# Scale Test Report

North Dakota Public Service Commission  
 Testing and Safety Division  
 SFN 19243 (5/2002)

Invoice# **WG 58954**

Operator# 465

Inspected Date: 5/26/2016 12:30:44

Date Last Tested: 2/24/2016

**Cenex Harvest States**  
**225 E A Avenue Box 339**  
**Lakota ND 58344-0339**

Phone: 701 247-2917

Variance Posted:

Contact:

E Mail:

Item Code: 3V6 1

Location Of Device: inbound; enclosed

Manufacturer: RLWS

Indicating Element: 920i

Capacity: 160,000 x 20

Serial #: **1731900003/5WRT**

Class: IILL

Lever System/Weighing Elements: 12 - 75K Load Cells

Platform Size: 11 x 117

Pit Depth: 12.5"

Approaches: 12'/28' Concrete

Printer: Computer System

Printer SN: NA

Environmental Conditions:

## Test Results

Initial Zero Test: 0

### Section Test

Section	Test Weight	Error
1	20,000	0
2	20,000	0
3	20,000	0
4	20,000	0/20
5	20,000	0/20
6	20,000	0

Balance Check: 0

### Sensitivity Response OR Discrimination Test

ZeroLoad	NA
Loaded	NA

### Strain Test

### Shift Test

Load Pt.	Test Weight	Error
1	10,000	0
2	10,000	0
3	10,000	0
4	10,000	0
5	10,000	0
6	10,000	0
7	10,000	0
8	10,000	0
9	10,000	0
10	10,000	0
11	10,000	0
12	10,000	0

Balance Check: 0

	Wt Full	Test Weight	Wt Empty	Error
1	52140	20,000	32160	20
2	52140	20,000	32160	20
3		20,000		
4		20,000		

Remarks:

**9 WM-15-427** Filed: 5/26/2016 Pages: 2  
**Test report**

Public Service Commission

# Scale Test Report

North Dakota Public Service Commission

Testing and Safety Division

SFN 19243 (5/2002)

Invoice# **WG 58954**

---

Remarks.

Scale is: **QC - Passed**

Inspected By **Grangaard**