

**Invoice**

North Dakota Public Service Commission - Tax ID 450309764

Testing and Safety Division

SFN 19242 (5/2002)

Invoice# **WG 59109**

Operator# 465

**Cenex Harvest States**

**225 E A Avenue Box 339**

**Lakota ND 58344-0339**

Phone: 701 247-2917

Contact:

E Mail:

Billed To: 465

**Cenex Harvest States**

**225 E A Avenue Box 339**

**Lakota ND 58344-0339**

Phone: 701 247-2917

Contact:

E Mail:

Time In: 4/20/2017 10:04 am Out: 11:53 am

Approval Seals: 18283

Inspection Record					Device Comments
	Status	Fee	Qty.	Fee Total	
Vehicle 6 Section	QC-Passed	0.00	1	\$0.00	
				<b>\$0.00</b>	

**Amount Due: \$0.00**

**Note: qa-pass(429)**

Accounts are due in thirty days. This account is due on May 20, 2017.

Inspected By: **Grangaard**

11 **WM-15-427** Filed: 4/20/2017 Pages: 3  
Test report

Public Service Commission

**Please remit payment to :** Public Service Commission  
600 E. Boulevard Ave Dept 408  
Bismarck ND 58505-0480

# Scale Test Report

North Dakota Public Service Commission  
Testing and Safety Division  
SFN 19243 (5/2002)

Invoice# **WG 59109**

Operator# 465

Inspected Date: 4/20/2017 11:32:53

Date Last Tested: 2/15/2017

**Cenex Harvest States**

**225 E A Avenue Box 339**

**Lakota ND 58344-0339**

Phone: 701 247-2917

Contact:

E Mail:

Item Code: 3V6 1

Location Of Device: inbound; enclosed

Manufacturer: RLWS

Indicating Element: 920i Capacity: 160,000 x 20

Serial #: **1731900003/5WRT** Class: III L

Lever System/Weighing Elements: 12 - 75K Load Cells

Platform Size: 11 x 117 Pit Depth: 12.5"

Approaches: 12'/28' Concrete

Printer: Computer System Printer SN: NA

Environmental Conditions:

Variance Posted:

## Test Results

Initial Zero Test: 0

### Section Test

Section	Test Weight	Error
1	20,000	-20
2	20,000	0
3	20,000	0
4	20,000	20
5	20,000	20
6	20,000	-20

Balance Check: 0

### Sensitivity Response OR Discrimination Test

ZeroLoad	NA
Loaded	NA

Strain Test

### Shift Test

Load Pt.	Test Weight	Error
1	10,000	0
2	10,000	0
3	10,000	0
4	10,000	0
5	10,000	0
6	10,000	0
7	10,000	0
8	10,000	0
9	10,000	0
10	10,000	0
11	10,000	0
12	10,000	0

Balance Check: 0

	Wt Full	Test Weight	Wt Empty	Error
1	51660	20,000	31680	20
2	51640	20,000	31660	20
3		20,000		
4		20,000		

Remarks:

# Scale Test Report

North Dakota Public Service Commission

Testing and Safety Division

SFN 19243 (5/2002)

Invoice# **WG 59109**

---

Remarks.

Scale is: **QC - Passed**

Inspected By **Grangaard**