

Scale Test Report

North Dakota Public Service Commission
Testing and Safety Division
SFN 19243 (5/2002)

Invoice# **WG 59063**

Operator# 9159 Inspected Date: 1/18/2017 15:55:45 Date Last Tested: 8/8/2016

Remington Seeds, LLC
1241 W 12th St
Grafton ND 58237

Phone: 701-379-1000

Variance Posted:

Contact: Kevin Wick

E Mail: kevin.wick@remingtonseeds.com

Item Code: 3V5 2

Location Of Device:

Manufacturer: Cardinal

Indicating Element: 225 Capacity: 200,000 x 20

Serial #: **E30114-0083** Class: IIII

Lever System/Weighing Elements: FLC

Platform Size: 70' Pit Depth: ~6"

Approaches: 12'+ concrete

Printer: Epson Printer SN: J9KF141824

Environmental Conditions:

Test Results

Initial Zero Test: 0

Section Test

Section	Test Weight	Error
1	20,000	
2	20,000	-100/120
3	20,000	20
4	20,000	20
5	20,000	20

Balance Check: 0

Sensitivity Response OR Discrimination Test

ZeroLoad	NA
Loaded	NA

Strain Test

Shift Test

Load Pt.	Test Weight	Error
1	10,000	
2	10,000	
3	10,000	
4	10,000	
5	10,000	
6	10,000	
7	10,000	
8	10,000	
9	10,000	
10	10,000	

Balance Check: 0

	Wt Full	Test Weight	Wt Empty	Error
1	51580	20,000	31540	40
2		20,000		
3		20,000		
4		20,000		

13 **WM-15-562** Filed: 1/18/2017 Pages: 3
Test report

Public Service Commission

Remarks: **Red Tag: Cannot be used commercially until repaired (>1D out of tol.); NO LEAD SEAL**

Scale is: **W**

Scale Test Report

North Dakota Public Service Commission

Testing and Safety Division

SFN 19243 (5/2002)

Invoice# **WG 59063**

Inspected By **Grangaard**

Invoice

North Dakota Public Service Commission - Tax ID 450309764

Testing and Safety Division

SFN 19242 (5/2002)

Invoice# **WG 59063**

Operator# 9159

Remington Seeds, LLC**1241 W 12th St****Grafton ND 58237**

Phone: 701-379-1000

Contact: Kevin Wick

E Mail: kevin.wick@remingtonseeds.com

Billed To: 9159

Remington Seeds, LLC**1241 W 12th St****Grafton ND 58237**

Phone: 701-379-1000

Contact: Kevin Wick

E Mail: kevin.wick@remingtonseeds.com

Time In: 1/18/2017 1:41 pm Out: 4:29 pm

Approval Seals: 18252

Inspection Record					Device Comments		
	Status	Fee	Qty.	Fee Total		Serial Number:	Seal Number:
Vehicle 5 Section	W	0.00	1	\$0.00	Vehicle 5 Section	E30114-0083	W
Hanging	QC-Passed	0.00	1	\$0.00	Red Tag: Cannot be used commercially until repaired (>1D out of tol.); NO LEAD SEAL		
				\$0.00			

Amount Due: \$0.00

Accounts are due in thirty days. This account is due on February 17, 2017.

Note: qa-fail / tolerance (3V5-2)(390)No Lead Seal &Red Tag: Cannot be used commercially until repaired (>1D out of tol.); qa-pass(9HA-1)(389)Inspected By: **Grangaard**

Please remit payment to : Public Service Commission
600 E. Boulevard Ave Dept 408
Bismarck ND 58505-0480