

Date 01/26/2017	PSC Device Code 3V5	No. of Sections 5	Complete each item with Y (Yes), N (No), or NA (Not Applicable)	
Name of Business Remington Seeds, LLC			<input type="checkbox"/> New Installation (w/ RFI Check)	<input type="checkbox"/> Performed Calibration
Mailing Address 1421 West 12th St. Grafton ND 58237			<input type="checkbox"/> Modified Equipment	<input type="checkbox"/> Use as a Reference Scale
Device Location Grafton PO Box 605 1421 12th St			<input type="checkbox"/> Replaced Existing Equipment	<input type="checkbox"/> Non-Commercial
City Grafton	State ND	Zip Code 58237	<input checked="" type="checkbox"/> Scale Indoors	
County Walsh	Telephone Number 701-379-1000		<input checked="" type="checkbox"/> Variance Permit Posted; Expiration Date: 08/31/2017	
Device Contact/Manager Kevin Wick	Cell Number		<input checked="" type="checkbox"/> Stored/Recalled Weights meet NDAC 69-10-02-23 & 69-10-02-24	
Email Address kevin.wick@remingtonseeds.com			<input checked="" type="checkbox"/> Software is NTEP Approved	
			<input checked="" type="checkbox"/> Built-In Standards; Date Certified:	
			<input type="checkbox"/> Multiple Decks/Single Indicator	<input checked="" type="checkbox"/> Customer Has Clear View
			<input checked="" type="checkbox"/> Clearance Below Scale Clear	<input checked="" type="checkbox"/> Video Camera Working
			<input checked="" type="checkbox"/> Approach Requirements Met	<input type="checkbox"/> View Distance > 200'
			<input checked="" type="checkbox"/> Pit Coping and Crush Strip Good	<input checked="" type="checkbox"/> Two-Way Audio Working
			<input type="checkbox"/> In response to PSC Quality Assurance Inspection	

Scale Information - East - 011147

Scale Manufacturer Cardinal	Indicator Manufacturer/Model 225 Navigator	Indicator Serial No. E30014-0300	Weighing Elements EL.
Capacity/Divisions/Units 200,000 x 20 LB	Legible Label X Y _ N	Class III	Printer Manufacturer/Model Epson TM-U295
Deck Size 80' X 11'	Clearance (inches) 6"	Approach - 12' Concrete Level 12'	Printer Serial No. J9KF116147
SR or Discrimination Test Zero Load = Loaded =	Motion Detection Range = 3d	Hard Surface Approach (Length/Slope/Condition) Good ,Concrete	
	AZSM (Auto Zero) Range = 3d		

Test Data

LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-	
			As Found	As Left				As Found	As Left
Description: Corner Test					Description: Section Test				
LBP									
1	10,000 lb		0	0	1	20,000 lb		0	0
2	10,000 lb		0	0	2	20,000 lb		0	0
3	10,000 lb		0	0	3	20,000 lb		0	0
4	10,000 lb		0	0	4	20,000 lb		0	0
5	10,000 lb		0	0	5	20,000 lb		0	0
6	10,000 lb		0	0					
7	10,000 lb		0	0					
8	10,000 lb		0	0		BC=0			
9	10,000 lb		0	0					
10	10,000 lb		0	0					
	BC=0								

Strain Load Test

Section (Increasing or Decreasing)	4 (Inc)	Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary. Semi-Annual test of the East truck scale. No adjustments were made, scale was found in tolerance. Variance is posted.
Full Truck Weight	59,680	
Test Weight	20,000	
Empty Truck Weight	39,680	
Errors	0	

Physical Seal Y N Electronic Audit Trail: Y N
 Seal Date: 01/26/2017 Audit Trail Information:
 Seal Type: Lead

Meets tolerances in PSC adopted NIST Handbook 44? Maintenance_ Acceptance
 Approved Rejected Taken out of Service Sticker Applied

By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.

M. Q. Smith
 Permit Holder Signature _____ 1514
 Permit Number
Kevin Wick
 Operator Signature _____ 01/26/2017
 Date