

5/23/2017

Invoice

North Dakota Public Service Commission - Tax ID 450309764

Testing and Safety Division

SFN 19242 (5/2002)

Invoice# **WG 59125**

Operator# 9159

Remington Seeds, LLC

1241 W 12th St

Grafton ND 58237

Phone: 701-379-1000

Contact: Kevin Wick

E Mail: kevin.wick@remingtonseeds.com

Billed To: 9159

Remington Seeds, LLC

1241 W 12th St

Grafton ND 58237

Phone: 701-379-1000

Contact: Kevin Wick

E Mail: kevin.wick@remingtonseeds.com

Time In: 5/23/2017 2:31 pm Out: 4:49 pm

Approval Seals: 18298 to 18299

Inspection Record					Device Comments
	Status	Fee	Qty.	Fee Total	
Vehicle 5 Section	QC-Passed	0.00	2	\$0.00	
				\$0.00	

Amount Due: \$0.00

Accounts are due in thirty days. This account is due on June 22, 2017.

Note: qa-pass(452,453) ; Recommend variances be set to permanent (ending 8/31/17)

Inspected By: **Grangaard**

16 WM-15-562 Filed: 5/23/2017 Pages: 3
Test reports

Public Service Commission

Please remit payment to : Public Service Commission
600 E. Boulevard Ave Dept 408
Bismarck ND 58505-0480

Scale Test Report

North Dakota Public Service Commission
Testing and Safety Division
SFN 19243 (5/2002)

Invoice# **WG 59125**

Operator# 9159

Inspected Date: 5/23/2017 16:19:00

Date Last Tested: 1/26/2017

Remington Seeds, LLC

1241 W 12th St

Grafton ND 58237

Phone: 701-379-1000

Variance Posted:

Contact: Kevin Wick

E Mail: kevin.wick@remingtonseeds.com

Item Code: 3V5 1

Location Of Device: east scale

Manufacturer: Cardinal

Indicating Element: 225

Capacity: 200,000 x 20

Serial #: **E30014-0300**

Class: III L

Lever System/Weighing Elements: FLC

Platform Size: 70'

Pit Depth: 6"

Approaches: good 12' concrete

Printer: Epson

Printer SN: J9KF116147

Environmental Conditions:

Test Results

Initial Zero Test: 0

Section Test

Section	Test Weight	Error
1	20,000	0
2	20,000	0
3	20,000	-20
4	20,000	0
5	20,000	0

Balance Check: 0

Sensitivity Response OR Discrimination Test

ZeroLoad	NA
Loaded	NA

Strain Test

Shift Test

Load Pt.	Test Weight	Error
1	10,000	0
2	10,000	0
3	10,000	0
4	10,000	0
5	10,000	0
6	10,000	0
7	10,000	0
8	10,000	0
9	10,000	0
10	10,000	0

Balance Check: 0

	Wt Full	Test Weight	Wt Empty	Error
1	51460	20,000	31460	0
2		20,000		
3		20,000		
4		20,000		

Remarks:

Scale is: **QC - Passed**

Scale Test Report

North Dakota Public Service Commission
Testing and Safety Division
SFN 19243 (5/2002)

Invoice# **WG 59125**

Operator# 9159

Inspected Date: 5/23/2017 16:21:39

Date Last Tested: 1/19/2017

Remington Seeds, LLC

1241 W 12th St

Grafton ND 58237

Phone: 701-379-1000

Variance Posted:

Contact: Kevin Wick

E Mail: kevin.wick@remingtonseeds.com

Item Code: 3V5 2

Location Of Device: west scale

Manufacturer: Cardinal

Indicating Element: 225

Capacity: 200,000 x 20

Serial #: **E30114-0083**

Class: III L

Lever System/Weighing Elements: FLC

Platform Size: 70'

Pit Depth: ~6"

Approaches: 12'+ concrete

Printer: Epson

Printer SN: J9KF141824

Environmental Conditions:

Test Results

Initial Zero Test: 0

Section Test

Section	Test Weight	Error
1	20,000	0
2	20,000	0
3	20,000	-20
4	20,000	0
5	20,000	0

Balance Check: 0

Sensitivity Response OR Discrimination Test

ZeroLoad	NA
Loaded	NA

Strain Test

Shift Test

Load Pt.	Test Weight	Error
1	10,000	0
2	10,000	0
3	10,000	0
4	10,000	0
5	10,000	0
6	10,000	0
7	10,000	0
8	10,000	0
9	10,000	0
10	10,000	0

Balance Check: 0

	Wt Full	Test Weight	Wt Empty	Error
1	51500	20,000	31500	0
2		20,000		
3		20,000		
4		20,000		

Remarks:

Scale is: **QC - Passed**