

PRAIRIE SCALE SYSTEMS, INC.
 P.O. Box 69 Horace, ND 58047-0069 701-281-9591



Invoice: CR-101335
 Scale ID: 011145

Date 08/07/2017	PSC Device Code 3V2	No. of Sections 2	Complete each item with Y (Yes), N (No), or NA (Not Applicable)	
Name of Business Remington Seeds, LLC			N New Installation (w/ RFI Check)	N Performed Calibration
Mailing Address 1421 West 12th St. Grafton ND 58237			N Modified Equipment	N Use as a Reference Scale
Device Location Grafton PO Box 605 1421 12th St			N Replaced Existing Equipment	N Non-Commercial
City Grafton	State ND	Zip Code 58237	Y Variance Permit Posted; Expiration Date: 08/31/2017	
County Walsh	Telephone Number 701-379-1000		NA Stored/Recalled Weights meet NDAC 69-10-02-23 & 69-10-02-24	
Device Contact/Manager Kevin Wick		Cell Number	NA Software is NTEP Approved	
Email Address kevin.wick@remingtonseeds.com			NA Built-In Standards; Date Certified:	
			Y Multiple Decks/Single Indicator	Y Customer Has Clear View
			NA Clearance Below Scale Clear	NA Video Camera Working
			Y Approach Requirements Met	NA View Distance > 200'
			Y Pit Coping and Crush Strip Good	NA Two-Way Audio Working
			NA In response to PSC Quality Assurance Inspection	

Scale Information - West #1 -011145

Scale Manufacturer Cardinal	Indicator Manufacturer/Model 225 Navigator	Indicator Serial No. E30114-0083 / E17415-0365	Weighing Elements EL.
Capacity/Divisions/Units 200,000 x 20 LB	Legible Label X Y _ N	Class III	Printer Manufacturer/Model Epson TM-U295
Deck Size 20' x 11' South	Clearance (inches) 6"	Approach - 12' Concrete Level 12'	Printer Serial No. J9KF141825 / J9KF112051
SR or Discrimination Test Zero Load = Loaded =	Motion Detection Range = 60 lbs.	Hard Surface Approach (Length/Slope/Condition) Good, concrete	
		AZSM (Auto Zero) Range = 60 lbs.	

Test Data

LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-	
			As Found	As Left				As Found	As Left
Description: Corner Test					Description: Section Test				
1	10,000 lbs.		0	0	1	20,000 lbs.		0	0
2	"		0	0	2	20,000 lbs.		0	0
3	"		0	0					
4	10,000 lbs.		0	0		BC=0			
	BC=0								

Strain Load Test

Section (Increasing or Decreasing)	Increasing Two	Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary. Inspected and tested scale, found no adjustments needed at this time.
Full Truck Weight	35,700	
Test Weight	20,000	
Empty Truck Weight	15,700	
Errors	0	
Physical Seal <input checked="" type="checkbox"/> Y <input type="checkbox"/> N	Electronic Audit Trail: <input type="checkbox"/> Y <input type="checkbox"/> N	
Seal Date: 08/07/2017	Audit Trail Information:	
Seal Type: Lead		
Meets tolerances in PSC adopted NIST Handbook 44? <input checked="" type="checkbox"/> Maintenance <input type="checkbox"/> Acceptance		
<input checked="" type="checkbox"/> Approved <input type="checkbox"/> Rejected <input type="checkbox"/> Taken out of Service <input checked="" type="checkbox"/> Sticker Applied		
By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.		
Cady Bennett	1718	
Permit Holder Signature	Permit Number	
	08/07/2017	
Operator Signature	Date	



PRAIRIE SCALE SYSTEMS, INC.
 P.O. Box 69 Horace, ND 58047-0069 701-281-9591

Invoice: CR-101335
 Scale ID: 011146

Date 08/07/2017	PSC Device Code 3V3	No. of Sections 3	Complete each item with Y (Yes), N (No), or NA (Not Applicable)	
Name of Business Remington Seeds, LLC			N New Installation (w/ RFI Check)	N Performed Calibration
Mailing Address 1421 West 12th St. Grafton ND 58237			N Modified Equipment	N Use as a Reference Scale
Device Location Grafton PO Box 605 1421 12th St			N Replaced Existing Equipment	N Non-Commercial
City Grafton	State ND	Zip Code 58237	Y Variance Permit Posted; Expiration Date: 08/31/2017	
County Walsh	Telephone Number 701-379-1000		NA Stored/Recalled Weights meet NDAC 69-10-02-23 & 69-10-02-24	
Device Contact/Manager Kevin Wick		Cell Number	NA Software is NTEP Approved	
Email Address kevin.wick@remingtonseeds.com			NA Built-In Standards; Date Certified:	
			Y Multiple Decks/Single Indicator	Y Customer Has Clear View
			NA Clearance Below Scale Clear	NA Video Camera Working
			Y Approach Requirements Met	NA View Distance > 200'
			Y Pit Coping and Crush Strip Good	NA Two-Way Audio Working
NA In response to PSC Quality Assurance Inspection				

Scale Information - West #2 - 011146

Scale Manufacturer Cardinal	Indicator Manufacturer/Model 225 Navigator	Indicator Serial No. E30114-0083 / E17415-0365	Weighing Elements EL
Capacity/Divisions/Units 200,000 x 20 LB	Legible Label X Y _ N	Class III	Printer Manufacturer/Model Epson TM-U295
Deck Size 40' x 11' North	Clearance (inches) 6"	Approach - 12' Concrete Level 12'	Printer Serial No. J9KF141825 / J9KF112051
SR or Discrimination Test Zero Load = Loaded =	Motion Detection Range = 60 lbs.		Hard Surface Approach (Length/Slope/Condition)
AZSM (Auto Zero) Range = 60 lbs.			

Test Data

LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-	
			As Found	As Left				As Found	As Left
Description: Corner Test					Description: Section Test				
1	10,000 lbs.		0	0	1	20,000 lbs.		0	0
2	"		0	0	2	"		0	0
3	"		0	0	3	20,000 lbs.		0	0
4	"		0	0					
5	"		0	0		BC=0			
6	10,000 lbs.		0	0					
	BC=0								

Strain Load Test

Section (Increasing or Decreasing)	Increasing Three	Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary. Inspected and tested scale, found no adjustments needed at this time. All three scales have a variance for clearance between I beam and finished work floor.
Full Truck Weight	51,700	
Test Weight	20,000	
Empty Truck Weight	31,700	
Errors	0	
Physical Seal X Y _ N	Electronic Audit Trail: _ Y _ N	
Seal Date: 08/07/2017	Audit Trail Information:	
Seal Type: Lead		
Meets tolerances in PSC adopted NIST Handbook 44? X Maintenance_ Acceptance		
X Approved _ Rejected _ Taken out of Service X Sticker Applied		
By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.		
Cody Binnick	1718	
Permit Holder Signature	Permit Number	
	08/07/2017	
Operator Signature	Date	



PRAIRIE SCALE SYSTEMS, INC.
 P.O. Box 69 Horace, ND 58047-0069 701-281-9591

Invoice: CR-101335
Scale ID: 011147

Date 08/07/2017	PSC Device Code 3V5	No. of Sections 5	Complete each item with Y (Yes), N (No), or NA (Not Applicable) <input type="checkbox"/> New Installation (w/ RFI Check) <input type="checkbox"/> Performed Calibration <input type="checkbox"/> Modified Equipment <input type="checkbox"/> Use as a Reference Scale <input type="checkbox"/> Replaced Existing Equipment <input type="checkbox"/> Non-Commercial <input type="checkbox"/> Scale Indoors <input type="checkbox"/> Variance Permit Posted; Expiration Date: 08/31/2017 <input type="checkbox"/> NA Stored/Recalled Weights meet NDAC 69-10-02-23 & 69-10-02-24 <input type="checkbox"/> NA Software is NTEP Approved <input type="checkbox"/> NA Built-In Standards; Date Certified: <input type="checkbox"/> Multiple Decks/Single Indicator <input type="checkbox"/> Customer Has Clear View <input type="checkbox"/> NA Clearance Below Scale Clear <input type="checkbox"/> NA Video Camera Working <input type="checkbox"/> Y Approach Requirements Met <input type="checkbox"/> NA View Distance > 200' <input type="checkbox"/> Y Pit Coping and Crush Strip Good <input type="checkbox"/> NA Two-Way Audio Working <input type="checkbox"/> NA In response to PSC Quality Assurance Inspection
Name of Business Remington Seeds, LLC			
Mailing Address 1421 West 12th St. Grafton ND 58237			
Device Location Grafton PO Box 605 1421 12th St			
City Grafton	State ND	Zip Code 58237	
County Walsh	Telephone Number 701-379-1000		
Device Contact/Manager Kevin Wick		Cell Number	
Email Address kevin.wick@remingtonseeds.com			

Scale Information - East - 011147

Scale Manufacturer Cardinal	Indicator Manufacturer/Model 225 Navigator	Indicator Serial No. E30014-0300	Weighing Elements EL
Capacity/Divisions/Units 200,000 x 20 LB	Legible Label X Y _ N	Class IIIL	Printer Manufacturer/Model Epson TM-U295
Deck Size 80' X 11'	Clearance (inches) 6"	Approach - 12' Concrete Level 12'	Printer Serial No. J9KF116147
SR or Discrimination Test Zero Load = Loaded =	Motion Detection Range = 3d		Hard Surface Approach (Length/Slope/Condition) Good ,Concrete AZSM (Auto Zero) Range = 3d

Test Data

LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-		LBP / Section / Product Wt.	Value of Test Weights Used	Serial/Test Number	Errors +/-	
			As Found	As Left				As Found	As Left
Description: Corner Test					Description: Section Test				
1	10,000 lb		-20	0	1	20,000 lb		-20	0
2	10,000 lb		0/-	0	2	20,000 lb		-40	0
3	10,000 lb		0/-	0	3	20,000 lb		-40	0
4	10,000 lb		0/-	0	4	20,000 lb		-40	0
5	10,000 lb		+20	0	5	20,000 lb		-20	0
6	10,000 lb		0	0					
7	10,000 lb		0	0					
8	10,000 lb		-20	0		BC=0			
9	10,000 lb		-40	0					
10	10,000 lb		-20	0					

Strain Load Test

Section (Increasing or Decreasing)	Increasing Five	Remarks (include environmental conditions, if applicable). Attach additional sheets as necessary. Inspected and tested scale, found all section running minus. Re-calibrated and adjusted corners as needed. Found no other adjustments needed at this time.
Full Truck Weight	59,960 lbs	
Test Weight	20,000 lbs	
Empty Truck Weight	39,960 lbs	
Errors	0	
Physical Seal <input checked="" type="checkbox"/> Y <input type="checkbox"/> N Electronic Audit Trail: <input type="checkbox"/> Y <input type="checkbox"/> N		
Seal Date: 08/07/2017 Audit Trail Information:		
Seal Type: Lead		
Meets tolerances in PSC adopted NIST Handbook 44? <input checked="" type="checkbox"/> Maintenance <input type="checkbox"/> Acceptance		
<input checked="" type="checkbox"/> Approved <input type="checkbox"/> Rejected <input type="checkbox"/> Taken out of Service <input checked="" type="checkbox"/> Sticker Applied		
By signing this test report, I declare that I have examined this report and to the best of my knowledge and belief, the report is complete and the facts stated are sufficient, true and correct. I also declare that, except for conditions noted in "Remarks," the device meets the minimum requirements of the State laws and rules, including NIST Handbook 44, for use of the device in commerce.		
Cody Bannick	1718	
Permit Holder Signature	Permit Number	
	08/07/2017	
Operator Signature	Date	