



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY

Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company Kennedy Scales, Inc	Email Address RBaker@KennedyScales.com	Application Date 12/22/2015	
Mailing Address 11485 Xeon Street NW	City Coon Rapids	State MN	Zip Code 55448
Telephone Number 763-784-1022	Cell Phone Number 612-965-3914	Fax Number 763-784-1206	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input checked="" type="checkbox"/> 1. Rail <input checked="" type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input checked="" type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
e.g. 1001	e.g. John Doe	e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6
1295	Tim Seifert	Scales 1, 2, 4, 6, 7



List below all field standards (attach current calibration reports):

Class F 25 LB Cast Weight	25121-25160
30 LB Kit Decimal/Fractional	A6
Metric Kit	M6
1000 LB Cast Cube	KSHD 1 - KSHD 18
Tiffin Weight Cart - 3000 LB	W60885
Class F 50 LB Cast Weight	L1 - L20


Additional Application Items (initial where appropriate):

Standardized Test Report	<input checked="" type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input checked="" type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input checked="" type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am Rob Baker, and have authority to represent this company.
By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.


Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410



Receipt Date: October 21, 2015
Test Date: October 21, 2015
Report Date: October 21, 2015

State Test No.: 334968
Set Serial No.: W60885
Barcode: 202397

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448

Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 11

Item(s) Submitted: 3000 lb Weight Cart
Manufacturer: Tiffin
ASTM E617 Type: NA
Equipment ID#: None
Condition: Good
Temperature: 20.5 °C
Pressure: 738.6 mmHg
Relative Humidity: 45. %

Nominal Value	Serial No.	Correction (g)		NIST HB105-8 Tol		Unc. (g) (k=2)
		As Found	As Left	As Found	As Left	
3000 lb	W60885	100	100	Meets	Meets	60.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Pete Whebbe

Pete Whebbe
Metrologist

Reviewed by:

Mark Nicollet

Mark Nicollet
Quality Manager



Receipt Date: October 19, 2015
Test Date: October 19, 2015
Report Date: October 19, 2015

State Test No.: 334949
Set Serial No.: KSHD 1 to KSHD 18
Barcode: 200647

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448

Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 11

Item(s) Submitted: Cast Cube Weights
Manufacturer: Toledo
ASTM E617 Type: II
Equipment ID#: None
Condition: Good
Temperature: 20.2 °C
Pressure: 731.7 mmHg
Relative Humidity: 35. %

Nominal Value	lb	Serial No.	Correction (g)		NIST HB105-1 Class		Unc. (g) (k=2)
			As Found	As Left	As Found	As Left	
1000	lb	KSHD1	7.	7.	F	F	5.
1000	lb	KSHD2	-13.	-13.	F	F	5.
1000	lb	KSHD3	-13.	-13.	F	F	5.
1000	lb	KSHD4	-5.	-5.	F	F	5.
1000	lb	KSHD5	-17.	-17.	F	F	5.
1000	lb	KSHD6	0.	0.	F	F	5.
1000	lb	KSHD7	5.	5.	F	F	5.
1000	lb	KSHD8	6.	6.	F	F	5.
1000	lb	KSHD9	-22.	-22.	F	F	5.
1000	lb	KSHD10	1.	1.	F	F	5.
1000	lb	KSHD11	-26.	-26.	F	F	5.
1000	lb	KSHD12	-16.	-16.	F	F	5.
1000	lb	KSHD13	-12.	-12.	F	F	5.
1000	lb	KSHD14	-10.	-10.	F	F	5.
1000	lb	KSHD15	-9.	-9.	F	F	5.
1000	lb	KSHD16	-1.	-1.	F	F	5.
1000	lb	KSHD17	-15.	-15.	F	F	5.
1000	lb	KSHD18	15.	15.	F	F	5.

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Pete Whebbe

Pete Whebbe
Metrologist

Reviewed by:

Mark Nicollet

Mark Nicollet
Quality Manager



Receipt Date: October 19, 2015
Test Date: October 20, 2015
Report Date: October 20, 2015

State Test No.: 334950
Set Serial No.: L1-L20
Barcode: 200648

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448

Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 11

Item(s) Submitted: Cast Hand Weights
Manufacturer: Assorted
ASTM E617 Type: II
Equipment ID#: None
Condition: Excellent
Temperature: 20.3 °C
Pressure: 734.2 mmHg
Relative Humidity: 39. %

Nominal Value	lb	Serial No.	Correction (mg)		NIST HB105-1 Class		Unc. (mg) (k=2)
			As Found	As Left	As Found	As Left	
50	lb	L1	160.	160.	F	F	110.
50	lb	L2	1480.	1480.	F	F	110.
50	lb	L3	1910.	1910.	F	F	110.
50	lb	L4	-370.	-370.	F	F	110.
50	lb	L5	1380.	1380.	F	F	110.
50	lb	L6	1600.	1600.	F	F	110.
50	lb	L7	1160.	1160.	F	F	110.
50	lb	L8	260.	260.	F	F	110.
50	lb	L9	640.	640.	F	F	110.
50	lb	L10	780.	780.	F	F	110.
50	lb	L11	410.	410.	F	F	110.
50	lb	L12	1030.	1030.	F	F	110.
50	lb	L13	1400.	1400.	F	F	110.
50	lb	L14	890.	890.	F	F	110.
50	lb	L15	2750.	-20.	*	F	110.
50	lb	L16	1000.	1000.	F	F	110.
50	lb	L17	-2800.	280.	*	F	110.
50	lb	L18	1830.	1830.	F	F	110.
50	lb	L19	820.	820.	F	F	110.
50	lb	L20	520.	520.	F	F	110.

* Weight(s) as found exceed NIST HB 105-1 Class F tolerance.

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Pete Whebbe

Pete J. Whebbe
Metrologist

Reviewed by:

Mark Nicollet

Mark Nicollet
Quality Manager



Receipt Date: November 3, 2015
Test Date: November 10, 2015
Report Date: November 10, 2015

State Test No.: 335028
Set Serial No.: 25121 - 25160
Barcode: 200651

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448
Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 11

Item(s) Submitted: Cast Hand Weights
Manufacturer: Rice Lake
ASTM E617 Type: II
Equipment ID#: None
Condition: Excellent
Temperature: 20.1 °C
Pressure: 736.2 mmHg
Relative Humidity: 38.9 %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Uncertainty (mg) (k = 2)
		As Found	As Left	As Found	As Left	
25 lb	25121	720	720	F	F	75
25 lb	25122	1720	460	*	F	75
25 lb	25123	1210	80	*	F	75
25 lb	25124	100	100	F	F	75
25 lb	25125	1060	150	*	F	75
25 lb	25126	2030	10	*	F	75
25 lb	25127	540	540	F	F	75
25 lb	25128	380	380	F	F	75
25 lb	25129	580	580	F	F	75
25 lb	25130	730	730	F	F	75
25 lb	25131	980	980	F	F	75
25 lb	25132	1370	-10	*	F	75
25 lb	25133	870	870	F	F	75
25 lb	25134	1300	370	*	F	75
25 lb	25135	1500	160	*	F	75
25 lb	25136	2310	160	*	F	75
25 lb	25137	220	220	F	F	75
25 lb	25138	1100	210	*	F	75
25 lb	25139	2140	60	*	F	75
25 lb	25140	620	620	F	F	75
25 lb	25141	340	340	F	F	75
25 lb	25142	890	890	F	F	75
25 lb	25143	1440	130	*	F	75
25 lb	25144	680	680	F	F	75
25 lb	25145	1690	460	*	F	75
25 lb	25146	1170	120	*	F	75
25 lb	25147	410	410	F	F	75
25 lb	25148	700	700	F	F	75
25 lb	25149	730	730	F	F	75
25 lb	25150	1430	210	*	F	75

* Weight(s) as found exceed NIST HB105-1 Class F Tolerance.



Receipt Date: November 3, 2015
Test Date: November 10, 2015
Report Date: November 10, 2015

State Test No.: 335028
Set Serial No.: 25121 - 25160
Barcode: 200651

Continued,

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448
Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 11

Item(s) Submitted: Cast Hand Weights
Manufacturer: Rice Lake
ASTM E617 Type: II
Equipment ID#: None
Condition: Excellent
Temperature: 20.1°C
Pressure: 736.2 mmHg
Relative Humidity: 38.9 %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Uncertainty (mg) (k = 2)
		As Found	As Left	As Found	As Left	
25 lb	25151	320	320	F	F	75
25 lb	25152	1090	520	*	F	75
25 lb	25153	610	610	F	F	75
25 lb	25154	1860	10	*	F	75
25 lb	25155	1650	120	*	F	75
25 lb	25156	680	680	F	F	75
25 lb	25157	970	970	F	F	75
25 lb	25158	560	560	F	F	75
25 lb	25159	1580	10	*	F	75
25 lb	25160	1490	40	*	F	75

* Weight(s) as found exceed NIST HB105-1 Class F Tolerance.

When used as a set these weights meet NIST HB105-1 Class F Tolerance.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). Results apply to items identified in this report only.

Pete Whebbe

Metrologist

Reviewed by:

Mark Nicollet

Quality Manager



Receipt Date: November 3, 2015
Test Date: November 16, 2015
Report Date: November 16, 2015

State Test No.: 335027
Set Serial No.: A6
Barcode: 200640

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448
Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 09

Item(s) Submitted: 30 lb kit w/ decimals & fractions
Manufacturer: Rice Lake
ASTM E617 Type: I & II
Equipment ID#: None
Condition: Good
Temperature: 20.5 °C
Pressure: 733.5 mmHg
Relative Humidity: 48.1 %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Uncertainty (mg) (k = 2)
		As Found	As Left	As Found	As Left	
10 lb		64	64	F	F	15
10 lb		141	141	F	F	15
5 lb		55	55	F	F	10.
1 lb		12.4	12.4	F	F	6.0
1 lb		11.4	11.4	F	F	6.0
1 lb		16.2	16.2	F	F	6.0
1 lb		11.7	11.7	F	F	6.0
1 lb		1.2	1.2	F	F	6.0
0.2 lb		-1.340	-1.340	F	F	0.070
0.2 lb		8.000	8.000	F	F	0.070
0.1 lb		0.060	0.060	F	F	0.070
0.05 lb		-0.410	-0.410	F	F	0.070
0.02 lb		-0.620	-0.620	F	F	0.070
0.02 lb		-0.280	-0.280	F	F	0.070
0.01 lb		0.230	0.230	F	F	0.070



Continued,

Receipt Date: November 3, 2015
Test Date: November 16, 2015
Report Date: November 16, 2015

State Test No.: 335027
Set Serial No.: A6
Barcode: 200640

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448
Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 09

Item(s) Submitted: 30 lb kit w/ decimals & fractions
Manufacturer: Rice Lake
ASTM E617 Type: I & II
Equipment ID#: None
Condition: Good
Temperature: 20.5°C
Pressure: 733.5 mmHg
Relative Humidity: 48.1 %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Uncertainty (mg) (k = 2)
		As Found	As Left	As Found	As Left	
4 oz		9.720	9.720	F	F	0.070
4 oz		3.220	3.220	F	F	0.070
4 oz		11.230	11.230	F	F	0.070
1 oz		0.990	0.990	F	F	0.070
1 oz		1.420	1.420	F	F	0.070
1 oz		1.620	1.620	F	F	0.070
1/2 oz		1.110	1.110	F	F	0.070
1/2 oz		1.100	1.100	F	F	0.070
1/4 oz		0.790	0.790	F	F	0.070
1/4 oz		0.510	0.510	F	F	0.070

When used as a set these weights meet NIST HB105-1 Class F Tolerance.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). Results apply to items identified in this report only.

Heidi Jones
Heidi Jones
Laboratory Administrator

Reviewed by:
Mark Nicollet
Mark Nicollet
Quality Manager



Receipt Date: November 4, 2015
Test Date: November 12, 2015
Report Date: November 13, 2015

State Test No.: 335026
Set Serial No.: M6
Barcode: 018947

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448

Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 09

Item(s) Submitted: Metric weight set
Manufacturer: Rice Lake
ASTM E617 Type: I & II
Equipment ID#: None
Condition: Good
Temperature: 20.8 °C
Pressure: 731.2 mmHg
Relative Humidity: 50.8 %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Uncertainty (mg) (k = 2)
		As Found	As Left	As Found	As Left	
5000	09	158	158	F	F	25
2000	09	64.8	64.8	F	F	10.
2000	09	78.7	78.7	F	F	10.
1000	09	38.3	38.3	F	F	6.0
500	09	25.1	25.1	F	F	4.6
200	09	7.1	7.1	F	F	1.0
200	09	6.1	6.1	F	F	1.0
100	09	3.18	3.18	F	F	0.45
50	09	4.14	4.14	F	F	0.25
20	09	1.66	1.66	F	F	0.25
20	09	0.62	0.62	F	F	0.25
10	09	0.69	0.69	F	F	0.12
5	09	0.90	0.90	F	F	0.10
2	09	0.480	0.480	F	F	0.070
2	09	0.510	0.510	F	F	0.070
1	09	0.550	0.550	F	F	0.070

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). Results apply to items identified in this report only.

Heidi Jones
Heidi Jones
Laboratory Administrator

Reviewed by:
Mark Nicollet
Mark Nicollet
Quality Manager





Receipt Date: October 19, 2015
Test Date: October 19, 2015
Report Date: October 20, 2015

State Test No.: 334948
Set Serial No.: None
Barcode: 202400

Calibration Report

KENNEDY SCALE, INC.
11485 XEON STREET NW
COON RAPIDS, MN 55448

Contact: ROB BAKER
Phone: 763-784-1022
PO Number: NONE
SOP: 12
Technician ID: 19

Item(s) Submitted: 8 oz disk weights
Manufacturer: RICE LAKE
ASTM E617 Type: II
Equipment ID#: None
Condition: Excellent
Temperature: 20.4 °C
Pressure: 730.9 mmHg
Relative Humidity: 43. %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Unc. (mg) (k=2)
		As Found	As Left	As Found	As Left	
8 OZ		22.	22.	F	F	6.
8 OZ		15.2	15.2	F	F	6.
8 OZ		17.9	17.9	F	F	6.
8 OZ		19.3	19.3	F	F	6.
8 OZ		15.	15.	F	F	6.
8 OZ		18.9	18.9	F	F	6.
8 OZ		20.5	20.5	F	F	6.
8 OZ		12.7	12.7	F	F	6.
8 OZ		22.4	22.4	F	F	6.
8 OZ		12.8	12.8	F	F	6.
8 OZ		18.8	18.8	F	F	6.
8 OZ		14.5	14.5	F	F	6.
8 OZ		7.3	7.3	F	F	6.
8 OZ		19.3	19.3	F	F	6.

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Technical Note 1297. Results apply to items identified in this report only.

Erik Alfvín

Metrologist

Reviewed by:

Mark Nicollet

Quality Manager

United States Department of Commerce
National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Minnesota

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon II	Weight Carts	Volume Gravimetric, I
50 kg to 1 mg	10 000 lb to 2000 lb	20 L to 1 ml
1000 lb to 0.001 lb	Wheel Load Weighers	100 gal to 0.25 qt
4 oz to 0.03125 oz	20 000 lb to 2000 lb	Volume Transfer, II
Mass Echelon III	Railroad Test Cars	1500 gal to 5 gal
50 kg to 1 mg	110 000 lb to 80 000 lb	100 gal to 25 gal LPG
5000 lb to 0.001 lb		
4 oz to 0.03125 oz		



2015

Carol T. Hockett, Chief
NIST Office of Weights and Measures

Effective Dates: 2015-01-01 to 2015-12-31

North Dakota

nd.gov Official Portal for
North Dakota State GovernmentNorth Dakota
LEGENDARY

SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

KENNEDY SCALES, INC.

Corporation Details

System ID: 15125100**Phone:** (763) 784-1022**Type:** FOREIGN BUSINESS CORPORATION**Status:** Active & Good Standing**Original File Date:** 11/01/1999**Effective Date:** 11/01/1999**State of Origin:** Minnesota

Nature of Business

SERVICE & SALES OF WEIGHING EQUIPMENT

Principal Office

11485 XEON ST NW COON RAPIDS, MN 55448-3150

Registered Agent

SEARCH COMPANY OF NORTH DAKOTA LLC

1501 N 12TH ST STE 1

BISMARCK, ND 58501-2713

Established Date: Jan 24, 2014

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2015](#) (generates a forms-fillable pdf in a new pop-up window)[Return to Search Results](#)[Contact Us](#)[Disclaimer](#) | [Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.