



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY

Public Service Commission
SFN 61277 (2/2014)

TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company <i>Lake Country Scale Works, Inc.</i>	Email Address <i>lcswdodge@gmail.com</i>	Application Date <i>12-15-15</i>	
Mailing Address <i>2511-60th St. NW</i>	City <i>Willmar</i>	State <i>MN</i>	Zip Code <i>56201</i>
Telephone Number <i>320-235-6863</i>	Cell Phone Number <i>320-979-0214</i>	Fax Number <i>320-235-2194</i>	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input checked="" type="checkbox"/> 1. Rail	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute)
<input checked="" type="checkbox"/> 2. Truck	<input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater)
<input checked="" type="checkbox"/> 3. Livestock	<input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____
<input checked="" type="checkbox"/> 4. Hopper: Max. Capacity: <i>250,000 #</i>	<input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____
<input type="checkbox"/> 5. Belt	<input type="checkbox"/> 5. LPG
<input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: <i>250,000 #</i>	<input type="checkbox"/> 6. Stationary LPG
<input checked="" type="checkbox"/> 7. 30 lbs. or less	<input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____
<input type="checkbox"/> 8. Class II (Indicate on your calibration report which weight kit is Class II certified)	<input type="checkbox"/> 8. Chemical
<input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 9. Anhydrous
	<input type="checkbox"/> 10. Loading Rack
	<input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
<i>e.g. 1001</i>	<i>e.g. John Doe</i>	<i>e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6</i>
<i>0430</i>	<i>Doug Martin - Owner</i>	<i>1, 2, 3, 4, 6, 7</i>

Application for Registration as a Registered Service Company
Page 2

List below all field standards (attach current calibration reports):

8 - 1000 # wts	1 - 30 # gram kit
12 - 500 # wts	1 - 200 # stainless steel basket
36 - 50 # wts	1 - 400 # weight cart
2 - 25 # wts	
5 - 5 # wts	
5 - 1 # wts	
1 - 0.5 # wts	
11 - Decimal wts - from 0.01-0.1#	
9 - Fraction wts from 1/32oz to 4oz	

Additional Application Items (Initial where appropriate):

Standardized Test Report	<u>DM</u> Copy enclosed No change in report filed previously
Tested and Approved Sticker	<u>DM</u> Copy enclosed No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<u>DM</u> Copy enclosed No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.

Yes No

I am Douglas Martin, and have authority to represent this company.
By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.

Douglas D Martin
Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410



14305 SOUTHCROSS DRIVE #150
BURNSVILLE, MN 55306-7008
MN.GOV/COMMERCE/
651.539.1555 FAX 952.435.4040
AN EQUAL OPPORTUNITY EMPLOYER

Receipt Date: December 7, 2015
Test Date: December 8, 2015
Report Date: December 8, 2015

State Test No.: 335171
Set Serial No.: None
Barcode: 017617

Calibration Report

LAKE COUNTRY SCALE WORKS
2511 60TH ST NW
WILLMAR, MN 56201-8747
Contact: Doug Martin
Phone: 320-235-6863
PO Number: NONE
SOP: 12
Technician ID: 11

Item(s) Submitted: 4000 lb Weight Cart w/rack
Manufacturer: Lake Country
ASTM E617 NA
Equipment ID#: None
Condition: Good
Temperature: 19. °C
Pressure: 729.8 mmHg
Relative Humidity: 38.1 %

Nominal Value	Serial No.	Correction (g)		NIST HB105-8 Tolerance		Uncertainty (g) (k = 2)
		As Found	As Left	As Found	As Left	
4000 lb		-2130.	-50.	*	Meets	60.

* Weight Cart as found exceeds NIST HB 105-8 tolerance.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-8 (2003). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). Results apply to items identified in this report only.

Pete Whebbe

Pete J. Whebbe
Metrologist

Reviewed by:

Mark Nicollet

Mark Nicollet
Quality Manager





14305 SOUTHCROSS DRIVE #150
 BURNSVILLE, MN 55306-7008
 MN.GOV/COMMERCE/
 651.539.1555 FAX 651.539.1553
 AN EQUAL OPPORTUNITY EMPLOYER

Receipt Date: December 7, 2015
 Test Date: December 8, 2015
 Report Date: December 8, 2015

State Test No.: 335172
 Set Serial No.: None
 Barcode: 201289

Calibration Report

LAKE COUNTRY SCALE WORKS
 2511 60TH ST NW
 WILLMAR, MN 56201-8747

Contact: Doug Martin
 Phone: 320-235-6863
 PO Number: NONE
 SOP: 12
 Technician ID: 11

Item(s) Submitted: Cast Cube Weights
 Manufacturer: Rice Lake
 ASTM E617 Type: II
 Equipment ID#: None
 Condition: Good
 Temperature: 18.9 °C
 Pressure: 729.7 mmHg
 Relative Humidity: 48.4 %

Nominal Value	Serial No.	Correction (g)		NIST HB 105-1 Class		Uncertainty (g) (k = 2)
		As Found	As Left	As Found	As Left	
1000 lb		5	5	F	F	5
1000 lb		-23	-23	F	F	5
1000 lb		16	16	F	F	5
1000 lb		20	20	F	F	5
1000 lb		-2	-2	F	F	5
1000 lb		-3	-3	F	F	5
1000 lb		-4	-4	F	F	5
1000 lb		24	24	F	F	5

When used as a set these weights meet NIST HB 105-1 Class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). Results apply to items identified in this report only.

Pete Whebbe

Pete J. Whebbe
 Metrologist

Reviewed by:

Mark Nicollet

Mark Nicollet
 Quality Manager



14305 SOUTHCROSS DRIVE #150
 BURNSVILLE, MN 55306-7008
 MN.GOV/COMMERCE/
 651.539.1555 FAX 651.539.1553
 AN EQUAL OPPORTUNITY EMPLOYER

Receipt Date: December 7, 2015 State Test No.: 335173
 Test Date: December 8, 2015 Set Serial No.: None
 Report Date: December 8, 2015 Barcode: 201290

Calibration Report

LAKE COUNTRY SCALE WORKS
 2511 60TH ST NW
 WILLMAR, MN 56201-8747
 Contact: Doug Martin
 Phone: 320-235-6863
 PO Number: NONE
 SOP: 12
 Technician ID: 11

Item(s) Submitted: Cast Cube Weights
 Manufacturer: Howe/P.U.C.
 ASTM E617 Type: II
 Equipment ID#: None
 Condition: Good
 Temperature: 18.9 °C
 Pressure: 729.6 mmHg
 Relative Humidity: 48.5 %

Nominal Value	Serial No.	Correction (g)		NIST HB105-1 Class		Uncertainty (g) (k = 2)
		As Found	As Left	As Found	As Left	
500 lb		-11.5	-11.5	F	F	1.1
500 lb		2.6	2.6	F	F	1.1
500 lb		-13.8	-13.8	F	F	1.1
500 lb		-18.8	-18.8	F	F	1.1
500 lb		4.9	4.9	F	F	1.1
500 lb		-8.8	-8.8	F	F	1.1
500 lb		-17.7	-17.7	F	F	1.1
500 lb		-14.8	-14.8	F	F	1.1
500 lb		-4.5	-4.5	F	F	1.1
500 lb		4.7	4.7	F	F	1.1
500 lb		-5.9	-5.9	F	F	1.1
500 lb		-1.1	-1.1	F	F	1.1

When used as a set these weights meet NIST HB 105-1 Class F tolerances.
 The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). Results apply to items identified in this report only.

Pete Wiebbe

 Metrologist

Reviewed by:
 Mark Njcollet

 Quality Manager



14305 SOUTHGROSS DRIVE #150
BURNSVILLE, MN 55306-7008
MN.GOV/COMMERCE/
651.539.1555 FAX 952.435.4040
AN EQUAL OPPORTUNITY EMPLOYER

Receipt Date: December 7, 2015
Test Date: December 8, 2015
Report Date: December 8, 2015

State Test No.: 335174
Set Serial No.: None
Barcode: 201292

Calibration Report

LAKE COUNTRY SCALE WORKS
2511 60TH ST NW
WILLMAR, MN 56201-8747

Contact: Doug Martin
Phone: 320-235-6863
PO Number: NONE
SOP: 12
Technician ID: 11

Item(s) Submitted: 200 lb Basket
Manufacturer: Lake Country
ASTM E617 Type: NA
Equipment ID#: None
Condition: Good
Temperature: 19. °C
Pressure: 728.6 mmHg
Relative Humidity: 51.2 %

Nominal Mass (lb)	Serial No.	Assigned Value (lb)	Assigned Value (g)	Uncertainty (g) (k = 2)
200		200.014	90725.0	1.1

The values given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to ASTM E617 (2013). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). Results apply to items identified in this report only.

Pete Whebbe

Pete Whebbe
Metrologist

Reviewed by:

Mark Nicollet

Mark Nicollet
Quality Manager



14305 SOUTHCROSS DRIVE #150
 BURNSVILLE, MN 55306-7008
 MN.GOV/COMMERCE/
 651.539.1555 FAX 651.539.1553
 AN EQUAL OPPORTUNITY EMPLOYER

Receipt Date: December 7, 2015
 Test Date: December 8, 2015
 Report Date: December 8, 2015

State Test No.: 335175
 Set Serial No.: None
 Barcode: 201291

Calibration Report

LAKE COUNTRY SCALE WORKS
 2511 60TH ST NW
 WILLMAR, MN 56201-8747
 Contact: Doug Martin
 Phone: 320-235-6863
 PO Number: None
 SOP: 12
 Technician ID: 19

Item(s) Submitted: (50, 30, 25, 20) lb Hand Weights
 Manufacturer: Rice Lake
 ASTM E617 Type: II
 Equipment ID#: None
 Condition: Good
 Temperature: 19.1 °C
 Pressure: 728.6 mmHg
 Relative Humidity: 51.4 %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Uncertainty (mg) (k = 2)
		As Found	As Left	As Found	As Left	
50 lb	280	-1870	-1870	F	F	110
50 lb		10	10	F	F	110
50 lb		3930	30	*	F	110
50 lb		-2310	180	*	F	110
50 lb		2500	80	*	F	110
50 lb		1720	1720	F	F	110
50 lb		2500	80	*	F	110
50 lb		2760	90	*	F	110
50 lb		2930	150	*	F	110
50 lb		1540	1540	F	F	110
50 lb		1110	1110	F	F	110
50 lb		-110	-110	F	F	110
50 lb		4190	60	*	F	110
50 lb		2170	140	F	F	110
50 lb		2500	50	*	F	110
50 lb		1540	1540	F	F	110
50 lb		1710	1710	F	F	110
50 lb		1200	1200	F	F	110
50 lb		6840	160	*	F	110
50 lb		5900	20	*	F	110
50 lb		-860	-860	F	F	110
50 lb		1770	1770	F	F	110
50 lb		4070	170	*	F	110
50 lb		1460	1460	F	F	110
50 lb		-3610	170	*	F	110
50 lb		2500	170	*	F	110
50 lb		1690	1690	F	F	110
50 lb		690	690	F	F	110
50 lb		-2470	120	*	F	110
50 lb		740	740	F	F	110

* Weight(s) as found exceed NIST HB105-1 Class F Tolerance. Page 1 of 2



14305 SOUTHCROSS DRIVE #150
BURNSVILLE, MN 55306-7008
MN.GOV/COMMERCE/
651.539.1555 FAX 651.539.1553
AN EQUAL OPPORTUNITY EMPLOYER

Continued,

Receipt Date: December 7, 2015
Test Date: December 8, 2015
Report Date: December 8, 2015

State Test No.: 335175
Set Serial No.: None
Barcode: 201291

Calibration Report

LAKE COUNTRY SCALE WORKS
2511 60TH ST NW
WILLMAR, MN 56201-8747
Contact: Doug Martin
Phone: 320-235-6863
PO Number: None
SOP: 12
Technician ID: 19

Item(s) Submitted: (50, 30, 25, 20) lb Hand Weights
Manufacturer: Rice Lake
ASTM E617 Type: II
Equipment ID#: None
Condition: Good
Temperature: 19.1°C
Pressure: 728.6 mmHg
Relative Humidity: 51.4 %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Uncertainty (mg) (k = 2)
		As Found	As Left	As Found	As Left	
50 lb		1760	1760	F	F	110
50 lb		3740	110	*	F	110
50 lb		1890	1890	F	F	110
50 lb		1890	1890	F	F	110
50 lb		1590	1590	F	F	110
50 lb		-2940	100	*	F	110
50 lb	Silver	390	390	F	F	110
30 lb		1180.	1180.	F	F	75
25 lb	Silver	1650.	190.	*	F	75
25 lb	279	2150.	20.	*	F	75
20 lb	175	1380.	20.	*	F	75

* Weight(s) as found exceed NIST HB105-1 Class F Tolerance.

When used as a set these weights meet NIST HB105-1 Class F Tolerance.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). Results apply to items identified in this report only.

Erik Alfvén

Metrologist

Reviewed by:

Mark Nicollet

Quality Manager



14305 SOUTHCROSS DRIVE #150
 BURNSVILLE, MN 55306-7008
 MN.GOV/COMMERCE/
 651.539.1555 FAX 952.435.4040
 AN EQUAL OPPORTUNITY EMPLOYER

Receipt Date: December 7, 2015
 Test Date: December 8, 2015
 Report Date: December 8, 2015

State Test No.: 335188
 Set Serial No.: 129
 Barcode: 202917

Calibration Report

LAKE COUNTRY SCALE WORKS
 2511 60TH ST NW
 WILLMAR, MN 56201-8747
 Contact: Doug Martin
 Phone: 320-235-6863
 PO Number: None
 SOP: 12
 Technician ID: 19

Item(s) Submitted: 10 lb Weight
 Manufacturer: Rice Lake
 ASTM E617 Type: II
 Equipment ID#: None
 Condition: Good
 Temperature: 20.7 °C
 Pressure: 727.8 mmHg
 Relative Humidity: 49.7 %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Uncertainty (mg) (k = 2)
		As Found	As Left	As Found	As Left	
10 lb	129	95	95	F	F	15

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). Results apply to items identified in this report only.

Erik Alfvén

Metrologist

Reviewed by:

Mark Nicollet

Quality Manager





14305 SOUTHCROSS DRIVE #150
BURNSVILLE, MN 55306-7008
MN.GOV/COMMERCE/
651.539.1555 FAX 651.539.1553
AN EQUAL OPPORTUNITY EMPLOYER

Receipt Date: December 7, 2015
Test Date: December 9, 2015
Report Date: December 10, 2015

State Test No.: 335169
Set Serial No.: None
Barcode: 017616

Calibration Report

LAKE COUNTRY SCALE WORKS
2511 60TH ST NW
WILLMAR, MN 56201-8747
Contact: Doug Martin
Phone: 320-235-6863
PO Number: None
SOP: 12
Technician ID: 19

Item(s) Submitted: AVDP Weight Kit - F
Manufacturer: Rice Lake
ASTM E617 Type: I & II
Equipment ID#: None
Condition: Good/Contaminated**
Temperature: 20.4 °C
Pressure: 729.8 mmHg
Relative Humidity: 48.3 %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Uncertainty (mg) (k = 2)
		As Found	As Left	As Found	As Left	
5 lb		-7	-7	F	F	10.
5 lb		28	28	F	F	10.
5 lb		102	102	F	F	10.
5 lb		93	93	F	F	10.
5 lb		-25	-25	F	F	10.
1 lb		25.9	25.9	F	F	6.0
1 lb		12.1	12.1	F	F	6.0
1 lb		13.3	13.3	F	F	6.0
1 lb		27.3	27.3	F	F	6.0
1 lb		11.8	11.8	F	F	6.0
8 oz		16.1	16.1	F	F	6.0
4 oz		7.020	7.020	F	F	0.070
2 oz		-0.640	-0.640	F	F	0.070
1 oz		2.020	2.020	F	F	0.070
1/2 oz		1.930	1.930	F	F	0.070
1/4 oz		1.300	1.300	F	F	0.070
1/8 oz		0.150	0.150	F	F	0.070
1/16 oz		0.770	0.770	F	F	0.070
1/32 oz		0.330	0.330	F	F	0.070
1/32 oz		0.340	0.340	F	F	0.070
0.2 lb		6.450	6.450	F	F	0.070
0.2 lb		5.480	5.480	F	F	0.070
0.1 lb		2.730	2.730	F	F	0.070
0.05 lb		0.340	0.340	F	F	0.070
0.02 lb		0.780	0.780	F	F	0.070
0.02 lb		0.140	0.140	F	F	0.070
0.01 lb		-0.710	-0.710	F	F	0.070

** Weights and padding arrived wet. Weights in good condition after drying.

When used as a set these weights meet NIST HB105-1 Class F Tolerance.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm^3 density and an air density of 1.2 mg/cm^3 . The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). Results apply to items identified in this report only.

Eric Alfvin
Eric Alfvin
Metrologist

Reviewed by:
Mark Nicollet
Mark Nicollet
Quality Manager



14305 SOUTHGROSS DRIVE #150
 BURNSVILLE, MN 55306-7008
 MN.GOV/COMMERCE/
 651.539.1555 FAX 651.539.1553
 AN EQUAL OPPORTUNITY EMPLOYER

Receipt Date: December 7, 2015
 Test Date: December 8, 2015
 Report Date: December 8, 2015

State Test No.: 335170
 Set Serial No.: None
 Barcode: 201274

Calibration Report

LAKE COUNTRY SCALE WORKS
 2511 60TH ST NW
 WILLMAR, MN 56201-8747
 Contact: Doug Martin
 Phone: 320-235-6863
 PO Number: None
 SOP: 12
 Technician ID: 19

Item(s) Submitted: Metric Weight Kit - F
 Manufacturer: Rice Lake
 ASTM E617 Type: I & II
 Equipment ID#: None
 Condition: Good / Excellent
 Temperature: 20.1 °C
 Pressure: 729.7 mmHg
 Relative Humidity: 48.9 %

Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Uncertainty (mg) (k = 2)
		As Found	As Left	As Found	As Left	
2000 g		99	99	F	F	10.
1000 g		47.2	47.2	F	F	6.0
500 g		16.4	16.4	F	F	6.0
200 g		10.2	10.2	F	F	1.0
200 g		13.0	13.0	F	F	1.0
100 g		6.38	6.38	F	F	0.45
50 g		3.50	3.50	F	F	0.25
20 g		1.80	1.80	F	F	0.25
20 g		1.71	1.71	F	F	0.25
10 g		0.57	0.57	F	F	0.12
5 g		0.52	0.52	F	F	0.10
2 g		0.450	0.450	F	F	0.070
2 g		0.090	0.090	F	F	0.070
1 g		-0.020	-0.020	F	F	0.070
0.5 g		0.290	0.290	F	F	0.070
0.2 g		0.200	0.200	F	F	0.070
0.2 g		0.210	0.210	F	F	0.070
0.1 g		0.150	0.150	F	F	0.070
0.05 g		0.080	0.080	F	F	0.070
0.02 g		0.070	0.070	F	F	0.070
0.02 g		0.080	0.080	F	F	0.070
0.01 g		0.120	0.120	F	F	0.070

When used as a set these weights meet NIST HB 105-1 class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). Results apply to items identified in this report only.

Erik Alfvén

Metrologist

Reviewed by:

Mark Nicollet

Quality Manager

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Minnesota




This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

2015

Scope

Mass Echelon II	Weight Carts	Volume Gravimetric, I
50 kg to 1 mg	10 000 lb to 2000 lb	20 L to 1 ml
1000 lb to 0.001 lb	Wheel Load Weighers	100 gal to 0.25 qt
4 oz to 0.03125 oz	20 000 lb to 2000 lb	Volume Transfer, II
Mass Echelon III	Railroad Test Cars	1500 gal to 5 gal
50 kg to 1 mg	110 000 lb to 80 000 lb	100 gal to 25 gal LPG
5000 lb to 0.001 lb		
4 oz to 0.03125 oz		


Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2015-01-01 to 2015-12-31

North Dakota

nd.gov Official Portal for
North Dakota State GovernmentNorth Dakota
LEGENDARY

SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

LAKE COUNTRY SCALE WORKS, INC.

Corporation Details

System ID: 21369500**Phone:** (320) 235-6863**Type:** FOREIGN BUSINESS CORPORATION**Status:** Active & Good Standing**Original File Date:** 08/10/2005**Effective Date:** 08/10/2005**State of Origin:** Minnesota

Nature of Business

SELL, SERVICE, TEST SCALES

Principal Office

2511 60TH ST NW WILLMAR, MN 56201-9736

Registered Agent

SEARCH COMPANY OF NORTH DAKOTA LLC

1501 N 12TH ST STE 1

BISMARCK, ND 58501-2713

Established Date: Oct 10, 2008

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2015](#) (generates a forms-fillable pdf in a new pop-up window)[Return to Search Results](#)[Contact Us](#)[Disclaimer](#) | [Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.