



Public Service Commission State of North Dakota

COMMISSIONERS

Julie Fedorchak
Randy Christmann
Brian P. Kalk

Executive Secretary
Darrell Nitschke

600 East Boulevard, Dept. 408
Bismarck, North Dakota 58505-0480
Web: www.psc.nd.gov
E-mail: ndpsc@nd.gov
Phone: 701-328-2400
ND Toll Free: 1-877-245-6685
Fax: 701-328-2410
TDD: 800-366-6888 or 711

PERMANENT VARIANCE PERMIT APPROVAL

**Northwood Equity Elevator Company
Variance Permit
Application**

Case No. WM-15-706

Issued To:

Northwood Equity Elevator Company
PO Box 380
Northwood ND 58267

Scale Location:

Northwood Equity Elevator Company
802 Old Highway 15
Northwood ND 58267

Scale Information:

Waconia, Survivor 420-E-1A, 20,000 x 10 lb., Serial Number 1724600006

Request To:

Operate a hopper scale that deviates from North Dakota Weights and Measures rules as follows:

- 1) Less unobstructed clearance than required (3') on all four sides of the scale to facilitate testing with large weights [N.D. Admin. Code § 69-10-02-26].

Permanent Variance Permit Conditions:

- The scale must be retested and adjusted once every twelve months by a registered service company to maintain the required tolerance;

- The registered service company test report must be submitted to the Commission within seven working days of the test date and a copy retained by Northwood Equity Elevator Company;
- The service company test report must include both the "as found" and "as left" readings in addition to any maintenance notes;
- Except for the specific variance granted by Permit No. 503, the scale must be operated according to all applicable North Dakota Laws and Rules;
- The enclosed variance permit must be posted near the designated scale in a place clearly visible to the public; and
- Failure to comply with the variance permit conditions may result in revocation.

If you have any further questions, or require additional information, please feel free to contact me at (701) 328-4097.

Dated November 30, 2016



Konrad Crockford, Director
Compliance Division



Public Service Commission

PERMANENT VARIANCE PERMIT

Issued To:

Northwood Equity Elevator Company
PO Box 380
Northwood ND 58267

Permit No. 503

Expiration Date: N/A

Scale Location:

Northwood Equity Elevator Company
802 Old Highway 15
Northwood ND 58267

Scale Information:

Waconia, Survivor 420-E-1A, 20,000 x 10 lb., Serial Number 1724600006

This Permit allows the above referenced scale to be operated in commerce with the following deviation from North Dakota Weights and Measures rules as follows:

1. Less unobstructed clearance than required (3') on all four sides of the scale to facilitate testing with large weights [N.D. Admin. Code § 69-10-02-26].

This Permit is conditioned upon Northwood Equity Elevator Company adhering to the terms of the November 30, 2016 Permanent Variance Permit Approval.

Dated November 30, 2016

A handwritten signature in black ink, appearing to read "K. Crockford", is written over a horizontal line.

Konrad Crockford, Director
Compliance Division