



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY
Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company Cothorn's Tanker Inspections, LLC	Email Address Cothornstankerinspect@yahoo.com	Application Date 1/20/16	
Mailing Address PO Box 1018	City Caldwell	State TX	Zip Code 77834
Telephone Number (979) 272-1400	Cell Phone Number (979) 324-3335	Fax Number (979) 272-1500	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input checked="" type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: <u>100</u> <input checked="" type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: <u>100</u> <input checked="" type="checkbox"/> 5. LPG <input checked="" type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
e.g. 1001	e.g. John Doe	e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6
1749	Jeffrey Smith	Liquid - 3, 4, 5, 6

Application for Registration as a Registered Service Company
Page 2

List below all field standards (attach current calibration reports):


Additional Application Items (initial where appropriate):

Standardized Test Report	<input checked="" type="checkbox"/> Copy enclosed <input type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact. <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No


I am Riley B. Ootherm, and have authority to represent this company.
By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.


Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410

North Dakota

nd.gov Official Portal for
North Dakota State GovernmentNorth Dakota
LEGENDARY

SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

COTHERN'S TANKER INSPECTIONS, LLC

Corporation Details

System ID: 37692200**Phone:** (800) 553-1770**Type:** FOREIGN LIMITED LIABILITY COMPANY**Status:** Active & Good Standing**Original File Date:** 10/06/2014**Effective Date:** 10/06/2014**State of Origin:** Texas

Nature of Business

METER CALIBRATIONS AND METER REPAIRS

Principal Office

8660 CR 201 SOMERVILLE TX PO BOX 1018 CALDWELL, TX 77836-0902

Registered Agent

ADVANCED CORPORATE SERVICES, INC.

720 MAIN AVE

PO BOX 2105

FARGO, ND 58107-2105

Established Date: Oct 06, 2014

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2015](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)[Disclaimer](#)[Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



Test Number
 G-000005270

REPORT OF CALIBRATION

FOR

1-100 Gallon Bulk Test Prover



SUBMITTED BY

Cothorn's Tanker Inspections LLC
 P.O. Box 1018
 Caldwell, Texas 77836

The standards of Texas are traceable to the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The data below applies only to the artifacts identified in this report at the time of test.

Test Date: 08/31/2015
Calibration Due: 08/31/2016


Temperature Range: 15 °C to 30 °C
Relative Humidity Range: 40 % to 60 %
Procedure: NISTIR 7383, SOP 19, Volume Transfer Method
Standard(s): Giddings Metrology Laboratory Echelon II Volume Transfer Standards


The expanded standard uncertainty includes the standard uncertainty reported for the standard, the standard uncertainty for the measurement process, and a component of uncertainty to account for any observed deviations that have a significant effect on the calibration. The expanded uncertainty given here is in compliance with NIST Technical Note 1297 ("Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results") with a $k = 2$, representing a 95.45 % confidence level.

This report is not to be used to claim product endorsement by the Texas Department of Agriculture or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of the Texas Department of Agriculture Metrology Laboratory.

Note:

To convert from gallon to cubic inch: Multiply gallon by 231
 To convert from cubic inch to gallon: Divide cubic inch by 231
 To convert from cubic inch to cubic meter: Multiply cubic inch by 0.000016387064


 for Philip Wright
 Director For Consumer Product Protection


 Daniel Gibbons
 Metrologist



REPORT OF CALIBRATION
 For

Test Number
 G-000005270

Test Completed:
 08/31/2015

1-100 Gallon Bulk Test Prover

Calibration Due:
 08/31/2016



Submitted by
 Cothorn's Tanker Inspections LLC
 P.O. Box 1018
 Caldwell, Texas 77836

The volumetric provers described below have been compared to the standards of the State of Texas and were found to deliver as follows:

Temperature Range: 15 °C to 30 °C

Relative Humidity Range: 40 % to 60 %

Procedure: NISTIR 7383, SOP 19, Volume Transfer Method

Standard(s): Giddings Metrology Laboratory Echelon II
 Volume Transfer Standards

Volume (Gallon)	Serial / ID #	As Found Volume Delivered @ 60 °F (in ³)	As Left Volume Delivered @ 60 °F (in ³)	Expanded Uncertainty ± (in ³)	Coefficient of Expansion (1 / °F)	Average Water Temp. at Time of Test (°F)	Manufacturer
100	1416	23101.0	23101.4	4.4	0.0000265	72.26	GSS

To convert from gallon to cubic inch: multiply by 231

To convert from cubic inch to gallon: divide by 231

Above values apply when prover is level and the 30-second drainage method is used after cessation of the main flow.

The above values do not apply if the calibration seals are tampered with or broken.

The expanded uncertainty given here is in compliance with NIST Technical Note 1297 ("Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results") with a coverage factor of two, representing a 95.45 % confidence level.

This report is not to be used to claim product endorsement by the Texas Department of Agriculture or any agency of the the U.S. Government. This document shall not be reproduced, except in full, without the written approval of the Texas Department of Agriculture Metrology Laboratory.

for Philip Wright
 Director For Consumer Product Protection

Daniel Gibbons
 Metrologist



**TEXAS DEPARTMENT OF AGRICULTURE
COMMISSIONER SID MILLER**

Metrology Laboratory - 1258 CR 226 / P.O. Box 1518 - Giddings, Texas 78942
Phone: (979) 542-3231 - Fax: (888) 205-7741

Test Number
G-000005270

REPORT OF CALIBRATION

FOR
100 Gallon LPG Prover
Serial Number: 2100



SUBMITTED BY
Cothorn's Tanker Inspections LLC
P.O. Box 1018
Caldwell, Texas 77836

The standards of Texas are traceable to the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The data below applies only to the artifacts identified in this report at the time of test.

Test Date: 08/31/2015
Calibration Due: 08/31/2016

Temperature Range: 15 °C to 30 °C
Relative Humidity Range: 40 % to 60 %

Procedure: NISTIR 7383, SOP 21, Volume Transfer Method
Standard(s)¹: Giddings Metrology Laboratory Echelon II Volume Transfer Standards

¹State standard(s) calibration and due dates are available upon request from the Giddings Metrology Laboratory.

The expanded standard uncertainty includes the standard uncertainty reported for the standard, the standard uncertainty for the measurement process, and a component of uncertainty to account for any observed deviations that have a significant effect on the calibration. The expanded uncertainty given here is in compliance with NIST Technical Note 1297 ("Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results") with a k = 2, representing a 95.45 % confidence level.

This report is not to be used to claim product endorsement by the Texas Department of Agriculture or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of the Texas Department of Agriculture Metrology Laboratory.

Note: Calibration status of the LPG pressure gauge has not been determined.

Note:

To convert from gallon to cubic inch: Multiply gallon by 231
To convert from cubic inch to gallon: Divide cubic inch by 231
To convert from cubic inch to cubic meter: Multiply cubic inch by 0.00016387064

for Philip Wright
Director For Consumer Product Protection

Daniel Gibbons
Metrologist



TEXAS DEPARTMENT OF AGRICULTURE
COMMISSIONER SID MILLER

Metrology Laboratory - 1258 CR 226 / P.O. Box 1518 - Giddings, Texas 78942

Phone: (979) 542-3231 - Fax: (888) 205-7741

REPORT OF CALIBRATION

Test Number
G-000005270

For

100 Gallon LPG Prover

Serial Number: 2100

Date Tested: 08/31/2015

Date Due: 08/31/2016

Submitted by

Cothern's Tanker Inspections LLC

P.O. Box 1018

Caldwell, Texas 77836

Temperature Range: 18 °C to 27 °C Relative Humidity Range: 40 % to 60 %

Calibration Procedure: NIST NISTIR 7383, SOP 21

The standard described above has been compared with the standards of the State of Texas and found to deliver as follows:

P.S.I.G.	Nominal Volume (Gallon)	As Found Volume Delivered at 60 °F (Gallon)	As Left Volume Delivered at 60 °F (Gallon)	Expanded Uncertainty ± (Gallon)
0	100	□ 99.994	□ 99.939	0.026
10	100	100.000	99.945	0.026
20	100	100.006	99.952	0.026
30	100	100.012	99.958	0.026
40	100	100.018	99.964	0.026
50	100	100.024	□ 99.971	0.026
60	100	100.030	99.976	0.026
70	100	100.036	99.982	0.026
80	100	100.043	99.988	0.026
90	100	100.049	99.994	0.026
100	100	□ 100.055	□ 100.000	0.026
110	100	100.061	100.001	0.026
120	100	100.067	100.003	0.026
130	100	100.073	100.005	0.026
140	100	100.079	100.007	0.026
150	100	100.085	□ 100.009	0.026
160	100	100.091	100.012	0.026
170	100	100.097	100.016	0.026
180	100	100.103	100.020	0.026
190	100	100.109	100.024	0.026
200	100	100.115	□ 100.028	0.026

Above values apply when prover is level, and when liquid reaches the top of the lower gage glass it is allowed to drain from the interior of the prover into the lower neck for 30-seconds prior to setting the bottom zero graduation. The above values do not apply if the levels are damaged or removed. Prover was calibrated at pressures indicated by a "□". All other values were determined by interpolation.

The expanded uncertainty given here is in compliance with NIST Technical Note 1297 ("Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results") with a coverage factor of two, representing a 95.45 % confidence level. This report is not to be used to claim product endorsement by the Texas Department of Agriculture or any agency of the U.S. Government. This document shall not be reproduced, except in full, without prior written approval of the Texas Department of Agriculture Metrology Laboratory.

for Philip Wright
Director For Consumer Product Protection

Daniel Gibbons
Metrologist



TEXAS DEPARTMENT OF AGRICULTURE
COMMISSIONER SID MILLER

Metrology Laboratory - 1258 CR 226 / P.O. Box 1518 - Giddings, Texas 78942

Phone: (979) 542-3231 - Fax: (888) 205-7741

REPORT OF CALIBRATION

Test Number

G-000005270

For

100 Gallon LPG Prover

Serial Number: 2100

Date Tested: 08/31/2015

Date Due: 08/31/2016

Submitted by

Cothem's Tanker Inspections LLC

P.O. Box 1018

Caldwell, Texas 77836

Temperature Range: 18 °C to 27 °C Relative Humidity Range: 40 % to 60 % Temperature Correction = $\Delta V = V(t - t_{ref}) \beta$

Calibration Procedure: NIST NISTIR 7383, SOP 21

Nominal Volume = 100 gal


Nominal Volume = 23100 in³

V = Volume in³

β = Prover Material Coefficient of Expansion / °F

t_{ref} = 60 °F

Table 3: Volume Corrections for Thermal Expansion or Contraction for a 100 (Gallon) GSS Steel LPG		
β : 0.0000160 / °F		
Temperature (°F)	ΔV (gal)	ΔV (in ³)
10	-0.0800	-18.480
20	-0.0640	-14.784
30	-0.0480	-11.088
40	-0.0320	-7.392
50	-0.0160	-3.696
60	0.0000	0.000
70	0.0160	3.696
80	0.0320	7.392
90	0.0480	11.088
100	0.0640	14.784
110	0.0800	18.480
120	0.0960	22.176



This document shall not be reproduced, except in full, without prior written approval of the Texas Department of Agriculture Metrology Laboratory.

Philip Wright

for Philip Wright
Director For Consumer Product Protection

Daniel Gibbons

Daniel Gibbons
Metrologist

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Texas

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon III
1000 kg to 1 mg
3750 lb to 0.001 lb
12 oz to 0.03125 oz

Weight Carts
6000 lb to 2500 lb

Volume Transfer, II
1000 gal to 5 gal
300 gal to 25 gal LPG



2015 to 2016

A handwritten signature in blue ink, appearing to read "Carol T. Hockert".

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2015-01-01 to 2016-12-31

Amended: 2015-12-31
Reduced upper limit of Mass
Echelon III Scope.