



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY

Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company <i>FAIRBANKS SCALES</i>	Email Address <i>Denver@fairbanks.com</i>	Application Date <i>2/24/16</i>	
Mailing Address <i>4850 Broadway</i>	City <i>Denver</i>	State <i>CO</i>	Zip Code <i>80216</i>
Telephone Number <i>303-209-5691</i>	Cell Phone Number <i>720-839-1792</i>	Fax Number <i>303-296-0269</i>	

Select below all device types your company will certify:

Weights (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input checked="" type="checkbox"/> 1. Rail	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute)
<input checked="" type="checkbox"/> 2. Truck	<input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater)
<input checked="" type="checkbox"/> 3. Livestock	<input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____
<input checked="" type="checkbox"/> 4. Hopper: Max. Capacity: _____	<input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____
<input type="checkbox"/> 5. Belt	<input type="checkbox"/> 5. LPG
<input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____	<input type="checkbox"/> 6. Stationary LPG
<input type="checkbox"/> 7. 30 lbs. or less	<input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____
<input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified)	<input type="checkbox"/> 8. Chemical
<input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 9. Anhydrous
	<input type="checkbox"/> 10. Loading Rack
	<input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
<i>e.g. 1001</i>	<i>e.g. John Doe</i>	<i>e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6</i>
	<i>Mark Conway</i>	<i>1, 2, 3, 4, 7</i>
	<i>Glen Meyer</i>	<i>1, 2, 3, 4, 7</i>

Continued on Page 2

Application for Registration as a Registered Service Company
Page 2



List below all field standards (attach current calibration reports):


Additional Application Items (initial where appropriate):

Standardized Test Report	<input checked="" type="checkbox"/> Copy enclosed <input type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input checked="" type="checkbox"/> Copy enclosed <input type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input checked="" type="checkbox"/> Copy enclosed <input type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am MANK CONWAY, and have authority to represent this company.
By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.



Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410

Bauske, Shelly A.

From: LOURDES VIDAURRI <LVIDAURRI@fairbanks.com>
Sent: Wednesday, March 16, 2016 10:24 AM
To: Bauske, Shelly A.
Subject: RE: Question - ND Application for Registration as a Registered Service Company

Shelly,

Thomas Becker is no longer with Fairbanks Scales. Tim Tobel however is still will Fairbanks but is not attached to the Denver location.

Ttobel@fairbanks.com

Lourdes Vidaurri | Service Administrator | Office: 800-435-4660 or 303-209-5691 | Fax:303-296-0269
| lvidaurri@fairbanks.com Fairbanks Scales Inc. | 4850 Broadway | Denver, CO 80216 |

Fairbanks Scales Inc. | 1525 South Gladiola | SLC, UT 84104 | www.fairbanksservice.com
Office: 800-453-8303 or 801-972-1491 | Fax:801-972-4389

Whitepaper: What every truck scale owner should know about scale maintenance.

“This email may contain confidential information intended only for the individual or organization named. If you received this email in error, please contact the above sender and or return or destroy the email as instructed. Thank you for your help in this matter.”

-----Original Message-----

From: Bauske, Shelly A. [<mailto:sbauske@nd.gov>]
Sent: Wednesday, March 16, 2016 9:21 AM
To: LOURDES VIDAURRI <LVIDAURRI@fairbanks.com>
Subject: Question - ND Application for Registration as a Registered Service Company


Good Morning Luli

Just one more question. I noticed that Tim Tobel and Wm. Thomas Becker Sr. are not listed as employees on your Application for Registration as a Registered Service Company. According to my records, they each have a current Registered Service Person Permit. Are they still employed with Fairbanks?

Thank you!

Shelly Bauske
Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
701-328-4070
701-328-2410 (fax)
sbauske@nd.gov

North Dakota

nd.gov Official Portal for
North Dakota State Government

SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

FAIRBANKS SCALES INC.

Corporation Details

System ID: 2987100 **Phone:** (816) 471-0231
Type: FOREIGN BUSINESS CORPORATION
Status: Active & Good Standing
Original File Date: 03/31/1988 **Effective Date:** 03/31/1988
State of Origin: Kansas

Nature of Business

SALES AND SERVICE OF SCALES AND RELATED EQUIPMENT

Principal Office

821 LOCUST ST KANSAS CITY, MO 64106-1908

Registered Agent

CORPORATION SERVICE COMPANY
1709 N 19TH ST STE 3
BISMARCK, ND 58501-2121
Established Date: Mar 01, 2010

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2015](#) [2016](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)

[Disclaimer](#)

[Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163

Company Name & Address: John Robinson Fairbanks Scales 2879 North 16th Road Worden, MT 59088	Date of Test: 4/7/2015	Test Number: 2015-060
Serial Number: 10 lb x 16		

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 4/6/2015
Description and condition of artifacts received: Items were fairly clean with significant signs of use.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
22.54	23.21	668.53	41.68	42.56

Conventional Mass Value: Assumed Density of Artifacts: 7.2 g/cm³

Nominal	Serial No.	As Found ±(g)	As Left ±(g)	Uncertainty ±(g)	NIST 105-1 Class F ±(g)	k factor
1000 lb	113-27	-8.7	-8.7	6.2	45	2.28
1000 lb	113-04	2.2	2.2	6.2	45	2.28
1000 lb	113-05	-55.7	21.2	6.2	45	2.28
1000 lb	133-23	26.2	26.2	6.2	45	2.28
1000 lb	133-06	1.2	1.2	6.2	45	2.28
1000 lb	113-25	19.3	19.3	6.2	45	2.28
1000 lb	113-38	-0.8	-0.8	6.2	45	2.28
5000 lb	Cart 113	-6078	4	17	680	2.28

Standards and Procedures used for testing:

The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7

All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:

The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:

The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows *NISTIR 6969*, issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: Dave Fraser

David Fraser

Email: dfraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

RECEIVED

MAR - 9 2016

ND Public
Service
Commission**Certificate of Mass Calibration**Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163**Company Name & Address:**

John Robinson

Fairbanks Scales

2879 North 16th Road

Worden, MT 59088

Date of Test:

4/7/2015

Test Number:

2015-060

Serial Number:

1000 lb x 16

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received:

4/6/2015

Description and condition of artifacts received:

Items were fairly clean with significant signs of use.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
22.54	23.21	668.53	41.68	42.56

Conventional Mass Value:**Assumed Density of Artifacts:**7.2 g/cm³

Nominal	Serial No.	As Found ±(g)	As Left ±(g)	Uncertainty ±(g)	NIST 105-1 Class F ±(g)	k factor
1000 lb	113-29	-30.7	-30.7	6.2	45	2.28
1000 lb	113-01	-65.7	19.2	6.2	45	2.28
1000 lb	113-02	-6.7	-6.7	6.2	45	2.28
1000 lb	113-28	7.2	7.2	6.2	45	2.28
1000 lb	113-21	-27.8	-27.8	6.2	45	2.28
1000 lb	113-31	-38.9	-2.7	6.2	45	2.28
1000 lb	113-37	-14.7	-14.7	6.2	45	2.28
1000 lb	113-03	-11.7	-11.7	6.2	45	2.28
1000 lb	113-32	-18.8	-18.8	6.2	45	2.28

Standards and Procedures used for testing:

The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7

All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:

The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:

The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows *NISTIR 6969*, issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated *k* factor, for a confidence level of 95.45%.

State Metrologist: Dave Fraser

David Fraser

Email: dafraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163



Company Name & Address: John Robinson Fairbanks Scales 2879 North 16th Road Worden, MT 59088	Date of Test: 4/7/2015	Test Number: 2015-059
	Kit Number No Kit #	

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 4/6/2015
Description and condition of artifacts received: Items were in poor condition, surface rust, not painted.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
22.2	22.65	663.70	40.21	42.85

Conventional Mass Value:		Assumed Density of Artifacts:			7.2 g/cm ³	
Nominal	Serial No.	As Found ± (g)	As Left ± (g)	Uncertainty ± (g)	NIST 105-1 Class F ± (g)	k factor
50 lb	614-317	-5.81	0.33	0.30	2.3	2.28
50 lb	11	-0.31	-0.31	0.30	2.3	2.28
50 lb	D-0130	1.31	1.31	0.30	2.3	2.28
50 lb	4	-4.27	-0.11	0.30	2.3	2.28
50 lb	5	-3.26	0.00	0.30	2.3	2.28
50 lb	FW-C-31	-6.17	0.43	0.30	2.3	2.28
50 lb	8	1.42	1.42	0.30	2.3	2.28
50 lb	1	-2.28	0.03	0.30	2.3	2.28
50 lb	HM-D-255	-8.03	0.38	0.30	2.3	2.28

Standards and Procedures used for testing:

The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7

All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:

The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:

The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows *NISTIR 6969*, Issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: Dave Fraser

David Fraser

Email: dafraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.



Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
 2801 N Cooke St. Helena, Montana 59601
 (406)449-2582 FAX (406)443-8163

Company Name & Address: John Robinson Fairbanks Scales 2879 North 16th Road Worden, MT 59088	Date of Test: 4/7/2015	Test Number: 2015-059
	Kit Number No Kit #	

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 4/6/2015
 Description and condition of artifacts received: Items were in poor condition, surface rust, not painted.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
22.2	22.65	663.70	40.21	42.85

Conventional Mass Value:		Assumed Density of Artifacts:			7.2 g/cm ³	
Nominal	Serial No.	As Found ± (g)	As Left ± (g)	Uncertainty ± (g)	NIST 105-1 Class F ± (g)	k factor
50 lb	10	-2.81	-0.08	0.30	2.3	2.28

Standards and Procedures used for testing:
 The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7
 All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, Issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:
 The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:
 The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows *NISTIR 6969*, issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: Dave Fraser

David Fraser

Email: dafraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.



Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
 2801 N Cooke St. Helena, Montana 59601
 (406)449-2582 FAX (406)443-8163

Company Name & Address: John Robinson Fairbanks Scales 2879 North 16th Road Worden, MT 59088	Date of Test: 4/8/2015	Test Number: 2015-061
	Kit Number FB01G	

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 4/6/2015
 Description and condition of artifacts received: Items were in good condition with no discernable defects.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
20.44	20.9	661.92	40.43	40.56

Conventional Mass Value:			Assumed Density of Artifacts:		7.84 g/cm ³	
Nominal	Serial No.	As Found ±(g)	As Left ±(g)	Uncertainty ±(g)	NIST 105-1 Class F ±(g)	k factor
1 kg	FB01G	0.038	0.038	0.012	0.1	2.28
500 g	FB01G	0.008 4	0.008 4	0.008 3	0.07	2.28
200 g - 1	FB01G	0.017 1	0.017 1	0.005 1	0.04	2.28
200 g - 2	FB01G	0.006 6	0.006 6	0.005 1	0.04	2.28
100 g	FB01G	0.006 9	0.006 9	0.000 7	0.02	2.28
50 g	FB01G	0.001 8	0.001 8	0.000 7	0.01	2.28
20 g - 1	FB01G	0.001 6	0.001 6	0.000 1	0.004	2.28
20 g - 2	FB01G	0.001 0	0.001 0	0.000 1	0.004	2.28
10 g	FB01G	0.000 5	0.000 5	0.000 1	0.002	2.28

Standards and Procedures used for testing:

The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7

All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:

The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:

The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows *NISTIR 6969*, issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: Dave Fraser

David Fraser

Email: dafraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.



Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
 2801 N Cooke St. Helena, Montana 59601
 (406)449-2582 FAX (406)443-8163

Company Name & Address:	Date of Test:	Test Number:
John Robinson	4/8/2015	2015-061
Fairbanks Scales		Kit Number
2879 North 16th Road		FB01G
Worden, MT 59088		

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 4/6/2015
 Description and condition of artifacts received: Items were in good condition with no discernable defects.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
20.44	20.90	661.92	40.43	40.56

Conventional Mass Value:			Assumed Density of Artifacts:		7.84 g/cm ³	
Nominal	Serial No.	As Found ±(g)	As Left ±(g)	Uncertainty ± (g)	NIST 105-1 Class F ± (g)	k factor
5 g	FB01G	0.000 21	0.000 21	0.000 27	0.001 5	2.28
2 g -1	FB01G	0.000 14	0.000 14	0.000 27	0.001 1	2.28
2 g -2	FB01G	0.000 49	0.000 49	0.000 27	0.001 1	2.28
1 g	FB01G	0.000 35	0.000 35	0.000 04	0.000 9	2.28

Standards and Procedures used for testing:
 The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7
 All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:
 The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:
 The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows *NISTIR 6969*, issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: *Dave Fraser*

Email: dafraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

United States Department of Commerce
National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Montana

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon III	Volume Transfer, II
30 kg to 1 mg	1500 gal to 5 gal
3000 lb to 0.001 lb	100 gal LPG to 25 gal LPG
8 oz to 0.03125 oz	
Weight Carts	
5000 lb to 2000 lb	



2015

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2015-01-01 to 2015-12-31