



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY
Public Service Commission
SFN 61277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company <i>Jeffrey D. Doyle, dba: Hobart Sale Service</i>		Email Address <i>Hobartfargo@yahoo.com</i>		Application Date <i>9/9/16</i>	
Mailing Address <i>18 14th St S</i>		City <i>Fargo</i>		State <i>ND</i>	Zip Code <i>58103</i>
Telephone Number <i>(701) 237-5240</i>		Cell Phone Number		Fax Number <i>(701) 235-0772</i>	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute)
<input type="checkbox"/> 2. Truck	<input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater)
<input type="checkbox"/> 3. Livestock	<input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____
<input type="checkbox"/> 4. Hopper: Max. Capacity: _____	<input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____
<input type="checkbox"/> 5. Belt	<input type="checkbox"/> 5. LPG
<input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____	<input type="checkbox"/> 6. Stationary LPG
<input checked="" type="checkbox"/> 7. 30 lbs. or less	<input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____
<input type="checkbox"/> 8. Class II (Indicate on your calibration report which weight kit is Class II certified)	<input type="checkbox"/> 8. Chemical
<input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 9. Anhydrous
	<input type="checkbox"/> 10. Loading Rack
	<input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
<i>e.g. 1001</i>	<i>e.g. John Doe</i>	<i>e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6</i>
<i>1463</i>	<i>Clayton Schauer</i>	<i>Scales 6+7</i>
<i>1750</i>	<i>Monty Brown</i>	<i>Scales 6+7</i>



Application for Registration as a Registered Service Company
Page 2

List below all field standards (attach current calibration reports):

50# weights Qty 7	
25# weights Qty 6	
Troemner 31# weight kit 21 decimals + fractionals	
Rice Lake weight kit 31.11# 10#-.001#	


Additional Application Items (initial where appropriate):

Standardized Test Report	J.D. <input checked="" type="checkbox"/> Copy enclosed No change in report filed previously
Tested and Approved Sticker	J.D. <input checked="" type="checkbox"/> Copy enclosed No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed <input type="checkbox"/> No change in crimped lead wire seal filed previously J.D. <input checked="" type="checkbox"/> Do not have one

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am Jeff Doyle, President, and have authority to represent this company.
 By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.



 Signature


Send Completed Application and Related Documents To:

Public Service Commission
 600 E Boulevard Ave Dept 408
 Bismarck ND 58505-0480
 Telephone: (701) 328-2400
 Fax: (701) 328-2410



North Dakota

nd.gov Official Portal for North Dakota State Government



SECRETARY OF STATE NORTH DAKOTA



[Home](#) | [Business Records Search](#)

HOBART SALES AND SERVICE

Trade Name

System ID: 26828700

Phone: (701) 237-5240

Status: Active

Original File Date: 03/11/2010

Last Renewal Date: 03/11/2015

Nature of Business

SALES & SERVICE OF COMMERCIAL FOOD EQUIPMENT

Owners

JEFFREY D. DOYLE INC.

18 14TH ST S

FARGO, ND 58103-1620

[Return to Search Results](#)

[Contact Us](#)

[Disclaimer](#)

[Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



Receipt Date: October 26, 2016
Cal. Date: October 27, 2016
Report Date: October 27, 2016

Report No.: 336675
Set Serial No.: None
Barcode: 200923

Calibration Certificate

HOBART SALES & SERVICE
18 SOUTH 14TH STREET
FARGO, ND 58103-1620
Contact: JEFF DOYLE
Phone: 701-237-5240
PO Number: NONE
SOP: 8
Technician ID: 11

Item(s) Submitted: Cast Hand Weights
Manufacturer: Rice Lake and Hobart
Weight Type: II
Equipment ID: None
Condition: Good
Temperature: 20.9 °C
Pressure: 742.6 mmHg
Relative Humidity: 49.6 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
50 lb		-630	-630	F	F	2.01	120
50 lb		80	80	F	F	2.01	120
50 lb		290	290	F	F	2.01	120
50 lb		1670	1670	F	F	2.01	120
50 lb		260	260	F	F	2.01	120
50 lb		-14620	110	*	F	2.01	120
50 lb		-10430	50	*	F	2.01	120
25 lb		139	139	F	F	2.01	67
25 lb		149	149	F	F	2.01	67
25 lb		-181	-181	F	F	2.01	67
25 lb		109	109	F	F	2.01	67
25 lb		59	59	F	F	2.01	67
25 lb		159	159	F	F	2.01	67

* Weight(s) as found exceed NIST Class F Tolerance.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³ at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Pete Whebbe

Pete Whebbe
Metrologist

Reviewed by:

Erik Alfvin
Erik Alfvin
Metrologist



Receipt Date: October 26, 2016
 Cal. Date: October 27, 2016
 Report Date: October 27, 2016

Report No.: 336673
 Set Serial No.: 4760
 Barcode: 019210

Calibration Certificate

HOBART SALES & SERVICE
 18 SOUTH 14TH STREET
 FARGO, ND 58103
 Contact: JEFF DOYLE
 Phone: 701-237-5240
 PO Number: NONE
 SOP: 8
 Technician ID: 09

Item(s) Submitted: 30 lb kit w/ decimals & fractions
 Manufacturer: Troemner
 Weight Type: I & II
 Equipment ID: None
 Condition: Good
 Temperature: 21.1 °C
 Pressure: 742.9 mmHg
 Relative Humidity: 45.5 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
5 lb		175	175	F	F	2.02	12
5 lb		165	165	F	F	2.02	12
5 lb		183	183	F	F	2.02	12
5 lb		79	79	F	F	2.02	12
5 lb		189	189	F	F	2.02	12
1 lb		34.0	34.0	F	F	2.02	2.8
1 lb		34.7	34.7	F	F	2.02	2.8
1 lb		36.4	36.4	F	F	2.02	2.8
1 lb		32.0	32.0	F	F	2.02	2.8
1 lb		44.1	44.1	F	F	2.02	2.8
0.2 lb		4.50	4.50	F	F	2.03	0.37
0.2 lb		3.50	3.50	F	F	2.03	0.37
0.1 lb		7.11	7.11	F	F	2.03	0.30
0.05 lb		2.60	2.60	F	F	2.03	0.30
0.02 lb		0.83	0.83	F	F	2.03	0.16
0.02 lb		0.83	0.83	F	F	2.03	0.16
0.01 lb		0.69	0.69	F	F	2.03	0.12
0.005 lb		0.58	0.58	F	F	2.03	0.10
0.002 lb		0.589	0.589	F	F	2.03	0.066
0.001 lb		-0.450	-0.450	F	F	2.03	0.057



Receipt Date: October 26, 2016
 Cal. Date: October 27, 2016
 Report Date: October 27, 2016

Report No.: 336673
 Set Serial No.: 4760
 Barcode: 019210

Continued,

Calibration Certificate

HOBART SALES & SERVICE
 18 SOUTH 14TH STREET
 FARGO, ND 58103
 Contact: JEFF DOYLE
 Phone: 701-237-5240
 PO Number: NONE
 SOP: 8
 Technician ID: 09

Item(s) Submitted: 30 lb kit w/ decimals & fractions
 Manufacturer: Troemner
 Weight Type: I & II
 Equipment ID: None
 Condition: Good
 Temperature: 21.1 °C
 Pressure: 742.9 mmHg
 Relative Humidity: 45.5 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
0.5 lb		3.9	3.9	F	F	2.02	2.3
4 oz		9.41	9.41	F	F	2.00	0.60
2 oz		6.02	6.02	F	F	2.00	0.36
1 oz		0.93	0.93	F	F	2.00	0.30
1/2 oz		1.41	1.41	F	F	2.00	0.24
1/4 oz		0.94	0.94	F	F	2.00	0.12
1/8 oz		-0.527	-0.527	F	F	2.00	0.090
1/16 oz		0.529	0.529	F	F	2.00	0.060
1/32 oz		0.496	0.496	F	F	2.00	0.049
1/32 oz		0.306	0.306	F	F	2.00	0.049

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³ at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Heidi Jones

 Laboratory Administrator

Reviewed by:
 Mark Nicollet

 Quality Manager



Receipt Date: October 26, 2016
Cal. Date: October 27, 2016
Report Date: October 27, 2016

Report No.: 336674
Set Serial No.: None
Barcode: 200897

Calibration Certificate

HOBART SALES & SERVICE
18 SOUTH 14TH STREET
FARGO, ND 58103
Contact: JEFF DOYLE
Phone: 701-237-5240
PO Number: NONE
SOP: 8
Technician ID: 09

Item(s) Submitted: 30 lb kit w/ decimals
Manufacturer: Rice Lake
Weight Type: I & II
Equipment ID: None
Condition: Good
Temperature: 21.1 °C
Pressure: 742.7 mmHg
Relative Humidity: 45.5 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
10 lb		144	144	F	F	2.02	15
10 lb		129	129	F	F	2.02	15
5 lb		74	74	F	F	2.02	12
2 lb		25.9	25.9	F	F	2.02	4.8
2 lb		29.2	29.2	F	F	2.02	4.8
1 lb		28.1	28.1	F	F	2.02	2.8
0.5 lb		15.4	15.4	F	F	2.02	2.3
0.2 lb		4.83	4.83	F	F	2.03	0.37
0.2 lb		4.78	4.78	F	F	2.03	0.37
0.1 lb		4.07	4.07	F	F	2.03	0.30
0.05 lb		1.91	1.91	F	F	2.03	0.30
0.02 lb		0.47	0.47	F	F	2.03	0.16
0.02 lb		0.59	0.59	F	F	2.03	0.16
0.01 lb		0.67	0.67	F	F	2.03	0.12
0.005 lb		0.23	0.23	F	F	2.03	0.10
0.002 lb		0.517	0.517	F	F	2.03	0.066
0.002 lb		0.213	0.213	F	F	2.03	0.066
0.001 lb		0.433	0.433	F	F	2.03	0.057

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³ at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Heidi Jones
Heidi Jones
Laboratory Administrator

Reviewed by:
Mark Nicollet
Mark Nicollet
Quality Manager

United States Department of Commerce
National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Minnesota

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon II	Weight Carts	Volume Gravimetric, I
50 kg to 1 mg	10 000 lb to 2000 lb	20 L to 10 mL
1000 lb to 0.001 lb	Wheel Load Weighers	100 gal to 0.25 qt
4 oz to 0.03125 oz	20 000 lb to 2000 lb	Volume Transfer, II
Mass Echelon III	Railroad Test Cars	1500 gal to 5 gal
50 kg to 1 mg	110 000 lb to 80 000 lb	100 gal to 25 gal LPG
5000 lb to 0.001 lb		
4 oz to 0.03125 oz		



2016 to 2017

Carol T. Hoekert, Chief
NIST Office of Weights and Measures

Effective Dates: 2016-01-01 to 2017-12-31