



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY

Public Service Commission
SFN 51277 (4/2013)

TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company Kinder Morgan	Email Address eugene_drewlow@kindermorgan.com	Application Date 5/1/2013	
Mailing Address Box 34	City Carrington	State ND	Zip Code 58421
Telephone Number 701-674-3318	Cell Phone Number 701-652-5227	Fax Number 701-674-3423	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input checked="" type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
<i>e.g. 1001</i>	<i>e.g. John Doe</i>	<i>e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6</i>
1603	Eugene M. Drewlow	Liquid-10
1673	Eugene M. Hartl	Liquid-10
1674	Paul R. Reimer	Liquid-10

Application for Registration as a Registered Service Company
Page 2

List below all field standards (attach current calibration reports):

See inclosed report	


Include the following additional items with this application:

- A sample of your standardized test report as required by North Dakota Administrative Code Section 69-10-04-05.
- A sample of your "tested and approved" sticker as required by North Dakota Administrative Code Section 69-10-01-03.
- A photocopy of each side of your crimped lead wire seal.

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am Eugene M. Drub, and have authority to represent this company.
By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.


Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410

Meter Engineers, Inc.



Customer | Kinder Morgan
Location | Carrington, ND
Designation | Carrington Terminal Propane
Serial | E-6409
Prover Type | Unidirectional



Certificate of Calibration

Jun 19, 2013

This is to certify that Meter Engineers, Inc. calibrated subject Unidirectional meter prover and established the volume at 60 degrees fahrenheit and atmospheric pressure for a one way trip of the displacer to be:

Average One Way Volume: 7.2736 bbl


The calibrating medium used was water and the calibration was performed in accordance with the latest edition of the API Standards Chapters 4, 11, & 12 by Gary Clark and certified by the undersigned. Test performed using trailer unit number 134.

Certified Test Measures Used:

Measure #	Gal	NIST #
1	100	6996
2	100	6997
3	50	6998
6	5	7002

Signed





Customer	Kinder Morgan
Location	Carrington, ND
Designation	Carrington Terminal Propane
Serial	E-6409
Prover Type	Unidirectional

UniDirectional Run Comparisons


Run 1 = 70,567.6734		
	0.0014 %	
Run 2 = 70,566.6881		0.0047 %
	0.0061 %	
Run 3 = 70,570.9943		

Previous Volume was 7.27191 barrels Previous Volume was 70552.07082 inches

This WaterDraw 7.2736 barrels .0232 % Diff from Previous Volume

This WaterDraw 70,568.4519 inches

(1.156390 cubic meters)



Customer	Kinder Morgan
Location	Carrington, ND
Designation	Carrington Terminal Propane
Serial	E-6409
Prover Type	Unidirectional

Prover Information

Prover Type	Unidirectional
Designation	Carrington
Serial	E-6409
Pipe I.D.	7.981
ANSI	
Ball Percent Oversize	3 %
Ball Composition	Yellow
Pipe Wall Thickness	.322
Mfg of Prover	F. H. Maloney
Mfg of Interchange	F. H. Maloney
Mfg of SwitchBase	F. H. Maloney
Mfg of Switches	F. H. Maloney
First Detector Switch Seal Number	
Second Detector Switch Seal Number	
Coated?	Yes
Insulated?	No
Metallic Composition	Carbon Steel
Coefficient of Cubical Expansion	.0000186
Modulus of Elasticity	30000000
Unit Number	134
Above or Below Ground	Above
Notes	new switch kit installed this water draw

Methodology

CTL Temperature Correction Factor is based on API Petroleum Measurement manual Chapter 11.2.3


CTS Metal Correction Factor = $(1 + ((\text{Can Temp in F} - 60) \text{ times the Can Coefficient of Expansion}))$
 divided by $(1 + ((\text{Prover Temp in F} - 60) \text{ times the Prover Coefficient of Expansion}))$

CPS = $1 + (\text{Pressure times the Pipe ID} \text{ divided by the Modulus of Elasticity times the Prover Pipewall Thickness})$

CPL = $1 \text{ divided by } \{1 - (\text{Compressibility Constant of Water times Prover Pressure})\}$

File Source: Waterdraws

Meter Engineers, Inc.



Customer	Kinder Morgan
Location	Carrington, ND
Designation	Carrington Terminal Propane
Serial	E-6409
Prover Type	Unidirectional

Run Calculation Summary

Gary Clark

Run # 1 Elapsed Time 5 : 23 min GPM 56.76 Pressure 42

Measure	Certified Volume	Scale Reading	Actual Volume	Prover Temp	Can Temp	CTL Temp Correction	CTS Metal Correction	Adjusted Volume
Can # 3 50gal	11,549.47	100	11,649.47	57.4	57.5	0.999992	0.999982	11,649.1671
Can # 1 100gal	23,090.34	0	23,090.34	57.4	57.5	0.999992	0.999982	23,089.7397
Can # 2 100gal	23,091.12	1	23,092.12	57.4	57.6	0.999984	0.999984	23,091.3811
Can # 3 50gal	11,549.47	47	11,596.47	57.4	57.8	0.999969	0.999990	11,595.9945
Can # 6 5gal	1,153.39	0	1,153.39	57.4	58.2	0.999937	1.000000	1,153.3173
Volume This Run								70,579.5997
Divided by CPS								1.000035
Divided by CPL								1.000134

Pressure & Temperature Adjusted Volume This Run **70,567.6734**

Run # 2 Elapsed Time 7 : 18 min GPM 41.85 Pressure 42

Measure	Certified Volume	Scale Reading	Actual Volume	Prover Temp	Can Temp	CTL Temp Correction	CTS Metal Correction	Adjusted Volume
Can # 3 50gal	11,549.47	100	11,649.47	57.8	58	0.999984	0.999988	11,649.1438
Can # 1 100gal	23,090.34	0	23,090.34	57.8	58	0.999984	0.999988	23,089.6935
Can # 2 100gal	23,091.12	0	23,091.12	57.8	58.3	0.999959	0.999996	23,090.0809
Can # 3 50gal	11,549.47	47	11,596.47	57.8	58.4	0.999951	0.999999	11,595.8902
Can # 6 5gal	1,153.39	.5	1,153.89	57.8	58.8	0.999918	1.000009	1,153.8058

Run Calculation Summary (cont)

Volume This Run	70,578.6142
Divided by CPS	1.000035
Divided by CPL	1.000134

Pressure & Temperature Adjusted Volume This Run **70,566.6881**

Run # 3

Elapsed Time 5 : 22 min

GPM 56.94

Pressure 42

Measure	Certified Volume	Scale Reading	Actual Volume	Prover Temp	Can Temp	CTL Temp Correction	CTS Metal Correction	Adjusted Volume
Can # 3 50gal	11,549.47	100	11,649.47	59.1	58.9	1.000017	0.999988	11,649.5282
Can # 1 100gal	23,090.34	0	23,090.34	59.1	59.1	1.000000	0.999993	23,090.1784
Can # 2 100gal	23,091.12	1	23,092.12	59.1	59.3	0.999983	0.999998	23,091.6813
Can # 3 50gal	11,549.47	49	11,598.47	59.1	59.4	0.999975	1.000001	11,598.1916
Can # 6 5gal	1,153.39	0	1,153.39	59.1	59.7	0.999949	1.000009	1,153.3416

Volume This Run	70,582.9211
Divided by CPS	1.000035
Divided by CPL	1.000134

Pressure & Temperature Adjusted Volume This Run **70,570.9943**

Meter Engineers, Inc.



WaterDraw Calibration Worksheet

Customer: Kinder Morgan
 Location: Carrington, ND
 Designation: Carrington Terminal Propane
 Serial: E-6409
 Prover Type: Unidirectional

Slow

Technician: Gary Clark
 Scale Reading: 118.3 cubic inches
 Tolerance: 14.1 cubic inches

Date: 6-19-13 1 57.4 42 5:23 2 57.8 42 7:18 3 59.1 42 5:22

Run Prv Temp Pres Time Run Prv Temp Pres Time Run Prv Temp Pres Time

Measure	Scale Reading	Temp	Scale Reading	Temp	Scale Reading	Temp
Can # 3 50 gal	3 <u>100</u>	<u>57.5</u>	3 <u>100</u>	<u>58</u>	3 <u>100</u>	<u>58.9</u>
Can # 1 100 gal	1 <u>0</u>	<u>57.5</u>	1 <u>0</u>	<u>58</u>	1 <u>0</u>	<u>59.1</u>
Can # 2 100 gal	2 <u>+1</u>	<u>57.6</u>	2 <u>0</u>	<u>58.3</u>	2 <u>+1</u>	<u>59.3</u>
Can # 3 50 gal	3 <u>47</u>	<u>57.8</u>	3 <u>47</u>	<u>58.4</u>	3 <u>49</u>	<u>59.4</u>
Can # 6 5 gal	6 <u>0</u>	<u>58.2</u>	6 <u>.5</u>	<u>58.8</u>	6 <u>0</u>	<u>59.7</u>
	<u>148</u> Total		<u>147.5</u> Total		<u>150</u> Total	

Thermometer Numbers: 4T1555 1630273

Witnesses: _____ (Company) _____
[Signature] MEI _____
[Signature] Kinder Morgan _____



National Institute of Standards and Technology
Technology Administration, U.S. Department of Commerce

REPORT OF CALIBRATION

FOR

A ONE-HUNDRED (100) GALLON VOLUME PROVER
(Graduated Neck Type)

March 31, 2009

Manufacturer: Seraphin
Rancocas, NJ

NIST Seal Number: 6996
Material: Stainless Steel
Serial Number: 2470-A

submitted by

Meter Engineers, Inc.
7718 W. 53rd St. N.
Mazie, KS 67101

(Reference: Purchase Order Number 003680, dated January 15, 2009)

The internal volume of the prover described above has been determined by the gravimetric method [1]. The gravimetric method requires weighing the vessel dry and empty and re-weighing it when filled with a fluid of known density. The internal or contained volume was determined in this way and the value is given in Table 1 using the requested units. The fluid used was distilled water and the prover was leveled before determining the volume.

To determine the delivered volume, the contained volume is poured from the prover by opening the valve at the bottom of the vessel. When this flow finishes, the valve is held open for 30 seconds to complete the drain procedure. Subsequent re-weighing completes the gravimetric procedure and enables calculation of the delivered volume, also given in Table 1. Both the contained and delivered volumes are given for the scale reading of zero (0) and have been corrected for the reference temperature of 15.56 °C (60 °F), assuming a volumetric coefficient of expansion of 0.0000477 per °C (0.0000265 per °F).

[1] Bean, V. E., Espina, R. I., Wright, J. D., Houser, J. F., Sheckels, S. D., and Johnson, A. N., "NIST Calibration Services for Liquid Volume," NIST Special Publication 250-72, National Institute of Standards and Technology, March 19, 2006.

Table 1. Contained and delivered volumes for the tested vessel for a scale reading* of zero.

	Volume Contained	Volume Delivered
gal at 60 °F	99.9755	99.9582
in ³ at 60 °F	23094.35	23090.34

The volume measurement procedure was repeated 5 times with the neck scale filled approximately to zero each time. The repeatability of the 5 measurements was 22 parts in 10⁶ and the expanded uncertainty in the measured volume is $\pm 0.007\%$. It was calculated according to References [1] and [2] with a 95 % confidence level† and is traceable to NIST mass, temperature, pressure, and humidity standards, and a NIST distilled water density determination.





Figure 1. Photographs of the volume prover.

The input data used for calculation of the prover volume are given in the spreadsheet attached to this report.

* The scale reading is determined by the intersection of the horizontal plane, tangent to the bottom of the meniscus reading on the gauge tube. For this vessel, the scale range was from -175 and +200 and each division is equivalent to 5 in³.

[2] Taylor, B. N. and Kuyatt, C. E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results," NIST Technical Note 1297, National Institute of Standards and Technology (January 1993).

† Coverage factor of 2.09 for 20 effective degrees of freedom.




Figure 2. Calibration control chart for 100 gallon graduated neck test measure SN 2470-A.

Table 2. Results of prior calibrations for the delivered volume.

Date	Delivered Volume [in ³]	Difference from Prior [in ³]	Degree of Equivalence [-]
03/23/09	23090.34	-0.43	-0.07
02/21/06	23090.77	-1.20	-0.20
04/07/03	23091.97	-3.43	-0.58
03/22/99	23095.40	1.30	0.22
06/06/94	23094.10	-	-

For the Director,
National Institute of Standards and Technology

Dr. John D. Wright
Project Leader, Fluid Metrology Group
Process Measurements Division
Chemical Science and Technology Laboratory
National Institute of Standards and Technology

Sherry Sheckels
Calibration Technician, Fluid Metrology Group
Process Measurements Division
Chemical Science and Technology Laboratory
National Institute of Standards and Technology



National Institute of Standards and Technology
Technology Administration, U.S. Department of Commerce

REPORT OF CALIBRATION

FOR

A ONE-HUNDRED (100) GALLON VOLUME PROVER
(Graduated Neck Type)

March 31, 2009

Manufacturer: Seraphin
Rancocas, NJ

NIST Seal Number: 6997
Material: Stainless Steel
Serial Number: 2470-B

submitted by

Meter Engineers, Inc.
7718 W. 53rd St. N.
Mazie, KS 67101

(Reference: Purchase Order Number 003680, dated January 15, 2009)

The internal volume of the prover described above has been determined by the gravimetric method [1]. The gravimetric method requires weighing the vessel dry and empty and re-weighing it when filled with a fluid of known density. The internal or contained volume was determined in this way and the value is given in Table 1 using the requested units. The fluid used was distilled water and the prover was leveled before determining the volume.

To determine the delivered volume, the contained volume is poured from the prover by opening the valve at the bottom of the vessel. When this flow finishes, the valve is held open for 30 seconds to complete the drain procedure. Subsequent re-weighing completes the gravimetric procedure and enables calculation of the delivered volume, also given in Table 1. Both the contained and delivered volumes are given for the scale reading of zero (0) and have been corrected for the reference temperature of 15.56 °C (60 °F), assuming a volumetric coefficient of expansion of 0.0000477 per °C (0.0000265 per °F).

[1] Bean, V. E., Espina, P. I., Wright, J. D., Houser, J. F., Sheckels, S. D., and Johnson, A. N., "NIST Calibration Services for Liquid Volume," NIST Special Publication 250-72, National Institute of Standards and Technology, March 24, 2006.

Table 1. Contained and delivered volumes for the tested vessel for a scale reading* of zero.

	Volume Contained	Volume Delivered
gal at 60 °F	99.9812	99.9615
in ³ at 60 °F	23095.67	23091.12

The volume measurement procedure was repeated 5 times with the neck scale filled approximately to zero each time. The repeatability of the 5 measurements was 28 parts in 10⁶ and the expanded uncertainty in the measured volume is $\pm 0.008\%$. It was calculated according to References [1] and [2] with a 95 % confidence level† and is traceable to NIST mass, temperature, pressure, and humidity standards, and a NIST distilled water density determination.




Figure 1. Photographs of the volume prover.

The input data used for calculation of the prover volume are given in the spreadsheet attached to this report.

* The scale reading is determined by the intersection of the horizontal plane, tangent to the bottom of the meniscus reading on the gauge tube. For this vessel, the scale range was from -200 and +200 and each division is equivalent to 5 in³.

[2] Taylor, B. N. and Kuyatt, C. E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results," NIST Technical Note 1297, National Institute of Standards and Technology (January 1993).

† Coverage factor of 2.14 for 15 effective degrees of freedom.




Figure 2. Calibration control chart for 100 gallon graduated neck test measure SN 2470-B.

Table 2. Results of prior calibrations for the delivered volume.

Date	Delivered Volume [in ³]	Difference from Prior [in ³]	Degree of Equivalence [-]
03/25/09	23091.12	0.39	0.07
02/22/06	23090.74	-0.19	-0.03
04/08/03	23090.93	-4.87	-0.83
03/22/99	23095.80	2.50	0.42
06/06/94	23093.30	-	-

For the Director,
National Institute of Standards and Technology

Dr. John D. Wright
Project Leader, Fluid Metrology Group
Process Measurements Division
Chemical Science and Technology Laboratory
National Institute of Standards and Technology

Sherry Sheckels
Calibration Technician, Fluid Metrology Group
Process Measurements Division
Chemical Science and Technology Laboratory
National Institute of Standards and Technology



National Institute of Standards and Technology
Technology Administration, U.S. Department of Commerce

REPORT OF CALIBRATION

FOR

A FIFTY (50) GALLON VOLUME PROVER
(Graduated Neck Type)

March 31, 2009

Manufacturer: Seraphin
Rancocas, NJ

NIST Seal Number: 6998
Material: Stainless Steel
Serial Number: 7293-C

submitted by

Meter Engineers, Inc.
7718 W. 53rd St. N.
Mazie, KS 67101

(Reference: Purchase Order Number 003680, dated January 15, 2009)

The internal volume of the prover described above has been determined by the gravimetric method [1]. The gravimetric method requires weighing the vessel dry and empty and re-weighing it when filled with a fluid of known density. The internal or contained volume was determined in this way and the value is given in Table 1 using the requested units. The fluid used was distilled water and the prover was leveled before determining the volume.

To determine the delivered volume, the contained volume is poured from the prover by opening the valve at the bottom of the vessel. When this flow finishes, the valve is held open for 30 seconds to complete the drain procedure. Subsequent re-weighing completes the gravimetric procedure and enables calculation of the delivered volume, also given in Table 1. Both the contained and delivered volumes are given for the scale reading of zero (0) and have been corrected for the reference temperature of 15.56 °C (60 °F), assuming a volumetric coefficient of expansion of 0.0000477 per °C (0.0000265 per °F).

[1] Bean, V. E., Espina, P. I., Wright, J. D., Houser, J. F., Sheckels, S. D., and Johnson, A. N., "NIST Calibration Services for Liquid Volume," NIST Special Publication 250-72, National Institute of Standards and Technology, March 24, 2006.

Table 1. Contained and delivered volumes for the tested vessel for a scale reading* of zero.

	Volume Contained	Volume Delivered
gal at 60 °F	50.0078	49.9977
in ³ at 60 °F	11551.80	11549.47

The volume measurement procedure was repeated 5 times with the neck scale filled approximately to zero each time. The repeatability of the 5 measurements was 8 parts in 10⁶ and the expanded uncertainty in the measured volume is $\pm 0.009\%$. It was calculated according to References [1] and [2] with a 95 % confidence level[†] and is traceable to NIST mass, temperature, pressure, and humidity standards, and a NIST distilled water density determination.




Figure 1. Photographs of the volume prover.

The input data used for calculation of the prover volume are given in the spreadsheet attached to this report.

* The scale reading is determined by the intersection of the horizontal plane, tangent to the bottom of the meniscus reading on the gauge tube. For this vessel, the scale range was from -110 and +100 and each division is equivalent to 2 in³.

[2] Taylor, B. N., and Kuyatt, C. E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results," NIST Technical Note 1297, National Institute of Standards and Technology (January 1993).

[†] Coverage factor of 1.96 for 5227 effective degrees of freedom.




Figure 2. Calibration control chart for 50 gallon graduated neck test measure SN 7293-C.

Table 2. Results of prior calibrations for the delivered volume.

Date	Delivered Volume [in ³]	Difference from Prior [in ³]	Degree of Equivalence [-]
03/24/09	11549.47	1.57	0.46
02/21/06	11547.90	-0.56	-0.16
04/07/03	11548.46	-0.44	-0.13
03/22/99	11548.90	-2.70	-0.78
06/06/94	11551.60	-	-

For the Director,
National Institute of Standards and Technology

Dr. John D. Wright
Project Leader, Fluid Metrology Group
Process Measurements Division
Chemical Science and Technology Laboratory
National Institute of Standards and Technology

Sherry Sheckels
Calibration Technician, Fluid Metrology Group
Process Measurements Division
Chemical Science and Technology Laboratory
National Institute of Standards and Technology



National Institute of Standards and Technology
Technology Administration, U.S. Department of Commerce

REPORT OF CALIBRATION

FOR

A FIVE (5) GALLON VOLUME PROVER
(Graduated Neck Type)

March 31, 2009

Manufacturer: Seraphin
Rancocas, NJ

NIST Seal Number: 7002
Material: Stainless Steel
Serial Number: 7293F

submitted by

Meter Engineers, Inc.
7718 W. 53rd St. N.
Mazie, KS 67101

(Reference: Purchase Order Number 003680, dated January 15, 2009)

The internal volume of the prover described above has been determined by the gravimetric method [1]. The gravimetric method requires weighing the vessel dry and empty and re-weighing it when filled with a fluid of known density. The internal or contained volume was determined in this way and the value is given in Table 1 using the requested units. The fluid used was distilled water and the prover was leveled before determining the volume.

To determine the delivered volume, the contained volume is poured from the vessel by tilting it through a continuously increasing angle so that the poured flow is smooth and not oscillating. When this flow finishes, the vessel is held for 10 seconds at an angle of 70° with the horizontal to complete the drain procedure. Subsequent re-weighing completes the gravimetric procedure and enables calculation of the delivered volume, also given in Table 1. Both the contained and delivered volumes are given for the scale reading of zero (0) and have been corrected for the reference temperature of 15.56 °C (60 °F), assuming a volumetric coefficient of expansion of 0.0000477 per °C (0.0000265 per °F).

[1] Bean, V. E., Espina, P. I., Wright, J. D., Houser, J. F., Sheckels, S. D., and Johnson, A. N., "NIST Calibration Services for Liquid Volume," NIST Special Publication 250-72, National Institute of Standards and Technology, March 24, 2006.

Table 1. Contained and delivered volumes for the tested vessel for a scale reading* of zero.

	Volume Contained	Volume Delivered
gal at 60 °F	4.9965	4.9930
in ³ at 60 °F	1154.20	1153.39

The volume measurement procedure was repeated 5 times with the neck scale filled approximately to zero each time. The repeatability of the 5 measurements was 272 parts in 10⁶ and the expanded uncertainty in the measured volume is $\pm 0.075\%$. It was calculated according to References [1] and [2] with a 95 % confidence level[†] and is traceable to NIST mass, temperature, pressure, and humidity standards, and a NIST distilled water density determination.




Figure 1. Photographs of the volume prover.

The input data used for calculation of the prover volume are given in the spreadsheet attached to this report.

* The scale reading is determined by the intersection of the horizontal plane, tangent to the bottom of the meniscus reading on the gauge tube. For this vessel, the scale range was from -20 and +20 and each division is equivalent to 1 in³.

[2] Taylor, B. N. and Kuyatt, C. E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results," NIST Technical Note 1297, National Institute of Standards and Technology (January 1993).

[†] Coverage factor of 2.45 for 6 effective degrees of freedom.




Figure 2. Calibration control chart for 5 gallon graduated neck test measure SN 7293F.

Table 2. Results of prior calibrations for the delivered volume.

Date	Delivered Volume [in ³]	Difference from Prior [in ³]	Degree of Equivalence [-]
03/26/09	1153.3879	-0.36	-0.29
02/22/06	1153.7442	0.48	0.39
04/08/03	1153.2600	-0.19	-0.15
03/24/99	1153.4500	-0.33	-0.27
06/06/94	1153.7800	-0.25	-0.20
02/13/89	1154.0300	-0.30	-0.24
12/07/83	1154.3300	-0.36	-0.29
05/09/78	1154.6900	-	-

For the Director,
National Institute of Standards and Technology

Dr. John D. Wright
Project Leader, Fluid Metrology Group
Process Measurements Division
Chemical Science and Technology Laboratory
National Institute of Standards and Technology

Sherry Sheckels
Calibration Technician, Fluid Metrology Group
Process Measurements Division
Chemical Science and Technology Laboratory
National Institute of Standards and Technology



Phone 316.944.6256 Fax 316.944.6256

Calibration Lab
Cert #2820-01

Certificate of Calibration

Report Number: 122131B
Customer: METER ENGINEERS, INC.

Page 1 of 1

Form 2511
REV C 12-01-11

7718 W. 53rd ST. NORTH
MAIZE KS 67101

Item: MERCURY THERMOMETER
Unit Number: 4J1555
Model Number: MILLER & WEBER T-3426
Serial Number: 4J1555
Location: PRODUCTION

Inspector: WM
Date Calibrated: Jan 17, 2013
Recalibration Date: Jan 17, 2014
Temperature: 69.4 °F
Relative Humidity: 32 %

Item Range: 30°F/124°F Resolution: 0.05°F
Tolerance: ±0.2°F Class: N/A
Controlling Specification(s): ISO 10012, ANSI/NCSL Z540-1
Calibration/Inspection Procedure(s): WP4001 Tolerance Source: R See Below

Standard Value Test Result Deviation Uncertainty @k=2

Standard Value	Test Result	Deviation	Uncertainty @k=2
32.00°F	32.00°F	NONE	0.12°F
76.00°F	76.00°F	NONE	0.12°F
120.00°F	120.00°F	NONE	0.12°F

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%. This estimate was performed in accordance with guidelines set forth in ANSI/NCSL Z540-2. The acceptance or rejection of the item(s) is based on the actual test values shown, without adjustment for measurement uncertainty. All testing performed using standards traceable to NIST or to intrinsic standards. This document shall not be reproduced except in full, without the written permission of Johnson Gage and Inspection. Tolerance source key: C = Customer; S = Specification; M = Manufacturer; R = JGI Recommended.

Comments: FOUND IN TOLERANCE. NO ADJUSTMENT REQUIRED.

ADDITIONAL STANDARD: HP34420A SN:US36001770; CERTIFIED:12-18-12; DUE:12-18-13; TN:58056AT

Condition: USED
RESULTS: ACCEPTED

CALIBRATION STANDARD

Manufacturer/type: BURNS ENGINEERING PRT Model Number: 3925
Date Certified: Jul 27, 2011 Serial Number: 496142
TRACEABILITY NUMBER: 40187470



Phone 316.944.6252, Fax 316.944.6256

Calibration
Cert #2820/01

Certificate of Calibration

Report Number: 122403N
Customer: METER ENGINEERS, INC.

Page 1 of 1

Form 2511
REV C 12-01-11

7718 W. 53rd ST. NORTH
MAIZE KS 67101

Item: DIGITAL THERMOMETER
Unit Number: 1630273
Model Number: MARTEL BETA PROBE T1
Serial Number: 1630273
Location: PRODUCTION

Inspector: WM
Date Calibrated: Jan 30, 2013
Recalibration Date: Jan 30, 2014
Temperature: 71.8 °F
Relative Humidity: 36 %

Item Range: -58°F/320°F Resolution: 0.01°F
Tolerance: ±0.2°F Class: N/A
Controlling Specification(s): ISO 10012, ANSI/NCSL Z540-1
Calibration/Inspection Procedure(s): WP4001 Tolerance Source: R See Below

Standard Value Test Result Deviation Uncertainty @k=2

Standard Value	Test Result	Deviation	Uncertainty @k=2
32.00°F	31.98°F	-0.02°F	0.12°F
76.00°F	76.00°F	NONE	0.12°F
120.00°F	120.00°F	NONE	0.12°F

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%. This estimate was performed in accordance with guidelines set forth in ANSI/NCSL Z540-2. The acceptance or rejection of the item(s) is based on the actual test values shown, without adjustment for measurement uncertainty. All testing performed using standards traceable to NIST or to intrinsic standards. This document shall not be reproduced except in full, without the written permission of Johnson Gage and Inspection. Tolerance source key: C = Customer; S = Specification; M = Manufacturer; R = JGI Recommended.

Comments: FOUND IN TOLERANCE. NO ADJUSTMENT REQUIRED. NOTE: TEST POINTS PER CUSTOMER REQUEST.

ADDITIONAL STANDARD: HP34420A SN:US36001770; CERTIFIED:12-18-12; DUE:12-18-13; TN:58056AT

RESOLUTION AVAILABLE FROM 0.001°F TO 0.1°F

Condition: USED

RESULTS: ACCEPTED AS NOTED

CALIBRATION STANDARD

Manufacturer/type: BURNS ENGINEERING PRT Model Number: 3925
Date Certified: Jul 27, 2011 Serial Number: 496142
Date Due: Jul 27, 2013 TRACEABILITY NUMBER: U187370

Unit 134

PO BOX 248
WICHITA KS 67201-0248
Phone: (316) 267-2893

ARROW LABORATORY, INC.

Metallurgical Analysis and Testing

1333 N MAIN ST
WICHITA KS 67203
Fax: (316) 267-0171

Attn: DEB JACOBS
METER ENGINEERS
7718 W 53rd St North
WICHITA KS 67101

12-28-12
Page 1 of 1

LABORATORY REPORT #d90mm654 PRESSURE GAGE CALIBRATION

One 4.5 inch WIKÁ XSEL pressure gage, S/N 586, 0-60 psi capacity with 0.5 psi subdivisions, was received in usable condition and calibrated as shown below. The gage was checked with our Ashcroft dead weight gage tester. Testing was performed in accordance with ASME B40.100-2005, ISO 10012-2003, and ANSI/NCSL Z540.3-2006.

Calibration Data:

LOAD psi	Average GAGE READING as Left
5	4.8 psi
15	14.6
25	24.5
35	34.4
45	44.3
55	54.3

Maximum Error center 1/2 of scale: 1.3% of span
 Max Error upper and lower 1/4 of scale: 1.3% of span
 Maximum Friction: 2.5% of span
 Repeatability: +/- 0.2% of span
 Accuracy: +/- 1.3% of span
 Adjustments: none Limitations of use: none
 Calibrated: 12-27-12 Temperature: 74°F
 Client Provided Confirmation Interval: Annual


These results meet ASME B40.100 grade B accuracy requirements.

Ashcroft Gage Tester: S/N DWT 11436, Uncertainty: +/- 0.07%
Weights: Tester weights were compared with weights calibrated on NIST.
Test #'s 822/262551-0 and 822/268214-03. Cal Due: 5-31-13

Piston Assembly:
The piston diameter and cylinder bore were measured by Dresser Industries Test# IR700547 and certified to +/- 0.05% using masters traceable to NIST. Cal Due: 2-28-13

DER/dr
Rec'd 12-27-12

Project Supervisor



Dale E. Roark

The sample or instrument was received in suitable condition for the tests performed except as noted above.
Chemical Analysis of Carbon & Sulfur: Combustion/IR, Silicon: grav., Others: ICP/AA.
Tensile per ASTM E 8-11, B 557-10, or A 370-12a. Hardness per ASTM E 10-12, E 18-08b, E 384-11.
The recording of false, fictitious, or fraudulent statements or entries on the certificate may be punished as a felony under federal law.
This report may not be reproduced except in full.

Kinder Morgan

July 9 2013

Carrington Terminal

Carrington, ND



Public Service Commission

Enclosed is the final report for the water draw at Carrington. This includes the reports of calibration and certificates of calibration. If there is anything else I can provide let me know.

Thank you

A handwritten signature in black ink, appearing to read "E. M. Drewlow".

Eugene M. Drewlow

Carrington Terminal Supervisor



METER PROVING REPORT

Kinder-Morgan Cochin Pipeline LLC.					Accumulators		
LOCATION		ELEV.	DATE	Amb. T.	Previous	Open	Close
CARRINGTON		1520					
PROVER				METER			
Make	MALONEY	Vol.	7.2736	Make	BROOKS	S/N 7709 - 14958 - 6A	
S/N	E-6409	Inside Dia.	7.981	Size"	6"	Design K-Factor	
T. coef.(GP)	0.0000186	W.T."	0.322	T. coef.(GM)		DKF	4200
b rod a (GD)		Elasticity	30000000	Temp. cmp.?		Meter I.D.	
b rod T				Last ATC Cal		T, 1 or 2 ?	1
PRODUCT (STD CONDITIONS)				TICKET			
TYPE	spgr@60F	Vapor Pres	Density	Batch No.		Ticket No.	
PROPANE	0.501	183	175.502805				
FIRST RUN DATA						E V P Prover	
run	Total Pulses	Prover Temp	Prover Pres	Meter Temp	Meter Press	Flow Rate	Fact A Prover
1							Fact B Prover
2							F Prover
3							E V P Meter
4							Fact A Meter
5							Fact B Meter
avg							F Meter
	CTSP	CTLP	CPSP	CPLP	CTLM	CPLM	Average Meter Factor
							MF
	Net Prover Volume		Net Meter Volume		New Meter Factor		@ Oper. Conditions
							MF*CPLM
SECOND RUN DATA						E V P Prover	
run	Total Pulses	Prover Temp	Prover Pres	Meter Temp	Meter Press	Flow Rate	Fact A Prover
1							Fact B Prover
2							F Prover
3							E V P Meter
4							Fact A Meter
5							Fact B Meter
avg							F Meter
	CTSP	CTLP	CPSP	CPLP	CTLM	CPLM	NEW K FACTOR
							DKF/MF
	Net Prover Volume		Net Meter Volume		New Meter Factor		LINEARITY
							(<0.16%)
HISTORICAL DATA					REPEATABILITY		
date				Long Aver	5 PASS @ <0.025%		
fact							
flow				Aver Factor	10 PASS @ <0.05%		
temp							
COMMENTS							
Report by _____				Checked by _____			

