



**APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY**  
Public Service Commission  
SFN 51277 (2/2014)



**TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED**

<b>Name of Company</b> <i>Petrokom Calibration Service</i>	<b>Email Address</b> <i>PetrokomCalibrationService@gmail.com</i>	<b>Application Date</b> <i>3/4/16</i>	
<b>Mailing Address</b> <i>203 E. Hwy 61</i>	<b>City</b> <i>Esko</i>	<b>State</b> <i>ND</i>	<b>Zip Code</b> <i>55733</i>
<b>Telephone Number</b> <i>(218) 213 6536</i>	<b>Cell Phone Number</b> <i>(218) 213 6536</i>	<b>Fax Number</b>	

Select below all device types your company will certify:

<b>Scales (include maximum capacity, if applicable)</b>	<b>Liquid (include maximum flow rate, if applicable)</b>
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input checked="" type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

<b>Permit No.</b>	<b>Employee</b>	<b>Device Types Registered to Certify (list using device type numbers from above)</b>
<i>e.g. 1001</i>	<i>e.g. John Doe</i>	<i>e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6</i>
<i>1625</i>	<i>Edgenc H Lewis</i>	<i>RETA motor fuel (less than 20 gal/min)</i>

Continued on Page 2

List below all field standards (attach current calibration reports):

Scraphin 5 Gal. S/N 49433-A	
Scraphin 5 Gal. S/N 49433-B	
Scraphin 5 Gal. S/N 49433-C	
Scraphin 5 Gal. S/N 00-13651-15	


Additional Application Items (initial where appropriate):

Standardized Test Report	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.  
 Yes     No

I am Ernest H Lewis, and have authority to represent this company.  
 By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.

  
 \_\_\_\_\_  
 Signature

Send Completed Application and Related Documents To:

Public Service Commission  
 600 E Boulevard Ave Dept 408  
 Bismarck ND 58505-0480  
**Telephone:** (701) 328-2400  
**Fax:** (701) 328-2410

Receipt Date: July 7, 2015  
Test Date: July 7, 2015  
Report Date: July 7, 2015

State Test No.: 334484  
Serial No.: 49433-A  
Barcode: 201323

## Calibration Report

PETROLEUM CALIBRATION SERVICE  
203 E HWY 61  
ESKO, MN 55733  
Contact: GENE LEWIS  
Phone: 218-213-6556  
PO Number: NONE  
SOP: 32  
Technician ID: 11

Item(s) Submitted: 5 Gallon Measure  
Manufacturer: Seraphin  
Material: Stainless Steel  
Equipment Number: None  
Condition: Good  
Temperature: 21.8 °C  
Pressure: 741. mmHg  
Relative Humidity: 39. %  
Standard H<sub>2</sub>O Temp.: 24.1 °C  
Artifact H<sub>2</sub>O Temp.: 24.3 °C

Nominal Volume (gal)		Error (in <sup>3</sup> )	Volume at Zero Line (gal)	Uncertainty (in <sup>3</sup> )	Coefficient of Expansion (1/°F)
5	As Found	0.07	5.0003	0.62	0.0000265
	As Left	0.07	5.0003	0.62	

Neck Calibration: No neck calibration was done at this time.


This measure or prover has been calibrated as a "to contain after wet down" vessel with a drain time of 30 seconds after cessation of full flow and at a reference temperature of 60 °F.

The measure or prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

The reported uncertainty conforms to NIST Technical Note 1297. The confidence interval is 95 %.

Results apply to item identified in this report only.


Pete Whebbe



Metrologist

Reviewed by:

Mark Nicollet



Quality Manager

Receipt Date: June 17, 2015  
 Test Date: June 22, 2015  
 Report Date: June 23, 2015

State Test No.: 334416  
 Serial No.: 49433-B  
 Barcode: 202262

## Calibration Report

**PETROLEUM CALIBRATION SERVICE**  
 203 E HWY 61  
 ESKO, MN 55733  
 Contact: GENE LEWIS  
 Phone: 218-213-6556  
 PO Number: NONE  
 SOP: 32  
 Technician ID: 11

Item(s) Submitted: 5 Gallon Prover  
 Manufacturer: Seraphin  
 Material: Stainless Steel  
 Equipment Number: None  
 Condition: Good  
 Temperature: 24.6 °C  
 Pressure: 722.2 mmHg  
 Relative Humidity: 47. %  
 Standard H<sub>2</sub>O Temp.: 24.7 °C  
 Artifact H<sub>2</sub>O Temp.: 24.5 °C

Nominal Volume (gal)		Error (in <sup>3</sup> )	Volume at Zero Line (gal)	Uncertainty (in <sup>3</sup> )	Coefficient of Expansion (°F)
5	As Found	0.79	5.0034	0.62	0.0000265
	As Left	-0.01	4.9999	0.62	

Neck Calibration: No neck calibration was done at this time.


This measure or prover has been calibrated as a "to contain after wet down" vessel with a drain time of 30 seconds after cessation of full flow and at a reference temperature of 60 °F.

The measure or prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

The reported uncertainty conforms to NIST Technical Note 1297. The confidence interval is 95 %.


Results apply to item identified in this report only.

Pete Whebbe

  
 Metrologist

Reviewed by:

Benjamin FitzPatrick

  
 Deputy Director

Receipt Date: June 17, 2015  
 Test Date: June 22, 2015  
 Report Date: June 23, 2015

State Test No.: 334414  
 Serial No.: 49433-C  
 Barcode: 201324

## Calibration Report

PETROLEUM CALIBRATION SERVICE  
 203 E HWY 61  
 ESKO, MN 55733  
 Contact: GENE LEWIS  
 Phone: 218-213-6556  
 PO Number: NONE  
 SOP: 32  
 Technician ID: 11

Item(s) Submitted: 5 Gallon Prover  
 Manufacturer: Seraphin  
 Material: Stainless Steel  
 Equipment Number: None  
 Condition: Good  
 Temperature: 24.6 °C  
 Pressure: 722.2 mmHg  
 Relative Humidity: 47. %  
 Standard H<sub>2</sub>O Temp.: 24.7 °C  
 Artifact H<sub>2</sub>O Temp.: 24.6 °C

Nominal Volume (gal)		Error (in <sup>3</sup> )	Volume at Zero		Coefficient of Expansion (1/°F)
			Line (gal)	Uncertainty (in <sup>3</sup> )	
5	As Found	-0.15	4.9994	0.62	0.0000265
	As Left	-0.15	4.9994	0.62	

Neck Calibration: No neck calibration was done at this time.


This measure or prover has been calibrated as a "to contain after wet down" vessel with a drain time of 30 seconds after cessation of full flow and at a reference temperature of 60 °F.

The measure or prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

The reported uncertainty conforms to NIST Technical Note 1297. The confidence interval is 95 %.

Results apply to item identified in this report only.


Pete Whebbe



Metrologist

Reviewed by:

Benjamin FitzPatrick



Deputy Director

Receipt Date: June 17, 2015  
Test Date: June 22, 2015  
Report Date: June 23, 2015

State Test No.: 334415  
Serial No.: 00-13651-15  
Barcode: 201325

## Calibration Report

PETROLEUM CALIBRATION SERVICE  
203 E HWY 61  
ESKO, MN 55733  
Contact: GENE LEWIS  
Phone: 218-213-6556  
PO Number: NONE  
SOP: 32  
Technician ID: 11

Item(s) Submitted: 5 Gallon Prover  
Manufacturer: Seraphin  
Material: Stainless Steel  
Equipment Number: None  
Condition: Good  
Temperature: 24.6 °C  
Pressure: 722.2 mmHg  
Relative Humidity: 47. %  
Standard H<sub>2</sub>O Temp.: 24.6 °C  
Artifact H<sub>2</sub>O Temp.: 24.7 °C

Nominal Volume (gal)		Error (in <sup>3</sup> )	Volume at Zero Line (gal)	Uncertainty (in <sup>3</sup> )	Coefficient of Expansion (°F)
5	As Found	-0.04	4.9998	0.62	0.0000265
	As Left	-0.04	4.9998	0.62	

Neck Calibration: No neck calibration was done at this time.


This measure or prover has been calibrated as a "to contain after wet down" vessel with a drain time of 30 seconds after cessation of full flow and at a reference temperature of 60 °F.

The measure or prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

The reported uncertainty conforms to NIST Technical Note 1297. The confidence interval is 95 %.

Results apply to item identified in this report only.


Pete Whebbe



Metrologist

Reviewed by:

Benjamin FitzPatrick



Deputy Director