



# Public Service Commission

## State of North Dakota

### COMMISSIONERS

Brian Kroshus  
Julie Fedorchak  
Randy Christmann

600 East Boulevard, Dept. 408  
Bismarck, North Dakota 58505-0480  
Web: [www.psc.nd.gov](http://www.psc.nd.gov)  
E-mail: [ndpsc@nd.gov](mailto:ndpsc@nd.gov)  
Phone: 701-328-2400  
ND Toll Free: 1-877-245-6685  
Fax: 701-328-2410  
TDD: 800-366-6888 or 711

### PERMANENT VARIANCE PERMIT APPROVAL

#### Proseed, Inc. Variance Permit - Amenia Application

Case No. WM-16-540

**Issued To:**  
Proseed, Inc.  
705 East Brewster  
Harvey, ND 58341

**Scale Location:**  
Proseed, Inc.  
Amenia, ND

**Truck Scale Information:**  
Mettler Toledo VTC100, SN B617392302, 120,000 lb., 11' x 80'

**Request To:**  
Operate in commerce a scale with a floating slab foundation.

#### Variance Permit Conditions:

- The foundation of the scale must be constructed on soil that has a uniform natural or engineered soil bearing of at least 1,500 pounds per square foot.
- A base of uniform, compacted, permeable aggregate, at least 24 inches deep must be installed. The concrete slab foundation must be poured on top of this base.
- The scale weighing element or elements must be supported on a monolithic, reinforced 12" thick concrete slab and must conform to the scale manufacturer's installation and reinforcing requirements.
- The top surface of the foundation slab must be installed above the surrounding grade level so that water will flow away from the foundation and away from the permeable aggregate base that supports the foundation.

Proseed, Inc.  
Page 2  
April 23, 2020

- The scale will not be used beyond its 120,000 lb. capacity.
- Once every twelve months, the scale must be retested and adjusted by a registered service company to maintain the required tolerance.
- The twelve-month test reports must be submitted to the Commission within seven working days of the test date and a copy retained by Proseed, Inc.
- The service company test reports must include both the "as found" and "as left" readings in addition to any maintenance notes.
- Except for the specific variance granted by Permit Number 517, the scale must be operated according to all applicable North Dakota Laws and Rules.
- The enclosed Permanent Variance Permit must be posted near the scale in a place clearly visible to the public.
- Failure to comply with the variance permit conditions may result in revocation.

If you have any further questions, or require additional information, please feel free to contact me at (701) 328-4097.

Dated April 23, 2020



Konrad Crockford, Director  
Compliance Division



Public Service Commission

Be Legendary.™

### PERMANENT VARIANCE PERMIT

**Issued To:**

Proseed, Inc.  
705 East Brewster  
Harvey, ND 58341

**Permit No. 517**

**Expiration Date: N/A**

**Scale Location:**

Proseed, Inc.  
Amenia, ND

**Scale Information:**

Mettler Toledo VTC100, SN B617392302, 120,000 lb., 11' x 80'

This Permit allows the above referenced **scale with a floating slab foundation to be operated in commerce.**

This Permit is conditioned upon Proseed, Inc. adhering to the terms of the April 23, 2020 Variance Permit Approval.

Dated April 23, 2020

Konrad Crockford, Director  
Compliance Division