



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY

Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company	Email Address	Application Date	
Binstock Inspection Service Inc	CSE#2001@msn.com	11-21-2016	
Mailing Address	City	State	Zip Code
2421 Crocus Drive N	Mandan	ND	58554
Telephone Number	Cell Phone Number	Fax Number	
701-400-0763 (cell#)	701-400-0763	N/A	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input checked="" type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input checked="" type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input checked="" type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: <u>500 gpm</u> <input checked="" type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: <u>500 gpm</u> <input checked="" type="checkbox"/> 5. LPG <input checked="" type="checkbox"/> 6. Stationary LPG <input checked="" type="checkbox"/> 7. Fertilizer: Max. Flow Rate: <u>500 gpm</u> <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input checked="" type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
e.g. 1001	e.g. John Doe	e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6
1459	Casey Binstock	Liquid 1,2,3,4,5,6,7,10
1624	Josh Levi	Liquid 1,2,3,4,5,6,7,10

Continued on Page 2



List below all field standards (attach current calibration reports):

5 gallon Refined Fuel Prover	
100 gallon Refined Fuel Prover	
500 gallon Refined Fuel Prover	
100 gallon LAG Prover	

Additional Application Items (initial where appropriate):

Standardized Test Report	<input checked="" type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am Cary Branstock, and have authority to represent this company.
 By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.

Cary Branstock
Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410

North Dakota

nd.gov Official Portal for
North Dakota State GovernmentNorth Dakota
LEGISLATURE

SECRETARY OF STATE NORTH DAKOTA


[Home](#) | [Business Records Search](#)

BINSTOCK INSPECTION SERVICE, INC.

Corporation Details

System ID: 21209700 **Phone:** (701) 663-7063
Type: BUSINESS CORPORATION
Status: Active & Good Standing
Original File Date: 06/27/2005 **Effective Date:** 06/27/2005
State of Origin: North Dakota

Nature of Business

INSPECTION & REPAIR

Principal Office

2421 CROCUS DR N MANDAN, ND 58554-8254

Registered Agent

CASEY BINSTOCK
 2421 CROCUS DR N
 MANDAN, ND 58554-8254
 Established Date: Jun 26, 2012

Authorized Shares

Class	Number	Par Value
	100.000000	\$1.000000

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2016](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)
[Disclaimer](#)
[Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



Receipt Date: June 20, 2016
Cal. Date: June 20, 2016
Report Date: June 20, 2016

Report No.: 336097
Serial No.: 00-50785
Barcode: 202324

Calibration Certificate

BINSTOCK INSPECTION SERVICE
2421 CROCUS DRIVE NORTH
MANDAN, ND 58554
Contact: Casey Binstock
Phone: 701-400-0763
PO Number: None
SOP: 19
Technician ID: 19

Item(s) Submitted: 5 Gallon Measure
Manufacturer: Seraphin
Material: Mild Steel
Type: Measure
Condition: Good
Temperature: 25.1 °C
Pressure: 740.6 mmHg
Relative Humidity: 47.9 %
Standard H₂O Temp.: 20.3 °C
Artifact H₂O Temp.: 20.4 °C

Nominal		Calibrated				
Volume (gal)		Volume (gal)	Error (in ³)	<i>k</i>	U (in ³)	CCE (1/°F)
5	As Found	4.9998	-0.05	2.06	0.24	0.0000186
	As Left	4.9998	-0.05			

Neck Calibration: No neck calibration was performed at this time.

This measure has been calibrated as a "to contain after wet down" vessel with a pour time of 30 seconds followed by a drain time of 10 seconds after cessation of full flow.

The vessel listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Erik Alfvin

Metrologist

Reviewed by:

Mark Nicollet

Quality Manager



Receipt Date: July 7, 2016
Cal. Date: July 7, 2016
Report Date: July 7, 2016

Report No.: 336172
Serial No.: 8861610
Barcode: 202326

Calibration Certificate

BINSTOCK INSPECTION SERVICE
2421 CROCUS DRIVE NORTH
MANDAN, ND 58554
Contact: Casey Binstock
Phone: 701-400-0763
PO Number: None
SOP: 34
Technician ID: 19

Item(s) Submitted: 100 Gallon LPG Prover
Manufacturer: Wheeler / Brownie
Material: Mild Steel
Description: Zero Bottom
Condition: Excellent
Temperature: 24.9 °C
Pressure: 731.3 mmHg
Relative Humidity: 54.5 %
Standard H₂O Temp. 16.9 °C
Artifact H₂O Temp.: 17.4 °C

Nominal		Calibrated				
Volume (gal)		Volume (gal)	Error (in ³)	<i>k</i>	U (in ³)	CCE (°F)
100	As Found (at 100 psig)	99.881	-27.5	2.02	5.3	0.0000186
	As Left (at 100 psig)	99.881	-27.5			

Neck Calibration: No neck calibration was performed at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a drain time of 30 seconds after cessation of full flow. The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-4 (2010). Uncertainty calculations contain the components in NIST SOP 21 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F.

Erik Alfvin

Metrologist

Reviewed by:
Benjamin FitzPatrick

Deputy Director



Receipt Date: July 7, 2016
Cal. Date: July 7, 2016
Report Date: July 7, 2016

Report No.: 336173
Serial No.: 2999731-04
Barcode: 202325

Calibration Certificate

BINSTOCK INSPECTION SERVICE
2421 CROCUS DRIVE NORTH
MANDAN, ND 58554
Contact: Casey Binstock
Phone: 701-400-0763
PO Number: None
SOP: 19
Technician ID: 19

Item(s) Submitted: 100 Gallon Prover
Manufacturer: Determan Brownie
Material: Stainless Steel
Type: No Bottom Zero
Condition: Excellent
Temperature: 25.5 °C
Pressure: 731.7 mmHg
Relative Humidity: 54.7 %
Standard H₂O Temp.: 16.1 °C
Artifact H₂O Temp.: 16.3 °C

Nominal		Calibrated				
Volume (gal)		Volume (gal)	Error (in ³)	<i>k</i>	U (in ³)	CCE (°F)
100	As Found	99.998	-0.5	2.01	3.3	0.0000265
	As Left	99.998	-0.5			

Neck Calibration: No neck calibration was performed at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a drain time of 30 seconds after cessation of full flow.

The vessel listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Erik Alfvin

Metrologist

Reviewed by:

Benjamin FitzPatrick

Deputy Director



Receipt Date: June 20, 2016
Cal. Date: June 20, 2016
Report Date: June 20, 2016

Report No.: 336096
Serial No.: 1101-1
Barcode: 202343

Calibration Certificate

BINSTOCK INSPECTION SERVICE
2421 CROCUS DRIVE NORTH
MANDAN, ND 58554
Contact: Casey Binstock
Phone: 701-400-0763
PO Number: None
SOP: 19
Technician ID: 19

Item(s) Submitted: 500 Gallon Prover
Manufacturer: Warner Lewis
Material: Mild Steel
Type: Bottom Zero
Condition: Good*
Temperature: 25.1 °C
Pressure: 740.5 mmHg
Relative Humidity: 48.8 %
Standard H₂O Temp.: 15.2 °C
Artifact H₂O Temp.: 15.1 °C

Nominal		Calibrated				
Volume (gal)		Volume (gal)	Error (in ³)	<i>k</i>	U (in ³)	CCE (°F)
500	As Found	499.99	-1	2.02	12	0.0000186
	As Left	499.99	-1			

Neck Calibration: No neck calibration was performed at this time.

***Please see addendum to certificate**

This prover has been calibrated as a "to contain after wet down" vessel with a drain time of 30 seconds after cessation of full flow. The vessel listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Erik Alfvín

Metrologist

Reviewed by:

Mark Nicollet

Quality Manager

Addendum to report number 336096.

500 Gallon Prover

SN: 1101-1

June 20, 2016

Prover does not meet NIST Handbook 105-3 specifications in its current condition. The following must be added by the next calibration cycle for prover to be tested:

Identification plate containing nominal capacity, construction material and cubical coefficient of expansion per °F – Handbook 105-3 section 4.3.

Scale plate must have units of measurement clearly marked – section 4.5.7.2.

Although not required by the next calibration cycle, thermometer wells will need to be installed at a future time that meet the requirements of section 4.5.11.

Erik Alfvín



Metrologist

United States Department of Commerce
National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Minnesota

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon II	Weight Carts	Volume Gravimetric, I
50 kg to 1 mg	10 000 lb to 2000 lb	20 L to 10 ml
1000 lb to 0.001 lb	Wheel Load Weighers	100 gal to 0.25 qt
4 oz to 0.03125 oz	20 000 lb to 2000 lb	Volume Transfer, II
Mass Echelon III	Railroad Test Cars	1500 gal to 5 gal
50 kg to 1 mg	110 000 lb to 80 000 lb	100 gal to 25 gal LPG
5000 lb to 0.001 lb		
4 oz to 0.03125 oz		



2016 to 2017

A handwritten signature in black ink, appearing to read "Carol T. Hockert".

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2016-01-01 to 2017-12-31