



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY

Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company Bison Food Equip. Service	Email Address bisonfoodequipment@hotmail.com	Application Date 11/22/16	
Mailing Address 17409 Beseau Lake Rd.	City Lake Park	State mn	Zip Code 56554
Telephone Number 218-238-6363	Cell Phone Number 701-261-2102	Fax Number 218-238-5427	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: <u>400 lbs</u> <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
e.g. 1001	e.g. John Doe	e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6
1462	Dale G. Bohner	Scale 6, 7

Continued on Page 2



List below all field standards (attach current calibration reports):


Additional Application Items (initial where appropriate):

Standardized Test Report	<input checked="" type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input checked="" type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input checked="" type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am Dale G. Bohner, and have authority to represent this company.
By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.



Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410

North Dakota

nd.gov Official Portal for
North Dakota State GovernmentNorth Dakota
LEGENDARY

SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

BISON FOOD EQUIPMENT SERVICE, INC.

Corporation Details

System ID: 19106600**Phone:** (218) 238-6363**Type:** FOREIGN BUSINESS CORPORATION**Status:** Active & Good Standing**Original File Date:** 04/03/2003**Effective Date:** 04/03/2003**State of Origin:** Minnesota

Nature of Business

FOOD EQUIPMENT SERVICE AND SALES

Principal Office

17409 BESEAU LAKE RD LAKE PARK, MN 56554-9645

Registered Agent

MICHAEL SCULLY

6620 SAPHIRE LANE

BISMARCK, ND 58501-

Established Date: May 18, 2005

Generate an Annual Report To File


To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2016](#) (generates a forms-fillable pdf in a new pop-up window)[Return to Search Results](#)[Contact Us](#)[Disclaimer](#)[Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



Receipt Date: March 29, 2016
Cal. Date: March 29, 2016
Report Date: March 29, 2016

Report No.: 335749
Set Serial No.: 14580/1 to 6
Barcode: 202472

Calibration Certificate


BISON FOOD EQUIPMENT SERVICE INC.
17409 BESEAU LAKE ROAD
LAKE PARK, MN 56554
Contact: Dale Bohner
Phone: 701-261-2102
PO Number: None
SOP: 8
Technician ID: 19

Item(s) Submitted: AVDP Weight Kit - F
Manufacturer: Troemner
Weight Type: II
Equipment ID: None
Condition: Good
Temperature: 20.6 °C
Pressure: 740.5 mmHg
Relative Humidity: 46.6 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
5 lb	1	112	112	F	F	2.06	12
5 lb	2	119	119	F	F	2.06	12
5 lb	3	102	102	F	F	2.06	12
5 lb	4	129	129	F	F	2.06	12
5 lb	5	109	109	F	F	2.06	12
5 lb	6	98	98	F	F	2.06	12

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³ at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.


Erik Alfvín



Metrologist

Reviewed by:

Mark Nicollet



Quality Manager



Receipt Date: March 29, 2016
Cal. Date: March 29, 2016
Report Date: March 29, 2016

Report No.: 335751
Set Serial No.: None
Barcode: 202268

Calibration Certificate

BISON FOOD EQUIPMENT SERVICE INC.
17409 BESEAU LAKE ROAD
LAKE PARK, MN 56554
Contact: Dale Bohner
Phone: 701-261-2102
PO Number: None
SOP: 8
Technician ID: 19


Item(s) Submitted: 50 lb, 25 lb Hand Weights
Manufacturer: Troemner
Weight Type: II
Equipment ID: None
Condition: Good
Temperature: 19.5 °C
Pressure: 740.5 mmHg
Relative Humidity: 47.0 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
50 lb	--	1320	1320	F	F	2.01	120
50 lb	--	2020	2020	F	F	2.01	120
50 lb	--	1690	1690	F	F	2.01	120
50 lb	--	410	410	F	F	2.01	120
50 lb	--	590	590	F	F	2.01	120
50 lb	--	1950	1950	F	F	2.01	120
25 lb	--	1038	128	*	F	2.01	67
25 lb	--	848	848	F	F	2.01	67
25 lb	--	798	798	F	F	2.01	67
25 lb	--	608	608	F	F	2.01	67

* Weight(s) as found exceed NIST Class F Tolerance.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³ at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Erik Alfvin




Metrologist


Page 1 of 1

Reviewed by:

Mark Nicollet



Quality Manager



Receipt Date: March 29, 2016
Cal. Date: March 29, 2016
Report Date: March 30, 2016


Report No.: 335750
Set Serial No.: None
Barcode: 201022

Calibration Certificate

BISON FOOD EQUIPMENT SERVICE INC.
17409 BESEAU LAKE ROAD
LAKE PARK, MN 56554
Contact: Dale Bohner
Phone: 701-261-2102
PO Number: None
SOP: 8
Technician ID: 19

Item(s) Submitted: AVDP Weight Kit - F
Manufacturer: Rice Lake
Weight Type: I & II
Equipment ID: None
Condition: Good
Temperature: 21.9 °C
Pressure: 740.5 mmHg
Relative Humidity: 46.3 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
5 lb	--	66	66	F	F	2.06	12
5 lb	--	118	118	F	F	2.06	12
5 lb	--	72	72	F	F	2.06	12
5 lb	--	54	54	F	F	2.06	12
5 lb	--	51	51	F	F	2.06	12
1 lb	--	27.7	27.7	F	F	2.06	2.8
1 lb	--	20.1	20.1	F	F	2.06	2.8
1 lb	--	19.7	19.7	F	F	2.06	2.8
1 lb	--	20.7	20.7	F	F	2.06	2.8
1 lb	--	26.8	26.8	F	F	2.06	2.8
8 oz	--	20.2	20.2	F	F	2.00	2.3
0.2 lb	--	5.08	5.08	F	F	2.06	0.37
0.2 lb	--	6.60	6.60	F	F	2.06	0.37
0.1 lb	--	2.14	2.14	F	F	2.07	0.30
0.05 lb	--	0.45	0.45	F	F	2.06	0.30
0.02 lb	--	0.75	0.75	F	F	2.06	0.16
0.02 lb	--	0.61	0.61	F	F	2.06	0.16
0.01 lb	--	0.16	0.16	F	F	2.06	0.12



Receipt Date: March 29, 2016
Cal. Date: March 29, 2016
Report Date: March 30, 2016

Report No.: 335750
Set Serial No.: None
Barcode: 201022

Continued,

Calibration Certificate

BISON FOOD EQUIPMENT SERVICE INC.
17409 BESEAU LAKE ROAD
LAKE PARK, MN 56554
Contact: Dale Bohner
Phone: 701-261-2102
PO Number: None
SOP: 8
Technician ID: 19

Item(s) Submitted: AVDP Weight Kit - F
Manufacturer: Rice Lake
Weight Type: I & II
Equipment ID: None
Condition: Good
Temperature: 21.9 °C
Pressure: 740.5 mmHg
Relative Humidity: 46.3 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
0.005 lb	--	0.30	0.30	F	F	2.06	0.10
0.002 lb	--	0.078	0.078	F	F	2.06	0.066
0.002 lb	--	0.238	0.238	F	F	2.06	0.066
0.001 lb	--	-0.105	-0.105	F	F	2.06	0.057

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³ at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Erik Alfvin
Erik Alfvin
Metrologist

Reviewed by:
Mark Nicollet
Mark Nicollet
Quality Manager

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Minnesota

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon II	Weight Carts	Volume Gravimetric, I
50 kg to 1 mg	10 000 lb to 2000 lb	20 L to 10 mL
1000 lb to 0.001 lb	Wheel Load Weighers	100 gal to 0.25 qt
4 oz to 0.03125 oz	20 000 lb to 2000 lb	Volume Transfer, II
Mass Echelon III	Railroad Test Cars	1500 gal to 5 gal
50 kg to 1 mg	110 000 lb to 80 000 lb	100 gal to 25 gal LPG
5000 lb to 0.001 lb		
4 oz to 0.03125 oz		



2016 to 2017

A handwritten signature in black ink, appearing to read "Carol T. Hockett".

Carol T. Hockett, Chief
NIST Office of Weights and Measures

Effective Dates: 2016-01-01 to 2017-12-31