




**APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY**  
 Public Service Commission  
 SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company <i>Farm Country Scale, LLC</i>	Email Address <i>farmcountry.scale@gmail.com</i>	Application Date <i>11-21-2016</i>	
Mailing Address <i>9449 County Road 81</i>	City <i>Fairmount</i>	State <i>ND</i>	Zip Code <i>58030</i>
Telephone Number <i>701-361-4483</i>	Cell Phone Number <i>701-361-4483</i>	Fax Number	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input checked="" type="checkbox"/> 1. Rail <input checked="" type="checkbox"/> 2. Truck <input checked="" type="checkbox"/> 3. Livestock <input checked="" type="checkbox"/> 4. Hopper: Max. Capacity: <i>60,000 lbs</i> <input checked="" type="checkbox"/> 5. Belt <input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
<i>1606</i>	<i>Joel Bonnersbach</i>	<i>1, 2, 3, 4, 5, 6 + 7</i>

Continued on Page 2



List below all field standards (attach current calibration reports):


Additional Application Items (initial where appropriate):

Standardized Test Report	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.  
 Yes     No

I am Joel Bommersbach, and have authority to represent this company.  
By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.

\_\_\_\_\_  
Signature

Send Completed Application and Related Documents To:  
Public Service Commission  
600 E Boulevard Ave Dept 408  
Bismarck ND 58505-0480  
Telephone: (701) 328-2400  
Fax: (701) 328-2410

**Bauske, Shelly A.**

---

**From:** Joel Bommersbach <farmcountryscale@gmail.com>  
**Sent:** Tuesday, January 24, 2017 4:11 PM  
**To:** Bauske, Shelly A.  
**Subject:** Re: Application for Registration as a Registered Service Company - Additional Information  
**Attachments:** state nd 1.pdf; state nd 2.pdf; weight cert 1.pdf; weight cert 2.pdf; weight cert 3.pdf; weight cert 4.pdf; weight cert 5.pdf; weight cert 6.pdf; weight cert 7.pdf

**CAUTION:** This email originated from an outside source. Do not click links or open attachments unless you know they are safe.

Hi Shelly,

Sorry about missing that section more new stuff to fill out. I have purchased most of my own weights now and am only using SP Specialty Products heavy capacity carts and weights and hopefully for only a couple months longer.

On Tue, Jan 24, 2017 at 10:08 AM, Bauske, Shelly A. <[sbauske@nd.gov](mailto:sbauske@nd.gov)> wrote:  
Good Morning Joel

Attached is the Application for Registration as a Registered Service Company. Please complete the employee and field standards sections and resubmit.

If you have any questions, please contact me. Thank you!


Shelly Bauske  
Public Service Commission  
600 E Boulevard Ave Dept 408  
Bismarck ND 58505-0480  
[701-328-4070](tel:701-328-4070)  
[701-328-2410](tel:701-328-2410) (fax)  
[sbauske@nd.gov](mailto:sbauske@nd.gov)

---

This transmission, email and any files transmitted with it, may be: (1) subject to the Attorney-Client Privilege, (2) an attorney work product, or (3) strictly confidential under federal or state law. If you are not the intended recipient of this message, you may not use, disclose, print, copy or disseminate this information. If you have received this transmission in error, notify the sender (only) and delete the message. This message may also be subject to disclosure under the North Dakota Open Records Laws.

--  
Joel Bommersbach  
Farm Country Scale, LLC.

North Dakota

nd.gov Official Portal for  
North Dakota State GovernmentNorth Dakota  
LEGENDARY

## SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

### FARM COUNTRY SCALE, LLC

#### Corporation Details

**System ID:** 41811200      **Phone:** Not on file  
**Type:** LIMITED LIABILITY COMPANY  
**Status:** Active & Good Standing  
**Original File Date:** 09/06/2016      **Effective Date:** 09/06/2016  
**State of Origin:** North Dakota

#### Nature of Business

SERVICE/REPAIR/SELL & INSTALL SCALES

#### Principal Office

9449 COUNTY ROAD 81 FAIRMOUNT, ND 58030-

#### Registered Agent

**JOEL M BOMMERSBACH**  
6302 17TH ST N  
FARGO, ND 58102-6027  
Established Date: Sep 06, 2016

#### Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2016](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)[Disclaimer](#)[Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.

Receipt Date: August 24, 2016  
Cal. Date: August 25, 2016  
Report Date: August 25, 2016

Report No.: 336427  
Set Serial No.: 11, 22  
Barcode: 201219

## Calibration Certificate

SPECIALTY PRODUCTS  
1420 N 4TH ST  
FARGO, ND 58102-2733  
Contact: PHIL MCINTYRE  
Phone: 701-235-2996  
PO Number: NONE  
SOP: 8  
Technician ID: 11

Item(s) Submitted: Two 5000 lb Weight Carts  
Manufacturer: Heinisch  
Weight Type: NA  
Equipment ID: None  
Condition: Good  
Temperature: 20.7 °C  
Pressure: 737.7 mmHg  
Relative Humidity: 58.2 %

Nominal Value	Serial No.	CM Correction (g)		NIST HB105-8 Tolerance		k	U (g)
		As Found	As Left	As Found	As Left		
5000 lb	11	-185	-185	Meets	Meets	2.10	60.
5000 lb	22	-95	-95	Meets	Meets	2.10	60.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm<sup>3</sup> density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-8 (2003). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Pete Whebbe

*Pete Whebbe*  
Metrologist

Page 1 of 1

Reviewed by:  
Mark Nicollet

*Mark Nicollet*  
Quality Manager

Receipt Date: August 24, 2016  
Cal. Date: August 24, 2016  
Report Date: August 24, 2016

Report No.: 336424  
Set Serial No.: 114, 119, 127, 128  
Barcode: 201216

### Calibration Certificate

SPECIALTY PRODUCTS  
1420 N 4TH ST  
FARGO, ND 58102-2733  
Contact: PHIL MCINTYRE  
Phone: 701-235-2996  
PO Number: NONE  
SOP: 8  
Technician ID: 11

Item(s) Submitted: Cast Cube Weights  
Manufacturer: Rice Lake  
Weight Type: II  
Equipment ID: None  
Condition: Good  
Temperature: 21.2 °C  
Pressure: 736.1 mmHg  
Relative Humidity: 51.7 %

Nominal Value	Serial No.	CM Correction (g)		NIST HB105-1 Class		k	U (g)
		As Found	As Left	As Found	As Left		
2500 lb	114	87.4	87.4	F	F	2.07	8.6
2500 lb	119	66.4	66.4	F	F	2.07	8.6
2500 lb	127	35.4	35.4	F	F	2.07	8.6
2500 lb	128	95.4	95.4	F	F	2.07	8.6

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm<sup>3</sup> density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Pete Whebbe  
*Pete Whebbe*  
Metrologist

Reviewed by:  
Mark Nicollet  
*Mark Nicollet*  
Quality Manager



Accredited by the National Voluntary Laboratory Accreditation Program for the specific scope of accreditation under lab code 105003. This report may not be used to claim product endorsement by NVLAP or any other government agency, and may not be reproduced, except in full, without written approval from the laboratory.

Receipt Date: January 17, 2017  
Cal. Date: January 17, 2017  
Report Date: January 17, 2017

Report No.: 337081  
Set Serial No.: 1 thru 4  
Barcode: 203144

## Calibration Certificate


FARM COUNTRY SCALE LLC  
9449 COUNTY ROAD 81  
FAIRMOUNT, ND 58030  
Contact: JOEL BOMMERSBACH  
Phone: 701-361-4483  
PO Number: NONE  
SOP: 8  
Technician ID: 11

Item(s) Submitted: Cast Cube Weights  
Manufacturer: Rice Lake  
Weight Type: II  
Equipment ID: None  
Condition: New  
Temperature: 19.2 °C  
Pressure: 736.3 mmHg  
Relative Humidity: 51.2 %

Nominal Value	Serial No.	CM Correction (g)		NIST HB105-1 Class		k	U (g)
		As Found	As Left	As Found	As Left		
1000 lb	1	-1.2	-1.2	F	F	2.01	1.5
1000 lb	2	-2.2	-2.2	F	F	2.01	1.5
1000 lb	3	4.8	4.8	F	F	2.01	1.5
1000 lb	4	0.8	0.8	F	F	2.01	1.5

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm<sup>3</sup> density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.


Pete Whebbe



Metrologist

Reviewed by:

Erik Alfvin



Metrologist

Receipt Date: January 17, 2017  
Cal. Date: January 17, 2017  
Report Date: January 17, 2017

Report No.: 337082  
Set Serial No.: 1 thru 20  
Barcode: 203143

## Calibration Certificate


FARM COUNTRY SCALE LLC  
9449 COUNTY ROAD 81  
FAIRMOUNT, ND 58030  
Contact: JOEL BOMMERSBACH  
Phone: 701-361-4483  
PO Number: NONE  
SOP: 8  
Technician ID: 11

Item(s) Submitted: Cast Hand Weights  
Manufacturer: Rice Lake  
Weight Type: II  
Equipment ID: None  
Condition: New  
Temperature: 19.2 °C  
Pressure: 736.5 mmHg  
Relative Humidity: 51.4 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
50 lb	1	970	970	F	F	2.01	56
50 lb	2	760	760	F	F	2.01	56
50 lb	3	680	680	F	F	2.01	56
50 lb	4	1040	1040	F	F	2.01	56
50 lb	5	750	750	F	F	2.01	56
50 lb	6	770	770	F	F	2.01	56
50 lb	7	960	960	F	F	2.01	56
50 lb	8	670	670	F	F	2.01	56
50 lb	9	910	910	F	F	2.01	56
50 lb	10	880	880	F	F	2.01	56
50 lb	11	890	890	F	F	2.01	56
50 lb	12	930	930	F	F	2.01	56
50 lb	13	880	880	F	F	2.01	56
50 lb	14	850	850	F	F	2.01	56
50 lb	15	840	840	F	F	2.01	56
50 lb	16	780	780	F	F	2.01	56
50 lb	17	940	940	F	F	2.01	56
50 lb	18	620	620	F	F	2.01	56
50 lb	19	690	690	F	F	2.01	56
50 lb	20	780	780	F	F	2.01	56

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm<sup>3</sup> density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.


Pete Whebbe



Metrologist

Reviewed by:

Erik Alfvin



Metrologist

Receipt Date: January 17, 2017  
Cal. Date: January 17, 2017  
Report Date: January 18, 2017

Report No.: 337083  
Set Serial No.: None  
Barcode: 203142

## Calibration Certificate


FARM COUNTRY SCALE LLC  
9449 COUNTY ROAD 81  
FAIRMOUNT, ND 58030  
Contact: Joel Bommersbach  
Phone: 701-361-4483  
PO Number: None  
SOP: 8  
Technician ID: 19

Item(s) Submitted: AVDP Weight Kit - F  
Manufacturer: Rice Lake  
Weight Type: I & II  
Equipment ID: None  
Condition: New  
Temperature: 19.8 °C  
Pressure: 736.3 mmHg  
Relative Humidity: 49.2 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
5 lb		64.4	64.4	F	F	2.01	6.0
5 . lb		61.4	61.4	F	F	2.01	6.0
5 .. lb		61.4	61.4	F	F	2.01	6.0
5 :: lb		67.4	67.4	F	F	2.01	6.0
5 ::: lb		62.4	62.4	F	F	2.01	6.0
1 lb		16.2	16.2	F	F	2.02	1.6
1 . lb		23.7	23.7	F	F	2.02	1.6
1 .. lb		23.0	23.0	F	F	2.02	1.6
1 :: lb		19.2	19.2	F	F	2.02	1.6
1 ::: lb		23.4	23.4	F	F	2.02	1.6
8 oz		15.1	15.1	F	F	2.02	1.3
4 oz		6.82	6.82	F	F	2.00	0.22
2 oz		4.69	4.69	F	F	2.00	0.14
1 oz		1.47	1.47	F	F	2.00	0.11
1/2 oz		0.898	0.898	F	F	2.00	0.092
1/4 oz		0.473	0.473	F	F	2.00	0.056
1/8 oz		0.365	0.365	F	F	2.00	0.046
1/16 oz		0.555	0.555	F	F	2.00	0.070
1/32 oz		0.272	0.272	F	F	2.00	0.044
1/32 . oz		0.464	0.464	F	F	2.00	0.044

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm<sup>3</sup> density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.


Erik Alfvin



Metrologist

Reviewed by:

Pete Whebbe



Metrologist

Receipt Date: January 17, 2017  
Cal. Date: January 17, 2017  
Report Date: January 18, 2017

Report No.: 337084  
Set Serial No.: None  
Barcode: 203141

## Calibration Certificate

FARM COUNTRY SCALE LLC  
9449 COUNTY ROAD 81  
FAIRMOUNT, ND 58030  
Contact: Joel Bommersback  
Phone: 701-361-4483  
PO Number: None  
SOP: 8  
Technician ID: 19

Item(s) Submitted: Metric Weight Kit - F  
Manufacturer: Rice Lake  
Weight Type: I & II  
Equipment ID: None  
Condition: New  
Temperature: 19.9 °C  
Pressure: 736.4 mmHg  
Relative Humidity: 46.5 %

Nominal Value	Serial No.	<u>CM Correction (mg)</u>		<u>NIST HB105-1 Class</u>		<i>k</i>	U (mg)
		As Found	As Left	As Found	As Left		
2000 g		59.1	59.1	F	F	2.02	5.7
1000 g		29.8	29.8	F	F	2.04	2.5
500 g		20.1	20.1	F	F	2.02	2.0
200 g		8.69	8.69	F	F	2.03	0.55
200 g		9.13	9.13	F	F	2.03	0.55
100 g		5.26	5.26	F	F	2.03	0.25
50 g		4.36	4.36	F	F	2.03	0.16
20 g		0.92	0.92	F	F	2.03	0.11
20 g		0.77	0.77	F	F	2.03	0.11
10 g		0.409	0.409	F	F	2.03	0.072
5 g		0.622	0.622	F	F	2.03	0.054
2 g		0.408	0.408	F	F	2.03	0.048
2 g		0.493	0.493	F	F	2.03	0.048
1 g		0.331	0.331	F	F	2.03	0.039

Receipt Date: January 17, 2017  
Cal. Date: January 17, 2017  
Report Date: January 18, 2017

Report No.: 337084  
Set Serial No.: None  
Barcode: 203141

Continued,


## Calibration Certificate


FARM COUNTRY SCALE LLC  
9449 COUNTY ROAD 81  
FAIRMOUNT, ND 58030  
Contact: Joel Bommersback  
Phone: 701-361-4483  
PO Number: None  
SOP: 8  
Technician ID: 19

Item(s) Submitted: Metric Weight Kit - F  
Manufacturer: Rice Lake  
Weight Type: I & II  
Equipment ID: None  
Condition: New  
Temperature: 19.9 °C  
Pressure: 736.4 mmHg  
Relative Humidity: 46.5 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
0.5 g		0.343	0.343	F	F	2.06	0.021
0.2 g		0.222	0.222	F	F	2.06	0.016
0.2 g		0.257	0.257	F	F	2.06	0.016
0.1 g		0.174	0.174	F	F	2.06	0.015
0.05 g		0.142	0.142	F	F	2.06	0.011
0.02 g		0.0255	0.0255	F	F	2.06	0.0092
0.02 g		0.1285	0.1285	F	F	2.06	0.0092
0.01 g		0.0984	0.0984	F	F	2.06	0.0086

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm<sup>3</sup> density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Erik Alfvin  
  
Metrologist

Reviewed by:  
Pete Whebbe  
  
Metrologist

United States Department of Commerce  
National Institute of Standards and Technology

Certificate of Metrological Traceability For:

# Minnesota

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

### Scope

<b>Mass Echelon II</b>	<b>Weight Carts</b>	<b>Volume Gravimetric, I</b>
50 kg to 1 mg	10 000 lb to 2000 lb	20 L to 10 mL
1000 lb to 0.001 lb	Wheel Load Weighers	100 gal to 0.25 qt
4 oz to 0.03125 oz	20 000 lb to 2000 lb	<b>Volume Transfer, II</b>
<b>Mass Echelon III</b>	Railroad Test Cars	1500 gal to 5 gal
50 kg to 1 mg	110 000 lb to 80 000 lb	100 gal to 25 gal LPG
5000 lb to 0.001 lb		
4 oz to 0.03125 oz		



2016 to 2017

A handwritten signature in black ink, appearing to read "Carol T. Hockert".

Carol T. Hockert, Chief  
NIST Office of Weights and Measures

Effective Dates: 2016-01-01 to 2017-12-31