



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY

Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED


Name of Company <i>Midwest Meter Provers</i>	Email Address	Application Date <i>12-5-16</i>	
Mailing Address <i>2965 85th Ave, S.E.</i>	City <i>Jamestown</i>	State <i>N.D.</i>	Zip Code <i>58401</i>
Telephone Number	Cell Phone Number <i>701-320-3698</i>	Fax Number	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input checked="" type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: <i>3000 lbs</i> <input type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input checked="" type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
<i>1503</i>	<i>Allen Marsolek</i>	<i>NH³ Meters</i>



List below all field standards (attach current calibration reports):

500gal NH ³ tank on a 3000 lbs scale	

Additional Application Items (initial where appropriate):

Standardized Test Report	<u>AM</u> Copy enclosed <u>AM</u> No change in report filed previously
Tested and Approved Sticker	<u>AM</u> Copy enclosed <u>AM</u> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<u>AM</u> Copy enclosed <u>AM</u> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am Allen Marsolek, and have authority to represent this company. By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.

Allen Marsolek
Signature

Send Completed Application and Related Documents To:
Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410

North Dakota

nd.gov Official Portal for North Dakota State Government



SECRETARY OF STATE NORTH DAKOTA



[Home](#) | [Business Records Search](#)

MIDWEST METER PROVER'S

Trade Name

System ID: 13488600	Phone: Not on file
Status: Active	
Original File Date: 04/17/1998	Last Renewal Date: 04/17/2013

Nature of Business

TEST ANHYDROUS METERS IN THE STATE OF NORTH DAKOTA; THESE TESTS DETERMINE THAT THE CUSTOMER IS RECEIVING THE APPROPRIATE AMOUNT OF PRODUCT

Owners

ALLEN MARSOLEK
2965 85TH AVE SE
JAMESTOWN, ND 58401-9604

[Return to Search Results](#)

[Contact Us](#)


[Disclaimer](#)

[Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



South Dakota Department of Public Safety
Office of Weights and Measures
Metrology Lab

Lab: 1500 N Garfield – E. Truck Bypass Phone: 605-773-3170
Office: 118 West Capitol Avenue Phone: 605-773-3697
Pierre SD 57501



REPORT OF CALIBRATION

LAB TEST NUMBER: MP3436
DATE OF REPORT: 03/31/2016
DATE RECEIVED: 03/31/2016
DATE OF TEST: 03/31/2016

Submitted By: Midwest Meter Provers
Contact: Allan Marsolek
Mailing Address: 2965 85th Ave SE
City, State, Zip: Jamestown, ND 58401
Phone: 701-320-3698
S/A Number:

Standards Submitted:


1 - Rice Lake 3000 lb truck mounted scale

Uncertainty Statement: The combined standard uncertainty includes the standard uncertainty reported for the standard and the standard uncertainty for the measurement process. The combined standard uncertainty is multiplied by a coverage factor k to provide an expanded uncertainty which defines an interval having a level of confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the 1995 ISO/IEC Guide to the Expression of Uncertainty in Measurement. The expanded uncertainty is not to be confused with a tolerance limit for the user during application.

Traceability statement:

The Standards of the SD Metrology Laboratory used for comparison are traceable to the International System of Units (SI) through the National Institute of Standards and Technology. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

This document does not represent or imply endorsement by NIST Office of Weights and Measures or any agency of the State and/or national governments. The reported test values relate only to the observations made at the time and conditions of the test. This report may not be reproduced, except in full without the written approval of this laboratory. The client must not use this document to claim product endorsement by this laboratory.


Ron Peterson, Metrologist



03/31/16
Date



SOUTH DAKOTA WEIGHTS AND MEASURES / METROLOGY LAB

Lab: 1500 N. Garfield-E. Truck Bypass Phone: 605-773-3170

Office: 118 West Capitol Avenue Phone: 605-773-3697

Pierre, SD 57501

Submitted by :	ALLAN MARSOLEK	Report Number:	MP3436
Mailing Address:	2965 85 th Ave SE	Date Received:	03/31/16
City, State, Zip:	Jamestown, ND 58401	Date Tested:	03/31/16
Make of Scale:	Rice Lake	Condition of Scale	Good
Serial Number:	0008689	Temperature (c):	21
Capacity:	3000 lb X 0.5 lb	Humidity:	50.0%
		Pressure (mm/Hg):	705.9

Test Load (lb)	As Found (lb)	Error (lb)	Allowable Tolerance (+/- lb)	PASS/FAIL	As Left (lb)	Error (lb)
0	0	0	0.5	P	0	0
500	500.0	0	0.5	P	500.0	0
1000	1000.0	0	1.5	P	1000.0	0
2000	1999.5	-0.5	1.5	P	2000.0	-0.5


A shift test was performed using 1000 lb standard and the results fell within tolerance.

Repeatability test was performed and scale repeated within tolerance requirements.

A sensitivity weight of 5 lb was used at each weight and the results were within tolerance.

Standards used were calibrated using the working standards of the SD Metrology lab which were tested by the Minnesota State Lab, Report #332339, on 04/03/14 and are traceable to NIST.

END OF REPORT


 Ron Peterson, Metrologist
 Office of Weights and Measures
 118 W Capitol
 Pierre, SD 57501-2000

03/31/16
 Date
 Phone: 605-773-3697
 FAX: 605-773-6131
www.state.sd.us/dps/wm



United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

South Dakota

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon III
30 kg to 100 mg
1000 lb to 0.001 lb
8 oz to 0.03125 oz

Weight Carts
4000 lb to 2000 lb

Volume Transfer, II
5 gal



2016

Carol T. Hocken, Chief
NIST Office of Weights and Measures

Effective Dates: 2016-01-01 to 2016-12-31