



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY

Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company NCR Corporation	Email Address dave.boykin@ncr.com	Application Date 12/6/2016	
Mailing Address 200 Highway 74 South	City Peachtree City	State GA	Zip Code 30269
Telephone Number 770-288-1556	Cell Phone Number 678-416-0208	Fax Number 678-329-5705	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
<i>e.g. 1001</i>	<i>e.g. John Doe</i>	<i>e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6</i>
1754	Jesse Simpson	Scales - 7
1800	David Ferre	Scales - 7

Continued on Page 2

2 WM-16-723 Filed: 2/9/2017
Application for permit

Pages: 10

Application for Registration as a Registered Service Company
Page 2



List below all field standards (attach current calibration reports):

30# kit containing 2-10#, 1-5#, 2-2# & 1-1#	
kit also contains 13 decimal weights	


Additional Application Items (initial where appropriate):

Standardized Test Report	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am David Boykin, and have authority to represent this company.
By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.


Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY
 Public Service Commission
 SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company NCR Corporation	Email Address dave.boykin@ncr.com	Application Date 12/6/2016	
Mailing Address 200 Highway 74 South	City Peachtree City	State GA	Zip Code 30269
Telephone Number 770-288-1556	Cell Phone Number 678-416-0208	Fax Number 678-329-5705	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
<i>e.g. 1001</i>	<i>e.g. John Doe</i>	<i>e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6</i>
1754	Jesse Simpson	Scales - 7
1800	David Ferre	Scales - 7



List below all field standards (attach current calibration reports):

30# kit containing 2-10#, 1-5#, 2-2# & 1-1#	
kit also contains 13 decimal weights	

Additional Application Items (initial where appropriate):

Standardized Test Report	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am David Boykin, and have authority to represent this company.
By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.


Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410

North Dakota

nd.gov Official Portal for North Dakota State Government



SECRETARY OF STATE NORTH DAKOTA



[Home](#) | [Business Records Search](#)

NCR CORPORATION

Corporation Details

System ID: 3206700	Phone: (770) 623-7249
Type: FOREIGN BUSINESS CORPORATION	
Status: Active & Good Standing	
Original File Date: 04/19/1926	Effective Date: 04/19/1926
State of Origin: Maryland	

Nature of Business

COMPUTER PROGRAMING, DATA PROCESSING & RELATED SERVICES

Principal Office

3097 SATELLITE BLVD DULUTH, GA 30096-1242

Registered Agent

CORPORATION SERVICE COMPANY
 1709 N 19TH ST STE 3
 BISMARCK, ND 58501-2121
 Established Date: Jul 02, 2014

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2016](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)

[Disclaimer](#) | [Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



Hugh E. Weathers
Commissioner

David Ferre

State of South Carolina
Department of Agriculture
Metrology Laboratory

Mailing Address
PO Box 11280
Columbia, SC 29211

Physical Location
237 Catawba Street
Columbia, SC 29201

803.253.4052 (Phone/Fax)
agriculture.sc.gov

CERTIFICATE OF CALIBRATION
FOR

One weight set ranging from 0.001 Pound through 10 Pounds consisting of 19 weights

Manufacturer
Troemner

Serial Number
Asset # 1100

State Test Number
SC131211-2-16

Submitted by

NCR Corporation
200 Highway 74 South
Peachtree City, GA 30269



The test weights described above have been compared with the standards of the State of South Carolina using procedure SOP 8 from NIST IR 6969 and were found (adjusted) to be within NIST Class F tolerances with the following exceptions:

None

Date of Test: 02/17/16

Robert L. McGee, Director
Metrology Laboratory

Comments:

- This calibration was performed within the requirements of the laboratory Quality Management System version 6, Jan 2010 which is compliant with the ISO/IEC 17025, 2005 standard. The laboratory has demonstrated measurement proficiency through training and interlaboratory comparisons and has been issued a NIST Certificate of Measurement Traceability effective through December 31, 2017.
- This calibration has metrological traceability to the International System of Units (SI) through South Carolina primary standards traceable to the National Institute of Standards and Technology through NIST Test Numbers 822/275141-07, 822/278993-10, 684/283739-13.
- Date received: 01/29/16
- Laboratory environment: Temperature: 22.19 °C Humidity: 54.07 %RH
- Density corrections are negligible at the applied tolerance but are included as a component in the uncertainty calculation.
- The calibration values reported on the following pages are Conventional Mass corrections based on a reference density of 8.0 g/cm³ and a reference temperature of 20 °C.
- If the standard was not adjusted the "As Found" and "As Left" values are the same.

Certificate of Calibration
 Metrology Laboratory
 State Test Number: SC131211-2-16



Date of Test: 02/17/16

Weight Identification	Nominal	"As Found" Correction (mg)	"As Left" Correction (mg)	Uncertainty (mg)	Tolerance (mg)
	10 lb	213.2		6.7	450
dot	10 lb	220.2		6.7	450
	5 lb	143.1		3.8	230
	2 lb	15.32		0.75	91
dot	2 lb	48.21		0.75	91
	1 lb	27.04		0.75	70
	0.3 lb	11.68		0.50	27
	0.2 lb	6.27		0.50	18
dot	0.2 lb	5.62		0.50	18
	0.1 lb	3.69		0.50	9.1
dot	0.1 lb	3.16		0.50	9.1
	0.05 lb	0.45		0.50	4.5
	0.02 lb	0.18		0.50	1.8
dot	0.02 lb	0.35		0.50	1.8
	0.01 lb	0.78		0.23	1.5
	0.005 lb	0.18		0.23	1.2
	0.003 lb	0.06		0.23	0.99
	0.002 lb	-0.10		0.23	0.87
dot	0.001 lb	0.29		0.23	0.70

Note: The measurement uncertainty is calculated at an approximate 95 percent confidence level.

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

South Carolina

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon I	Weight Carts	Volume Transfer, II
30 kg to 1 mg	6000 lb to 2500 lb	20 L to 1 L
Mass Echelon II	Wheel Load Weighers	1500 gal to 1 gal
50 kg to 1 mg	36 000 lb to 12 000 lb	100 gal to 20 gal LPG
1000 lb to 0.001 lb	Volume Gravimetric, I	Grain Moisture
8 oz to 0.015625 oz	20 L to 100 ml	23 % to 13 %
Mass Echelon III	5 gal to 1 gal	
1000 kg to 1 mg		
2500 lb to 0.001 lb		
8 oz to 0.015625 oz		



2016 to 2017

Carol T. Hockett, Chief
NIST Office of Weights and Measures

Effective Dates: 2016-01-01 to 2017-12-31

Receipt Date: March 4, 2016
Test Date: March 21, 2016
Report Date: March 21, 2016

State Test No.: 335655
Set Serial No.: Asset #2312
Barcode: 018130

Calibration Report

NCR
213 CIRCLE DR N
FARGO, ND 58102
Contact: JESSE SIMPSON
Phone: 770-288-1568
PO Number: NONE
SOP: 12
Technician ID: 09

Item(s) Submitted: 30 lb kit w/ decimals
Manufacturer: Troemner
ASTM E617 Type: I & II
Equipment ID#: None
Condition: Good
Temperature: 20.7 °C
Pressure: 734. mmHg
Relative Humidity: 49.0 %



Nominal Value	Serial No.	Correction (mg)		NIST HB105-1 Class		Uncertainty (mg) (k = 2)
		As Found	As Left	As Found	As Left	
10 lb		35	35	F	F	15
10 lb		34	34	F	F	15
5 lb		-13	-13	F	F	12
2 lb		-0.5	-0.5	F	F	4.8
2 lb		46.1	46.1	F	F	4.8
1 lb		-13.3	-13.3	F	F	2.8
0.3 lb		0.66	0.66	F	F	0.48
0.2 lb		-3.95	-3.95	F	F	0.37
0.2 lb		-4.91	-4.91	F	F	0.37
0.1 lb		7.43	7.43	F	F	0.30
0.1 lb		8.22	8.22	F	F	0.30
0.05 lb		-1.48	-1.48	F	F	0.30
0.02 lb		-0.73	-0.73	F	F	0.16
0.02 lb		0.01	0.01	F	F	0.16
0.01 lb		-0.06	-0.06	F	F	0.12
0.005 lb		0.73	0.73	F	F	0.10
0.002 lb		-0.120	-0.120	F	F	0.066
0.002 lb		0.110	0.110	F	F	0.066
0.001 lb		-0.430	-0.430	F	F	0.057

When used as a set these weights meet NIST HB 105-1 Class F tolerances.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³. The items listed above have been compared to the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). Results apply to items identified in this report only.

Heidi Jones
Heidi Jones
Laboratory Administrator

Reviewed by:
Mark Nicollet
Mark Nicollet
Quality Manager

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Minnesota

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon II	Weight Carts	Volume Gravimetric, I
50 kg to 1 mg	10 000 lb to 2000 lb	20 L to 10 mL
1000 lb to 0.001 lb	Wheel Load Weighers	100 gal to 0.25 qt
4 oz to 0.03125 oz	20 000 lb to 2000 lb	Volume Transfer, II
Mass Echelon III	Railroad Test Cars	1500 gal to 5 gal
50 kg to 1 mg	110 000 lb to 80 000 lb	100 gal to 25 gal LPG
5000 lb to 0.001 lb		
4 oz to 0.03125 oz		



2016 to 2017

A handwritten signature in black ink, appearing to read "Carol T. Hockett".

Carol T. Hockett, Chief
NIST Office of Weights and Measures

Effective Dates: 2016-01-01 to 2017-12-31