



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY
Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company <i>O'Day Equipment LLC</i>	Email Address <i>Alton.Graves@odayequipment.com</i>	Application Date	
Mailing Address <i>635 31st St. SE</i>	City <i>Minot</i>	State <i>ND</i>	Zip Code <i>58701</i>
Telephone Number <i>(701) 852-3145</i>	Cell Phone Number <i>(701) 818-4243</i>	Fax Number <i>(701) 852-6058</i>	

Select below all device types your company will certify:

Scales (Include maximum capacity, if applicable)	Liquid (Include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail	<input checked="" type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute)
<input type="checkbox"/> 2. Truck	<input checked="" type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater)
<input type="checkbox"/> 3. Livestock	<input checked="" type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: <i>150 gpm</i>
<input type="checkbox"/> 4. Hopper: Max. Capacity: _____	<input checked="" type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: <i>150 gpm</i>
<input type="checkbox"/> 5. Belt	<input checked="" type="checkbox"/> 5. LPG
<input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____	<input checked="" type="checkbox"/> 6. Stationary LPG
<input type="checkbox"/> 7. 30 lbs. or less	<input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____
<input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified)	<input type="checkbox"/> 8. Chemical
<input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 9. Anhydrous
	<input type="checkbox"/> 10. Loading Rack
	<input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
e.g. 1001	e.g. John Doe	e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6
<i>1476</i>	<i>James Rogers</i>	<i>1, 2, 4, 5, 6</i>
<i>1678</i>	<i>Kurt Helseth</i>	<i>1, 2, 4, 5, 6</i>
<i>1634</i>	<i>Vernon Patterson</i>	<i>1, 2, 4, 5, 6</i>
<i>1668</i>	<i>John Boehm</i>	<i>1, 2, 4, 5, 6</i>
<i>1740</i>	<i>Zebulan June</i>	<i>1, 2, 4, 5, 6</i>
<i>1770</i>	<i>Levi Jensen</i>	<i>1, 2, 4, 5, 6</i>
<i>1797</i>	<i>Russell Davis</i>	<i>1, 2, 4, 5, 6</i>

Continued on Page 2

Pages: 22

Filed: 1/12/2017

Application for permit - Minot

5

O'Day Equipment, LLC

Cont -

Cont



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY
Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company	Email Address	Application Date	
Mailing Address	City	State	Zip Code
Telephone Number	Cell Phone Number	Fax Number	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute)
<input type="checkbox"/> 2. Truck	<input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater)
<input type="checkbox"/> 3. Livestock	<input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____
<input type="checkbox"/> 4. Hopper: Max. Capacity: _____	<input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____
<input type="checkbox"/> 5. Belt	<input type="checkbox"/> 5. LPG
<input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____	<input type="checkbox"/> 6. Stationary LPG
<input type="checkbox"/> 7. 30 lbs. or less	<input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____
<input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified)	<input type="checkbox"/> 8. Chemical
<input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 9. Anhydrous
	<input type="checkbox"/> 10. Loading Rack
	<input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
e.g. 1001	e.g. John Doe	e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6
1798	Leif Benckley	1, 2, 4, 5, 6
1799	James Crawford	1, 2, 4, 5, 6



Application for Registration as a Registered Service Company
Page 2

List below all field standards (attach current calibration reports):

100 gal Prover ^{SN#} 041320840-0102	5 gal Seraphim ^{SN #1}
" " ^{SN#} 051320917-0103	" " ^{SN} 09-06136
100 gal LP Prover ^{SN} 44753	^{SN} 40549
5 gal Seraphim ^{SN} 09-6135	42653
" " ^{SN} 09-45256-07	15-92443
^{SN} 45401	15-92450
09-06131	15-92411
11-88799	
07-08144	


Additional Application Items (initial where appropriate):

Standardized Test Report	<input checked="" type="checkbox"/> Copy enclosed
	<input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input checked="" type="checkbox"/> Copy enclosed
	<input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input checked="" type="checkbox"/> Copy enclosed
	<input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am Allen J. Graves SRVC MGR and have authority to represent this company. By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.


Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410

North Dakota

nd.gov Official Portal for
North Dakota State GovernmentNorth Dakota
LEGENDARY

SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

O'DAY EQUIPMENT, LLC

Corporation Details

System ID: 22745300**Phone:** (800) 654-6329**Type:** LIMITED LIABILITY COMPANY**Status:** Active & Good Standing**Original File Date:** 12/12/2006**Effective Date:** 12/12/2006**State of Origin:** North Dakota

Nature of Business

OPERATE A PETROLEUM EQUIPMENT BUSINESS

Principal Office

1301 40TH ST NW PO BOX 2706 FARGO, ND 58108-2706

Registered Agent

D JAMES O'DAY

1301 40TH ST N

PO BOX 2706

FARGO, ND 58108-2706

Established Date: Dec 12, 2006

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2016](#) (generates a forms-fillable pdf in a new pop-up window)[Return to Search Results](#)[Contact Us](#)[Disclaimer](#)[Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



Receipt Date: March 18, 2016
Cal. Date: March 24, 2016
Report Date: March 25, 2016

Report No.: 335719
Serial No.: 09-06135
Barcode: 202245

Calibration Certificate

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, ND 58701
Contact: AJ GRAVES
Phone: 701-356-8472
PO Number: NONE
SOP: 19
Technician ID: 07

Item(s) Submitted: 5 Gallon Measure
Manufacturer: Seraphin
Material: Stainless Steel
Type: Measure
Condition: Good
Temperature: 19.1 °C
Pressure: 733.3 mmHg
Relative Humidity: 35.3 %
Standard H₂O Temp.: 12.7 °C
Artifact H₂O Temp.: 12.8 °C

Nominal		Calibrated				
Volume (gal)		Volume (gal)	Error (in ³)	<i>k</i>	U (in ³)	CCE (°F)
5	As Found	4.9995	-0.11	2.06	0.24	0.0000265
	As Left	4.9995	-0.11			

Neck Calibration: No neck calibration was performed at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a pour time of 30 seconds followed by a drain time of 10 seconds after cessation of full flow.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Mark Nicollet

Mark Nicollet
Quality Manager

Reviewed by:

Pete Whebbe

Pete Whebbe
Metrologist



Receipt Date: March 18, 2016
Cal. Date: March 23, 2016
Report Date: March 23, 2016

Report No.: 335712
Serial No.: 07-45256-07
Barcode: 202242

Calibration Certificate

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, ND 58701

Contact: AJ GRAVES
Phone: 701-356-8472
PO Number: NONE
SOP: 19
Technician ID: 11

Item(s) Submitted: 5 Gallon Prover
Manufacturer: Seraphin
Material: Stainless Steel
Description: No Bottom Zero
Condition: Good
Temperature: 19.1 °C
Pressure: 736.5 mmHg
Relative Humidity: 39.2 %
Standard H₂O Temp.: 17.2 °C
Artifact H₂O Temp.: 17.2 °C

Nominal		Calibrated				
Volume (gal)		Volume (gal)	Error (in ³)	<i>k</i>	U (in ³)	CCE (°F)
5	As Found	5.0004	0.10	2.06	0.24	0.0000265
	As Left	5.0004	0.10			

Neck Calibration: No neck calibration was performed at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a drain time of 30 seconds after cessation of full flow.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Pete Whebbe

Peter J. Whebbe
Metrologist

Reviewed by:

Mark Nicollet

Mark Nicollet
Quality Manager



Receipt Date: March 21, 2016
Cal. Date: March 24, 2016
Report Date: March 25, 2016

Report No.: 335718
Serial No.: 45401
Barcode: 202729

Calibration Certificate

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, ND 58701
Contact: AJ GRAVES
Phone: 701-356-8472
PO Number: NONE
SOP: 19
Technician ID: 07

Item(s) Submitted: 5 Gallon Measure
Manufacturer: Seraphin
Material: Mild Steel
Type: Measure
Condition: Fair/Dirty
Temperature: 19.1 °C
Pressure: 733.3 mmHg
Relative Humidity: 35.3 %
Standard H₂O Temp.: 12.5 °C
Artifact H₂O Temp.: 12.5 °C

Nominal		Calibrated		<i>k</i>	U (in ³)	CCE (°F)
Volume (gal)		Volume (gal)	Error (in ³)			
5	As Found	4.9991	-0.21	2.06	0.24	0.0000186
	As Left	4.9991	-0.21			

Neck Calibration: No neck calibration was performed at this time.


This prover has been calibrated as a "to contain after wet down" vessel with a pour time of 30 seconds followed by a drain time of 10 seconds after cessation of full flow.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.


CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Mark Nicollet


Quality Manager

Reviewed by:

Pete Whebbe


Metrologist



Receipt Date: March 18, 2016
Cal. Date: March 23, 2016
Report Date: March 23, 2016

Report No.: 335713
Serial No.: 09-06131
Barcode: 202247

Calibration Certificate

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, ND 58701
Contact: AJ GRAVES
Phone: 701-356-8472
PO Number: NONE
SOP: 19
Technician ID: 11

Item(s) Submitted: 5 Gallon Measure
Manufacturer: Seraphin
Material: Stainless Steel
Description: Measure
Condition: Good
Temperature: 19.1 °C
Pressure: 736.5 mmHg
Relative Humidity: 39.2 %
Standard H₂O Temp.: 17.3 °C
Artifact H₂O Temp.: 17.3 °C

Nominal		Calibrated				
Volume (gal)		Volume (gal)	Error (in ³)	<i>k</i>	U (in ³)	CCE (°F)
5	As Found	4.9991	-0.22	2.06	0.24	0.0000265
	As Left	4.9991	-0.22			

Neck Calibration: No neck calibration was performed at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a pour time of 30 seconds followed by a drain time of 10 seconds after cessation of full flow.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Pete Whebbe

Pete J. Whebbe
Metrologist

Reviewed by:

Mark Nicollet

Mark Nicollet
Quality Manager



Receipt Date: March 18, 2016
Cal. Date: March 24, 2016
Report Date: March 25, 2016

Report No.: 335717
Serial No.: 11-88799
Barcode: 202244

Calibration Certificate

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, ND 58701

Contact: AJ GRAVES
Phone: 701-356-8472
PO Number: NONE
SOP: 19
Technician ID: 07

Item(s) Submitted: 5 Gallon Measure
Manufacturer: Seraphin
Material: Stainless Steel
Type: Measure
Condition: Good/Dirty
Temperature: 19.1 °C
Pressure: 733.3 mmHg
Relative Humidity: 35.3 %
Standard H₂O Temp.: 14.1 °C
Artifact H₂O Temp.: 14.1 °C

Nominal		Calibrated				
Volume (gal)		Volume (gal)	Error (in ³)	<i>k</i>	U (in ³)	CCE (°F)
5	As Found	4.9994	-0.14	2.06	0.24	0.0000265
	As Left	4.9994	-0.14			

Neck Calibration: No neck calibration was performed at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a pour time of 30 seconds followed by a drain time of 10 seconds after cessation of full flow.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Mark Nicollet

Quality Manager

Reviewed by:

Pete Whebbe

Metrologist



Receipt Date: March 18, 2016
Cal. Date: March 24, 2016
Report Date: March 25, 2016

Report No.: 335720
Serial No.: 07-03144
Barcode: 202246

Calibration Certificate

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, ND 58701

Contact: AJ GRAVES
Phone: 701-356-8472
PO Number: NONE
SOP: 19
Technician ID: 07

Item(s) Submitted: 5 Gallon Measure
Manufacturer: Seraphin
Material: Stainless Steel
Type: Measure
Condition: Good/Dirty
Temperature: 19.1 °C
Pressure: 733.3 mmHg
Relative Humidity: 35.3 %
Standard H₂O Temp.: 12.8 °C
Artifact H₂O Temp.: 12.8 °C

Nominal Volume (gal)		Calibrated		<i>k</i>	U (in ³)	CCE (°F)
		Volume (gal)	Error (in ³)			
5	As Found	5.0021	0.49	2.06	0.24	0.0000265
	As Left	5.0006	0.14			

Neck Calibration: No neck calibration was performed at this time.


This prover has been calibrated as a "to contain after wet down" vessel with a pour time of 30 seconds followed by a drain time of 10 seconds after cessation of full flow.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.


CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Mark Nicollet


Quality Manager

Reviewed by:

Pete Whebbe


Metrologist



Receipt Date: March 18, 2016
Cal. Date: March 24, 2016
Report Date: March 25, 2016

Report No.: 335715
Serial No.: 1
Barcode: 202467

Calibration Certificate

O'DAY EQUIPMENT LLC

635 31ST ST SE

MINOT, ND 58701

Contact: AJ GRAVES

Phone: 701-356-8472

PO Number: NONE

SOP: 19

Technician ID: 07

Item(s) Submitted: 5 Gallon Measure

Manufacturer: SERAPHIN

Material: Mild Steel

Type: Measure

Condition: Fair

Temperature: 19.1 °C

Pressure: 733.3 mmHg

Relative Humidity: 35.5 %

Standard H₂O Temp.: 14.6 °C

Artifact H₂O Temp.: 14.6 °C

Nominal Volume (gal)		Calibrated		<i>k</i>	U (in ³)	CCE (°F)
		Volume (gal)	Error (in ³)			
5	As Found	4.9997	-0.08	2.06	0.24	0.0000186
	As Left	4.9997	-0.08			

Neck Calibration: No neck calibration was performed at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a pour time of 30 seconds followed by a drain time of 10 seconds after cessation of full flow.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Mark Nicollet

Mark Nicollet
Quality Manager

Reviewed by:

Pete Whebbe

Pete Whebbe
Metrologist



Receipt Date: March 18, 2016
Cal. Date: March 23, 2016
Report Date: March 23, 2016

Report No.: 335714
Serial No.: 09-06136
Barcode: 202243

Calibration Certificate

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, ND 58701

Contact: AJ GRAVES
Phone: 701-356-8472
PO Number: NONE
SOP: 19
Technician ID: 11

Item(s) Submitted: 5 Gallon Measure
Manufacturer: Seraphin
Material: Stainless Steel
Description: Measure
Condition: Good
Temperature: 19.1 °C
Pressure: 736.5 mmHg
Relative Humidity: 39.2 %
Standard H₂O Temp.: 17.5 °C
Artifact H₂O Temp.: 17.5 °C

Nominal		Calibrated				
Volume (gal)		Volume (gal)	Error (in ³)	<i>k</i>	U (in ³)	CCE (°F)
5	As Found	4.9974	-0.61	2.06	0.24	0.0000265
	As Left	4.9997	-0.06			

Neck Calibration: No neck calibration was performed at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a pour time of 30 seconds followed by a drain time of 10 seconds after cessation of full flow.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Pete Whebbe

Pete Whebbe
Metrologist

Reviewed by:

Mark Nicollet

Mark Nicollet
Quality Manager



Receipt Date: March 18, 2016
Cal. Date: March 24, 2016
Report Date: March 25, 2016

Report No.: 335716
Serial No.: 40549
Barcode: 202248

Calibration Certificate

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, ND 58701
Contact: AJ GRAVES
Phone: 701-356-8472
PO Number: NONE
SOP: 19
Technician ID: 07

Item(s) Submitted: 5 Gallon Measure
Manufacturer: Seraphin
Material: Mild Steel
Type: Measure
Condition: Good
Temperature: 19.1 °C
Pressure: 733.3 mmHg
Relative Humidity: 35.3 %
Standard H₂O Temp.: 15.2 °C
Artifact H₂O Temp.: 15.3 °C

Nominal		Calibrated				
Volume (gal)		Volume (gal)	Error (in ³)	<i>k</i>	U (in ³)	CCE (°F)
5	As Found	5.0036	0.84	2.06	0.24	0.0000186
	As Left	5.0003	0.06			

Neck Calibration: No neck calibration was performed at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a pour time of 30 seconds followed by a drain time of 10 seconds after cessation of full flow.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Mark Nicollet

Quality Manager

Reviewed by:

Pete Whebbe

Metrologist



Receipt Date: March 18, 2016
Cal. Date: March 24, 2016
Report Date: March 25, 2016

Report No.: 335721
Serial No.: 42653
Barcode: 202970

Calibration Certificate

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, ND 58701
Contact: AJ GRAVES
Phone: 701-356-8472
PO Number: NONE
SOP: 19
Technician ID: 07

Item(s) Submitted: 5 Gallon Measure
Manufacturer: Seraphin
Material: Mild Steel
Type: Measure
Condition: Good
Temperature: 19.1 °C
Pressure: 733.3 mmHg
Relative Humidity: 35.3 %
Standard H₂O Temp.: 12.4 °C
Artifact H₂O Temp.: 12.5 °C

Nominal	Calibrated				
Volume (gal)	Volume (gal)	Error (in ³)	<i>k</i>	U (in ³)	CCE (°F)
5	As Found	4.9976	-0.55	2.06	0.24 0.0000186
	As Left	5.0003	0.08		


Neck Calibration: No neck calibration was performed at this time.


This prover has been calibrated as a "to contain after wet down" vessel with a pour time of 30 seconds followed by a drain time of 10 seconds after cessation of full flow.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Mark Nicollet

Quality Manager

Reviewed by:
Pete Whebbe

Metrologist



Receipt Date: March 18, 2016
Cal. Date: March 24, 2016
Report Date: March 25, 2016

Report No.: 335722
Serial No.: 15-92443
Barcode: 202811

Calibration Certificate

O'DAY EQUIPMENT LLC

635 31ST ST SE

MINOT, ND 58701

Contact: AJ GRAVES

Phone: 701-356-8472

PO Number: NONE

SOP: 19

Technician ID: 07

Item(s) Submitted: 5 Gallon Measure

Manufacturer: Seraphin

Material: Stainless Steel

Type: Measure

Condition: Good/Dirty

Temperature: 19.1 °C

Pressure: 733.3 mmHg

Relative Humidity: 35.3 %

Standard H₂O Temp.: 15.0 °C

Artifact H₂O Temp.: 14.9 °C

Nominal		Calibrated				
Volume (gal)		Volume (gal)	Error (in ³)	<i>k</i>	U (in ³)	CCE (/°F)
5	As Found	5.0005	0.11	2.06	0.24	0.0000265
	As Left	5.0005	0.11			

Neck Calibration: No neck calibration was performed at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a pour time of 30 seconds followed by a drain time of 10 seconds after cessation of full flow.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Mark Nicollet

Quality Manager

Reviewed by:

Pete Whebbe

Metrologist



Receipt Date: March 18, 2016
Cal. Date: March 24, 2016
Report Date: March 25, 2016

Report No.: 335723
Serial No.: 15-92450
Barcode: 202816

Calibration Certificate

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, ND 58701
Contact: AJ GRAVES
Phone: 701-356-8472
PO Number: NONE
SOP: 19
Technician ID: 07

Item(s) Submitted: 5 Gallon Measure
Manufacturer: Seraphin
Material: Stainless Steel
Type: Measure
Condition: Good
Temperature: 19.1 °C
Pressure: 733.3 mmHg
Relative Humidity: 35.3 %
Standard H₂O Temp.: 13.4 °C
Artifact H₂O Temp.: 13.4 °C

Nominal		Calibrated		<i>k</i>	U (in ³)	CCE (°F)
Volume (gal)		Volume (gal)	Error (in ³)			
5	As Found	4.9993	-0.16	2.06	0.24	0.0000265
	As Left	4.9993	-0.16			

Neck Calibration: No neck calibration was performed at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a pour time of 30 seconds followed by a drain time of 10 seconds after cessation of full flow.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Mark Nicollet

Quality Manager

Reviewed by:

Pete Whebbe

Metrologist



Receipt Date: March 18, 2016
Cal. Date: March 24, 2016
Report Date: March 25, 2016

Report No.: 335724
Serial No.: 15-92411
Barcode: 202817

Calibration Certificate

O'DAY EQUIPMENT LLC

635 31ST ST SE

MINOT, ND 58701

Contact: AJ GRAVES

Phone: 701-356-8472

PO Number: NONE

SOP: 19

Technician ID: 07

Item(s) Submitted: 5 Gallon Measure

Manufacturer: Seraphin

Material: Stainless Steel

Type: Measure

Condition: Good

Temperature: 19.1 °C

Pressure: 733.3 mmHg

Relative Humidity: 35.3 %

Standard H₂O Temp.: 12.1 °C

Artifact H₂O Temp.: 12.1 °C

Nominal Volume (gal)		Calibrated		<i>k</i>	U (in ³)	CCE (°F)
		Volume (gal)	Error (in ³)			
5	As Found	5.0002	0.04	2.06	0.24	0.0000265
	As Left	5.0002	0.04			

Neck Calibration: No neck calibration was performed at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a pour time of 30 seconds followed by a drain time of 10 seconds after cessation of full flow.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-3 (2010). Uncertainty calculations contain the components in NIST SOP 19 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F

Mark Nicollet

Quality Manager

Reviewed by:

Pete Whebbe

Metrologist



Receipt Date: March 18, 2016
Test Date: March 22, 2016
Report Date: March 22, 2016

State Test No.: 335711
Serial No.: 041320840-0102
Barcode: 202317

Calibration Report

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, ND 58701

Contact: AJ Graves
Phone: 701-356-8472
PO Number: None
SOP: 33
Technician ID: 19

Item(s) Submitted: 100 Gallon Prover
Manufacturer: Westmor Fluid Solution
Material: Stainless Steel (304)
Description: Dry Bottom
Condition: Good
Temperature: 19.2 °C
Pressure: 728. mmHg
Relative Humidity: 42.7 %
Standard H₂O Temp.: 7.5 °C
Artifact H₂O Temp.: 7.6 °C

Nominal Volume (gal)		Tested		Uncertainty (in ³)	Coefficient of Expansion(°F)
		Volume (gal)	Error (in ³)		
100	As Found	99.971	-6.8	2.4	0.0000288
	As Left	100.009	2.1	2.4	

Neck Calibration: No neck calibration was done at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a drain time of 30 seconds after cessation of full flow and at a reference temperature of 60 °F.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

The reported uncertainty conforms to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). The confidence interval is 95 %.

Erik Alfvin

Metrologist

Reviewed by:

Mark Nicollet

Quality Manager



Receipt Date: March 18, 2016
Test Date: March 22, 2016
Report Date: March 22, 2016

State Test No.: 335710
Serial No.: 051320997-0103
Barcode: 202318

Calibration Report

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, MN 58701
Contact: AJ Graves
Phone: 701-356-8472
PO Number: None
SOP: 33
Technician ID: 19

Item(s) Submitted: 100 Gallon Prover
Manufacturer: Determan Brownie Inc.
Material: Stainless Steel (304)
Description: Dry Bottom
Condition: Good
Temperature: 19.3 °C
Pressure: 728.1 mmHg
Relative Humidity: 39.0 %
Standard H₂O Temp.: 7.6 °C
Artifact H₂O Temp.: 7.6 °C

Nominal Volume (gal)		Tested Volume (gal)	Error (in ³)	Uncertainty (in ³)	Coefficient of Expansion(°F)
100	As Found	99.991	-2.1	2.4	0.0000288
	As Left	99.991	-2.1	2.4	

Neck Calibration: No neck calibration was done at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a drain time of 30 seconds after cessation of full flow and at a reference temperature of 60 °F.

The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

The reported uncertainty conforms to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008). The confidence interval is 95 %.

Erik Alfvin

Metrologist

Reviewed by:
Mark Nicollet

Quality Manager



Receipt Date: March 18, 2016
Cal. Date: March 23, 2016
Report Date: March 24, 2016

Report No.: 335709
Serial No.: 44753
Barcode: 202355

Calibration Certificate

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, ND 58701
Contact: AJ Graves
Phone: 701-356-8472
PO Number: None
SOP: 34
Technician ID: 19

Item(s) Submitted: 100 Gallon LPG Prover
Manufacturer: Arrow Tank
Material: Mild Steel
Description: Zero Bottom
Condition: Good
Temperature: 18.8 °C
Pressure: 736.8 mmHg
Relative Humidity: 39.2 %
Standard H₂O Temp. 8.7 °C
Artifact H₂O Temp.: 9.7 °C

Nominal		Calibrated				
Volume (gal)		Volume (gal)	Error (in ³)	<i>k</i>	U (in ³)	CCE (°F)
100	As Found (at 100 psig)	99.901	-22.8	2.02	5.3	0.0000186
	As Left (at 100 psig)	100.015	3.5			

Neck Calibration: No neck calibration was performed at this time.

This prover has been calibrated as a "to contain after wet down" vessel with a drain time of 30 seconds after cessation of full flow. The prover listed above has been compared by volumetric transfer methods to the standards of the State of Minnesota using water as the calibration medium. The standards are traceable to the SI through NIST. Statistical process control charts indicate standards are currently in control. All gauges were sealed in place.

All tolerances and specifications were evaluated according to NIST Handbook 105-4 (2010). Uncertainty calculations contain the components in NIST SOP 21 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to item identified in this report only.

CCE is the cubical coefficient of thermal expansion, and the reference temperature is 60 °F.

Erik Alfvin

Metrologist

Reviewed by:
Mark Nicollet

Quality Manager



Receipt Date: March 18, 2016
Cal. Date: March 23, 2016
Report Date: March 24, 2016

Report No.: 335709
Serial No.: 44753
Barcode: 202355

Pressure Correction Chart

O'DAY EQUIPMENT LLC
635 31ST ST SE
MINOT, ND 58701
Contact: AJ Graves
Phone: 701-356-8472
PO Number: None
SOP: 34
Technician ID: 19

Item(s) Submitted: 100 Gallon LPG Prover
Manufacturer: Arrow Tank
Material: Mild Steel
Description: Zero Bottom
Condition: Good
Temperature: 18.8 °C
Pressure: 736.8 mmHg
Relative Humidity: 39.2 %

Pressure Gauge Reading (psig)	Corrected Volume (gal)
0	99.870
10	99.888
20	99.907
30	99.925
40	99.944
50	99.962
60	99.973
70	99.983
80	99.994
90	100.004
100	100.015
110	100.024
120	100.033
130	100.042
140	100.051
150	100.060
160	100.069
170	100.078
180	100.087
190	100.096
200	100.105

Erik Alfvén

Metrologist

United States Department of Commerce
National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Minnesota

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon II	Weight Carts	Volume Gravimetric, I
50 kg to 1 mg	10 000 lb to 2000 lb	20 L to 10 mL
1000 lb to 0.001 lb	Wheel Load Weighers	100 gal to 0.25 qt
4 oz to 0.03125 oz	20 000 lb to 2000 lb	Volume Transfer, II
Mass Echelon III	Railroad Test Cars	1500 gal to 5 gal
50 kg to 1 mg	110 000 lb to 80 000 lb	100 gal to 25 gal LPG
5000 lb to 0.001 lb		
4 oz to 0.03125 oz		



2016 to 2017

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2016-01-01 to 2017-12-31