



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY
 Public Service Commission
 SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company Pomeroy IT Solutions Sales Co Inc	Email Address WandM@Pomeroy.com	Application Date 12/15/16	
Mailing Address 69009 Northpark Blvd	City Charlotte	State NC	Zip Code 28214
Telephone Number 704-943-5343	Cell Phone Number	Fax Number	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____ <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
e.g. 1001	e.g. John Doe	e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6
1790	Brandon Elenspeter	Scales 7
1791	Eric Gault	Scales 7

Continued on Page 2

2 WM-16-729 Filed: 12/19/2016 Pages: 6
 Application for permit



List below all field standards (attach current calibration reports):

02 ZS	
A051D	


Additional Application Items (initial where appropriate):

Standardized Test Report	<input checked="" type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.
 Yes No

I am Cass Utter, and have authority to represent this company.
By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.




Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410

North Dakota

nd.gov Official Portal for North Dakota State Government



SECRETARY OF STATE NORTH DAKOTA



[Home](#) | [Business Records Search](#)

POMEROY IT SOLUTIONS SALES COMPANY, INC.

Corporation Details

System ID: 26645300	Phone: (310) 712-1850
Type: FOREIGN BUSINESS CORPORATION	
Status: Active & Good Standing	
Original File Date: 02/04/2010	Effective Date: 02/04/2010
State of Origin: Delaware	

Nature of Business

TECHNOLOGY SERVICES

Principal Office

1020 PETERSBURG RD HEBRON, KY 41048-8222

Registered Agent

REGISTERED AGENT SOLUTIONS, INC.
 1709 N 19TH ST STE 3
 BISMARCK, ND 58501-2121
 Established Date: Apr 12, 2016

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2016](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)

[Disclaimer](#)

[Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



Virginia Department of Agriculture and Consumer Services
 Office of Weights & Measures • Metrology Lab
 600 North 5th St. Rm: 162 • Richmond, Va. 23219 • USA
 Telephone: (804)786-0479 • Fax: (804)371-0206



NVLAP Lab Code 105007-0

CALIBRATION CERTIFICATE

Tested for: TOLT SERVICE GROUP
 6900 NORTH PARK BLVD, SUITE C
 CHARLOTTE, NC 28216

Date Received: April 8, 2016
 Date of Test: April 19, 2016
 Test Number: VA-16-10112 P
 Tested By: WSWHL
 Procedure: NISTIR 6969 SOP #8

Description of items calibrated: STAINLESS STEEL WEIGHT KIT - #A051D
 Class: NIST HB105-1 (Class F) Condition: GOOD
 MEETS NIST CLASS F SPECIFICATION

* The uncertainty represents an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of k=2.

Environmental condition taken during testing: Temp: 21°C Hum: 47%

Nominal Value	ID	As Found** (error) (mg)	As Left (error) (mg)	Tolerance (mg)	Uncertainty* (±mg)	Balance Used	ID of Standard Set Used
5 lb	A	47	47	230	13	PR5003	W5LB
5 lb	B	61	61	230	13	PR5003	W5LB
5 lb	C	72	72	230	13	PR5003	W5LB
5 lb	D	53	53	230	13	PR5003	W5LB
5 lb	E	34	34	230	13	PR5003	W5LB
1 lb	A	16.6	16.6	70	4.5	CC1200	W1LB
1 lb	B	12.6	12.6	70	4.5	CC1200	W1LB
1 lb	C	10.3	10.3	70	4.5	CC1200	W1LB
1 lb	D	11.5	11.5	70	4.5	CC1200	W1LB
1 lb	E	23.6	23.6	70	4.5	CC1200	W1LB
0.5 lb		18.3	18.3	45	2.5	CC1200	W0.5LB
0.2 lb		6.45	6.45	18	0.22	MSE225S	W0.2LB
0.2 lb	dot	5.76	5.76	18	0.22	MSE225S	W0.2LB
0.1 lb		3.62	3.62	9.1	0.13	MSE225S	W0.1LB
0.05 lb		0.82	0.82	4.5	0.13	MSE225S	W0.05LB
0.02 lb		0.73	0.73	1.8	0.077	MSE225S	W0.02LB
0.02 lb	dot	0.72	0.72	1.8	0.077	MSE225S	W0.02LB
0.01 lb		0.25	0.25	1.5	0.077	MSE225S	W0.01LB
0.005 lb		0.65	0.65	1.2	0.077	MSE225S	W0.005LB
0.002 lb		0.81	0.81	0.87	0.066	MSE225S	W0.002LB
0.002 lb	dot	0.24	0.24	0.87	0.066	MSE225S	W0.002LB
0.001 lb		0.29	0.29	0.70	0.066	MSE225S	W0.001LB

**1 US pound equals 453.5924 grams.

The results of the calibration are metrological traceable to the SI. The SI of the Va lab is through the National Institute of Standards and Technology (NIST) through an unbroken chain of traceable standards and controlled processes.

* The uncertainty represents an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of k=2.

This report shall not be reproduced except in full without the written approval of the Commonwealth of Virginia Metrology Lab and may not be used to claim product endorsement by this laboratory, NVLAP, NIST, or any agencies of the Federal Government.

This certificate is issued pursuant to Virginia Code § 3.2-5604. The calibration results relate only to the calibrated items identified above.

Reviewed By: [Signature]

[Signature]
 Authorized Signature

4/19/16
 Date



Virginia Department of Agriculture and Consumer Services
 Office of Weights & Measures • Metrology Lab
 600 North 5th St. Rm: 162 • Richmond, Va. 23219 • USA
 Telephone: (804)786-0479 • Fax: (804)371-0206

NVLAP Lab Code 105007-0



CALIBRATION CERTIFICATE

Tested for: TOLT SERVICE GROUP
 6900 NORTH PARK BLVD, SUITE C
 CHARLOTTE, NC 28216

Date Received: June 9, 2016
 Date of Test: June 15, 2016
 Test Number: VA-16-10210 T
 Tested By: WS/WHL
 Procedure: NISTIR 6969 SOP #8

Description of items calibrated: STAINLESS STEEL WEIGHT KIT - #02ZS
 Class: NIST HB105-1 (Class F) Condition: GOOD
 MEETS NIST CLASS F SPECIFICATION

* The uncertainty represents an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of k=2.

Environmental condition taken during testing: Temp: 21°C Hum: 53%

Nominal Value	ID	As Found** (error) (mg)	As Left (error) (mg)	Tolerance (mg)	Uncertainty* (±mg)	Balance Used	ID of Standard Set Used
5 lb	A	93	93	230	13	PR5003	W5LB
5 lb	B	92	92	230	13	PR5003	W5LB
5 lb	C	95	95	230	13	PR5003	W5LB
5 lb	D	88	88	230	13	PR5003	W5LB
5 lb	E	103	103	230	13	PR5003	W5LB
1 lb	A	25.1	25.1	70	4.5	CC1200	W1LB
1 lb	B	27.4	27.4	70	4.5	CC1200	W1LB
1 lb	C	30.5	30.5	70	4.5	CC1200	W1LB
1 lb	D	14.0	14.0	70	4.5	CC1200	W1LB
1 lb	E	13.0	13.0	70	4.5	CC1200	W1LB
0.5 lb		14.6	14.6	45	2.5	CC1200	W0.5LB
0.2 lb		6.81	6.81	18	0.22	MSE225S	W0.2LB
0.2 lb	dot	8.34	8.34	18	0.22	MSE225S	W0.2LB
0.1 lb		3.41	3.41	9.1	0.13	MSE225S	W0.1LB
0.05 lb		1.97	1.97	4.5	0.13	MSE225S	W0.05LB
0.02 lb	dot	0.59	0.59	1.8	0.077	MSE225S	W0.02LB
0.02 lb	2 dots	0.46	0.46	1.8	0.077	MSE225S	W0.02LB
0.01 lb		0.71	0.71	1.5	0.077	MSE225S	W0.01LB
0.005 lb		0.43	0.43	1.2	0.077	MSE225S	W0.005LB
0.002 lb		0.17	0.17	0.87	0.066	MSE225S	W0.002LB
0.002 lb	dot	-0.03	-0.03	0.87	0.066	MSE225S	W0.002LB
0.001 lb		0.24	0.24	0.70	0.066	MSE225S	W0.001LB

**1 US pound equals 453.5924 grams.

The results of the calibration are metrological traceable to the SI. The SI of the Va lab is through the National Institute of Standards and Technology (NIST) through an unbroken chain of traceable standards and controlled processes.

* The uncertainty represents an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of k=2. This report shall not be reproduced except in full without the written approval of the Commonwealth of Virginia Metrology Lab and may not be used to claim product endorsement by this laboratory, NVLAP, NIST, or any agencies of the Federal Government. This certificate is issued pursuant to Virginia Code § 3.2-5604. The calibration results relate only to the calibrated items identified above.

Reviewed By: WHL

WHL
 Authorized Signature

6-16-16
 Date

United States Department of Commerce
National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Virginia

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon II	Volume Transfer, II
20 kg to 1 mg	10 gal to 5 gal
Mass Echelon III	Tuning Forks
25 kg to 1 mg	10 KHz to 1 KHz
2500 lb to 0.001 lb	
8 oz to 0.03125 oz	
Weight Carts	
6000 lb to 3000 lb	



2016 to 2017

A handwritten signature in black ink, appearing to read "Carol T. Hockert".

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2016-01-01 to 2017-12-31