



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY

Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company <i>Rocky Mountain Scale Works</i>	Email Address <i>Dylan@Scales911.com</i>	Application Date <i>12-20-16</i>	
Mailing Address <i>2935 Stockyard rd Suite M3</i>	City <i>Missoula</i>	State <i>MT</i>	Zip Code <i>59808</i>
Telephone Number <i>406-543-5181</i>	Cell Phone Number <i>406-239-0228</i>	Fax Number <i>406-543-5182</i>	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input checked="" type="checkbox"/> 1. Rail <input checked="" type="checkbox"/> 2. Truck <input checked="" type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: <i>200,000</i> <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
<i>e.g. 1001</i>	<i>e.g. John Doe</i>	<i>e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6</i>
	<i>Timothy Dylan Moses</i>	<i>1, 2, 3, 6, 7</i>

Continued on Page 2



List below all field standards (attach current calibration reports):

See attached Reports	

Additional Application Items (initial where appropriate):

Standardized Test Report	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:


Include my company information on your registered service company list for public contact.
 Yes No

I am Timothy Dylan Moses, and have authority to represent this company. By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.

Signature

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410



SECRETARY OF STATE NORTH DAKOTA



[Home](#) | [Business Records Search](#)

ROCKY MOUNTAIN SCALE WORKS

Trade Name

System ID: 33561000 **Phone:** (406) 543-5181
Status: Active
Original File Date: 02/05/2013 **Last Renewal Date:**

Nature of Business

SCALE SALES AND SERVICE

Owners

RATHER ENTERPRISES, INC.
6360 HWY 12 W
PO BOX 1056
LOLO, MT 59847-1056

[Return to Search Results](#)

[Contact Us](#)

[Disclaimer](#)

[Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.

North Dakota

nd.gov Official Portal for
North Dakota State GovernmentNorth Dakota
LEGENDARY

SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

RATHER ENTERPRISES, INC.

Corporation Details

System ID: 33560900 **Phone:** (800) 700-5048
Type: FOREIGN BUSINESS CORPORATION
Status: Active & Good Standing
Original File Date: 02/05/2013 **Effective Date:** 02/05/2013
State of Origin: Montana

Nature of Business

SCALE SALES AND SERVICE

Principal Office

6360 HWY 12 W PO BOX 1056 LOLO, MT 59847-1056

Registered Agent

JIM ROBERTS
1217 HIGHWAY 85 SW
FAIRFIELD, ND 58627-9433
Established Date: Jul 17, 2014

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2016](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)[Disclaimer](#) | [Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163

Company Name & Address: Date of Test: Test Number:

 Serial Number:

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received:
 Description and condition of artifacts received:

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
22.65	24.49	670.81	40.57	42.10

Conventional Mass Value: Assumed Density of Artifacts:

Nominal	Serial No.	As Found (g)	As Left (g)	Uncertainty ± (g)	NIST 105-1 Class F ± (g)	k factor
1000 lb	4N91	3.1	3.1	5.6	45	2.18
1000 lb	8	-5.8	-5.8	5.6	45	2.18
1000 lb	54EP	8.1	8.1	5.6	45	2.18
1000 lb	4N8W	1.1	1.1	5.6	45	2.18
1000 lb	4N8X	-0.9	-0.9	5.6	45	2.18
1000 lb	4N90	-13.9	-13.9	5.6	45	2.18
1000 lb	7	-6.9	-6.9	5.6	45	2.18
1000 lb	4N8Y	-21.9	-21.9	5.6	45	2.18
1000 lb	54ER	18.1	18.1	5.6	45	2.18

Standards and Procedures used for testing:
 The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7
 All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:
 The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:
 The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows *NISTIR 6969*, issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: *Dave Fraser*

Email: dafraser@mt.gov

This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.



Certificate of Mass Calibration

Montana Department of Labor & Industry Metrology Laboratory
2801 N Cooke St. Helena, Montana 59601
(406)449-2582 FAX (406)443-8163

Company Name & Address:	Date of Test:	Test Number:
Rocky Mountain Scale Works	4/26-27/2016	2016-044
6360 U.S. Hwy 12		
Lolo, MT 59847	Serial Number:	1000's
(406) 543-5181		

All results contained within this report only relate to the item(s) listed in this report. This calibration report must not be used to claim product endorsement by the State of Montana or any other government agency.

Date these weights were received: 4/26/2016
Description and condition of artifacts received: Cart was in good condition. 1000lb weights were rusty w/paint chips.

Environmental Conditions at Time of Test:

Temperature °C		Pressure mmHg	Relative Humidity %	
Start	End	Duration of Test	Start	End
22.65	24.49	670.81	40.57	42.10

Conventional Mass Value: Assumed Density of Artifacts: 7.2 g/cm³

Nominal	Serial No.	As Found (g)	As Left (g)	Uncertainty ± (g)	NIST 105-1 Class F ± (g)	k factor
1000 lb	54EK	12.1	12.1	5.6	45	2.18
1000 lb	54EM	11.1	11.1	5.6	45	2.18
1000 lb	54ES	20.1	20.1	5.6	45	2.18
1000 lb	54EN	15.1	15.1	5.6	45	2.18
1000 lb	54EO	7.1	7.1	5.6	45	2.18
1000 lb	54EL	9.1	9.1	5.6	45	2.18
1000 lb	4N8Z	4.1	4.1	5.6	45	2.18
4500 lb	Cart	29.3	29.3	12.0	567.0	2.18

Standards and Procedures used for testing:
The Standards used for this comparison are continuously monitored by a measurement control program for ensuring continued accuracy and traceability within the level of uncertainty reported. These standards were calibrated by a nationally accredited laboratory on 10/2009 (Reports on File) and are traceable to the SI. The test number listed above is traceable to National Standards through an unbroken chain of comparison each having stated uncertainties. This information is on file and available upon request.

Procedure Used: SOP-7
All procedures used in this laboratory are in accordance to National Institute of Standards and Technology Intermediate Report (NISTIR) 6969, issue February 2012, and the *Quality Assurance of Metrological Measurements*.

Traceability Statement:
The equipment in this report has been compared to the standards of the State of Montana. The States equipment complies with the specifications and tolerances listed in NIST 105-1 Class F tolerances. The standards of the State of Montana are traceable to the SI through the National Institute of Standards and Technology.

Uncertainty Statement:
The expanded uncertainty presented in this report is consistent with the 1993 *ISO Guide to Expression of Uncertainty in Measurement* and follows *NISTIR 6969*, issue February 2012, SOP-29. The reported uncertainty is calculated by combining the uncertainty of the standard used, with the uncertainty of the measurement process in a root sum square formula using a calculated k factor, for a confidence level of 95.45%.

State Metrologist: Dave Fraser
David Fraser

Email: dafrazer@mt.gov
This document shall not be reproduced except in full without prior written agreement given by the State Metrologist.

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Montana

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon III	Volume Transfer, II
30 kg to 1 mg	1500 gal to 5 gal
3000 lb to 0.001 lb	100 gal LPG to 25 gal LPG
8 oz to 0.03125 oz	
Weight Carts	
5000 lb to 2000 lb	



2016

A handwritten signature in black ink, appearing to read "Carol Hoekert".

Carol T. Hoekert, Chief
NIST Office of Weights and Measures

Effective Dates: 2016-01-01 to 2016-12-31