



Receipt Date: August 23, 2017  
Cal. Date: August 24, 2017  
Report Date: August 24, 2017

Report No.: 338134  
Set Serial No.: 11, 22  
Barcode: 201219

## Calibration Certificate

**SPECIALTY PRODUCTS**

1420 N 4TH ST  
FARGO, ND 58102-2733  
Contact: PHIL MCINTYRE  
Phone: 701-235-2996  
PO Number: NONE  
Procedure: NIST SOP 8  
Technician ID: 11

Item(s) Submitted: 2 - 5000 lb Weight Carts  
Manufacturer: Heinisch  
Weight Type: NA  
Equipment ID: None  
Condition: Good  
Temperature: 19.8 °C  
Pressure: 740.4 mmHg  
Relative Humidity: 49.5 %

| Nominal Value | Serial No. | CM Correction (g) |         | NIST HB105-8 Tolerance |         | k    | U (g) |
|---------------|------------|-------------------|---------|------------------------|---------|------|-------|
|               |            | As Found          | As Left | As Found               | As Left |      |       |
| 5000 lb       | 11         | -90.              | -90.    | Meets                  | Meets   | 2.08 | 60.   |
| 5000 lb       | 22         | -300.             | -300.   | Meets                  | Meets   | 2.08 | 60.   |

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm<sup>3</sup> density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-8 (2003). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Pete Whebbe

*Peter J. Whebbe*  
Metrologist

Reviewed by:

Benjamin FitzPatrick

*Benjamin FitzPatrick*

Deputy Director



Receipt Date: August 23, 2017  
Cal. Date: August 23, 2017  
Report Date: August 23, 2017

Report No.: 338135  
Set Serial No.: 114, 119, 127, 128  
Barcode: 201216

## Calibration Certificate

**SPECIALTY PRODUCTS**

1420 N 4TH ST  
FARGO, ND 58102-2733  
Contact: PHIL MCINTYRE  
Phone: 701-235-2996  
PO Number: NONE  
Procedure: NIST SOP 8  
Technician ID: 11

Item(s) Submitted: Cast Cube Weights  
Manufacturer: Rice Lake  
Weight Type: II  
Equipment ID: None  
Condition: Good  
Temperature: 20.1 °C  
Pressure: 738.2 mmHg  
Relative Humidity: 49.8 %

| Nominal Value |    | Serial No. | CM Correction (g) |         | NIST HB105-1 Class |         | k    | U (g) |
|---------------|----|------------|-------------------|---------|--------------------|---------|------|-------|
|               |    |            | As Found          | As Left | As Found           | As Left |      |       |
| 2500          | lb | 114        | 105.3             | 74.3    | *                  | F       | 2.05 | 8.0   |
| 2500          | lb | 119        | 80.3              | 80.3    | F                  | F       | 2.05 | 8.0   |
| 2500          | lb | 127        | 56.3              | 56.3    | F                  | F       | 2.05 | 8.0   |
| 2500          | lb | 128        | 115.3             | 4.3     | *                  | F       | 2.05 | 8.0   |

\* Weight(s) as found exceed NIST HB105-1 Class F tolerance.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm<sup>3</sup> density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Pete Whebbe

*Pete Whebbe*  
Metrologist

Reviewed by:

Heidi Jones  
*Heidi Jones*  
Laboratory Administrator



Receipt Date: August 23, 2017  
Cal. Date: August 24, 2017  
Report Date: August 24, 2017

Report No.: 338136  
Set Serial No.: None  
Barcode: 201217

## Calibration Certificate

SPECIALTY PRODUCTS  
1420 N 4TH ST  
FARGO, ND 58102-2733

Contact: PHIL MCINTYRE  
Phone: 701-235-2996  
PO Number: NONE  
Procedure: NIST SOP 8  
Technician ID: 11

Item(s) Submitted: Cast Cube Weights  
Manufacturer: Howe  
Weight Type: II  
Equipment ID: None  
Condition: Good  
Temperature: 20.1 °C  
Pressure: 740.7 mmHg  
Relative Humidity: 49.2 %

| Nominal Value | Serial No. | CM Correction (g) |         | NIST HB105-1 Class |         | k    | U (g) |
|---------------|------------|-------------------|---------|--------------------|---------|------|-------|
|               |            | As Found          | As Left | As Found           | As Left |      |       |
| 500 lb        |            | 6.52              | 6.52    | F                  | F       | 2.01 | 0.90  |
| 500 lb        |            | 4.32              | 4.32    | F                  | F       | 2.01 | 0.90  |
| 500 lb        |            | -2.28             | -2.28   | F                  | F       | 2.01 | 0.90  |
| 500 lb        |            | -4.48             | -4.48   | F                  | F       | 2.01 | 0.90  |

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm<sup>3</sup> density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Pete Whebbe

*Pete Whebbe*  
Metrologist

Reviewed by:

Heidi Jones

*Heidi Jones*  
Laboratory Administrator



Receipt Date: August 23, 2017  
Cal. Date: August 23, 2017  
Report Date: August 23, 2017

Report No.: 338137  
Set Serial No.: None  
Barcode: 201218

## Calibration Certificate

SPECIALTY PRODUCTS  
1420 N 4TH ST  
FARGO, ND 58102-2733

Contact: PHIL MCINTYRE  
Phone: 701-235-2996  
PO Number: NONE  
Procedure: NIST SOP 8  
Technician ID: 11

Item(s) Submitted: Cast Hand Weights  
Manufacturer: Rice Lake  
Weight Type: II  
Equipment ID: None  
Condition: Good  
Temperature: 19.7 °C  
Pressure: 738.1 mmHg  
Relative Humidity: 49.1 %

| Nominal Value | Serial No. | CM Correction (mg) |         | NIST HB105-1 Class |         | k    | U (mg) |
|---------------|------------|--------------------|---------|--------------------|---------|------|--------|
|               |            | As Found           | As Left | As Found           | As Left |      |        |
| 50 lb         |            | 294                | 294     | F                  | F       | 2.01 | 56     |
| 50 lb         |            | 644                | 644     | F                  | F       | 2.01 | 56     |

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm<sup>3</sup> density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Pete Whebbe

Metrologist

Reviewed by:

Heidi Jones

Laboratory Administrator



**DEPARTMENT OF COMMERCE**  
WEIGHTS & MEASURES DIVISION



14305 Southcross Drive #150  
Burnsville, MN 55306-7008  
mn.gov/commerce/  
651.539.1555 FAX 952.435.4040  
An equal opportunity employer

Receipt Date: August 23, 2017  
Cal. Date: August 24, 2017  
Report Date: August 24, 2017

Report No.: 338141  
Set Serial No.: NONE  
Barcode: 203255

## Calibration Certificate

SPECIALTY PRODUCTS  
1420 N 4TH ST  
FARGO, ND 58102

Contact: PHIL MCINTYRE  
Phone: 701-235-2996  
PO Number: NONE  
Procedure: NIST SOP 8  
Technician ID: 09

Item(s) Submitted: 30 lb kit w/ decimals & fractions  
Manufacturer: RICE LAKE  
Weight Type: I & II  
Equipment ID: None  
Condition: Good  
Temperature: 19.2 °C  
Pressure: 740.4 mmHg  
Relative Humidity: 52.2 %

| Nominal Value | Serial No. | CM Correction (mg) |         | NIST HB105-1 Class |         | k    | U (mg) |
|---------------|------------|--------------------|---------|--------------------|---------|------|--------|
|               |            | As Found           | As Left | As Found           | As Left |      |        |
| 5 lb          |            | 19.2               | 19.2    | F                  | F       | 2.01 | 6.0    |
| 5 lb          |            | 19.2               | 19.2    | F                  | F       | 2.01 | 6.0    |
| 5 lb          |            | 31.2               | 31.2    | F                  | F       | 2.01 | 6.0    |
| 5 lb          |            | 95.2               | 95.2    | F                  | F       | 2.01 | 6.0    |
| 5 lb          |            | 38.2               | 38.2    | F                  | F       | 2.01 | 6.0    |
| 1 lb          |            | 35.0               | 35.0    | F                  | F       | 2.01 | 1.6    |
| 1 lb          |            | 33.7               | 33.7    | F                  | F       | 2.01 | 1.6    |
| 1 lb          |            | 23.9               | 23.9    | F                  | F       | 2.01 | 1.6    |
| 1 lb          |            | 26.6               | 26.6    | F                  | F       | 2.01 | 1.6    |
| 1 lb          |            | 23.3               | 23.3    | F                  | F       | 2.01 | 1.6    |
| 0.2 lb        |            | 1.93               | 1.93    | F                  | F       | 2.02 | 0.22   |
| 0.2 lb        |            | 1.37               | 1.37    | F                  | F       | 2.02 | 0.22   |
| 0.1 lb        |            | 1.42               | 1.42    | F                  | F       | 2.02 | 0.14   |
| 0.05 lb       |            | 1.84               | 1.84    | F                  | F       | 2.02 | 0.11   |
| 0.02 lb       |            | 0.592              | 0.592   | F                  | F       | 2.02 | 0.066  |
| 0.02 lb       |            | 0.980              | 0.980   | F                  | F       | 2.02 | 0.066  |
| 0.01 lb       |            | 0.806              | 0.806   | F                  | F       | 2.02 | 0.052  |
| 0.005 lb      |            | 1.461              | 0.551   | *                  | F       | 2.02 | 0.073  |
| 0.002 lb      |            | 0.877              | 0.371   | *                  | F       | 2.02 | 0.047  |
| 0.002 lb      |            | 0.868              | 0.271   | *                  | F       | 2.02 | 0.047  |
| 0.001 lb      |            | 0.641              | 0.197   | F                  | F       | 2.02 | 0.041  |

\* Weight(s) as found exceed NIST HB105-1 Class F tolerance.



Receipt Date: August 23, 2017  
Cal. Date: August 24, 2017  
Report Date: August 24, 2017

Report No.: 338141  
Set Serial No.: NONE  
Barcode: 203255

Continued,

## Calibration Certificate

**SPECIALTY PRODUCTS**

1420 N 4TH ST  
FARGO, ND 58102  
Contact: PHIL MCINTYRE  
Phone: 701-235-2996  
PO Number: NONE  
Procedure: NIST SOP 8  
Technician ID: 09

Item(s) Submitted: 30 lb kit w/ decimals & fractions  
Manufacturer: RICE LAKE  
Weight Type: I & II  
Equipment ID: None  
Condition: Good  
Temperature: 19.2 °C  
Pressure: 740.4 mmHg  
Relative Humidity: 52.2 %

| Nominal Value | Serial No. | CM Correction (mg) |         | NIST HB105-1 Class |         | k    | U (mg) |
|---------------|------------|--------------------|---------|--------------------|---------|------|--------|
|               |            | As Found           | As Left | As Found           | As Left |      |        |
| 8 oz          |            | 8.3                | 8.3     | F                  | F       | 2.01 | 1.3    |
| 4 oz          |            | 10.03              | 10.03   | F                  | F       | 2.00 | 0.22   |
| 2 oz          |            | 2.67               | 2.67    | F                  | F       | 2.00 | 0.14   |
| 1 oz          |            | 0.34               | 0.34    | F                  | F       | 2.00 | 0.11   |
| 1/2 oz        |            | -1.412             | -1.412  | F                  | F       | 2.00 | 0.092  |
| 1/4 oz        |            | 0.800              | 0.800   | F                  | F       | 2.00 | 0.056  |
| 1/8 oz        |            | 0.288              | 0.288   | F                  | F       | 2.00 | 0.046  |
| 1/16 oz       |            | 0.642              | 0.642   | F                  | F       | 2.00 | 0.070  |
| 1/32 oz       |            | -0.132             | -0.132  | F                  | F       | 2.00 | 0.044  |
| 1/32 oz       |            | 0.220              | 0.220   | F                  | F       | 2.00 | 0.044  |

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm<sup>3</sup> density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Heidi Jones  
*Heidi Jones*  
Laboratory Administrator

Reviewed by:  
Pete Whebbe  
*Pete Whebbe*  
Metrologist



Receipt Date: August 23, 2017  
Cal. Date: August 24, 2017  
Report Date: August 24, 2017

Report No.: 338139  
Set Serial No.: NONE  
Barcode: 201209

## Calibration Certificate

**SPECIALTY PRODUCTS**

1420 N 4TH ST  
FARGO, ND 58102  
Contact: PHIL MCINTYRE  
Phone: 701-235-2996  
PO Number: NONE  
Procedure: NIST SOP 8  
Technician ID: 09

Item(s) Submitted: Metric weight set  
Manufacturer: Rice Lake  
Weight Type: I & II  
Equipment ID: None  
Condition: Good  
Temperature: 18.9 °C  
Pressure: 740.4 mmHg  
Relative Humidity: 47.4 %

| Nominal Value | Serial No. | CM Correction (mg) |         | NIST HB105-1 Class |         | k    | U (mg) |
|---------------|------------|--------------------|---------|--------------------|---------|------|--------|
|               |            | As Found           | As Left | As Found           | As Left |      |        |
| 2000 g        |            | 40.1               | 40.1    | F                  | F       | 2.02 | 5.7    |
| 1000 g        |            | 30.3               | 30.3    | F                  | F       | 2.02 | 2.5    |
| 500 g         |            | 28.2               | 28.2    | F                  | F       | 2.02 | 2.0    |
| 200 g         |            | 8.63               | 8.63    | F                  | F       | 2.02 | 0.55   |
| 200 g         |            | 7.52               | 7.52    | F                  | F       | 2.02 | 0.55   |
| 100 g         |            | 5.58               | 5.58    | F                  | F       | 2.02 | 0.25   |
| 50 g          |            | 2.06               | 2.06    | F                  | F       | 2.03 | 0.16   |
| 20 g          |            | 1.93               | 1.93    | F                  | F       | 2.02 | 0.11   |
| 20 g          |            | 2.10               | 2.10    | F                  | F       | 2.02 | 0.11   |
| 10 g          |            | 0.691              | 0.691   | F                  | F       | 2.02 | 0.072  |
| 5 g           |            | 0.586              | 0.586   | F                  | F       | 2.03 | 0.054  |
| 2 g           |            | 0.351              | 0.351   | F                  | F       | 2.02 | 0.048  |
| 2 g           |            | 0.468              | 0.468   | F                  | F       | 2.02 | 0.048  |
| 1 g           |            | 0.503              | 0.503   | F                  | F       | 2.03 | 0.039  |

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm<sup>3</sup> density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Heidi Jones  
*Heidi Jones*  
Laboratory Administrator

Reviewed by:  
Pete Whebbe  
*Pete Whebbe*  
Metrologist



# United States Department of Commerce

## National Institute of Standards and Technology

Certificate of Metrological Traceability For:

# Minnesota

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.


The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

### Scope

|                              |  |  |
|------------------------------|--|--|
| <b>Mass Echelon I</b>        |  |  |
| 10 kg to 1 mg                |  |  |
| <b>Mass Echelon II</b>       |  |  |
| 50 kg to 1 mg                |  |  |
| 1000 lb to 0.001 lb          |  |  |
| 4 oz to 0.03125 oz           |  |  |
| <b>Mass Echelon III</b>      |  |  |
| 50 kg to 1 mg                |  |  |
| 5000 lb to 0.001 lb          |  |  |
| 4 oz to 0.03125 oz           |  |  |
| <b>Volume Gravimetric, I</b> |  |  |
| 20 L to 10 mL                |  |  |
| 100 gal to 0.25 qt           |  |  |
| <b>Volume Transfer, II</b>   |  |  |
| 1500 gal to 5 gal            |  |  |
| 100 gal to 25 gal LPG        |  |  |
| <b>Weight Carts</b>          |  |  |
| 10 000 lb to 2000 lb         |  |  |
| <b>Wheel Load Weighers</b>   |  |  |
| 20 000 lb to 2000 lb         |  |  |
| <b>Railroad Test Cars</b>    |  |  |
| 110 000 lb to 80 000 lb      |  |  |



2017

  
Georgia L. Harris, Acting Chief  
NIST Office of Weights and Measures

Effective Dates: 2017-01-01 to 2017-12-31

Amended: 2016-12-31

Scope modified for 2017.