



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY
 Public Service Commission
 SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company	Email Address	Application Date	
Bison Food Equipment	bisonfoodequipment ^{hotmail.} @com	12/15/17	
Mailing Address	City	State	Zip Code
17409 Beseau Lake Rd	Lake Park	mn	56554
Telephone Number	Cell Phone Number	Fax Number	
218-238-6363	701-261-2102	218-238-	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: <u>400</u> <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
e.g. 1001	e.g. John Doe	e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6
1462	Dale Bohner	6, 7

Application for Registration as a Registered Service Company
Page 2


List below all field standards (attach current calibration reports):

Additional Application Items (initial where appropriate):

Standardized Test Report	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed <input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously


Public Company Listing:

Include my company information on your registered service company list for public contact. <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No

I am <u>Dale G. Boner</u> , and have authority to represent this company. By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete. <div style="text-align: right;"> _____ Signature</div>
--

Send Completed Application and Related Documents To:

Public Service Commission
600 E Boulevard Ave Dept 408
Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410



SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

BISON FOOD EQUIPMENT SERVICE, INC.

Corporation Details

System ID: 19106600 **Phone:** (218) 238-6363
Type: FOREIGN BUSINESS CORPORATION
Status: Active & Good Standing
Original File Date: 04/03/2003 **Effective Date:** 04/03/2003
State of Origin: Minnesota

Nature of Business

FOOD EQUIPMENT SERVICE AND SALES

Principal Office

17409 BESEAU LAKE RD LAKE PARK, MN 56554-9645

Registered Agent

MICHAEL SCULLY
6620 SAPHIRE LANE
BISMARCK, ND 58501-
Established Date: May 18, 2005


Generate an Annual Report To File


To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2017](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)[Disclaimer](#) | [Privacy Policy](#)

 We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

 Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.

Receipt Date: March 23, 2017
Cal. Date: March 23, 2017
Report Date: March 23, 2017

Report No.: 337469
Set Serial No.: None
Barcode: 202268

Calibration Certificate

BISON FOOD EQUIPMENT SERVICE INC.
17409 BESEAU LAKE ROAD
LAKE PARK, MN 56554
Contact: Dale Bohner
Phone: 701-261-2102
PO Number: NONE
Procedure: NIST SOP 8
Technician ID: 11

Item(s) Submitted: Cast Handle Weights
Manufacturer: Troemner
Weight Type: II
Equipment ID: None
Condition: Fair*
Temperature: 21.1 °C
Pressure: 738.8 mmHg
Relative Humidity: 49.8 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
25 lb		609	609	F	F	2.01	52
25 lb		129	129	F	F	2.01	52
25 lb		769	769	F	F	2.01	52
25 lb		869	869	F	F	2.01	52
50 lb		2140	2140	F	F	2.01	56
50 lb		580	580	F	F	2.01	56
50 lb		1790	1790	F	F	2.01	56
50 lb		1500	1500	F	F	2.01	56
50 lb		650	650	F	F	2.01	56
50 lb		2170	2170	F	F	2.01	56

* Weights must be cleaned and repainted prior to next calibration.

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³ at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Pete Whebbe

Peter J. Whebbe
Metrologist

Reviewed by:

Mark Nicollet

Mark Nicollet
Quality Manager

Receipt Date: March 23, 2017
 Cal. Date: March 23, 2017
 Report Date: March 23, 2017

Report No.: 337468
 Set Serial No.: 14580/1-6
 Barcode: 202472

Calibration Certificate

BISON FOOD EQUIPMENT SERVICE INC.
 17409 BESEAU LAKE ROAD
 LAKE PARK, MN 56554

Contact: Dale Bohner
 Phone: 701-261-2102
 PO Number: NONE
 Procedure: NIST SOP 8
 Technician ID: 11

Item(s) Submitted: Satin Stainless Cylinder Weights
 Manufacturer: Troemner
 Weight Type: II
 Equipment ID: None
 Condition: Good
 Temperature: 18.8 °C
 Pressure: 738.1 mmHg
 Relative Humidity: 53.1 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
5 lb	1	119.4	119.4	F	F	2.01	6.0
5 lb	2	121.4	121.4	F	F	2.01	6.0
5 lb	3	105.4	105.4	F	F	2.01	6.0
5 lb	4	136.4	136.4	F	F	2.01	6.0
5 lb	5	111.4	111.4	F	F	2.01	6.0
5 lb	6	100.4	100.4	F	F	2.01	6.0

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³ at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (k) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Pete Whebbe



Metrologist

Reviewed by:
 Mark Nicollet

Quality Manager

Receipt Date: March 23, 2017
Cal. Date: March 23, 2017
Report Date: March 23, 2017

Report No.: 337467
Set Serial No.: None
Barcode: 201022

Calibration Certificate

BISON FOOD EQUIPMENT SERVICE INC.
17409 BESEAU LAKE ROAD
LAKE PARK, MN 56554
Contact: Dale Bohner
Phone: 701-261-2102
PO Number: NONE
Procedure: NIST SOP 8
Technician ID: 11

Item(s) Submitted: 30 lb Weight Kit
Manufacturer: Rice Lake
Weight Type: I & II
Equipment ID: None
Condition: Good
Temperature: 18.7 °C
Pressure: 737.9 mmHg
Relative Humidity: 53.1 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
5 lb		58.4	58.4	F	F	2.01	6.0
5 lb		67.4	67.4	F	F	2.01	6.0
5 lb		49.4	49.4	F	F	2.01	6.0
5 lb		45.4	45.4	F	F	2.01	6.0
5 lb		116.4	116.4	F	F	2.01	6.0
1 lb		20.9	20.9	F	F	2.01	1.6
1 lb		28.1	28.1	F	F	2.01	1.6
1 lb		20.3	20.3	F	F	2.01	1.6
1 lb		20.4	20.4	F	F	2.01	1.6
1 lb		27.3	27.3	F	F	2.01	1.6
8 oz		19.9	19.9	F	F	2.01	1.3
0.2 lb		6.39	6.39	F	F	2.02	0.22
0.2 lb		4.78	4.78	F	F	2.02	0.22
0.1 lb		2.02	2.02	F	F	2.02	0.14
0.05 lb		0.39	0.39	F	F	2.02	0.11
0.02 lb		0.618	0.618	F	F	2.02	0.066
0.02 lb		0.765	0.765	F	F	2.02	0.066
0.01 lb		0.156	0.156	F	F	2.02	0.052

Receipt Date: March 23, 2017
Cal. Date: March 23, 2017
Report Date: March 23, 2017

Report No.: Continued,
Set Serial No.: 337467
Barcode: None
201022

Calibration Certificate

BISON FOOD EQUIPMENT SERVICE INC.
17409 BESEAU LAKE ROAD
LAKE PARK, MN 56554

Contact: Dale Bohner
Phone: 701-261-2102
PO Number: NONE
Procedure: NIST SOP 8
Technician ID: 11

Item(s) Submitted: 30 lb Weight Kit
Manufacturer: Rice Lake
Weight Type: I & II
Equipment ID: None
Condition: Good
Temperature: 18.7 °C
Pressure: 737.9 mmHg
Relative Humidity: 53.1 %

Nominal Value	Serial No.	CM Correction (mg)		NIST HB105-1 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
0.005 lb		0.358	0.358	F	F	2.02	0.073
0.002 lb		0.246	0.246	F	F	2.02	0.047
0.002 lb		0.070	0.070	F	F	2.02	0.047
0.001 lb		-0.102	-0.102	F	F	2.02	0.041

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³ at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to NIST Handbook 105-1 (1990). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to items identified in this report only.

Pete Whebbe

Metrologist



Reviewed by:

Mark Nicollet

Quality Manager

Receipt Date: May 22, 2017
Cal. Date: May 30, 2017
Report Date: May 30, 2017

Report No.: 337761
Set Serial No.: X7670
Barcode: 019695

Calibration Certificate

BISON FOOD EQUIPMENT SERVICE
17409 BESEAU LAKE ROAD
LAKE PARK, MN 56554
Contact: Dale Bohner
Phone: 701-261-2102
PO Number: None
Procedure: NIST SOP 8
Technician ID: 19

Item(s) Submitted: AVDP Weight Kit - ASTM 4, 5, 6
Manufacturer: Troemner
Weight Type: I & II
Equipment ID: None
Condition: Good
Temperature: 19.6 °C
Pressure: 735.9 mmHg
Relative Humidity: 49.0 %

Nominal Value	Serial No.	CM Correction (mg)		ASTM E617 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
10 lb	1	119	119	5	5	2.01	12
10 lb	2	159	159	5	5	2.01	12
10 lb	3	147	147	5	5	2.01	12
5 lb	1	-24.6	-24.6	4	4	2.01	6.0
5 lb	2	-31.6	-31.6	4	4	2.01	6.0
5 lb	3	-21.6	-21.6	4	4	2.01	6.0
1 lb	1	0.9	0.9	4	4	2.01	1.6
1 lb	2	0.3	0.3	4	4	2.01	1.6
1 lb	3	-0.7	-0.7	4	4	2.01	1.6
1 lb	4	0.6	0.6	4	4	2.01	1.6
1 lb	5	1.2	1.2	4	4	2.01	1.6
0.5 lb		0.8	0.8	4	4	2.01	1.3
0.2 lb		2.66	2.66	5	5	2.02	0.22
0.2 lb		1.02	1.02	4	4	2.02	0.22
0.1 lb		1.08	1.08	5	5	2.02	0.14
0.05 lb		0.26	0.26	4	4	2.02	0.11
0.02 lb		0.698	0.698	5	5	2.02	0.066
0.02 lb		0.955	0.955	5	5	2.02	0.066
0.01 lb		0.083	0.083	4	4	2.02	0.052

Receipt Date: May 22, 2017
Cal. Date: May 30, 2017
Report Date: May 30, 2017

Report No.: 337761
Set Serial No.: X7670
Barcode: 019695

Continued,

Calibration Certificate

BISON FOOD EQUIPMENT SERVICE
17409 BESEAU LAKE ROAD
LAKE PARK, MN 56554

Contact: Dale Bohner
Phone: 701-261-2102
PO Number: None
Procedure: NIST SOP 8
Technician ID: 19

Item(s) Submitted: AVDP Weight Kit - ASTM 4, 5, 6
Manufacturer: Troemner
Weight Type: I & II
Equipment ID: None
Condition: Good
Temperature: 19.6 °C
Pressure: 735.9 mmHg
Relative Humidity: 49.0 %

Nominal Value	Serial No.	CM Correction (mg)		ASTM E617 Class		k	U (mg)
		As Found	As Left	As Found	As Left		
0.005 lb		0.367	0.367	5	5	2.02	0.073
0.002 lb		0.737	0.737	6	6	2.02	0.047
0.002 lb		0.647	0.647	6	6	2.02	0.047
0.001 lb		0.378	0.378	6	6	2.02	0.041


The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³ at 20 °C. The items listed above have been calibrated using the Standards of the State of Minnesota which are currently in control. These standards are traceable to the SI through NIST. Calibration processes were monitored and found to be in control. All of the tolerances and specifications were evaluated according to ASTM E617 (2013). Uncertainty calculations contain the components in NIST SOP 8 and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Results apply to

Erik Alfvín


Metrologist

Reviewed by:

Pete Whebbe


Metrologist

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Minnesota




This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

2017

Mass Echelon I 10 kg to 1 mg	Mass Echelon III 50 kg to 1 mg 5000 lb to 0.001 lb 4 oz to 0.03125 oz	Volume Gravimetric, I 20 L to 10 mL 100 gal to 0.25 qt
Mass Echelon II 50 kg to 1 mg 1000 lb to 0.001 lb 4 oz to 0.03125 oz	Weight Carts 10 000 lb to 2000 lb 20 000 lb to 2000 lb	Volume Transfer, II 1500 gal to 5 gal 100 gal to 25 gal LPG
	Wheel Load Weighers 20 000 lb to 2000 lb	
	Railroad Test Cars 110 000 lb to 80 000 lb	


Georgia L. Harris, Acting Chief
NIST Office of Weights and Measures

Effective Dates: 2017-01-01 to 2017-12-31

Amended: 2016-12-31
Scope modified for 2017.