



**APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY**  
Public Service Commission  
SFN 51277 (2/2014)



**TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED**

<b>Name of Company</b>	<b>Email Address</b>	<b>Application Date</b>	
Pomeroy IT Solutions Sales Co. Inc.	WandM@pomeroy.com	12/18/17	
<b>Mailing Address</b>	<b>City</b>	<b>State</b>	<b>Zip Code</b>
13534 Plaza Rd. Extension, #102	Charlotte	NC	28215
<b>Telephone Number</b>	<b>Cell Phone Number</b>	<b>Fax Number</b>	
	704-650-4794		

**Select below all device types your company will certify:**

<b>Scales (include maximum capacity, if applicable)</b>	<b>Liquid (include maximum flow rate, if applicable)</b>
<input type="checkbox"/> 1. Rail	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute)
<input type="checkbox"/> 2. Truck	<input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater)
<input type="checkbox"/> 3. Livestock	<input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____
<input type="checkbox"/> 4. Hopper: Max. Capacity: _____	<input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____
<input type="checkbox"/> 5. Belt	<input type="checkbox"/> 5. LPG
<input type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: _____	<input type="checkbox"/> 6. Stationary LPG
<input checked="" type="checkbox"/> 7. 30 lbs. or less	<input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____
<input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified)	<input type="checkbox"/> 8. Chemical
<input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 9. Anhydrous
	<input type="checkbox"/> 10. Loading Rack
	<input type="checkbox"/> 11. Other: Please List:

**List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):**

<b>Permit No.</b>	<b>Employee</b>	<b>Device Types Registered to Certify (list using device type numbers from above)</b>
<i>e.g. 1001</i>	<i>e.g. John Doe</i>	<i>e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6</i>
1790	Brandon Elsenpeter	Scales 7
1791	Eric Gault	Scales 7

Continued on Page 2

Application for Registration as a Registered Service Company  
Page 2

List below all field standards (attach current calibration reports):

attached	

Additional Application Items (initial where appropriate):

Standardized Test Report	<input checked="" type="checkbox"/> Copy enclosed
	<input type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input type="checkbox"/> Copy enclosed
	<input checked="" type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input type="checkbox"/> Copy enclosed
	<input checked="" type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:

Include my company information on your registered service company list for public contact.  
 Yes     No


I am Cass Utter, and have authority to represent this company. By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.

  
 \_\_\_\_\_  
 Signature

Send Completed Application and Related Documents To:

Public Service Commission  
 600 E Boulevard Ave Dept 408  
 Bismarck ND 58505-0480  
 Telephone: (701) 328-2400  
 Fax: (701) 328-2410





# SECRETARY OF STATE NORTH DAKOTA

[Home](#) | [Business Records Search](#)

## POMEROY IT SOLUTIONS SALES COMPANY, INC.

### Corporation Details

**System ID:** 26645300 **Phone:** (310) 712-1850  
**Type:** FOREIGN BUSINESS CORPORATION  
**Status:** Active & Good Standing  
**Original File Date:** 02/04/2010 **Effective Date:** 02/04/2010  
**State of Origin:** Delaware

### Nature of Business

TECHNOLOGY SERVICES

### Principal Office

1020 PETERSBURG RD HEBRON, KY 41048-8222

### Registered Agent

**REGISTERED AGENT SOLUTIONS, INC.**  
1709 N 19TH ST STE 3  
BISMARCK, ND 58501-2121  
Established Date: Apr 12, 2016

### Generate an Annual Report To File


To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.


[2017](#) [2018](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)

[Disclaimer](#) | [Privacy Policy](#)

 We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

 Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.

# CALIBRATION CERTIFICATE

## Virginia Department of Agriculture and Consumer Services



Office of Weights & Measures • Metrology Lab  
600 North 5th St. Rm: 162 • Richmond, Va. 23219 • USA  
Telephone: (804)786-0479 • Fax: (804)371-0206

NVLAP Lab Code 105007-0

Tested for: POMEROY  
13534 PLAZA RD EXTENSION, STE 102  
CHARLOTTE, NC 28215

Date Received: November 17, 2017  
Date of Test: November 28, 2017  
Test Number: VA-18-11011 H  
Tested By: William Scott  
Procedure: NISTIR 6969 SOP #8

Description of items calibrated: STAINLESS STEEL WEIGHT KIT #37902  
Class: NIST HB105-1 (Class F) Condition: GOOD

MEETS CLASS F SPECIFICATION

The Standards described below have been compared with the Standards of the Commonwealth of Virginia.

Environmental condition taken during testing: Temperature: 20.8°C Humidity: 38.2%

Nominal Value	ID	As Found** (error) (mg)	As Left (error) (mg)	Tolerance (mg)	Uncertainty* (±mg)	Balance Used	ID of Standard Set Used
10 lb		192	192	450	58	PR5003	W10LB
10 lb	dot	211	211	450	58	PR5003	W10LB
5 lb		105	105	230	31	PR5003	W5LB
2 lb		56.5	56.5	91	12	CC1200	W2LB
2 lb	dot	47.0	47.0	91	12	CC1200	W2LB
1 lb		33.3	33.3	70	8.7	CC1200	W1LB

\*\*1 US pound equals 453.5924 grams.

The results of the calibration are metrological traceable to the SI. The SI of the Va lab is through the National Institute of Standards and Technology (NIST) through an unbroken chain of traceable standards and controlled processes.

\* The uncertainty represents an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of  $k=2$ .

This report shall not be reproduced except in full without the written approval of the Commonwealth of Virginia Metrology Lab and may not be used to claim product endorsement by this laboratory, NVLAP, NIST, or any agencies of the Federal Government.

This certificate is issued pursuant to Virginia Code § 3.2-5604. The calibration results relate only to the calibrated items identified above.

Reviewed By W.S.

William Scott  
Authorized Signature

11/30/2017  
Date

# CALIBRATION CERTIFICATE



## Virginia Department of Agriculture and Consumer Services

Office of Weights & Measures • Metrology Lab  
600 North 5th St. Rm: 162 • Richmond, Va. 23219 • USA  
Telephone: (804)786-0479 • Fax: (804)371-0206

NVLAP Lab Code 105007-0

Tested for: **POMEROY**  
13534 PLAZA RD EXTENSION. STE 2  
CHARLOTTE, NC 28215

Date Received: November 17, 2017  
Date of Test: November 28, 2017  
Test Number: VA-18-11011 U  
Tested By: William Scott  
Procedure: NISTIR 6969 SOP #8

Description of items calibrated: **STAINLESS STEEL WEIGHT KIT - #27439**  
Class: **NIST HB105-1 (Class F)** Condition: **GOOD**

### MEETS NIST CLASS F SPECIFICATION

The Standards described below have been compared with the Standards of the Commonwealth of Virginia.

Environmental condition taken during testing: Temperature: 20.8°C Humidity: 38.2%

Nominal Value	ID	As Found** (error) (mg)	As Left (error) (mg)	Tolerance (mg)	Uncertainty* (±mg)	Balance Used	ID of Standard Set Used
5 lb	1	100	100	230	31	PR5003	W5LB
5 lb	2	81	81	230	31	PR5003	W5LB
5 lb	3	99	99	230	31	PR5003	W5LB
5 lb	4	91	91	230	31	PR5003	W5LB
5 lb	5	123	123	230	31	PR5003	W5LB
1 lb	1	21.5	21.5	70	8.7	CC1200	W1LB
1 lb	2	31.7	31.7	70	8.7	CC1200	W1LB
1 lb	3	18.2	18.2	70	8.7	CC1200	W1LB
1 lb	4	21.7	21.7	70	8.7	CC1200	W1LB
1 lb	5	13.1	13.1	70	8.7	CC1200	W1LB
0.5 lb		16.2	16.2	45	6.0	CC1200	W0.5LB

\*\*1 US pound equals 453.5924 grams.

The results of the calibration are metrological traceable to the SI. The SI of the Va lab is through the National Institute of Standards and Technology (NIST) through an unbroken chain of traceable standards and controlled processes.

\* The uncertainty represents an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of  $k=2$ .

This report shall not be reproduced except in full without the written approval of the Commonwealth of Virginia Metrology Lab and may not be used to claim product endorsement by this laboratory, NVLAP, NIST, or any agencies of the Federal Government.

This certificate is issued pursuant to Virginia Code § 3.2-5604. The calibration results relate only to the calibrated items identified above.

Reviewed By: WSS

William Scott  
Authorized Signature

11/30/2017  
Date

# United States Department of Commerce National Institute of Standards and Technology

## Certificate of Metrological Traceability For:

# Virginia



This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

2016 to 2017

### Scope

<b>Mass Echelon II</b> 20 kg to 1 mg	<b>Volume Transfer, II</b> 10 gal to 5 gal
<b>Mass Echelon III</b> 25 kg to 1 mg 2500 lb to 0.001 lb 8 oz to 0.03125 oz	<b>Tuning Forks</b> 10 KHz to 1 KHz
<b>Weight Carts</b> 6000 lb to 3000 lb	

A handwritten signature in black ink, appearing to read "Carol T. Hockett".

Carol T. Hockett, Chief  
NIST Office of Weights and Measures

Effective Dates: 2016-01-01 to 2017-12-31