



APPLICATION FOR REGISTRATION AS A REGISTERED SERVICE COMPANY

Public Service Commission
SFN 51277 (2/2014)



TYPE OR PRINT - AN INCOMPLETE OR ILLEGIBLE APPLICATION WILL BE REJECTED

Name of Company Bizerba USA	Email Address michael.turner@bizerba.com	Application Date 3/15/18	
Mailing Address 239 Old New Brunswick Rd. Suite B	City Discataway	State NJ	Zip Code 08854
Telephone Number 732-565-6505	Cell Phone Number N/A	Fax Number 732-819-0581	

Select below all device types your company will certify:

Scales (include maximum capacity, if applicable)	Liquid (include maximum flow rate, if applicable)
<input type="checkbox"/> 1. Rail <input type="checkbox"/> 2. Truck <input type="checkbox"/> 3. Livestock <input type="checkbox"/> 4. Hopper: Max. Capacity: _____ <input type="checkbox"/> 5. Belt <input checked="" type="checkbox"/> 6. Over 30 lbs.: Max. Capacity: <u>60</u> <input checked="" type="checkbox"/> 7. 30 lbs. or less <input type="checkbox"/> 8. Class II (indicate on your calibration report which weight kit is Class II certified) <input type="checkbox"/> 9. Other: Please List:	<input type="checkbox"/> 1. Retail Fuel (less than 20 gal. per minute) <input type="checkbox"/> 2. High Flow Retail Fuel (20 gal. per minute or greater) <input type="checkbox"/> 3. Vehicle Tank: Max. Flow Rate: _____ <input type="checkbox"/> 4. Stationary Bulk (fuel or oil): Max. Flow Rate: _____ <input type="checkbox"/> 5. LPG <input type="checkbox"/> 6. Stationary LPG <input type="checkbox"/> 7. Fertilizer: Max. Flow Rate: _____ <input type="checkbox"/> 8. Chemical <input type="checkbox"/> 9. Anhydrous <input type="checkbox"/> 10. Loading Rack <input type="checkbox"/> 11. Other: Please List:

List below all persons employed by your company as a North Dakota Registered Service Person and the device types they are registered to certify (attach a separate sheet to list additional employees):

Permit No.	Employee	Device Types Registered to Certify (list using device type numbers from above)
e.g. 1001	e.g. John Doe	e.g. Scales - 2, 3, 6, 8; e.g. Liquid - 1, 2, 6
	Ben Thomsen	7, 6
		8 LPG
		7

Continued on Page 2

List below all field standards (attach current calibration reports):

New Mexico Dept of Agriculture P.O. Box 30005 Las Cruces New Mexico 88003-8005	Attached is a Certificate of Calibration


Additional Application Items (initial where appropriate):

Standardized Test Report	<input checked="" type="checkbox"/> Copy enclosed <input type="checkbox"/> No change in report filed previously
Tested and Approved Sticker	<input checked="" type="checkbox"/> Copy enclosed <input type="checkbox"/> No change in sticker filed previously
Photocopy of Crimped Lead Wire Seal	<input checked="" type="checkbox"/> Copy enclosed <input type="checkbox"/> No change in crimped lead wire seal filed previously

Public Company Listing:


Include my company information on your registered service company list for public contact.
 Yes No

I am Michael Turner, and have authority to represent this company.
 By signing this application, I declare that I have examined this form and accompanying documentation, and to the best of my knowledge and belief, the facts stated and documentation provided is true, correct, and complete.


 Signature

Send Completed Application and Related Documents To:

Public Service Commission
 600 E Boulevard Ave Dept 408
 Bismarck ND 58505-0480
Telephone: (701) 328-2400
Fax: (701) 328-2410



SECRETARY OF STATE NORTH DAKOTA



[Home](#) | [Business Records Search](#)

BIZERBA USA INC.

Corporation Details

System ID: 45161300
Type: FOREIGN BUSINESS CORPORATION
Status: Active & Good Standing
Original File Date: 06/20/2018
State of Origin: Delaware

Phone: (410) 879-8220

Effective Date: 06/20/2018



Nature of Business

SELL FOOD SLICERS, SCALES, LABELS, GLUE

Principal Office

1804 FASHION CT JOPPA, MD 21085-3237

Registered Agent

C T CORPORATION SYSTEM
314 E THAYER AVE
BISMARCK, ND 58501-4018
Established Date: Jun 20, 2018

Generate an Annual Report To File

To Generate a Annual Report form to be filed with the Secretary of State, select the appropriate year of the report you intend to file. This report does not contain details of a report previously filed with the Secretary of State. The annual report years reflected are an indication of the various report forms available in this site and is not an indication that an entity needs to file reports for all years. Missing years indicate that the forms for the missing year have not yet been deployed to the website, or have already been removed, and can be obtained by contacting the Secretary of State.

[2018](#) (generates a forms-fillable pdf in a new pop-up window)

[Return to Search Results](#)

[Contact Us](#)

[Disclaimer](#) | [Privacy Policy](#)

We use Secure Sockets Layer (SSL) encryption technology to ensure your information is secure and protected.

Will open a new window (pop-up).

W3C WAI AA, CSS, XHTML Compliant | Copyright 2006. All Rights Reserved. The State of North Dakota.



New Mexico Department of Agriculture
 Laboratory Division
 MSC 3-LD
 P.O. Box 30005
 Las Cruces, NM 88003-8005
 575-646-1551, fax: 575-646-2361



CALIBRATION CERTIFICATE

NM # 18 - 062

1040 AF

BIZERBA USA INC.
 239 OLD NEW BRUNSWICK RD STE B
 PISCATAWAY, NJ 08854

Artifacts: Rice Lake, 31 pcs, 1/32 oz-8 oz, .001 lb-5 lb
 Procedure: NM SLP 8
 Balance(s): AT 261 CC5001
 Condition of Artifact(s): Good

Laboratory Environmental Conditions at time of test:

Start	Temp °C	21.5	46.1 % Humidity	Pressure (mmHg)	664.4
Finish		21.1	47.4		664.3

Weight	Tolerance Class	ID#	As Found (mg)	Conventional Mass in (lb)	As Left corr (mg)	(X=accept)	Expanded Uncertainty (mg)
						Conventional Mass when adjusted in (lb)	
5 lb	A	230	52	5.000114		X	13
5 lb	B	230	49	5.000108		X	13
5 lb	C	230	89	5.000197		X	13
5 lb	D	230	77	5.000169		X	13
5 lb	E	230	46	5.000102		X	13
1 lb	A	70	14.7	1.000032		X	2.9
1 lb	B	70	18.2	1.000040		X	2.9
1 lb	C	70	11.2	1.000025		X	2.9
1 lb	D	70	7.2	1.000016		X	2.9
1 lb	E	70	28.7	1.000063		X	2.9

Date Calibrated 2/26/2018
 Date Due Recalibration 2/26/2019

Clay Ivey
 Clay Ivey, Metrologist

1040 AF

NOV 08 2013

Weight	Tolerance Class F (mg)	As Found (mg)	Conventional Mass in (oz)	As Left corr (mg)	(X=accept) Conventional Mass when adjusted in (oz)	Expanded Uncertainty (mg)
	ID#					
8 oz	45	17.6	8.000619		X	1.9
4 oz	23	8.1	4.000287		X	0.6
2 oz	11	3.4	2.000120		X	0.3
1 oz	5.4	2.4	1.000084		X	0.2
1/2 oz	2.8	1.37	0.500048		X	0.10
1/4 oz	1.7	0.750	0.2500265		X	0.085
1/8 oz	1.3	0.315	0.1250111		X	0.064
1/16 oz	1.1	0.353	0.0625124		X	0.055
1/32 oz	0.87	0.426	0.0312650		X	0.053
1/32 oz	* 0.87	0.366	0.0312629		X	0.053
			(lb)		(lb)	
0.2 lb	18	6.4	0.200014		X	0.5
0.2 lb	* 18	7.5	0.200017		X	0.5
0.1 lb	9.1	4.2	0.100009		X	0.2
0.05 lb	4.5	2.16	0.050005		X	0.14
0.02 lb	1.8	0.837	0.020002		X	0.070
0.02 lb	* 1.8	0.787	0.020002		X	0.070
0.01 lb	1.5	0.448	0.010001		X	0.066
0.005 lb	1.2	0.154	0.005000		X	0.056
0.002 lb	0.87	0.244	0.002001		X	0.053
0.002 lb	* 0.87	-0.036	0.002000		X	0.053
0.001 lb	0.70	0.222	0.001000		X	0.051

The standards of New Mexico are traceable to the U.S. prototypes through NIST, as required by State law, and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory is accredited by NVLAP, Lab Code number 200647-0. The unique report number above is to be used in referencing measurement traceability for artifacts identified in this report only. This certificate shall not be reproduced, except in full, without the written approval of the NMDA. This report must not be used by the client to claim product certification, approval, or endorsement by NVLAP, NIST or any agency of the federal government. The data obtained applies only to the artifacts identified in this report. Artifact classification derived from NIST105-1 or ASTM. All measurements are performed in the laboratory located at 3190 S. Espina, Las Cruces, NM. Measurements performed by Ryan Rust.

The laboratory's measurement uncertainty includes components for the standard, the measurement process, for any uncorrected errors associated with buoyancy corrections, and for differences in the assumed density of the weights. The combined standard uncertainty is multiplied by a coverage factor, k , of ≈ 2 to give an expanded uncertainty, which defines an interval having a confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the 2012 ISO Guide to the Expression of Uncertainty in Measurement. The expanded uncertainty is not to be confused with a tolerance limit for the user during application. Balances are for comparisons only, no calibration required.

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

New Mexico

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon II	Mass Echelon III	Volume Gravimetric, I
30 kg to 1 mg	500 kg to 1 mg	1 L
50 lb to 0.5 lb	5000 lb to 0.001 lb	5 gal to 1 dram
	8 oz to 0.03125 oz	Volume Transfer, II
	100 oz t	20 L
	500 gr to 200 gr	500 gal to 5 gal
	10 N to 1 N	100 gal to 25 gal LPG
	Weight Carts	
	5000 lb to 1000 lb	



2018

Douglas A. Olson

Douglas A. Olson, Chief
NIST Office of Weights and Measures

Effective Dates: 2018-01-01 to 2018-12-31

RECEIVED
APR 30 2018
ND Public
Service
Commission

