

PUBLIC SERVICE COMMISSION  
STATE OF NORTH DAKOTA

M E M O R A N D U M

TO: Memorandum for Record

FROM: Konrad Crockford *KSC*  
Director, Compliance Division

DATE: October 19, 2020

RE: Capital Scale Weights and Measures Rulemaking Verbal Comments  
Case No. WM-20-369

While Capital Scale Co., Bismarck, ND did not provide formal written comments, Staff wanted to memorialize comments provided during a telephone call by Travis Will, on behalf of Capital Scale Co. on September 3, 2020 for Weights and Measures Rulemaking, Case No. WM-20-369.

**69-10-02-05.1.(1.a.) & 69-10-02-05.1.(1.a.)**

- Capitol Scale Co. does not object to the reduction of the scale clearance from twelve inches to ten inches but prefers twelve inches.

**69-10-02-05.1.(1.b.)**

- Capitol Scale Co. recommends adding a four-foot minimum depth to the pier foundation requirement since frost lines can vary in different locations.

**69-10-02-26.**

- Capitol Scale Co. recommended increasing the minimum three feet of unobstructed clearance on all sides of a hopper scale.

**69-10-03-02.**

- Capital Scale Co. preferred the field standard certification requirement to remain at twelve months and believed it may be controversial for the automatic bulk weighing system five-year certification interval.